

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333

Accuracy check date/time: 12/04/2014 08:55

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3130	
Check sum	43295	Download date/time:	12/04/2014	08:58		Record ID	7MG4	
Record imported (Y/N)				N	Source Diagnostic			

Instrument serial No: 68-012333

Accuracy check date/time: 12/04/2014 08:56

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .112

Solution lot#: IRS021314

Soln temp C: 34.00

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	51815	Download date/time:	12/04/2014	08:58		Record ID	7MF1	
Record imported (Y/N)				N	Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 12/04/2014 08:58

First calibration check: .102

Interferent detected: No

Second calibration check: .104

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12985	Download date/time: 12/04/2014	09:00	Record ID 7MH1
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012333 Accuracy check date/time: 12/11/2014 10:03

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3129
Check sum 51488	Download date/time: 12/11/2014	10:06	Record ID 7MQ9				
Record imported (Y/N) N		Source Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 12/11/2014 10:04

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 53402	Download date/time: 12/11/2014	10:06	Record ID 7MP2
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012333 Accuracy check date/time: 12/11/2014 10:08

First calibration check: .102

Interferent detected: No

Second calibration check: .103

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C:
Simulator Ser No. DRXXXX

Remote calibration check by COBRA

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 14297	Download date/time: 12/11/2014	10:12	Record ID 7MR8
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 12/11/2014 10:10

First calibration check: .104

Interferent detected: No

Second calibration check: .103

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 62880	Download date/time: 12/11/2014	10:12	Record ID 7MS5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012333 Accuracy check date/time: 12/18/2014 11:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3104
Check sum 36598	Download date/time: 12/18/2014	11:05	Record ID 7NC3
Record imported (Y/N) N		Source Diagnostic	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 12/18/2014 11:04

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64721	Download date/time: 12/18/2014	11:05	Record ID 7NB0
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012333 Accuracy check date/time: 12/18/2014 11:06

First calibration check: .104

Interferent detected: No

Second calibration check: .106

Interferent detected: No

Concentration (%W/V) .112

Solution lot#: IRS021314

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 56833	Download date/time: 12/18/2014	11:07	Record ID 7ND4
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333

Accuracy check date/time: 12/19/2014 01:14

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 14756	Download date/time: 01/02/2015	10:32	Record ID 7OB1
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012333

Accuracy check date/time: 12/26/2014 09:46

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3103
Check sum 33488	Download date/time: 01/02/2015	10:32	Record ID 7O99				
Record imported (Y/N) N		Source Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012333 Accuracy check date/time: 12/26/2014 09:47

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 002

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 30471	Download date/time: 01/02/2015	10:32	Record ID 7071
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012333 Accuracy check date/time: 12/26/2014 09:49

First calibration check: .104

Interferent detected: No

Second calibration check: .105

Interferent detected: No

Concentration (%W/V) .112

Solution lot#: IRS021314

Soln temp C: 34.00

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 36973	Download date/time: 01/02/2015	10:32	Record ID 7082
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.