

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/07/2013 09:47

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3387
Check sum	62180	Download date/time:	02/07/2013 09:49	Record ID	6FP2		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012340

Accuracy check date/time: 02/07/2013 09:50

First calibration check: .143

Interferent detected: No

Second calibration check: .144

Interferent detected: No

Concentration (%W/V) .152

Solution lot#: IRS042012

Soln temp C: 34.03

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum	52197	Download date/time:	02/07/2013 09:49
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/14/2013 08:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3356
Check sum	19590	Download date/time:	02/21/2013 11:32	Record ID	6G10		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012340

Accuracy check date/time: 02/14/2013 08:44

First calibration check: .144

Interferent detected: No

Second calibration check: .144

Interferent detected: No

Concentration (%W/V) .152

Solution lot#: IRS042012

Soln temp C: 34.02

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum	48457	Download date/time:	02/21/2013 11:32
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/21/2013 11:29

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3353
Check sum	11631	Download date/time:	02/21/2013 11:32	Record ID	6G23		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012340

Accuracy check date/time: 02/21/2013 11:32

First calibration check: .142

Interferent detected: No

Second calibration check: .144

Interferent detected: No

Concentration (%W/V) .152

Solution lot#: IRS042012

Soln temp C: 33.99

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum	59777	Download date/time:	02/21/2013 11:32
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/28/2013 09:20

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3358	
Check sum	32623	Download date/time:	03/01/2013	08:35		Record ID	6GO9	
Record imported (Y/N)				N	Source Diagnostic			

Instrument serial No: 68-012340

Accuracy check date/time: 02/28/2013 09:22

First calibration check: RFI detected

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .152

Solution lot#: IRS042012

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	50300	Download date/time:	03/01/2013	08:35		Record ID	6GM1	
Record imported (Y/N)				N	Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/28/2013 09:22

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Radio frequency interference detected

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 53667	Download date/time: 03/01/2013	08:35	Record ID 6GP7
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012340

Accuracy check date/time: 02/28/2013 09:24

First calibration check: .146

Interferent detected: No

Second calibration check: .147

Interferent detected: No

Concentration (%W/V) .152

Solution lot#: IRS042012

Soln temp C: 34.04

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22670	Download date/time: 03/01/2013	08:35	Record ID 6GN2
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.