

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/26/2015 08:27

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .100 Solution lot#: 3131

Soln temp C: 34.00
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator 3131

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 34725	Download date/time: 03/05/2015	08:42	Record ID 7S95
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340

Accuracy check date/time: 02/26/2015 08:29

First calibration check: .119

Interferent detected: No

Second calibration check: .121

Interferent detected: No

Concentration (%W/V) .120 Solution lot#: IRS121014

Soln temp C: 34.00
Simulator Ser No. DR3707

LPACED INTO SERVICE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 32712	Download date/time: 03/05/2015	08:42	Record ID 7SA8
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/26/2015 08:31

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3595	
Check sum	25430	Download date/time:	03/05/2015	08:42		Record ID	7SC8	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340

Accuracy check date/time: 02/27/2015 08:57

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp	Fail	Internal standards		
Prom	Pass	Real Time clock		Motor		Printer		
Ram	Pass	Auto calibration		Range		Digital voltmeter	???	
Check sum	35922	Download date/time:	03/05/2015	08:42		Record ID	7SD3	
Record imported (Y/N)				N				Source Diagnostic

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340

Accuracy check date/time: 02/28/2015 02:51

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3614
Check sum	79170	Download date/time:	03/05/2015 08:42			Record ID	7SE0
		Record imported (Y/N)	N	Source	Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.