

**Long Beach Police Department**

**Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/15/2015 09:47

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3426
Check sum	39025	Download date/time:	01/21/2015 09:04	Record ID	7PQ6		
Record imported (Y/N) N				Source Diagnostic			

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/15/2015 12:09

First calibration check: .116

Interferent detected: No

Second calibration check: .118

Interferent detected: No

Concentration (%W/V) .120

Solution lot#: IRS121014

Soln temp C: 34.00

Simulator Ser No. DR3707

PLACED INTO SERVICE - BOOKING

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum	55010	Download date/time:	01/21/2015 09:04
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department**

**Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/15/2015 16:18

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3310
Check sum	28595	Download date/time:	01/21/2015 09:04	Record ID	7PR9		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012336

Accuracy check date/time: 01/21/2015 08:54

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3376
Check sum	11557	Download date/time:	01/21/2015 09:04	Record ID	7PS2		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department**

**Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336 Accuracy check date/time: 01/21/2015 08:56

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 024

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 15963	Download date/time: 01/21/2015	09:04	Record ID 7PO5
Record imported (Y/N) N		Source Calibration	

---

Instrument serial No: 68-012336 Accuracy check date/time: 01/21/2015 09:02

First calibration check: .124

Interferent detected: No

Second calibration check: .123

Interferent detected: No

Concentration (%W/V) .120

Solution lot#: IRS121014

Soln temp C: 34.00  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 024

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64911	Download date/time: 01/21/2015	09:04	Record ID 7PP3
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department**

**Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/29/2015 09:00

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3389
Check sum	47711	Download date/time:	01/29/2015 09:10	Record ID	7Q13		
Record imported (Y/N) N				Source Diagnostic			

---

Instrument serial No: 68-012336

Accuracy check date/time: 01/29/2015 09:03

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum	32340	Download date/time:	01/29/2015 09:10
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 01/29/2015 09:08

First calibration check: .122

Interferent detected: No

Second calibration check: .122

Interferent detected: No

Concentration (%W/V) .120

Solution lot#: IRS121014

Soln temp C: 34.00

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 018

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 55044	Download date/time: 01/29/2015	09:10	Record ID 7Q00
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.