

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338

Accuracy check date/time: 07/02/2014 10:11

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | |
|-------------------------|-------|---------------------|------------------|-------------------|------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3275 |
| Check sum | 59837 | Download date/time: | 07/02/2014 10:13 | Record ID | 7AL2 | | |
| Record imported (Y/N) N | | | | Source Diagnostic | | | |

Instrument serial No: 68-012338

Accuracy check date/time: 07/02/2014 10:13

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator _____

Diagnostics check

| | | | |
|-------------------------|------------------|---------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum | 55827 | Download date/time: | 07/02/2014 10:13 |
| Record imported (Y/N) N | | Source Calibration | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 07/02/2014 10:15

First calibration check: .103

Interferent detected: No

Second calibration check: .105

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 34.01
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 45747 | Download date/time: 07/02/2014 | 10:16 | Record ID 7AM1 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 07/10/2014 09:52

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|-------------------|-------------------------|
| Eeprom Pass | Serial No. Pass | Temp Pass | Internal standards Pass |
| Prom Pass | Real Time clock Pass | Motor Pass | Printer Pass |
| Ram Pass | Auto calibration Pass | Range Pass | Digital voltmeter 3283 |
| Check sum 63617 | Download date/time: 07/10/2014 | 09:55 | Record ID 7B26 |
| Record imported (Y/N) N | | Source Diagnostic | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 07/10/2014 09:53

First calibration check: Unstable reference

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 58669 | Download date/time: 07/10/2014 | 09:55 | Record ID 7B04 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 07/10/2014 09:53

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | |
|-------------------------|--------------------------------|------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 12505 | Download date/time: 07/10/2014 | 09:55 | Record ID 7B35 |
| Record imported (Y/N) N | | Source Exception | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 07/10/2014 09:54

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 29268 | Download date/time: 07/10/2014 | 09:55 | Record ID 7B13 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 07/10/2014 09:57

First calibration check: .103

Interferent detected: No

Second calibration check: .105

Interferent detected: No

Concentration (%W/V) .112

Solution lot#: IRS021314

Soln temp C: 0000
Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 63293 | Download date/time: 07/10/2014 | 09:58 | Record ID 7B42 |
| Record imported (Y/N) N | | Source Calibration | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338

Accuracy check date/time: 07/18/2014 09:26

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

| | | | | | | | |
|-------------------------|-------|---------------------|------------------|-------------------|------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3283 |
| Check sum | 72730 | Download date/time: | 07/18/2014 09:28 | Record ID | 7BI8 | | |
| Record imported (Y/N) N | | | | Source Diagnostic | | | |

Instrument serial No: 68-012338

Accuracy check date/time: 07/18/2014 09:27

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .112

Solution lot#: IRS021314

Soln temp C: 0000

Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|------------------|---------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum | 24300 | Download date/time: | 07/18/2014 09:28 |
| Record imported (Y/N) N | | Source Calibration | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 07/18/2014 09:30

First calibration check: .102

Interferent detected: No

Second calibration check: .104

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 0000
Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| | | | |
|-------------------------|--------------------------------|--------------------|--------------------|
| Eeprom | Serial No. | Temp | Internal standards |
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 21667 | Download date/time: 07/18/2014 | 09:30 | Record ID 7BJ3 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 07/26/2014 20:55

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

| | | | | | | | |
|-------------------------|--------------------------------|-------------------|----------------|-------|------|--------------------|------|
| Eeprom | Pass | Serial No. | Pass | Temp | Pass | Internal standards | Pass |
| Prom | Pass | Real Time clock | Pass | Motor | Pass | Printer | Pass |
| Ram | Pass | Auto calibration | Pass | Range | Pass | Digital voltmeter | 3294 |
| Check sum 39168 | Download date/time: 07/31/2014 | 11:01 | Record ID 7C38 | | | | |
| Record imported (Y/N) N | | Source Diagnostic | | | | | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012338 Accuracy check date/time: 07/26/2014 20:56

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 0000
Simulator Ser No. DRXXXX

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| Eeprom | Serial No. | Temp | Internal standards |
|-------------------------|--------------------------------|--------------------|--------------------|
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 36710 | Download date/time: 07/31/2014 | 11:01 | Record ID 7C12 |
| Record imported (Y/N) N | | Source Calibration | |

Instrument serial No: 68-012338 Accuracy check date/time: 07/26/2014 20:59

First calibration check: .106

Interferent detected: No

Second calibration check: .107

Interferent detected: No

Concentration (%W/V) .112 Solution lot#: IRS021314

Soln temp C: 34.01
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

| Eeprom | Serial No. | Temp | Internal standards |
|-------------------------|--------------------------------|--------------------|--------------------|
| Prom | Real Time clock | Motor | Printer |
| Ram | Auto calibration | Range | Digital voltmeter |
| Check sum 11179 | Download date/time: 07/31/2014 | 11:01 | Record ID 7C27 |
| Record imported (Y/N) N | | Source Calibration | |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.