

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336 Accuracy check date/time: 05/26/2016 09:16

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |            |       |      |                    |      |                   |
|-----------------------|-------|---------------------|------------|-------|------|--------------------|------|-------------------|
| Eeprom                | Pass  | Serial No.          | Pass       | Temp  | Pass | Internal standards | Pass |                   |
| Prom                  | Pass  | Real Time clock     | Pass       | Motor | Pass | Printer            | Pass |                   |
| Ram                   | Pass  | Auto calibration    | Pass       | Range | Pass | Digital voltmeter  | 3417 |                   |
| Check sum             | 62974 | Download date/time: | 06/02/2016 | 09:03 |      | Record ID          | 8MT7 |                   |
| Record imported (Y/N) |       |                     |            | N     |      |                    |      | Source Diagnostic |

Instrument serial No: 68-012336 Accuracy check date/time: 05/26/2016 10:40

First calibration check: .137  
Second calibration check: .139

Interferent detected: No  
Interferent detected: No

Concentration (%W/V) .146

Solution lot#: IRS112515

Soln temp C: 34.09  
Simulator Ser No. DR3707

PLACED INTO SERVICE - CSULB

Number of people tested since last periodic determination of accuracy: 000

Operator G. GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

|                       |       |                     |            |       |  |                    |      |                    |
|-----------------------|-------|---------------------|------------|-------|--|--------------------|------|--------------------|
| Eeprom                |       | Serial No.          |            | Temp  |  | Internal standards |      |                    |
| Prom                  |       | Real Time clock     |            | Motor |  | Printer            |      |                    |
| Ram                   |       | Auto calibration    |            | Range |  | Digital voltmeter  |      |                    |
| Check sum             | 55559 | Download date/time: | 06/02/2016 | 09:03 |  | Record ID          | 8MR5 |                    |
| Record imported (Y/N) |       |                     |            | N     |  |                    |      | Source Calibration |

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012336

Accuracy check date/time: 05/28/2016 05:36

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

|           |                       |                     |                  |            |      |                    |      |
|-----------|-----------------------|---------------------|------------------|------------|------|--------------------|------|
| Eeprom    | Pass                  | Serial No.          | Pass             | Temp       | Pass | Internal standards | Pass |
| Prom      | Pass                  | Real Time clock     | Pass             | Motor      | Pass | Printer            | Pass |
| Ram       | Pass                  | Auto calibration    | Pass             | Range      | Pass | Digital voltmeter  | 3499 |
| Check sum | 24091                 | Download date/time: | 06/02/2016 09:03 |            |      | Record ID          | 8MU8 |
|           | Record imported (Y/N) | N                   | Source           | Diagnostic |      |                    |      |

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.