

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 10/25/2013 11:33

First calibration check: .114

Interferent detected: No

Second calibration check: .116

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.08
Simulator Ser No. DR3707

PLACED INTO SERVICE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64427	Download date/time: 11/07/2013	09:56	Record ID 6TZ7
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012335 Accuracy check date/time: 10/25/2013 11:36

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3410
Check sum 20851	Download date/time: 11/07/2013	09:56	Record ID 6U93
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 10/25/2013 11:44

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3414
Check sum	31447	Download date/time:	11/07/2013 09:56	Record ID	6UA0		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012335

Accuracy check date/time: 10/25/2013 12:48

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3372
Check sum	33946	Download date/time:	11/07/2013 09:56	Record ID	6UB9		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335

Accuracy check date/time: 10/28/2013 01:34

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 26827	Download date/time: 11/07/2013	09:56	Record ID 6UK4
Record imported (Y/N) N	Source	Exception	

Instrument serial No: 68-012335

Accuracy check date/time: 10/31/2013 11:54

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3338
Check sum 42050	Download date/time: 11/07/2013	09:56	Record ID 6UC4				
Record imported (Y/N) N	Source	Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 10/31/2013 11:55

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 029

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 20887	Download date/time: 11/07/2013	09:56	Record ID 6U37
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012335 Accuracy check date/time: 10/31/2013 12:02

First calibration check: .122

Interferent detected: No

Second calibration check: .122

Interferent detected: No

Concentration (%W/V) .119

Solution lot#: IRS022113

Soln temp C: 33.99

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 16657	Download date/time: 11/07/2013	09:56	Record ID 6U04
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.