

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 21:41

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2615
Check sum	43515	Download date/time:	09/05/2013 08:07			Record ID	6QS9
		Record imported (Y/N)	N	Source	Diagnostic		

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 21:47

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2623
Check sum	26320	Download date/time:	09/05/2013 08:07			Record ID	6QT8
		Record imported (Y/N)	N	Source	Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2635
Check sum	64697	Download date/time:	09/05/2013 08:07			Record ID	6QU3
		Record imported (Y/N)	N	Source	Diagnostic		

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:15

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2647
Check sum	18853	Download date/time:	09/05/2013 08:07			Record ID	6QV4
		Record imported (Y/N)	N	Source	Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:16

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2648
Check sum	35721	Download date/time:	09/05/2013 08:07			Record ID	6QW9
	Record imported (Y/N)	N		Source	Diagnostic		

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:57

First calibration check: Unstable reference

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .119

Solution lot#: IRS022113

Soln temp C: 34.03

Simulator Ser No. DR3703

TAKEN OUT OF SERVICE - INT STD FAIL

Number of people tested since last periodic determination of accuracy: 030

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards	
Prom	Real Time clock	Motor	Printer	
Ram	Auto calibration	Range	Digital voltmeter	
Check sum	37096	Download date/time:	09/05/2013 08:07	
	Record imported (Y/N)	N	Source	Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:57

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 48059	Download date/time: 09/05/2013	08:07	Record ID 6R14
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:58

First calibration check: Unstable reference

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .119

Solution lot#: IRS022113

Soln temp C: 34.01

Simulator Ser No. DR3703

TAKEN OUT OF SERVICE - INT STD FAIL

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 17206	Download date/time: 09/05/2013	08:07	Record ID 6QP8
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 22:58

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 61413	Download date/time: 09/05/2013	08:07	Record ID 6R27
Record imported (Y/N) N		Source Exception	

Instrument serial No: 68-012323

Accuracy check date/time: 09/03/2013 23:02

First calibration check: .STD

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31074	Download date/time: 09/05/2013	08:07	Record ID 6QQ3
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 09/03/2013 23:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Internal standard out of tolerance

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 19455	Download date/time: 09/05/2013	08:07	Record ID 6R32
Record imported (Y/N) N	Source	Exception	

Instrument serial No: 68-012323 Accuracy check date/time: 09/03/2013 23:04

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Fail
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	2666
Check sum 38526	Download date/time: 09/05/2013	08:07	Record ID 6QX2				
Record imported (Y/N) N	Source	Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012323 Accuracy check date/time: 09/03/2013 23:06

First calibration check: .115

Interferent detected: No

Second calibration check: .117

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.01
Simulator Ser No. DR3703

TAKEN OUT OF SERVICE - INT STD FAIL

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 37976	Download date/time: 09/05/2013	08:07	Record ID 6QR4
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012323 Accuracy check date/time: 09/04/2013 17:31

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Fail
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 2601
Check sum 60756	Download date/time: 09/05/2013	08:07	Record ID 6QY7
Record imported (Y/N) N		Source Diagnostic	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.