

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 09/09/2013 10:03

First calibration check: .121

Interferent detected: No

Second calibration check: .121

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.01
Simulator Ser No. DR3703

PLACED INTO SERVICE - YOUTH SERVICES

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 34604	Download date/time: 10/03/2013	11:03	Record ID 6RZ5
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 09/09/2013 10:04

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.01
Simulator Ser No. DR3703

PLACED INTO SERVICE - YOUTH SERVICES

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49220	Download date/time: 10/03/2013	11:03	Record ID 6S06
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337

Accuracy check date/time: 09/09/2013 10:05

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3510
Check sum	62093	Download date/time:	10/03/2013 11:03	Record ID	6S95		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012337

Accuracy check date/time: 09/09/2013 12:04

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3419
Check sum	24225	Download date/time:	10/03/2013 11:03	Record ID	6SA4		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 09/12/2013 12:10

First calibration check: .106

Interferent detected: No

Second calibration check: .108

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.05
Simulator Ser No. DR3703

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 64866	Download date/time: 10/03/2013	11:03	Record ID 6S11
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 09/12/2013 12:12

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3573
Check sum 25401	Download date/time: 10/03/2013	11:03	Record ID 6SB7				
Record imported (Y/N) N		Source Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 09/12/2013 12:16

First calibration check: .110

Interferent detected: No

Second calibration check: .110

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.01
Simulator Ser No. DR3703

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 26563	Download date/time: 10/03/2013	11:03	Record ID 6S20
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 09/19/2013 11:44

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3568
Check sum 61646	Download date/time: 10/03/2013	11:03	Record ID 6SC8				
Record imported (Y/N) N		Source Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 09/19/2013 11:46

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 002

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12877	Download date/time: 10/03/2013	11:03	Record ID 6S35
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 09/19/2013 11:49

First calibration check: .108

Interferent detected: No

Second calibration check: .110

Interferent detected: No

Concentration (%W/V) .119

Solution lot#: IRS022113

Soln temp C: 34.05

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 002

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 41884	Download date/time: 10/03/2013	11:03	Record ID 6S44
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

Long Beach Police Department

Criminalistics Laboratory

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337 Accuracy check date/time: 09/19/2013 11:54

First calibration check: .111

Interferent detected: No

Second calibration check: .113

Interferent detected: No

Concentration (%W/V) .119 Solution lot#: IRS022113

Soln temp C: 34.01
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 57862	Download date/time: 10/03/2013	11:03	Record ID 6S59
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012337 Accuracy check date/time: 09/19/2013 12:02

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V) Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3503
Check sum 82010	Download date/time: 10/03/2013	11:03	Record ID 6SD5				
Record imported (Y/N) N		Source Diagnostic					

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337

Accuracy check date/time: 09/19/2013 12:07

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3497
Check sum	11372	Download date/time:	10/03/2013 11:03			Record ID	6SE6
		Record imported (Y/N)	N	Source	Diagnostic		

Instrument serial No: 68-012337

Accuracy check date/time: 09/26/2013 11:53

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3593
Check sum	30862	Download date/time:	10/03/2013 11:03			Record ID	6SF9
		Record imported (Y/N)	N	Source	Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Criminalistics Laboratory**

1400 Canal Ave

Long Beach, Ca 90813

PH: (562) 570-7680 Fax: (562) 570-7704

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012337

Accuracy check date/time: 09/26/2013 12:00

First calibration check: .110

Interferent detected: No

Second calibration check: .111

Interferent detected: No

Concentration (%W/V) .119

Solution lot#: IRS022113

Soln temp C: 34.06

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 31900	Download date/time: 10/03/2013	11:03	Record ID 6S68
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.