

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008501
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1258	34	13A28-13-5-10- KKF24C	.100	23		.096			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:01
 NS4 Serial no: 008501
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/30/09 15:40 .100
 Test Result
 0001 02/27/14 13:00 .096
 Push No. 3 to erase

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1301	34	13A28-13-5-10- KKF24C	.100	23		.100			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
Push No. 3 to erase														
Test Result 0001 02/27/14 13:03 .100														
Calibration 07/18/09 13:27 .100														
No. Date Time %BAC Version no: 4797 AS4 Serial no: 018785 02/27/14 13:03														
OR POWER DOWN TO RECALIB														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014			13A28-13-5-10- KKF24C	.100								Lee	17046	NOT IN LOCKER

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	082659

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	OUT FOR REPAIR
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	OUT for REPAIR
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008580
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	OUT FOR REPAIR
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 08332 083032
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)
02/27/2014	1305	34	13A28-13-5-10- KKF24C	.100	23		.100			X	Lee	17046	

Prepared by:

Douglas C. Talbot

01-28-13

Expires on:

01-28-15

or power down to retain

02/27/14 13:05
AS4 Serial no: 083032
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 09:01 .100

Test Result
0001 02/22/14 02:54 .097
0002 02/22/14 02:57 .101
0003 02/23/14 03:51 .081
0004 02/23/14 03:54 .082
0005 02/27/14 13:04 .100

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1300	34	13A28-13-5-10-KKF24C	.100	23		.098			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

02/27/14 13:14
AS4 Serial no: 008620
Version no: 4797
No. Date Time ZBAC
Calibration 01/30/13 14:34 .100
Test Result
0001 02/26/14 02:50 .131
0002 02/26/14 02:52 .134
0003 02/27/14 13:13 .098
Push No. 3 to erase,
or power down to retain
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1311	34	13A28-13-5-10- KKF24C	.100	24		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:11
 NS4 Serial no: 082957
 Version no: 4797

No. Date Time XMBC
 Calibration
 02/17/14 08:52 .100

Test Result
 0001 02/27/14 13:11 .101

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1314	34	13A28-13-5-10-KKF24C	.100	23		.099		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:14
AS4 Serial no: 077505
Version no: 4797

No. Date Time %MAC
Calibration
05/21/08 13:16 .108

Test Result
0001 02/27/14 13:13 .099

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1317	34	13A28-13-5-10- KKF24C	.100	22		.100			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Push No. 3 to erase,
or power down to retain

Test Result
0001 02/23/14 04:03 .045
0002 02/24/14 01:22 .091
0003 02/24/14 01:25 .088
0004 02/27/14 13:18 .100

Mo. Date Time NBAC
Calibration
02/17/14 09:10 .100

02/27/14 13:18
NS4 Serial no: 077479
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1320	34	13A28-13-5-10- KKF24C	.100	23		.100			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result
 0001 02/22/14 02:41 .087
 0002 02/22/14 03:27 .044
 0003 02/23/14 01:39 .178
 0004 02/23/14 01:41 .168
 0005 02/23/14 04:06 .098
 0006 02/23/14 04:09 .099
 0007 02/27/14 13:20 .100
 Calibration
 02/17/14 09:16 .100
 No. Date Time ZBAC
 Calibration
 02/17/14 09:16 .100
 Version no: 4797
 NS4 Serial no: 082660
 02/27/14 13:20

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1326	34	13A28-13-5-10- KKF24C	.100	23		.103			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Test Result
0001 02/27/14 13:30 .103
 No. Date Time NBAC
Calibration
03/11/04 14:05 .100
 02/27/14 13:30
NS4 Serial no: 008714
Version no: 4797
 Push No. 3 to erase,
or power down to retain

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1329	34	13A28-13-5-10-KKF24C	.100	23		.101			X	Lee	17046	
<div style="display: flex; justify-content: space-between; align-items: flex-start;"> <div style="width: 45%;"> <p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p> </div> <div style="width: 45%; border: 1px solid black; padding: 5px;"> <p style="text-align: center;">02/27/14 13:25 HS4 Serial no: 008888 Version no: 4797</p> <p style="text-align: center;">No. Date Time ZBAC Calibration 08/25/13 18:45 .100</p> <p style="text-align: center;">Test Result 0001 02/27/14 13:25 .101</p> <p style="text-align: center;">Push No. 3 to erase, or power down to retain</p> </div> </div>													

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)
02/27/2014	1332	34	13A28-13-5-10- KKF24C	.100	23		.100			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:33
 MS4 Serial no: 008783
 Version no: 4797
 No. Date Time %RNC
 Calibration 07/18/08 13:52 .100
 Test Result
 0001 02/27/14 13:32 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1335	34	13A28-13-5-10- KKF24C	.100	23		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Push No. 3 to erase,
or power down to retain
 Test Result
 0001 02/22/14 02:52 .042
 0002 02/27/14 13:34 .101
 No. Date Time ZMNC
 Calibration 02/17/14 09:50 .100
 02/27/14 13:34
 AS4 Serial no: 082657
 version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1330	36	13A28-13-5-10- KKF24C	.100	22		.097			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 02/20/14 23:47 .179 0002 02/20/14 23:50 .171 0003 02/25/14 08:49 .107 0004 02/25/14 08:51 .127 0005 02/27/14 13:37 .097 </p>													
<p style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time %BAC Calibration 18/30/03 09:53 .100 </p>													
<p style="writing-mode: vertical-rl; transform: rotate(180deg);"> 02/27/14 13:37 AS4 Serial no: 008810 Version no: 4797 </p>													
<p style="writing-mode: vertical-rl; transform: rotate(180deg);"> Push No. 3 to erase, or power down to retain </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	134)	34	13A28-13-5-10- KKF24C	.100	24		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										
			Test Result										
			0001 02/21/14 00:22 .208										
			0002 02/21/14 00:25 .196										
			0003 02/26/14 02:06 .000										
			0004 02/27/14 13:38 .101										
			Calibration										
			11/02/11 16:03 .100										
			No. Date Time %BAC										
			11/02/11 16:03 .100										
			Version no: 4797										
			AS4 Serial no: 018684										
			02/27/14 13:38										

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1344	34	13A28-13-5-10- KKF24C	.100	23		.101			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											
<p>02/27/14 13:42 AS4 Serial no: 008635 Version no: 4797</p> <p>No. Date Time %BNC Calibration 10/24/03 09:04 .100</p> <p>Test Result 0001 02/21/14 02:50 .123 0002 02/21/14 02:53 .128 0003 02/21/14 06:00 .105 0004 02/21/14 06:04 .096 0005 02/25/14 02:05 .291 0006 02/25/14 02:07 .279 0007 02/26/14 03:26 .143 0008 02/26/14 03:28 .149 0009 02/27/14 13:41 .101</p>														
			Push No. 3 to erase, or Power down to retain											
			or Power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
02/27/2014	1347	34	13A28-13-5-10-KKF24C	.100	23		.100		X		Lee	17046			
Prepared by:															
Douglas C. Talbot															
01-28-13															
Expires on:															
01-28-15															
									Test Result 0001 02/21/14 20:53 .000 0002 02/21/14 22:07 .000 0003 02/25/14 20:00 .000 0004 02/27/14 13:43 .106						
									No. Date Time ZMOC Calibration 12/24/03 11:01 .100						
									Push No. 3 to erase, or power down to retain						

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1350	34	13A28-13-5-10- KKF24C	.100	23		101		X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p>02/27/14 19:48 AS4 Serial no: 082653 Version no: 4797</p> <p>No. Date Time %DAC Calibration 02/17/14 10:08 .100</p> <p>Test Result 0001 02/22/14 01:41 .158 0002 02/22/14 01:45 .157 0003 02/22/14 21:37 .194 0004 02/22/14 21:39 .207 0005 02/22/14 23:00 .181 0006 02/22/14 23:11 .175 0007 02/23/14 09:53 .000 0008 02/23/14 09:55 .000 0009 02/23/14 21:27 .155 0010 02/23/14 21:30 .205 0011 02/23/14 21:33 .180 0012 02/24/14 02:00 .047 0013 02/24/14 02:02 .044 0014 02/27/14 13:47 .101</p> <p>Push No. 3 to erase, or Power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1353	34	13A28-13-5-10- KKF24C	.100	22		.103			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:53
AS4 Serial no: 008848
Version no: 4797

No. Date Time %MAC
Calibration
01/13/09 09:25 .100

Test Result
0001 02/27/14 13:52 .103

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1356	34	13A28-13-5-10- KKF24C	.100	22		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 13:53
 AS4 Serial no: 077591
 Version no: 4297
 No. Date Time XBAC
 Calibration 05/21/08 15:09 .100
 Test Result
 0001 02/27/14 09:13 .000
 0002 02/27/14 13:53 .102
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
02/27/2014	1359	34	13A28-13-5-10- KKF24C	.100	22	.102				X		Lee	17046		

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Test Result
 0001 02/24/14 23:02 .028
 0002 02/24/14 23:28 .021
 0003 02/25/14 00:49 .036
 0004 02/25/14 01:14 .055
 0005 02/25/14 01:34 .100
 0006 02/25/14 01:36 .100
 0007 02/27/14 13:59 .102
 No. Date Time ZBAC
 Calibration 03/02/06 09:07 .100
 Version no: 4797
 NS4 Serial no: 008850
 02/27/14 14:00

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	OUT FOR SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	140234		13A28-13-5-10- KKF24C	.100	23		.100		X		Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15										

02/27/14 13:57
 RS4 Serial no: 083051
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/25/13 14:00 .100
 Test Result
 0001 02/22/14 20:36 .000
 0002 02/22/14 20:37 .000
 0003 02/27/14 13:57 .100
 Push No. 3 to erase
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1405	34	13A28-13-5-10-KKF24C	.100	22		107			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result
 0001 02/21/14 23:20 .000
 0002 02/22/14 01:10 .130
 0003 02/22/14 01:13 .131
 0004 02/27/14 14:03 .107
 Push No. 3. To erase.

No. Date Time M8AC
 Calibration 01/16/12 16:13 .100
 02/27/14 14:03
 AS4 Serial no: 008897
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	LD. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	------------------------------------	---------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1411	34	13A28-13-5-10- KKF24C	.100	24		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

02/27/14 14:28
 NS4 Serial no: 018646
 Version no: 4797

No: Date Time %RAC
 Calibration
 02/29/13 13:32 .100

Test Result
 0001 02/27/14 14:20 .099

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1414	34	13A28-13-5-10- KKF24C	.100	22		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 14:23
 MS4 Serial no: 077488
 Version no: 4797
 No. Date Time ZPMC
 Calibration
 05/21/08 14:35 .100
 Test Result
 0001 02/27/14 14:23 .102
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1417	34	13A28-13-5-10-	.100	22		.105			X	Lee	17046	
			KKF24C										
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 14:29
 HS4 Serial no: 008877
 Version no: 4797
 No. Date Time %RAC
 Calibration 05/26/08 18:40 .100
 Test Result
 0001 02/27/14 14:28 .105
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1420	34	13A28-13-5-10- KKF24C	.100	24		.098			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

02/27/14 14:44
AS4 Serial no: 018615
Version no: 4797
No. Date Time ZBAC
Calibration
12/19/08 14:44 .100
Test Result
0001 02/27/14 14:43 .098
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1423	34	13A28-13-5-10-KKF24C	.100	24		.099			X	Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13										
			Expires on: 01-28-15										

02/27/14 14:44
 NS4 Serial no: 018626
 Version no: 4797
 No. Date Time %BAC
 Calibration 04/29/05 08:59 .101
 Test Result
 0001 02/27/14 14:43 .099
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1426	34	13A28-13-5-10- KKF24C	.100	23	.101				X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result
 0001 02/27/14 14:46 .101
 Calibration
 09/09/13 14:50 .100
 No. Date Time ZBHC
 02/27/14 14:46
 RS4 Serial no: 018696
 Version no: 4797

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
02/27/2014	1429	34	13A28-13-5-10- KKF24C	.100	23		.099			X	Lee	17046		
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											
<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p style="text-align: center;">02/27/14 14:48 NS4 Serial no: 083030 Version no: 4797</p> <p style="text-align: center;">No. Date Time BAC Calibration 01/29/09 10:39 .100</p> <p style="text-align: center;">Test Result 0001 02/27/14 14:47 .099</p> <p style="text-align: center;">Push No. 3 to erase. or power down to retain</p> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)	YES
02/27/2014	1432	34	13A28-13-5-10- KKF24C	.100	23		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

02/27/14 14:56
 AS4 Serial no: 018692
 Version no: 4797
 No. Date Time %BAC
 Calibration 08/12/08 14:35 .100
 Test Result
 0001 02/27/14 14:55 .099
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1435	34	13A28-13-5-10-KKF24C	.100	22		.102			X		Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Push No. 3 to erase.
Test Result
0001 02/27/14 14:59 .100
Calibration
02/27/09 15:52 .100
No. Date Time ZBAC
Version no: 479?
MS4 Serial no: 018762
02/27/14 15:00

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1438	34	13A28-13-5-10- KKF24C	.100	23							Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15											
			Push No. 3 to erase, or power down to retain											
			Test Result 0001 02/21/14 23:14 .000 0002 02/22/14 00:10 .136 0003 02/22/14 00:14 .131 0004 02/22/14 01:57 .110 0005 02/22/14 02:00 .092 0006 02/25/14 02:10 .164 0007 02/25/14 02:13 .155 0008 02/26/14 23:29 .114 0009 02/26/14 23:32 .111 0010 02/27/14 02:39 .166 0011 02/27/14 02:42 .163 0012 02/27/14 15:02 .096											
			No. Date Time ZBAC Calibration 03/12/12 12:30 .100											
			02/27/14 15:03 HS4 Serial no: 018766 Version no: 4797											
			or power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
02/27/2014	1441	34	13A28-13-5-10-KKF24C	.100	23		.104			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

02/27/14 15:08
 PAS4 Serial no: 018691
 Version no: 4797
 No. Date Time ZBAC
 Calibration 09/08/08 14:01 .100
 Test Result
 0001 02/22/14 07:01 .149
 0002 02/22/14 07:05 .141
 0003 02/22/14 20:01 .000
 0004 02/22/14 21:22 .000
 0005 02/23/14 01:12 .130
 0006 02/23/14 01:14 .122
 0007 02/27/14 15:00 .104
 Push No. 3 to erase
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
02/27/2014	1444	34	13A28-13-5-10- KKF24C	.100	23		.107			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			01-28-13												
			Expires on:												
			01-28-15												

02/27/14 15:09
 NS4 Serial no: 017990
 Version no: 4797
 No. Date Time ZBAC
 Calibration 11/26/03 11:10 .100
 Test Result
 0001 02/22/14 01:36 .161
 0002 02/22/14 01:38 .154
 0003 02/22/14 01:40 .000
 0004 02/22/14 03:07 .159
 0005 02/22/14 03:09 .177
 0006 02/23/14 01:25 .110
 0007 02/23/14 01:28 .107
 0008 02/23/14 01:35 .094
 0009 02/27/14 15:00 .107

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1448	34	13A28-13-5-10- KKF24C	.100	23		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											
<div style="text-align: right;"> <p>02/27/14 15:11 HS4 Serial no: 018774 Version no: 4797</p> <p>No. Date Time %BAC Calibration 01/13/09 10:25 .100</p> <p>Test Result 0001 02/21/14 01:27 .051 0002 02/21/14 01:29 .050 0003 02/21/14 02:13 .027 0004 02/21/14 03:08 .032 0005 02/25/14 02:58 .074 0006 02/25/14 03:02 .069 0007 02/26/14 00:51 .221 0008 02/26/14 00:54 .222 0009 02/27/14 15:11 .100</p> <p>Push No. 3 to erase, or Power down to retain</p> </div>														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1451	34	13A28-13-5-10- KKF24C	.100	23		.105			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

02/27/14 15:16
 NS4 Serial no: 041486
 Version no: 4797
 No. Date Time ZMAC
 Calibration 03/08/06 13:56 .100
 Test Result 0001 02/27/14 15:13 .105
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1500	34	13A28-13-5-10- KKF24C	.100	23		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

02/27/14 15:21
 PAS Serial no: 041495
 Version no: 4797
 No. Date Time %BAC
 Calibration
 02/20/14 11:28 .100
 Test Result
 0001 02/27/14 15:20 .099
 Push No. 3 to erase.
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1505	34	13A28-13-5-10- KKF24C	.100	23		.098			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test Result
0001 02/27/14 15:25 .098
 No. Date Time ZBNC
Calibration
05/05/09 08:03 .100
 02/27/14 15:25
 H54 Serial no: 041502
Version no: 4797
 02/27/14 15:25
 Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

SGT

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	------------------------------------	----------------------	----------------------------------------------------	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/27/2014	1509	34	13A28-13-5-10- KKF24C	.100	22		.103			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

02/27/14 15:28
 NS4 Serial no: 008874
 Version no: 4797
 No. Date Time ZBAC
 Calibration 06/25/13 10:32 .107

Test Result
 0001 02/21/14 00:34 .005
 0002 02/21/14 00:37 .090
 0003 02/25/14 00:15 .176
 0004 02/25/14 00:18 .165
 0005 02/25/14 22:32 .090
 0006 02/25/14 22:35 .009
 0007 02/27/14 15:28 .103

Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.