

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008501
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1213	34	13A28-13-5-10-KKF24C	.100	22		.093			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

03/06/14 12:19  
 AS4 Serial no: 008501  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration  
 03/30/09 15:40 .100  
 Test Result  
 0001 03/06/14 12:16 .093  
 Push No. 3 to erase  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	12:16	34	13A28-13-5-10- KKF24C	.100	22		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Push No. 3 to erase,  
or power down to retain  
 Test Result  
 0001 03/06/14 12:21 .099  
 No. Date Time %BAC  
 Calibration 07/18/08 13:27 .100  
 03/06/14 12:22  
 AS4 Serial no: 018785  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	12:19	34	13A28-13-5-10- KKF24C	.100	21		.101			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														

03/06/14 12:25  
AS4 Serial no: 008546  
Version no: 4797  
No. Date Time %BAC  
Calibration 06/28/13 16:36 .100  
Test Result  
0001 02/22/14 00:59 .000  
0002 02/22/14 01:26 .016  
0003 02/22/14 01:29 .015  
0004 02/22/14 02:04 .000  
0005 02/22/14 02:05 .000  
0006 02/22/14 02:31 .127  
0007 02/22/14 02:35 .134  
0008 02/22/14 02:40 .037  
0009 02/22/14 04:45 .166  
0010 02/22/14 04:48 .179  
0011 02/23/14 01:15 .000  
0012 02/23/14 02:32 .028  
0013 02/23/14 02:49 .000  
0014 02/23/14 03:14 .102  
0015 02/23/14 03:17 .102  
0016 02/23/14 19:29 .119  
0017 02/23/14 19:31 .116  
0018 02/24/14 01:35 .067  
0019 02/24/14 01:40 .069  
0020 02/24/14 02:15 .197  
0021 02/24/14 02:17 .195  
0022 02/24/14 03:29 .087  
0023 02/24/14 03:31 .086  
0024 03/01/14 01:34 .017  
0025 03/01/14 01:51 .017  
0026 03/01/14 02:30 .148  
0027 03/01/14 02:34 .140  
0028 03/01/14 03:15 .055  
0029 03/01/14 03:18 .055  
0030 03/01/14 22:00 .067  
0031 03/01/14 22:04 .077  
0032 03/01/14 22:39 .006  
0033 03/01/14 23:32 .006  
0034 03/02/14 00:14 .062  
0035 03/02/14 00:41 .152  
0036 03/02/14 00:44 .119  
0037 03/02/14 00:46 .133  
0038 03/02/14 02:43 .035  
0039 03/02/14 03:53 .155  
0040 03/02/14 03:58 .212  
0041 03/02/14 04:00 .194  
0042 03/02/14 22:49 .013  
0043 03/02/14 23:24 .072  
0044 03/02/14 23:30 .074  
0045 03/06/14 12:23 .101

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

**#5**

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1230	39	13A28-13-5-10- KKF24C	.100	22		.103			X	Lee	17046	
			Prepared by: Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

03/06/14 12:35  
 ASA Serial no: 008559  
 Version no: 4797  
  
 No. Date Time ZBAC  
 Calibration  
 11/30/11 08:55 .100  
  
 Test Result  
 0001 02/20/14 10:27 .000  
 0002 02/25/14 10:20 .114  
 0003 02/25/14 10:26 .113  
 0004 03/06/14 12:35 .103

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008580
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	12:33	34	13A28-13-5-10-KKF24C	.100	22	23	.103			X	.100	Lee	17046	CALIBRATED
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>														
<p>03/06/14 12:42 MS4 Serial no: 008580 Version no: 4797</p> <p>No. Date Time %BAC Calibration 03/06/14 12:39 .100</p> <p>Test Result 0001 03/06/14 12:41 .100</p> <p>Push No. 3 to erase, or Power down to Retain</p>														
<p>03/06/14 12:37 MS4 Serial no: 008580 Version no: 4797</p> <p>No. Date Time %BAC Calibration 12/01/06 08:14 .099</p> <p>Test Result 0001 01/22/14 01:03 .118 0002 01/22/14 01:06 .115 0003 01/23/14 21:37 .101 0004 01/23/14 21:40 .114 0005 05/09/18 20:40 .113 0007 05/18/18 22:49 .103</p> <p>Push No. 3 to erase, or Power down to Retain</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 093032
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	NOT IN LOCKER
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. <b>008620</b>
--------------------------------	-----------------	------------------------------------	----------------------	--	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1242	34	13A28-13-5-10- KKF24C	.100	23	.099			X		Lee	17046	

Prepared by:  
Douglas C. Talbot  
01-28-13  
Expires on:  
01-28-15

No. Date Time %BAC  
 Calibration 01/30/13 14:34 .100  
  
 Test Result  
 0001 03/05/14 03:11 .000  
 0002 03/05/14 03:46 .047  
 0003 03/06/14 01:24 .148  
 0004 03/06/14 01:26 .139  
 0005 03/06/14 01:30 .098  
 0006 03/06/14 12:47 .099  
  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



### PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1245	34	13A28-13-5-10- KKF24C	.100	23		.102			X		Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13											
			Expires on: 01-28-15											

03/06/14 12:44  
 NS4 Serial no: 082957  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 02/17/14 08:52 .100  
 Test Result  
 0001 03/01/14 08:11 .154  
 0002 03/01/14 08:15 .154  
 0003 03/06/14 12:44 .102  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

~~#10~~ KZ #10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1258	34	13A28-13-5-10- KKF24C	.100	23		.099			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

03/06/14 12:57  
 AS4 Serial no: 077505  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 05/21/08 13:16 .100  
 Test Result  
 0001 03/06/14 12:57 .099  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1301	34	13A28-13-5-10- KKF24C	.100	22		.100			X		Lee	17046	

Prepared by:  
Douglas C. Talbot  
01-28-13  
Expires on:  
01-28-15

03/06/14 13:01  
 AS4 Serial no: 672479  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 02/17/14 09:10 .100  
 Test Result  
 0001 02/27/14 22:05 .099  
 0002 02/27/14 22:08 .083  
 0003 03/01/14 08:00 .178  
 0004 03/01/14 08:03 .185  
 0005 03/01/14 02:11 .124  
 0006 03/01/14 02:13 .132  
 0007 03/01/14 02:57 .122  
 0008 03/01/14 03:00 .117  
 0009 03/06/14 13:01 .100  
 Push No. 3 to erase,  
 or Power down to retain

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1304	34	13A28-13-5-10- KKF24C	.100	22		.101			X	Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13													
Expires on: 01-28-15													
Test Result 0001 02/27/14 21:35 .121 0002 02/27/14 21:37 .122 0003 03/01/14 09:15 .147 0004 03/01/14 09:19 .154 0005 03/02/14 02:24 .138 0006 03/02/14 02:26 .128 0007 03/06/14 13:03 .101 Calibration 02/17/14 09:16 .100 No. Date Time %BNC 03/06/14 13:04 AS4 Serial no: 082660 Version no: 4797 Push No. 3 to erase, or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. <u>008709</u>
--------------------------------	-----------------	------------------------------------	----------------------	--	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1307	34	13A28-13-5-10-KKF24C	.100	23	.100				X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result  
 0001 03/06/14 13:09 .100  
 Calibration  
 02/17/14 09:27 .100  
 No. Date Time %BAC  
 Version no: 4797  
 AS4 Serial no: 008709  
 03/06/14 13:09

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	13:10	34	13A28-13-5-10- KKF24C	.100	23		.103			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result  
 0001 03/06/14 13:11 .103  
 No. Date Time 20MC  
 Calibration 03/11/04 14:05 .100  
 AS4 Serial no: 008714  
 Version no: 4797  
 03/06/14 13:12

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

### PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1313	34	13A28-13-5-10- KKF24C	.100	23		.101			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 13:07  
 AS4 Serial no: 008888  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 03/25/13 10:45 .100  
 Test Result  
 0001 03/06/14 13:06 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1310	34	13A28-13-5-10- KKF24C	.100	22		.100			X		Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15											

Push No. 3 to erase,  
or power down to retain

Test Result  
0001 03/04/14 00:28 .095  
0002 03/04/14 00:30 .110  
0003 03/04/14 02:42 .145  
0004 03/04/14 02:45 .146  
0005 03/05/14 01:52 .031  
0006 03/05/14 01:54 .031  
0007 03/06/14 13:14 .100

No. Date Time %MC  
Calibration  
07/10/08 13:52 .100

03/06/14 13:15  
AS4 Serial no: 008783  
Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1313	34	13A28-13-5-10- KKF24C	.100	22		1					Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 13:17  
 AS4 Serial no: 082657  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 02/17/14 09:50 .100  
 Test Result  
 0001 03/01/14 23:39 .000  
 0002 03/06/14 13:16 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1316	34	13A28-13-5-10- KKF24C	.100	22		.098			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
No. Date Time %BAC Calibration 10/30/03 09:53 .100 Test Result 0001 03/06/14 13:20 .098 Push No. 3 to erase.														
03/06/14 13:20 AS4 Serial no: 008810 Version no: 4797														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	13A	34	13A28-13-5-10-KKF24C	.100	24		101			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p>														
<p>Test Result 0001 03/06/14 02:38 .123 0002 03/06/14 02:40 .123 0003 03/06/14 13:20 .101</p> <p>No. Date Time ZBAC Calibration 11/03/11 16:03 .100</p> <p>Push No. 3 to erase, or power down to retain</p> <p>03/06/14 13:20 AS4 Serial no: 018684 Version no: 4797</p> <p>or power down to retain</p>														

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1326	34	13A28-13-5-10- KKF24C	.100	23		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

03/06/14 13:29  
 NS4 Serial no: 008635  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration  
 10/24/09 09:04 .100  
 Test Result  
 0001 03/06/14 13:28 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1329	34	13A28-13-5-10-KKF24C	.100	23		105			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p>Test Result 0001 03/04/14 02:37 .139 0002 03/04/14 02:38 .131 0003 03/04/14 02:45 .068 0004 03/06/14 13:30 .105</p> <p>Calibration 12/24/03 11:01 .100</p> <p>No. Date Time ZHAC Version no: 4797</p> <p>03/06/14 13:30 AS4 Serial no: 008845 Version no: 4797</p> <p>Push No. 3 to erase, or power down to retain</p>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	12:32	34	13A28-13-5-10-KKF24C	.100	22		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

03/06/14 13:33  
AS4 Serial no: 082653  
Version no: 4797

No. Date Time ZBMC  
Calibration 02/17/14 10:08 .100

Test Result  
0001 03/01/14 01:21 .000  
0002 03/06/14 13:33 .101

Push No. 3 to erase  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1335	34	13A28-13-5-10- KKF24C	.100	23		.101				X	Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
Push No. 3 to erase Test Result 0001 03/06/14 13:38 .101 No. Date Time %BAC Calibration 01/18/03 09:25 .100 Version not 4797 03/06/14 13:39 AS4 Serial no: 008848 Version not 4797														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1330	34	13A28-13-5-10-KKF24C	.100	22		1099				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

```

03/06/14 13:39
854 Serial no: 077591
Version: not 4797

No. Date Time ZBAC
Calibration 05/21/08 15:09 .100

Test Result
0001 03/06/14 13:39 .099

Push No. 3 to erase,
or power down to retain

or power down to retain
  
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1341	34	13A28-13-5-10- KKF24C	.100	22		.101			X		Lee	17046	

Prepared by:  
Douglas C. Talbot  
01-28-13  
Expires on:  
01-28-15

Test Result  
0001 03/04/14 02:38 .061  
0002 03/06/14 13:45 .101  
No. Date Time ZBNC  
Calibration  
03/02/06 09:07 .100  
03/06/14 13:46  
AS4 Serial no: 008850  
Version no: 4797  
Push No. 3 to erase,  
or power down to retain

\*: If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\*: If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

### PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1344	34	13A28-13-5-10- KKF24C	.100	22		.100				X	Lee	17046	
														Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15
														Push No. 3 to erase, or power down to retain  Test Result 0001 03/06/14 13:42 .100 Calibration 03/25/13 14:00 .100 No. Date Time %BAC Version no: 4797 NS4 Serial no: 083051 03/06/14 13:43

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1347	34	13A28-13-5-10- KKF24C	.100	22		.105			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

```

03/06/14 13:49
MS4 Serial no: 008897
Version no: 4797

No. Date Time
Calibration 01/18/12 16:13 .100

Test Result
0001 03/06/14 13:48 .105

Push No. 3 to erase,
or power down to retain
  
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1350	34	13A28-13-5-10-KKF24C	.100	24		.96			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 13:48  
 ASA Serial no: 018624  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 18/07/08 09:19 .100  
 Test Result  
 0001 03/06/14 13:48 .096  
 Push No. 3 to erase,  
 or power down to retain.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

### PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1353	34	13A28-13-5-10- KKF24C	.100	24		.090			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
Push No. 3 to erase: Test Result 0001 03/06/14 13:51 .090 No. Date Time %BAC Calibration 07/29/13 13:32 .100 Version no: 4797 AS4 Serial no: 018646 03/06/14 13:52														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077498
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1356	34	13A28-13-5-10-KKF24C	.100	22		.101				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 13:53  
 AS4 Serial no: 077498  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 05/21/08 14:35 .100  
 Test Result  
 0001 03/06/14 13:52 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1359	34	13A28-13-5-10- KKF24C	.100	22		.104			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 13:59  
 NS4 Serial no: 008877  
 Version no: 4797

No. Date Time %BAC  
 Calibration 05/26/05 10:40 .100

Test Result  
 0001 03/06/14 13:58 .104

Push No. 3 to erase,  
 or Power down to retain

or Power down to retain

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1400	34	13A28-13-5-10- KKF24C	.100	23		.097			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test Result  
 0001 03/06/14 14:05 .097  
 No. Date Time ZBAC  
 Calibration 12/19/08 14:44 .100  
 03/06/14 14:06  
 HSA Serial no: 018615  
 Version no: 4797  
 Push No. 3 to erase,  
 or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1423	34	13A28-13-5-10-KKF24C	.100	23		101				X	Lee	17046	

Prepared by:  
Douglas C. Talbot  
01-28-13  
Expires on:  
01-28-15

Push No. 3 to erase, or power down to retain  
 Test Result  
 0001 03/06/14 14:05 .101  
 No. Date Time XMAC  
 Calibration 04/29/05 08:59 .101  
 Version no: 4797  
 NS4 Serial no: 018626  
 03/06/14 14:05

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1406	34	13A28-13-5-10- KKF24C	.100	23		.102			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 14:08  
 NS4 Serial no: 018696  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 09/09/13 14:50 .100  
 Test Result  
 0001 03/06/14 14:07 .102  
 Push No. 3 to erase.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1409	34	13A28-13-5-10- KKF24C	.100	23		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

03/06/14 14:09  
 HSA Serial no: 083030  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/29/09 10:39 .100  
 Test Result  
 0001 03/06/14 14:09 .099  
 Push No. 3 to erase,  
 or power down to retain  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1412	34	13A28-13-5-10-KKF24C	.100	23		.098			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result: .098  
 Calibration: 08/12/08 14:35 .100  
 No. Date Time ZMAC  
 Version no: 4797  
 AS4 Serial no: 018692  
 03/06/14 14:14

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1415	34	13A28-13-5-10- KKF24C	.100	22		.102			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>														
<p style="text-align: center;">             Test Result              0001 03/01/14 17:23 .169              0002 03/01/14 17:25 .165              0003 03/06/14 14:16 .102              Push No. 3 to erase.           </p>														
<p style="text-align: center;">             No. Date Time %BAC              Calibration              03/30/09 15:52 .100           </p>														
<p style="text-align: center;">             03/06/14 14:16              AS4 Serial no: 018762              Version no: 4797              or power down to retain           </p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

# 43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1418	34	13A28-13-5-10- KKF24C	.100	23		.097			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>														
<p>Test Result 0001 03/06/14 14:17 .097</p> <p>No. Date Time %DMC Calibration 03/12/12 12:30 .100</p> <p>03/06/14 14:17 AS4 Serial no: 018766 Version no: 4797</p> <p>Push No. 3 to erase, or Power down to retain</p> <p>or Power down to retain</p>														

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
03/06/2014	1424	34	13A28-13-5-10- KKF24C	.100	23		.104		X			Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			01-28-13												
			Expires on:												
			01-28-15												

Push No. 3 to erase,  
or power down to retain

03/06/14 14:21  
AS4 Serial no: 018691  
Version no: 4797

No. Date Time XBAC  
Calibration  
09/08/08 14:01 .100

Test Result  
0001 02/27/14 22:00 .240  
0002 02/27/14 22:02 .238  
0003 02/28/14 03:21 .000  
0004 03/01/14 02:34 .000  
0005 03/01/14 02:35 .000  
0006 03/02/14 01:10 .070  
0007 03/02/14 01:12 .066  
0008 03/02/14 01:49 .171  
0009 03/02/14 01:51 .165  
0010 03/02/14 04:53 .031  
0011 03/02/14 04:56 .030  
0012 03/06/14 14:21 .104

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	LD. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------------	---------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1424	34	13A28-13-5-10-KKF24C	.100	23		.107		X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15													
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">                     Test Result                      0001 02/27/14 22:42 .060                      0002 02/27/14 22:44 .061                      0003 02/27/14 22:47 .000                      0004 03/01/14 02:42 .139                      0005 03/01/14 02:44 .000                      0006 03/01/14 17:36 .000                      0007 03/02/14 03:57 .000                      0008 03/03/14 00:30 .243                      0009 03/03/14 00:33 .232                      0010 03/03/14 00:44 .094                      0011 03/03/14 00:45 .082                      0012 03/03/14 04:14 .191                      0013 03/06/14 14:21 .187                 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">                     Push No. 3 to erase, or Power down to retain                 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">                     No. Date Time ZHMC                      Calibration 11/26/03 11:10 .100                 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">                     03/06/14 14:22                      AS4 Serial no: 017990                      Version no: 4797                 </div> </div>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1427	34	13A28-13-5-10- KKF24C	.100	23		.099		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Test Result: 0001 03/06/14 14:23 .099  
 No. Date Time ZBAC  
 Calibration 01/13/09 10:25 .100  
 Push No. 3 to erase,  
 or power down to retain

03/06/14 14:24  
 AS4 Serial no: 018774  
 Version no: 4797

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	1430	34	13A28-13-5-10- KKF24C	.100	23		102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result  
 0001 03/06/14 14:32 .102  
 Calibration  
 03/06/14 13:56 .100  
 No. Date Time %BAC  
 03/06/14 14:32  
 ASA Serial no: 041486  
 Version no: 4797  
 Push No. 3 - to erase

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/06/2014	3:14:33	34	I3A28-13-5-10- KKF24C	.100	22		.099			X	Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15													
Test Result 0001 03/01/14 06:25 .100 0002 03/01/14 06:27 .097 0003 03/06/14 14:32 .099													
Push No. 3 to erase, or power down to retain													
No. Date Time ZBAC Calibration 02/20/14 11:28 .100													
03/06/14 14:32 AS4 Serial no: 041495 Version no: 4797													
or power down to retain													

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082654
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014			13A28-13-5-10- KKF24C	.100								Lee	17046	OUT for SERVICE
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1436	34	13A28-13-5-10-KKF24C	.100	23		096			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
Test Result 0001 02/28/14 00:00 .203 0002 02/28/14 00:03 .215 0003 02/28/14 23:33 .000 0004 02/28/14 23:34 .000 0005 02/28/14 02:50 -.175 0006 03/02/14 02:54 .176 0007 03/03/14 04:21 .169 0008 03/03/14 04:23 .180 0009 03/06/14 14:36 .096 Push No. 3 to erase or power down to retain														
Calibration 05/05/09 08:03 .100 No. Date Time %BAC 03/06/14 14:36 AS4 Serial no: 041502 Version no: 4797														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

SGT.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 009874
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/06/2014	1439	34	13A28-13-5-10-KKF24C	.100	22		.102			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p>														
<p>Test Result                      0001 03/04/14 00:21 .000                      0002 03/04/14 02:28 .000                      0003 03/04/14 02:49 .061                      0004 03/04/14 02:50 .065                      0005 03/05/14 23:38 .080                      0006 03/05/14 23:34 .074                      0007 03/06/14 14:39 .102</p> <p>No. Date Time ZMAC                      Calibration 06/25/13 10:32 .107</p> <p>03/06/14 14:39                      RS4 Serial no: 0008974                      Version no: 4797</p> <p>Push No. 3 to erase,                      or power down to retain</p>														

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.