STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8501

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1610	34	14F27-09-5-10-	.100	20		.093			X		Flick	15063	Cook Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 15:11
AS4 Serial no: 008501
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/30/09	15:40	.100

Test Result

0001	03/03/15	02:21	.055
0002	03/03/15	02:23	.055
0003	03/04/15	01:59	.071
0004	03/04/15	02:01	.069
0005	03/04/15	04:27	.000
0006	03/11/15	00:27	.222
0007	03/11/15	00:32	.218
0008	03/12/15	01:17	.113
0009	03/12/15	01:20	.111
0010	03/12/15	22:06	.137
0011	03/12/15	22:09	.123
0012	03/16/15	15:10	.093

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

2

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18785

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1346	34	14F27-09-5-10-C1307252	.100	24		.095			X		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 12:46
ASA Serial no: 018785
Version no: 4797

No.  Date     Time  %BAC
Calibration
     07/18/08 13:27 .100

Test Result
0001 03/16/15 12:46 .095

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

3

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8546

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1607	34	14F27-09-5-10-	.100	22		.097			Y		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 15:08
AS4 Serial no: 008546
Version no: 4797

No.	Date	Time	%BAC
Calibration	06/28/13	16:36	.100

Test	Result		
0001	02/27/15	22:36	.011
0002	02/28/15	04:14	.178
0003	02/28/15	04:17	.166
0004	03/01/15	01:55	.081
0005	03/01/15	01:57	.078
0006	03/01/15	02:00	.000
0007	03/01/15	03:45	.123
0008	03/01/15	03:50	.127
0009	03/07/15	23:50	.131
0010	03/07/15	23:53	.124
0011	03/08/15	02:54	.080
0012	03/08/15	02:57	.088
0013	03/15/15	03:18	.090
0014	03/15/15	03:20	.095
0015	03/16/15	15:03	.097

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	82059

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1605	34	14F27-09-5-10-	.100	21		.095			P		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 15:06
AS4 Serial no: 082659
Version  no: 4797

No.  Date     Time  %BAC
Calibration
     03/28/14 09:04  .100

Test Result
0001 02/28/15 02:27  .042
0002 02/28/15 02:29  .043
0003 02/28/15 03:30  .095
0004 02/28/15 03:32  .093
0005 03/01/15 00:20  .114
0006 03/01/15 00:22  .106
0007 03/07/15 02:06  .109
0008 03/07/15 02:09  .110
0009 03/08/15 02:08  .118
0010 03/08/15 02:11  .112
0011 03/08/15 18:07  .190
0012 03/13/15 02:08  .000
0013 03/13/15 02:45  .000
0014 03/14/15 01:37  .098
0015 03/14/15 01:40  .102
0016 03/14/15 03:38  .143
0017 03/14/15 03:40  .144
0018 03/14/15 23:35  .013
0019 03/16/15 01:07  .075
0020 03/16/15 01:10  .069
0021 03/16/15 15:05  .095

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

5

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1553	34	14F27-09-5-10-	.100	21		.099			X		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:56
ASA Serial no: 008559
Version no: 4797

No. Date Time %BAC
Calibration
11/30/11 08:55 .100

Test Result
0001 03/07/15 23:30 .037
0002 03/07/15 23:31 .035
0003 03/16/15 14:56 .099

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

6

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8580

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1600	34	I4F27-09-5-10-	.100	21		.097			ρ		Flick	15063	CLOCK CHANGE
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 15:00
AS4 Serial no: 008580
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/06/14	12:39	.100

Test Result

0001	03/04/15	01:13	.032
0002	03/04/15	01:16	.031
0003	03/04/15	03:21	.247
0004	03/04/15	03:23	.256
0005	03/10/15	23:50	.182
0006	03/10/15	23:52	.148
0007	03/16/15	15:00	.097

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

7

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	83032

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1544	34	14F27-09-5-10-	.100	21		.097			Y		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 14:45
ASA Serial no: 083032
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 09:01 .100

Test Result
0001 03/08/15 14:36 .279
0002 03/09/15 14:38 .268
0003 03/14/15 06:18 .166
0004 03/14/15 06:20 .168
0005 03/14/15 15:29 .256
0006 03/14/15 15:30 .260
0007 03/16/15 14:45 .097

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

8

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8620

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1548	34	14F27-09-5-10-	.100	21		.097			✓		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:50
AS4 Serial no: 008620
Version no: 4797

No.	Date	Time	%BAC
Calibration	01/30/13	14:34	.100

Test Result

0001	03/03/15	01:14	.093
0002	03/03/15	01:15	.097
0003	03/03/15	20:01	.013
0004	03/03/15	20:38	.000
0005	03/06/15	00:02	.055
0006	03/06/15	00:04	.055
0007	03/11/15	00:01	.000
0008	03/11/15	00:48	.029
0009	03/16/15	14:49	.097

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

9

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	82957

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1543	34	14F27-09-5-10-	.100	21		.099			X		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:44
AS4 Serial no: 082957
Version no: 4797
No. Date Time %BAC
Calibration
02/17/14 08:52 .100
Test Result
0001 03/16/15 14:43 .099
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

10

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	77505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1538	34	14F27-09-5-10-	.100	21		.090			X		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:39
AS4 Serial no: 077505
Version no: 4797

No. Date Time %BAC
Calibration
09/18/14 07:21 .099

Test Result
0001 03/16/15 14:39 .096

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

11

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	77479

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1531	34	14F27-09-5-10-	.100	80		.096			X		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 14:31
AS4 Serial no: 077479
Version no: 4797

No.	Date	Time	%BAC
Calibration	02/17/14	09:10	.100

Test Result

0001	02/28/15	20:55	.170
0002	02/28/15	20:58	.157
0003	02/28/15	21:43	.202
0004	02/28/15	21:45	.203
0005	03/01/15	02:40	.175
0006	03/01/15	02:42	.175
0007	03/07/15	02:01	.129
0008	03/07/15	02:03	.120
0009	03/07/15	02:08	.000
0010	03/16/15	14:31	.096

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

12

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	82660

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015			14F27-09-5-10-	.100		Not in PAS Case						Flick	15063	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

14

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1526	34	14F27-09-5-10-	.100	21		.102			O		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:27
AS4 Serial no: 008714
Version nos: 4797

No.	Date	Time	%BAC
Calibration	03/11/04	14:05	.100

Test Result

0001 03/09/15 00:55 .129
0002 03/09/15 00:57 .125
0003 03/13/15 20:29 .166
0004 03/13/15 20:31 .186
0005 03/14/15 02:24 .007
0006 03/14/15 02:26 .000
0007 03/15/15 00:17 .181
0008 03/15/15 00:19 .176
0009 03/15/15 02:49 .090
0010 03/15/15 02:52 .090
0011 03/16/15 14:26 .102

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

13

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1535	34	14F27-09-5-10-	.100	21		.099			X		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 14:36
AS4 Serial no: 008709
Version  no: 4797

No.  Date     Time  %BAC
Calibration
     02/17/14 09:27 .100

Test Result
0001 03/02/15 04:56 .079
0002 03/02/15 04:59 .091
0003 03/10/15 23:22 .008
0004 03/10/15 23:25 .009
0005 03/11/15 01:21 .199
0006 03/11/15 01:23 .197
0007 03/12/15 20:17 .000
0008 03/13/15 00:17 .119
0009 03/13/15 00:19 .123
0010 03/15/15 22:28 .219
0011 03/15/15 22:31 .214
0012 03/16/15 14:35 .099

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

16

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8888

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1510	34	14F27-09-5-10-C1307252	.100	21		.097			✓		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 14:11
AS4 Serial no: 008888
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/25/13	10:45	.100

Test Result

0001	03/05/15	00:04	.286
0002	03/05/15	00:07	.296
0003	03/10/15	01:07	.063
0004	03/10/15	01:09	.063
0005	03/16/15	14:10	.097

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

18

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	82657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1517	34	14F27-09-5-10-	.100	21		.098			P		Flick	15063	Clear Intake
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 14:18
ASA Serial no: 082657
Version   no: 4797

No.   Date    Time  %BAC
Calibration
      02/17/14 09:50  .100

Test Result
0001 03/16/15 14:17  .098

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

17

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8783

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1524	34	14F27-09-5-10-C1307252	.100	20		.099			X		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 14:25
AS4 Serial no: 008783
Version    no: 4797

No.  Date     Time  %BAC
Calibration
     07/18/08 13:52 .100

Test Result
0001 02/28/15 02:03 .047
0002 02/28/15 02:06 .116
0003 02/28/15 02:08 .109
0004 03/01/15 01:11 .000
0005 03/01/15 02:37 .110
0006 03/01/15 02:38 .107
0007 03/07/15 03:33 .125
0008 03/14/15 20:33 .015
0009 03/14/15 20:34 .017
0010 03/15/15 00:00 .178
0011 03/15/15 00:02 .180
0012 03/15/15 22:38 .146
0013 03/15/15 22:41 .142
0014 03/15/15 23:17 .067
0015 03/15/15 23:19 .063
0016 03/16/15 02:32 .104
0017 03/16/15 02:33 .097
0018 03/16/15 14:25 .099

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

19

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8810

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015			14F27-09-5-10-	.100								Flick	15063	
			C1307252		Not in Locker									
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

20

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18084

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1517	34	14F27-09-5-10-	.100	21		.103			P		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:17
ASA Serial no: 018684
Version no: 4797

No.	Date	Time	%BAC
Calibration	11/03/11	16:03	.100

Test Result
0001 03/16/15 14:17 .103

Push No. 3 to erase
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1521	34	14F27-09-5-10-	.100	20		.101			P		Flick	15063	Clock Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 14:22
ASA Serial no: 008635
Version no: 4797

No.	Date	Time	%BAC
Calibration	10/24/03	09:04	.100

Test Result

0001	03/03/15	01:22	.106
0002	03/03/15	01:24	.116
0003	03/03/15	01:27	.000
0004	03/03/15	02:45	.069
0005	03/05/15	23:15	.208
0006	03/05/15	23:18	.210
0007	03/16/15	14:22	.101

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8845

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1514	34	14F27-09-5-10- C1307252	.100	20		.104			P		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 14:15
AS4 Serial no: 008845
Version  no: 4797

No.  Date   Time  %BAC
Calibration
12/24/03 11:01  .100

Test Result
0001 03/03/15 01:13  .095
0002 03/03/15 18:45  .000
0003 03/16/15 14:14  .104

Push No. 3 to erase
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

23

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	822653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1504	34	14F27-09-5-10-	.100	20		.100			✓		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 14:05
AS4 Serial no: 002653
Version no: 4797

No. Date Time %BAC
Calibration 02/17/14 10:08 .100

Test Result
0001 03/07/15 01:22 .167
0002 03/07/15 01:25 .156
0003 03/07/15 02:58 .068
0004 03/08/15 20:43 .175
0005 03/08/15 20:46 .107
0006 03/14/15 12:15 .033
0007 03/14/15 12:15 .000
0008 03/15/15 06:58 .000
0009 03/15/15 09:10 .028
0010 03/16/15 14:04 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1510	34	14F27-09-5-10- C1307252	.100	20		.102			P		Flick	15063	Check Change
			Prepared by: Douglas C. Talbot on 06-27-14. Expires on 06-27-16											

03/16/15 14:10
AS4 Serial no: 008848
Version no: 4797

No. Date Time %BAC
Calibration
01/13/03 09:25 .100

Test Result
0001 03/16/15 14:10 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

25

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	77591

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1504	34	14F27-09-5-10-C1307252	.100	19		.100			X		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 14:05
AS4 Serial no: 077591
Version no: 4797
No. Date Time %BAC
Calibration
05/21/08 15:09 .100
Test Result
0001 03/16/15 14:04 .100
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

26

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8850

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015			14F27-09-5-10-	.100		Not in Cruiser						Flick	15063	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

27

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	77573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1444	34	14F27-09-5-10-	.100	21		.096			✓		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:45
ASA Serial no: 077573
Version no: 4797

No. Date Time %BAC
Calibration
04/09/14 08:36 .080

Test Result
0001 03/03/15 01:48 .080
0002 03/03/15 01:50 .081
0003 03/09/15 22:23 .192
0004 03/09/15 22:26 .191
0005 03/11/15 01:05 .075
0006 03/11/15 01:07 .081
0007 03/11/15 01:28 .122
0008 03/11/15 01:30 .124
0009 03/16/15 13:44 .096

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	83051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1442	34	14F27-09-5-10-	.100	21		.098			0		Flick	15063	Clock Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:43
ASA Serial no: 083051
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/25/13	14:00	.100

Test	Result		
0001	02/27/15	19:03	.089
0002	02/27/15	19:05	.090
0003	02/28/15	19:02	.090
0004	02/28/15	19:03	.238
0005	02/28/15	19:05	.219
0006	03/16/15	13:43	.098

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

30

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8897

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1452	34	14F27-09-5-10-C1307252	.100	20		.106			✓		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 13:53
AS4 Serial no: 008897
Version   no: 4797

No.  Date     Time  %BAC
Calibration
     01/18/12 16:13 .100

Test Result
0001 03/01/15 03:01 .017
0002 03/01/15 04:38 .140
0003 03/01/15 04:41 .154
0004 03/08/15 05:12 .165
0005 03/08/15 05:15 .150
0006 03/09/15 00:47 .066
0007 03/09/15 01:09 .133
0008 03/09/15 01:12 .105
0009 03/09/15 01:14 .118
0010 03/13/15 21:47 .243
0011 03/13/15 21:51 .238
0012 03/15/15 02:06 .260
0013 03/15/15 02:08 .247
0014 03/15/15 02:11 .243
0015 03/15/15 02:55 .099
0016 03/15/15 02:58 .070
0017 03/15/15 03:01 .109
0018 03/16/15 13:53 .106

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

31

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1443	34	14F27-09-5-10-C1307252	.100	22		.094			Ø		Flick	15063	Clock Change

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 13:43
AS4 Serial no: 018624
Version no: 4797

No.	Date	Time	%BAC
Calibration	10/07/08	09:19	.100

Test Result

0001	03/04/15	00:48	.201
0002	03/04/15	00:49	.199
0003	03/10/15	23:44	.210
0004	03/10/15	23:48	.205
0005	03/12/15	02:42	.150
0006	03/12/15	02:45	.146
0007	03/16/15	13:43	.094

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

32

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18646

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1432	34	14F27-09-5-10- C1307252	.100	22		.098			0		Flick	15063	Change Clock
			Prepared by: Douglas C. Talbot on 06-27-14. Expires on 06-27-16											

03/16/15 13:32
AS4 Serial no: 018646
Version no: 4797

No.	Date	Time	%BAC
Calibration	07/29/13	13:32	.100

Test Result
0001 03/14/15 04:48 .133
0002 03/14/15 04:51 .137
0003 03/16/15 13:32 .098

Push No. 3 to erase.
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

33

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	77488

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1471	34	14F27-09-5-10-	.100	21		.102			✓		Flick	15063	Change Cable
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:21
AS4 Serial no: 077488
Version no: 4797

No. Date Time %BAC
Calibration
05/21/08 14:35 .100

Test Result
0001 03/16/15 13:21 .102

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

34

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1431	34	I4F27-09-5-10-	.100	21		.103			X		Flick	15063	CHANGE CLOCK
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:32
AS4 Serial no: 008877
Version no: 4797

No. Date Time %BAC
Calibration
05/26/05 10:40 .100

Test Result
0001 03/16/15 13:31 .103

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

36

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18615

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1422	34	14F27-09-5-10-	.100	27		.100			P		Flick	15063	Changed Cords
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 13:23
ASA Serial no: 018615
Version    no: 4797

No.   Date     Time   %BAC
Calibration
      12/19/08 14:44  .100

Test Result
0001 03/07/15 01:28  .030
0002 03/07/15 01:29  .029
0003 03/11/15 19:29  .148
0004 03/11/15 19:31  .149
0005 03/12/15 00:15  .173
0006 03/12/15 00:18  .139
0007 03/12/15 00:20  .162
0008 03/13/15 01:33  .134
0009 03/13/15 01:34  .125
0010 03/13/15 01:37  .145
0011 03/16/15 13:23  .100

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

37

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18626

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1414	34	14F27-09-5-10-	.100	22		.099			P		Flick	15063	Cuck Change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:15
AS4 Serial no: 018626
Version no: 4797

No.	Date	Time	%BAC
Calibration	04/29/05	08:59	.101

Test Result

0001	03/08/15	01:35	.064
0002	03/08/15	01:37	.063
0003	03/08/15	22:53	.023
0004	03/08/15	22:55	.024
0005	03/09/15	02:13	.209
0006	03/09/15	02:15	.202
0007	03/16/15	13:15	.099

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

38

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18696

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/16/2015			14F27-09-5-10-	.100	35.0							Flick	15063	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

39

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	83030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1411	34	14F27-09-5-10-	.100	21		.101			10		Flick	15063	Cook Change
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 13:12
AS4 Serial no: 083030
Version   no: 4797

No.  Date     Time  %BAC
Calibration
     01/29/09 10:39 .100

Test Result
0001 02/28/15 00:29 .167
0002 02/28/15 00:32 .169
0003 03/01/15 19:58 .192
0004 03/01/15 20:02 .129
0005 03/01/15 20:05 .186
0006 03/01/15 21:16 .091
0007 03/01/15 21:19 .096
0008 03/07/15 02:43 .094
0009 03/07/15 02:45 .092
0010 03/07/15 21:32 .203
0011 03/07/15 21:35 .211
0012 03/13/15 02:10 .152
0013 03/13/15 21:22 .000
0014 03/13/15 21:25 .000
0015 03/16/15 13:11 .101

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

4/6

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18692

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1412	34	14F27-09-5-10-	.100	21		.110			X		Flick	15063	CHANGE CLOCK
			C1307252											

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 13:14
AS4 Serial no: 018692
Version    no: 4797

No.   Date     Time   %BAC
Calibration
      08/12/08 14:35  .100

Test Result
0001 02/28/15 22:25  .000
0002 03/07/15 18:45  .006
0003 03/07/15 18:46  .005
0004 03/16/15 13:13  .100

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

4/2

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18762

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1409	34	14F27-09-5-10-	.100	21		.102			P		Flick	15063	Change Clock
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:10
ASA Serial no: 018762
Version no: 4797

No. Date Time %BAC
Calibration
03/30/09 15:52 .100

Test Result
0001 03/16/15 13:10 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

43

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	187669

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1406	34	14F27-09-5-10-	.100	21		.101			X		Flick	15063	CHANGE Clock
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:06
ASA Serial no: 018766
Version no: 4797

No. Date Time %BAC
Calibration
09/12/14 13:45 .100

Test Result
0001 03/08/15 04:25 .130
0002 03/08/15 04:28 .172
0003 03/08/15 04:32 .143
0004 03/16/15 13:06 .101

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

4/4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	14/08	34	14F27-09-5-10-	.100	21		.103			P		Flick	15063	CHANGE TIME
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:09
AS4 Serial no: 018691
Version no: 4797

No. Date Time %BAC
Calibration
09/08/08 14:01 .100

Test Result
0001 03/02/15 04:21 .073
0002 03/02/15 04:23 .072
0003 03/06/15 21:56 .218
0004 03/06/15 21:59 .191
0005 03/08/15 22:24 .160
0006 03/08/15 22:26 .145
0007 03/15/15 01:24 .222
0008 03/15/15 01:28 .178
0009 03/16/15 13:09 .103

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

45

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	17990

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1402	34	14F27-09-5-10-	.100	21		.107			X		Flick	15063	CHANGE TIME
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 13:03
ASA Serial no: 017990
Version no: 4797

No.	Date	Time	%BAC
Calibration	11/26/03	11:10	.100
Test Result			
0001	03/02/15	20:20	.037
0002	03/02/15	20:21	.035
0003	03/02/15	21:01	.000
0004	03/02/15	21:31	.176
0005	03/02/15	21:34	.164
0006	03/03/15	22:57	.000
0007	03/03/15	23:12	.103
0008	03/03/15	23:15	.103
0009	03/04/15	02:42	.112
0010	03/04/15	02:44	.124
0011	03/04/15	20:29	.000
0012	03/04/15	21:18	.000
0013	03/04/15	23:41	.041
0014	03/05/15	01:01	.057
0015	03/05/15	01:03	.059
0016	03/05/15	20:44	.150
0017	03/05/15	20:47	.149
0018	03/05/15	23:08	.010
0019	03/10/15	01:14	.047
0020	03/10/15	01:18	.042
0021	03/10/15	21:00	.150
0022	03/10/15	21:02	.146
0023	03/11/15	01:48	.000
0024	03/11/15	03:15	.204
0025	03/11/15	03:28	.211
0026	03/11/15	03:31	.201
0027	03/11/15	19:07	.000
0028	03/11/15	21:19	.104
0029	03/11/15	21:23	.114
0030	03/12/15	01:03	.128
0031	03/12/15	01:05	.125
0032	03/13/15	00:01	.073
0033	03/13/15	00:04	.060
0034	03/13/15	00:08	.065
0035	03/13/15	00:45	.000
0036	03/13/15	01:17	.011
0037	03/13/15	01:54	.111
0038	03/13/15	01:56	.106
0039	03/16/15	13:02	.107

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

46

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	18774

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1356	24	14F27-09-5-10-	.100	21°		.101			P		Flick	15063	CHANGE TIME
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 12:59
ASA Serial no: 018774
Version no: 4797

No. Date Time %BAC
Calibration
01/13/09 10:25 .100

Test Result
0001 03/03/15 02:59 .159
0002 03/03/15 03:03 .153
0003 03/04/15 02:30 .160
0004 03/04/15 23:34 .026
0005 03/04/15 23:35 .028
0006 03/05/15 04:12 .061
0007 03/10/15 01:32 .178
0008 03/10/15 01:35 .178
0009 03/11/15 01:16 .149
0010 03/11/15 01:19 .145
0011 03/12/15 01:38 .215
0012 03/12/15 01:40 .197
0013 03/16/15 12:57 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

50

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	41486

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1355	34	14F27-09-5-10-	.100	21		.104			X		Flick	15063	CHANGE TIME
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 12:57
AS4 Serial no: 041486
Version no: 4797

No. Date Time %BAC
Calibration
03/08/06 13:56 .100

Test Result
0001 03/16/15 12:56 .104

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

51

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	4/1/95

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1345	34	14F27-09-5-10-	.100	20		.099			X		Flick	15063	CHANGE TIME
		/	C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 12:48
AS4 Serial no: 041495
Version no: 4797
No. Date Time %BAC
Calibration
02/20/14 11:28 .100

Test Result
0001 03/16/15 12:46 .099

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

52

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	41502

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C) TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	TEST #2 (Optional)	CALIBRATION REQUIRED YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1353	34	14F27-09-5-10-C1307252	.100	21		.096			N		Flick	15063	CHANGE TIME

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

```
03/16/15 12:54
AS4 Serial no: 041502
Version    no: 4797

No.   Date    Time  %BAC
Calibration
      05/05/09 08:03 .100

Test Result
0001 02/28/15 02:35 .154
0002 02/28/15 02:37 .198
0003 02/28/15 02:40 .211
0004 03/01/15 00:00 .198
0005 03/01/15 00:03 .194
0006 03/01/15 02:20 .198
0007 03/01/15 02:23 .156
0008 03/01/15 02:26 .187
0009 03/07/15 02:04 .109
0010 03/07/15 02:06 .094
0011 03/07/15 21:25 .000
0012 03/07/15 21:57 .124
0013 03/07/15 22:00 .122
0014 03/08/15 01:46 .247
0015 03/08/15 01:49 .220
0016 03/08/15 01:52 .238
0017 03/08/15 02:51 .092
0018 03/08/15 02:54 .095
0019 03/14/15 21:45 .000
0020 03/15/15 02:21 .154
0021 03/15/15 02:23 .158
0022 03/16/15 12:54 .096

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

53

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	822654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1346	34 /	14F27-09-5-10- C1307252	.100	.21		.098			X		Flick	15063	CHANGE TIMES

Prepared by:
Douglas C. Talbot
on 06-27-14.
Expires on 06-27-16

03/16/15 12:47
AS4 Serial no: 082654
Version no: 4797

No. Date Time %BAC
Calibration
03/28/14 11:36 .100

Test Result
0001 03/13/15 17:15 .051
0002 03/16/15 12:47 .098

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

Sgt's PAS

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alcosensor IV	8874

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
03/16/2015	1614	34	14F27-09-5-10-	.100	21		.099			X		Flick	15063	Clock change
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			on 06-27-14.											
			Expires on 06-27-16											

03/16/15 15:15
ASA Serial no: 008874
Version no: 4797

No. Date Time %BAC
Calibration
03/28/14 11:40 .100

Test Result
0001 02/27/15 17:13 .044
0002 02/27/15 17:17 .043
0003 03/06/15 02:42 .081
0004 03/16/15 15:15 .099

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.