

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

SGT.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8874
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	1936	34	13A28-13-5-10 KKF24C	.100	22		.104			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 19:37  
 NS4 Serial no: 000874  
 Version no: 4797  
 No. Date Time NBAC  
 Calibration 06/25/13 10:32 .107  
 Test Result  
 0001 03/19/14 19:37 .104

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8501
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	1939	34	13A28-13-5-10 KKF24C	.100	21		0.98			Y	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 19:40  
 AS4 Serial no: 008501  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 03/30/09 15:40 .100  
 Test Result  
 0001 03/19/14 19:39 .098

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18785
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	1942	34	13A28-13-5-10 KKF24C	.100	21		.100			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

No. Date Time ZBAC  
 Calibration 07/18/08 13:27 .100  
 Test Result 0001 03/19/14 19:42 .100  
 03/19/14 19:43  
 ASA Serial no: 018785  
 Version no: 4797

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8546
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	1944	34	13A28-13-5-10	.100	22		103			X	Valadez	16071	
			KKF24C										
			Prepared by										
			Douglas C. Talbot										
			on 01-28-13										
			Expires 01-28-15										

Test Result  
 0001 03/19/14 19:45 .103  
 No. Date Time %BAC  
 Calibration 06/28/13 16:36 .100  
 Version no: 4797  
 NS4 Serial no: 000546  
 03/19/14 19:45

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#4  
82659

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82659
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2014	1844	34	13A28-13-5-10 KKF24C	.100	21		.098			X		Valadez	16071	TIME CHANGE
			Prepared by Douglas C. Talbot on 01-28-13  Expires 01-28-15											
			<pre> Test Result 0001 03/05/14 14:48 .099 0002 03/06/14 05:46 .099 0003 03/06/14 08:06 .100 0004 03/06/14 12:14 .097 0005 03/06/14 13:17 .098 0006 03/06/14 13:19 .097 0007 03/06/14 13:22 .097 0008 03/06/14 13:26 .000 0009 03/06/14 13:32 .000 0010 03/08/14 08:29 .095 0011 03/08/14 08:31 .097 0012 03/08/14 08:31 .097 0013 03/08/14 08:32 .000 0014 03/19/14 10:49 .098                     </pre>											
			<pre> No. Date Time ZBAC Calibration 03/06/14 13:09 .098                     </pre>											
			<pre> 03/19/14 18:50 AS4 Serial no: 002659 Version no: 4797                     </pre>											

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8580
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2014	34	13A28-13-5-10 KKF24C	.100	21		.102			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

```

03/19/14 20:17
AS4 Serial no: 000580
Version no: 4797
No. Date Time %BAC
Calibration 03/06/14 12:39 .106
Test Result
0001 03/15/14 23:28 .127
0002 03/15/14 23:31 .123
0003 03/16/14 01:46 .170
0004 03/16/14 01:50 .190
0005 03/16/14 03:26 .187
0006 03/16/14 03:29 .188
0007 03/16/14 22:40 .205
0008 03/16/14 22:43 .215
0009 03/19/14 20:16 .102
    
```

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83032
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	1954	34	13A28-13-5-10	.100	22		.101			X	Valadez	16071	
			KKF24C										
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 19:55  
 AS4 Serial no: 083032  
 Version no: 4797  
 No. Date Time ZDHC  
 Calibration 02/19/14 09:01 .100  
 Test Result  
 0001 03/16/14 02:25 .220  
 0002 03/16/14 02:27 .223  
 0003 03/17/14 08:37 .068  
 0004 03/17/14 08:39 .068  
 0005 03/19/14 19:54 .101

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

HP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8620
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
3/19/2014	1956	34	13A28-13-5-10 KKF24C	.100	21	.101			X		Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 19:50  
 164 Serial no: 008620  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/30/13 14:34 .108  
 Test Result  
 0001 03/13/14 23:08 .125  
 0002 03/13/14 23:10 .129  
 0003 03/14/14 01:14 .122  
 0004 03/14/14 01:16 .113  
 0005 03/17/14 23:49 .066  
 0006 03/17/14 23:54 .075  
 0007 03/18/14 01:40 .060  
 0008 03/18/14 01:41 .061  
 0009 03/18/14 02:36 .061  
 0010 03/18/14 18:44 .282  
 0011 03/19/14 18:57 .101

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82957
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	1959	34	13A28-13-5-10 KKF24C	.100	22		103			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:06  
 H24 Serial no: 082957  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 02/17/14 09:52 .100  
 Test Result  
 #001 03/19/14 20:00 .103

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77505
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2014	2003	34	13A28-13-5-10 KKF24C	.100	21		.099		X		Valadez	16071		
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											

No. Date Time %BAC  
 Calibration 05/21/08 13:16 .105  
 Test Result 0001 03/19/14 20:03 .099  
 03/19/14 20:04  
 MS4 Serial No: 072505  
 Version No: 4797

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

HP 202K (Rev. 8-08) OPI 051

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77479
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2007 34	34	13A28-13-5-10	.100	21		.100				Valadez	16071	
			KKF24C										
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:07  
 AC4 Serial no: 077479  
 Version no: 4797

No. Date Time %BMC  
 Calibration 02/17/14 09:10 .100

Test Result  
 0001 03/16/14 01:35 .077  
 0002 03/16/14 01:37 .073  
 0003 03/16/14 03:26 .200  
 0004 03/16/14 03:29 .166  
 0005 03/16/14 03:31 .197  
 0006 03/17/14 02:08 .170  
 0007 03/17/14 02:10 .190  
 0008 03/17/14 02:12 .173  
 0009 03/19/14 20:07 .100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 † If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#12

AREA NAME *	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	82660

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
3/19/2014	2009	34	13A28-13-5-10	.100	21		.102				X	Valadez	16071	
			KKF24C											
			Prepared by											
			Douglas C. Talbot											
			on 01-28-13											
			Expires 01-28-15											

03/19/14 20:10  
 AS4 Serial no: 082660  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 02/17/14 09:16 .100  
 Test Result  
 0001 03/13/14 23:25 .014  
 0002 03/14/14 04:10 .259  
 0003 03/14/14 04:13 .242  
 0004 03/16/14 01:19 .048  
 0005 03/16/14 01:22 .051  
 0006 03/19/14 20:09 .102

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8709
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
3/19/2014	2012	34	13A28-13-5-10 KKF24C	.100	22		.102				X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											

02/19/14 20:13  
 AS4 Serial no: 000209  
 Version no: 4797  
 No. Date Time 2BAC  
 Calibration 02/17/14 09:27 .100  
 Test Result 0001 02/19/14 20:13 .100

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8714
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2014 34		13A28-13-5-10 KKF24C	.100	21		.105			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										
				No. Date Time ZMAC Calibration 03/11/04 14:05 .106 Test Result 0001 03/19/14 20:15 .105 03/19/14 20:15 HSA Serial no: 000714 Version no: 4797									

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8888
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2019	34	13A28-13-5-10 KKF24C	.100	22		.100			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:20  
 ASA Serial no: 000008  
 Version no: 4797  
 No. Date Time ZBMC  
 Calibration 03/25/13 10:45 .100  
 Test Result  
 0001 03/19/14 20:19 .100

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#18

EA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82657
------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2024	34	13A28-13-5-10 KKF24C	.100	22		.101			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:22  
 PAS Serial no: 082657  
 Version no: 4792  
 No. Date Time %BAC  
 Calibration 02/17/14 09:50 .100  
 Test Result  
 0001 03/19/14 02:34 .120  
 0002 03/19/14 02:37 .119  
 0003 03/16/14 08:09 .112  
 0004 03/16/14 08:04 .110  
 0005 03/16/14 04:43 .115  
 0006 03/16/14 04:45 .117  
 0007 03/19/14 20:21 .101

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8810
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2014	1930	34	13A28-13-5-10	.100	21		.097				K	Valadez	16071	
			KKF24C											
			Prepared by											
			Douglas C. Talbot											
			on 01-28-13											
			Expires 01-28-15											

03/19/14 19:32  
MS4 Serial no: 008810  
Version no: 4797

No. Date Time ZBAC  
Calibration 10/30/03 09:53 .100

Test Result  
0001 03/19/14 19:29 .000  
0002 03/19/14 19:31 .097

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

HP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8845
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2628	34	13A28-13-5-10 KKF24C	.100	22		.105			X	Valadez	16071	
Prepared by Douglas C. Talbot on 01-28-13 Expires 01-28-15													
<pre> 03/19/14 20:29 AS4 Serial no: 008845 Version no: 4797  No. Date Time %BAC Calibration 12/24/03 11:01 .100  Test Result 0001 03/18/14 00:14 .000 0002 03/18/14 01:10 .000 0003 03/18/14 01:59 .000 0004 03/18/14 02:29 .000 0005 03/19/14 20:28 .105                     </pre>													

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

IP 202K (Rev. 8-08) OPI 051

#23

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82653
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																															
1/19/2014	2029	34	13A28-13-5-10 KKF24C	.100	21		102			X	Valadez	16071																																	
<p>Prepared by Douglas C. Talbot on 01-28-13</p> <p>Expires 01-28-15</p>																																													
<p>Test Result</p> <table border="1"> <tr><td>0001</td><td>03/15/14</td><td>03:36</td><td>.108</td></tr> <tr><td>0002</td><td>03/15/14</td><td>03:39</td><td>.116</td></tr> <tr><td>0003</td><td>03/16/14</td><td>05:12</td><td>.204</td></tr> <tr><td>0004</td><td>03/16/14</td><td>05:15</td><td>.207</td></tr> <tr><td>0005</td><td>03/17/14</td><td>04:02</td><td>.120</td></tr> <tr><td>0006</td><td>03/17/14</td><td>04:04</td><td>.155</td></tr> <tr><td>0007</td><td>03/17/14</td><td>04:08</td><td>.149</td></tr> <tr><td>0008</td><td>03/19/14</td><td>20:29</td><td>.102</td></tr> </table> <p>Calibration 02/17/14 10:08 .100</p> <p>No. Date Time ZBMC Calibration 02/17/14 10:08 .100</p> <p>03/19/14 20:30 HS4 Serial no: 082653 Version no: 4797</p> <p>Push No. 3 to erase, or Power down to Retain</p>														0001	03/15/14	03:36	.108	0002	03/15/14	03:39	.116	0003	03/16/14	05:12	.204	0004	03/16/14	05:15	.207	0005	03/17/14	04:02	.120	0006	03/17/14	04:04	.155	0007	03/17/14	04:08	.149	0008	03/19/14	20:29	.102
0001	03/15/14	03:36	.108																																										
0002	03/15/14	03:39	.116																																										
0003	03/16/14	05:12	.204																																										
0004	03/16/14	05:15	.207																																										
0005	03/17/14	04:02	.120																																										
0006	03/17/14	04:04	.155																																										
0007	03/17/14	04:08	.149																																										
0008	03/19/14	20:29	.102																																										

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

IP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8848
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
3/19/2014	2031	34	13A28-13-5-10 KKF24C	.100	21	.104			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13									
			Expires 01-28-15									

03/19/14 20:32  
 AS4 Serial no: 008848  
 Version no: 4797  
 No. Date Time YRAC  
 Calibration 01/13/09 09:25 .100  
 Test Result 0001 03/19/14 20:32 .104

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77591
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS/ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2034	34	13A28-13-5-10 KKF24C	.100	20		103			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:34  
 HSA Serial no: 077591  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 05/21/08 15:09 .100  
 Test Result  
 0001 03/19/14 20:34 .103

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

EA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8850
------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
3/19/2014	2037	34	13A28-13-5-10 KKF24C	.100	22		.103				X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											

03/19/14 20:38  
 AS4 Serial not 008850  
 Version not 4797  
 No. Date Time ZNHC  
 Calibration 03/02/06 09:07 .100  
 Test Result  
 0001 03/19/14 20:37 .103

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77573
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014			13A28-13-5-10 KKF24C	.100							Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										
<i>NOT WORKING</i>													

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#28

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83051
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2038	34	13A28-13-5-10	.100	22		.101			X	Valadez	16071	
			KKF24C										
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:39  
 HSA Serial no: 002051  
 Version no: 4797  
 No. Date Time 2BAC  
 Calibration 03/25/13 14:00 .100  
 Test Result 0001 03/19/14 20:39 .101  
 Push No. 3 to erase,  
 or power down to retain

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CP 202K (Rev. 8-08) OPI 051

#30

AREA NAME City of Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8897
----------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
1/19/2014	204034	34	13A28-13-5-10	.100							Valadez	16071	
			KKF24C										
			Prepared by										
			Douglas C. Talbot										
			on 01-28-13										
			Expires 01-28-15										

*NOT*

*WORKING*

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18624
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2014	2047	34	13A28-13-5-10 KKF24C	.100	22		099			X		Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											

03/19/14 20:48  
 MS4 Serial no: 018624  
 User: talbot no: 4797  
 NO. Date Time Z/DNC  
 Calibration 10/07/08 09:19 .100  
 Test Result  
 MS4 03/19/14 20:47 .096

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

IP 202K (Rev. 8-08) OPI 051

#32

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18644
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2051	34	13A28-13-5-10	.100	22		.100			X	Valadez	16071	
			KKF24C										
			Prepared by										
			Douglas C. Talbot										
			on 01-28-13										
			Expires 01-28-15										

03/19/14 20:53  
 AS4 Serial no: 018646  
 Version no: 4797  
 No. Date Time ZHMC  
 Calibration 07/29/13 13:32 .100  
 Test Result  
 0001 03/19/14 20:52 .100  
 Push No. 3 to erase  
 or power down to retain

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#33

HP 202K (Rev. 8-08) OPI 051

EA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77488
------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2053	34	13A28-13-5-10 KKF24C	.100	21		.103			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

No. Date Time ZHNC  
 Calibration 05/21/09 14:35 .100  
 Test Result 0001 03/19/14 20:53 .103  
 Push No. 3 to erase,  
 or power down to retain

03/19/14 20:53  
 RS4 Serial no: 072488  
 Version no: 4797

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

HP 202K (Rev. 8-08) OPI 051

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8877
--------------------------------	-----------------	--------------------------------------	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
3/19/2014	205 <sup>3</sup>	34	13A28-13-5-10 KKF24C	.100	21		105			Valadez	16071	
<p>Prepared by Douglas C. Talbot on 01-28-13</p> <p>Expires 01-28-15</p>												
<p>Push No. 3 to enter or power down to return to 0001</p> <pre> 03/19/14 20:54 MS4 Serial not 00000000 Version not 4797  No. Date Time ZRAC Calibration 05/26/05 10:40 .100  Test Result 0001 03/19/14 20:54 .105                     </pre>												

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#36

EA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18615
------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	2057	34	13A28-13-5-10	.100	22		.100			X	Valadez	16071	
			KKF24C										
			Prepared by										
			Douglas C. Talbot										
			on 01-28-13										
			Expires 01-28-15										

No. Date Time ZBAC  
 Calibration 12/19/08 14:44 .100  
 Test Result  
 01001 03/19/14 20:57 .100  
 03/19/14 20:58  
 NS4 Serial no: 018615  
 Version no: 4797

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#37

EA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18626
------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
13/19/2014	2058	34	13A28-13-5-10 KKF24C	.100	22		.102				Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

No. Date Time ZHAC  
 Calibration 04/29/05 08:59 .101  
 Test Result  
 0001 03/19/14 20:59 .102

03/19/14 20:59  
 MS4 Serial no: 018626  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18696
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
03/19/2014	2059	34	13A28-13-5-10 KKF24C	.100	21	22	.110	.104		X	Valadez	16071		
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											
							Test Result 0001 03/19/14 20:59 .110 0002 03/19/14 21:00 .104							
							No. Date Time ZONE Calibration 09/08/13 14:50 .100							
							03/19/14 21:01 654 Serial no: 018696 Version no: 4797							

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83030
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2102	34	13A28-13-5-10	.100	21		.102			X	Valadez	16071	
			KKF24C										
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

No. Date Time ZBAC  
 Calibration 01/29/09 10:39 .100  
 Test Result  
 0001 03/19/14 21:02 .102  
 Push No. 3 to erase  
 or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18692
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2014	2:03	34	13A28-13-5-10 KKF24C	.100	22		101			X		Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13											
			Expires 01-28-15											

03/19/14 21:04  
 H54 Serial no: 018692  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 08/12/08 14:35 .100  
 Test Result  
 0001 03/14/14 19:27 .110  
 0002 03/14/14 19:28 .102  
 0003 03/14/14 16:21 .018  
 0004 03/19/14 21:04 .101

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18762
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2107	34	13A28-13-5-10 KKF24C	.100	21		.102			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 21:07  
NSA Serial no: 010762  
Version no: 4797

No. Date Time ZIMC  
Calibration  
03/20/09 15:52 .100

Test Result  
RM01 03/19/14 21:07 .102

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#43

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18766
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/19/2014	2009	34	13A28-13-5-10	.100	22		.097			X	Valadez	16071	Time Change
			KKF24C										
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 20:11  
 AS4 Serial no: 018766  
 Version no: 4797  
 No. Date Time ZMAC  
 Calibration 03/12/12 12:30 .100  
 Test Result  
 0001 03/12/14 22:14 .145  
 0002 03/12/14 22:17 .134  
 0003 03/13/14 01:24 .138  
 0004 03/13/14 01:26 .133  
 0005 03/13/14 01:30 .000  
 0006 03/14/14 00:09 .120  
 0007 03/14/14 00:12 .121  
 0008 03/14/14 02:03 .103  
 0009 03/14/14 02:06 .103  
 0010 03/14/14 02:12 .090  
 0011 03/14/14 03:43 .126  
 0012 03/14/14 03:46 .116  
 0013 03/17/14 19:40 .000  
 0014 03/18/14 20:10 .097

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 17990
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
1/19/2014	2112	34	13A28-13-5-10 KKF24C	.100	22		108				X	Valadez	16071		
			Prepared by Douglas C. Talbot on 01-28-13												
			Expires 01-28-15												

02/19/14 21:13  
 AS4 Serial no: 017990  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 11/26/03 11:10 .100  
 Test Result  
 0001 02/15/14 02:35 .137  
 0002 02/15/14 02:38 .134  
 0003 02/15/14 02:55 .191  
 0004 02/15/14 02:58 .199  
 0005 02/19/14 21:12 .188  
 Push No. 3 to erase.

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#46

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18774
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
1/19/2014	2113	34	13A28-13-5-10	.100	22		.103			X		Valadez	16071	
			KKF24C											
			Prepared by Douglas C. Talbot on 01-28-13 Expires 01-28-15											

03/19/14 21:14  
 NS4 Serial no: 018774  
 Version no: 4797  
 No. Date Time RWAC  
 Calibration 01/13/09 10:25 .100  
 Test Result  
 0001 03/17/14 22:49 .106  
 0002 03/17/14 22:51 .103  
 0003 03/19/14 21:14 .103

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

P 202K (Rev. 8-08) OPI 051

#50

AREA NAME North Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41486
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
1/19/2014	2116	34	13A28-13-5-10 KKF24C	.100	22		.106			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 21:17  
 ASD Serial no: 041486  
 Version no: 4797  
 No. Date Time ZBMC  
 Calibration 03/08/06 13:56 .100  
 Test Result  
 M001 03/19/14 21:17 .106

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

HP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41495
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/19/2014	2117	34	13A28-13-5-10 KKF24C	.100	21		.102			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

Test Result  
 03/19/14 21:17 .102  
 Calibration  
 02/26/14 11:28 .100  
 No. Date Time ZIMC  
 Calibration  
 02/26/14 11:28 .100  
 Version no: 4797  
 03/19/14 21:18  
 NS4 Serial no: 041495  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CP 202K (Rev. 8-08) OPI 051

#52

OFFICER NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41502
-----------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/19/2014	2120	34	13A28-13-5-10 KKF24C	.100	21		.100			X	Valadez	16071	
			Prepared by Douglas C. Talbot on 01-28-13										
			Expires 01-28-15										

03/19/14 21:21  
 ASA Serial no: 041502  
 Version no: 4797  
 No. Date Time YRAC  
 Calibration 05/05/09 08:08 .100  
 Test Result  
 0001 03/17/14 01:56 .097  
 0002 03/17/14 01:53 .098  
 0003 03/19/14 21:20 .100

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**RELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

IP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82654
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
3/19/2014	0022	34	13A28-13-5-10 KKF24C	.100	22		.094			X	Valadez	16071	TIME CHANGE
<p>Prepared by Douglas C. Talbot on 01-28-13</p> <p>Expires 01-28-15</p>													
<p>Test Result 0001 03/07/14 06:53 .000 0002 03/08/14 08:34 .098 0003 03/08/14 08:35 .097 0004 03/08/14 08:36 .000 0005 03/19/14 28:22 .094</p> <p>No. Date Time ZMRC Calibration 03/06/14 12:07 .098</p> <p>03/19/14 28:23 AS4 Serial no: 882654 Version no: 4797</p> <p>Push No. 3 to erase, or power down to retain.</p>													

If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/20/2014	1308	34	13A28-13-5-10- KKF24C	.100	22		-100			X	Lee	17046	
													Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15
													Test Result 0001 03/18/14 01:32 .098 0002 03/18/14 02:53 .142 0003 03/20/14 02:24 .160 0004 03/20/14 02:28 .178 0005 03/20/14 13:08 .100 Push No. 3 to erase. or power down to retain
													No. Date Time ZBAC Calibration 10/24/03 09:04 .100
													03/20/14 13:09 AS4 Serial no: 008635 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 00B559
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/20/2014	1300	34	13A28-13-5-10-KKF24C	.100	21	.101			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p> <p>Test Result 0001 03/20/14 13:04 .101</p> <p>Calibration 11/30/11 00:55 .100</p> <p>No. Date Time %RNC 11/30/11 00:55 .100</p> <p>Version no: 4797 NS4 Serial no: 00B559 03/20/14 13:05</p> <p>Push No. 3 to erase, or power down to retain</p>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.