

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

SGT.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1411	30	13A28-13-5-10- KKF24C	.100	21		.100			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p>													
<p>Test Result 0001 04/15/14 23:52 .140 0002 04/15/14 23:55 .093 0003 04/15/14 23:57 .139 0004 04/16/14 02:34 .172 0005 04/16/14 02:37 .162 0006 04/16/14 14:59 .000 0007 04/18/14 14:15 .100</p>													
<p>No. Date Time %TMC Calibration 03/28/14 11:40 .100</p>													
<p>04/18/14 14:15 AS4 Serial no: 008874 Version no: 4797</p>													
<p>Push No. 3 to erase, or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1147	34	13A28-13-5-10- KKF24C	.100	20		.101			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test Result
 0001 04/18/14 11:49 .101
 Date Time %BAC
 Calibration 07/18/08 13:27 .100
 Version no: 4797
 04/18/14 11:49
 Test Serial no: 018785
 Version no: 4797
 Push No. 3 to erase
 or Power down for safety

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1150	34	13A28-13-5-10- KKF24C	.100	22		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result
0001 03/23/14 17:14 .070
0002 03/23/14 17:15 .112
0003 03/23/14 17:16 .109
0004 04/09/14 15:45 .066
0005 04/18/14 11:59 .101

Calibration
06/28/13 16:36 .100

No. Date Time ZBAC
Calibration

04/18/14 11:56
854 Serial no: 008546
Device no: 4797

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082659
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	12:09	34	13A28-13-5-10-KKF24C	.100	20		.100			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p> <p>Test Result 0001 04/12/14 02:10 .123 0002 04/12/14 02:12 .124 0003 04/12/14 23:55 .000 0004 04/18/14 12:10 .100</p> <p>No. Date Time ZMAC Calibration 03/28/14 09:04 .100</p> <p>04/18/14 12:11 AS4 Serial no: 082659 Version no: 4797</p> <p>Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014		34	13A28-13-5-10- KKF24C	.100							Lee	17046	NOT IN LOCKER

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083032
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1215	34	13A28-13-5-10-KKF24C	.100	20		.102			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

04/18/14 12:16
AS4 Serial no: 083032
Version no: 4797
No. Date Time %BAC
Calibration
02/17/14 09:01 .100
Test Result
0001 04/18/14 12:15 .102
Push Ho. 3 to erase.
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	12:18	34	13A28-13-5-10-KKF24C	.100	20	.101			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

No. Date Time ZONE
 Calibration 01/30/13 14:34 .100
 Test Result
 0001 04/15/14 03:16 .146
 0002 04/15/14 03:18 .145
 0003 04/16/14 02:52 .105
 0004 04/18/14 12:19 .101
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	17 20 34		13A28-13-5-10- KKF24C	.100	21		103		X		Lee	17046	

Prepared by:
 Douglas C. Talbot
 01-28-13
 Expires on:
 01-28-15

Test Result
 0001 04/17/14 22:30 .000
 0002 04/12/14 22:52 .000
 0003 04/13/14 00:27 .055
 0004 04/13/14 00:52 .065
 0005 04/13/14 02:17 .130
 0006 04/13/14 02:19 .126
 0007 04/14/14 03:07 .078
 0008 04/14/14 03:51 .194
 0009 04/14/14 03:54 .153
 0010 04/18/14 12:20 .103
 Calibration 02/17/14 08:52 .100
 No. Date Time ZINC
 Version not 4797
 04/18/14 12:20
 AS4 Serial no: 082957
 Version not 4797

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1223	34	13A28-13-5-10- KKF24C	.100	21		.099			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p style="text-align: center;"> Test Result 0001 04/18/14 12:23 .099 Calibration 05/21/08 13:16 .100 No. Date Time ZBAC Version no: 4797 04/18/14 12:23 NS4 Serial no: 077505 Version no: 4797 Push No. 3 to erase. </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1230	34	13A28-13-5-10- KKF24C	.100	20		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result
 0001 04/10/14 12:30 .102
 No. Date Time ZHNC
 Calibration 02/17/14 09:16 .100
 Version no: 4797
 HSA Serial no: 082660
 04/18/14 12:30

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																																																																																																																								
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																																																																																																																																																												
04/18/2014	1233	34	13A28-13-5-10- KKF24C	.100	21		.102			X		Lee	17046																																																																																																																																																									
<table style="width: 100%; font-size: small; border-collapse: collapse;"> <tr> <td colspan="5">Prepared by:</td> <td colspan="5"></td> <td colspan="5"></td> </tr> <tr> <td colspan="5">Douglas C. Talbot</td> <td colspan="5"></td> <td colspan="5"></td> </tr> <tr> <td colspan="5">01-28-13</td> <td colspan="5"></td> <td colspan="5"></td> </tr> <tr> <td colspan="5">Expires on:</td> <td colspan="5"></td> <td colspan="5"></td> </tr> <tr> <td colspan="5">01-28-15</td> <td colspan="5"></td> <td colspan="5"></td> </tr> <tr> <td colspan="15" style="text-align: center;"> <div style="display: flex; justify-content: space-between; align-items: center;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Push No. 3 to erase, or power down to retain</div> <div> <p>Yest Result</p> <table style="font-size: x-small;"> <tr><td>0001</td><td>04/12/14</td><td>01:08</td><td>.000</td></tr> <tr><td>0002</td><td>04/12/14</td><td>03:20</td><td>.148</td></tr> <tr><td>0003</td><td>04/12/14</td><td>03:32</td><td>.139</td></tr> <tr><td>0004</td><td>04/12/14</td><td>21:36</td><td>.115</td></tr> <tr><td>0005</td><td>04/12/14</td><td>21:38</td><td>.185</td></tr> <tr><td>0006</td><td>04/13/14</td><td>01:10</td><td>.208</td></tr> <tr><td>0007</td><td>04/13/14</td><td>01:19</td><td>.200</td></tr> <tr><td>0008</td><td>04/13/14</td><td>02:35</td><td>.112</td></tr> <tr><td>0009</td><td>04/13/14</td><td>02:38</td><td>.121</td></tr> <tr><td>0010</td><td>04/13/14</td><td>02:39</td><td>.026</td></tr> <tr><td>0011</td><td>04/13/14</td><td>04:16</td><td>.169</td></tr> <tr><td>0012</td><td>04/13/14</td><td>04:18</td><td>.150</td></tr> <tr><td>0013</td><td>04/18/14</td><td>12:34</td><td>.102</td></tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>04/18/14 12:35</td> <td>AS4 Serial no: 008709</td> </tr> <tr> <td>Version no: 4797</td> <td></td> </tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>No. Date Time</td> <td>XPAC</td> </tr> <tr> <td>Calibration</td> <td></td> </tr> <tr> <td>02/17/14 09:27</td> <td>.100</td> </tr> </table> </div> </div> </td> </tr> </table>															Prepared by:															Douglas C. Talbot															01-28-13															Expires on:															01-28-15															<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Push No. 3 to erase, or power down to retain</div> <div> <p>Yest Result</p> <table style="font-size: x-small;"> <tr><td>0001</td><td>04/12/14</td><td>01:08</td><td>.000</td></tr> <tr><td>0002</td><td>04/12/14</td><td>03:20</td><td>.148</td></tr> <tr><td>0003</td><td>04/12/14</td><td>03:32</td><td>.139</td></tr> <tr><td>0004</td><td>04/12/14</td><td>21:36</td><td>.115</td></tr> <tr><td>0005</td><td>04/12/14</td><td>21:38</td><td>.185</td></tr> <tr><td>0006</td><td>04/13/14</td><td>01:10</td><td>.208</td></tr> <tr><td>0007</td><td>04/13/14</td><td>01:19</td><td>.200</td></tr> <tr><td>0008</td><td>04/13/14</td><td>02:35</td><td>.112</td></tr> <tr><td>0009</td><td>04/13/14</td><td>02:38</td><td>.121</td></tr> <tr><td>0010</td><td>04/13/14</td><td>02:39</td><td>.026</td></tr> <tr><td>0011</td><td>04/13/14</td><td>04:16</td><td>.169</td></tr> <tr><td>0012</td><td>04/13/14</td><td>04:18</td><td>.150</td></tr> <tr><td>0013</td><td>04/18/14</td><td>12:34</td><td>.102</td></tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>04/18/14 12:35</td> <td>AS4 Serial no: 008709</td> </tr> <tr> <td>Version no: 4797</td> <td></td> </tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>No. Date Time</td> <td>XPAC</td> </tr> <tr> <td>Calibration</td> <td></td> </tr> <tr> <td>02/17/14 09:27</td> <td>.100</td> </tr> </table> </div> </div>															0001	04/12/14	01:08	.000	0002	04/12/14	03:20	.148	0003	04/12/14	03:32	.139	0004	04/12/14	21:36	.115	0005	04/12/14	21:38	.185	0006	04/13/14	01:10	.208	0007	04/13/14	01:19	.200	0008	04/13/14	02:35	.112	0009	04/13/14	02:38	.121	0010	04/13/14	02:39	.026	0011	04/13/14	04:16	.169	0012	04/13/14	04:18	.150	0013	04/18/14	12:34	.102	04/18/14 12:35	AS4 Serial no: 008709	Version no: 4797		No. Date Time	XPAC	Calibration		02/17/14 09:27	.100
Prepared by:																																																																																																																																																																						
Douglas C. Talbot																																																																																																																																																																						
01-28-13																																																																																																																																																																						
Expires on:																																																																																																																																																																						
01-28-15																																																																																																																																																																						
<div style="display: flex; justify-content: space-between; align-items: center;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Push No. 3 to erase, or power down to retain</div> <div> <p>Yest Result</p> <table style="font-size: x-small;"> <tr><td>0001</td><td>04/12/14</td><td>01:08</td><td>.000</td></tr> <tr><td>0002</td><td>04/12/14</td><td>03:20</td><td>.148</td></tr> <tr><td>0003</td><td>04/12/14</td><td>03:32</td><td>.139</td></tr> <tr><td>0004</td><td>04/12/14</td><td>21:36</td><td>.115</td></tr> <tr><td>0005</td><td>04/12/14</td><td>21:38</td><td>.185</td></tr> <tr><td>0006</td><td>04/13/14</td><td>01:10</td><td>.208</td></tr> <tr><td>0007</td><td>04/13/14</td><td>01:19</td><td>.200</td></tr> <tr><td>0008</td><td>04/13/14</td><td>02:35</td><td>.112</td></tr> <tr><td>0009</td><td>04/13/14</td><td>02:38</td><td>.121</td></tr> <tr><td>0010</td><td>04/13/14</td><td>02:39</td><td>.026</td></tr> <tr><td>0011</td><td>04/13/14</td><td>04:16</td><td>.169</td></tr> <tr><td>0012</td><td>04/13/14</td><td>04:18</td><td>.150</td></tr> <tr><td>0013</td><td>04/18/14</td><td>12:34</td><td>.102</td></tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>04/18/14 12:35</td> <td>AS4 Serial no: 008709</td> </tr> <tr> <td>Version no: 4797</td> <td></td> </tr> </table> </div> <div> <table style="font-size: x-small;"> <tr> <td>No. Date Time</td> <td>XPAC</td> </tr> <tr> <td>Calibration</td> <td></td> </tr> <tr> <td>02/17/14 09:27</td> <td>.100</td> </tr> </table> </div> </div>															0001	04/12/14	01:08	.000	0002	04/12/14	03:20	.148	0003	04/12/14	03:32	.139	0004	04/12/14	21:36	.115	0005	04/12/14	21:38	.185	0006	04/13/14	01:10	.208	0007	04/13/14	01:19	.200	0008	04/13/14	02:35	.112	0009	04/13/14	02:38	.121	0010	04/13/14	02:39	.026	0011	04/13/14	04:16	.169	0012	04/13/14	04:18	.150	0013	04/18/14	12:34	.102	04/18/14 12:35	AS4 Serial no: 008709	Version no: 4797		No. Date Time	XPAC	Calibration		02/17/14 09:27	.100																																																																																										
0001	04/12/14	01:08	.000																																																																																																																																																																			
0002	04/12/14	03:20	.148																																																																																																																																																																			
0003	04/12/14	03:32	.139																																																																																																																																																																			
0004	04/12/14	21:36	.115																																																																																																																																																																			
0005	04/12/14	21:38	.185																																																																																																																																																																			
0006	04/13/14	01:10	.208																																																																																																																																																																			
0007	04/13/14	01:19	.200																																																																																																																																																																			
0008	04/13/14	02:35	.112																																																																																																																																																																			
0009	04/13/14	02:38	.121																																																																																																																																																																			
0010	04/13/14	02:39	.026																																																																																																																																																																			
0011	04/13/14	04:16	.169																																																																																																																																																																			
0012	04/13/14	04:18	.150																																																																																																																																																																			
0013	04/18/14	12:34	.102																																																																																																																																																																			
04/18/14 12:35	AS4 Serial no: 008709																																																																																																																																																																					
Version no: 4797																																																																																																																																																																						
No. Date Time	XPAC																																																																																																																																																																					
Calibration																																																																																																																																																																						
02/17/14 09:27	.100																																																																																																																																																																					

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	12:36	34	13A28-13-5-10-KKF24C	.100	20		.107			X	Lee	17046	
			Prepared by: Douglas C. Talbot 01-28-13										
			Expires on: 01-28-15										

Push No. 3 to erase, or power down to retain.

Test Result
 0001 04/11/14 23:00 .009
 0002 04/11/14 23:04 .008
 0003 04/12/14 01:59 .031
 0004 04/12/14 03:21 .018
 0005 04/12/14 04:19 .103
 0006 04/12/14 04:24 .105
 0007 04/12/14 23:25 .000
 0008 04/13/14 02:40 .132
 0009 04/13/14 02:51 .137
 0010 04/13/14 04:18 .091
 0011 04/13/14 04:22 .089
 0012 04/14/14 01:07 .018
 0013 04/14/14 02:40 .244
 0014 04/14/14 02:44 .222
 0015 04/14/14 02:46 .215
 0016 04/18/14 12:37 .107

No. Date Time ZTMC
 Calibration
 04/18/14 12:37
 HSA Serial no: 008714
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 0098880
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*				CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1239	34	13A28-13-5-10- KKF24C	.100	21		.102			X		Lee	17046	
			Prepared by: Douglas C. Talbot											
			01-28-13											
			Expires on: 01-28-15											

Push No. 3 to erase,
 or power down to retain

 No. Date Time ZBAC
 Calibration
 03/25/13 10:45 .100

 Test Result
 0001 04/12/14 09:14 .095
 0002 04/12/14 09:16 .099
 0003 04/18/14 12:37 .102

 04/18/14 12:38
 NS4 Serial no: 008888
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1242	34	13A28-13-5-10- KKF24C	.100	21		.102			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Test Result
 0001 04/18/14 09:21 .106
 0002 04/18/14 09:24 .162
 0003 04/18/14 09:27 .169
 0004 04/18/14 12:41 .102
 Calibration
 07/18/08 13:52 .100
 No. Date Time 2BAC
 04/18/14 12:41
 HS4 Serial no: 008783
 Version no: 4797
 04/18/14 12:41
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)
04/18/2014	12:45	34	13A28-13-5-10-KKF24C	.100	20		.101			X	Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15													
Test Result 0001 04/11/14 02:19 .153 0002 04/11/14 02:27 .139 0003 04/18/14 12:43 .101 Calibration 02/17/14 09:50 .100 No. Date Time ZRNC 04/18/14 12:44 NS4 Serial no: 082657 Version no: 4797 Push No. 3 to erase, or Power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1248	34	13A28-13-5-10- KKF24C	.100	20		.099			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Test Result
0001 04/18/14 12:47 .099

Calibration
18/30/03 09:53 .100

No. Date Time ZHNC

Version no: 4797

04/18/14 12:48
654 Serial no: 008810

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1251	34	13A28-13-5-10-KKF24C	.100	21		.103			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p>Test Result 0001 04/18/14 12:50 .103</p> <p>No. Date Time %BAC Calibration 11/03/11 16:00 .100</p> <p>04/18/14 12:50 NS4 Serial no: 018684 Version no: 4797</p> <p>Push No. 3 to erase, or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	12:54	34	13A28-13-5-10- KKF24C	.100	20		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

No. Date Time ZPAC
 Calibration 18/24/03 09:04 .100
 Test Result 0001 04/18/14 12:55 .102
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. D08845
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	12:57	34	13A28-13-5-10- KKF24C	.100	20		.107			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
<div style="border: 1px solid black; padding: 5px; display: inline-block;"> <p>04/18/14 12:57 NS4 Serial no: 0000915 Version no: 4797</p> <p>No. Date Time ZBNC Calibration 12/24/03 11:01 .100</p> <p>Test Result</p> <p>0001 04/16/14 01:40 .081 0002 04/16/14 01:42 .081 0003 04/16/14 03:44 .114 0004 04/16/14 03:46 .111 0005 04/17/14 02:13 .094 0006 04/17/14 02:15 .103 0007 04/18/14 12:57 .107</p> </div> <p style="margin-top: 10px;">Push No. 3 to erase,</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1259	34	13A28-13-5-10-KKF24C	.100	20	.101				X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p> <p style="text-align: center;">Push No. 3 to erase, or power down to retain</p> <p style="text-align: center;">Test Result 0001 04/12/14 02:32 .066 0002 04/13/14 03:09 .185 0003 04/13/14 03:11 .167 0004 04/18/14 12:59 .101</p> <p style="text-align: center;">No. Date Time %BAC Calibration 02/17/14 10:08 .100</p> <p style="text-align: center;">04/18/14 12:59 AS4 Serial no: 082653 Version no: 4797</p> <p style="text-align: center;">or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
04/18/2014	1302	34	13A28-13-5-10- KKF24C	.100	21					Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Push No. 3 to erase or power down to retain

Test Result
0001 04/18/14 07:09 .146
0002 04/18/14 07:14 .138
0003 04/18/14 13:03 .104

No. Date Time %DAC
Calibration
01/18/03 09:25 .100

04/18/14 13:03
NS4 Serial no: 008848
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1:30	34	13A28-13-5-10-KKF24C	.100	20		.100			X	Lee	17046	
		Prepared by:											
		Douglas C. Talbot											
		01-28-13											
		Expires on:											
		01-28-15											
<div style="border: 1px solid black; padding: 5px; display: inline-block;"> <p style="margin: 0;">64/18/14 13:09 NS4 Serial no: 077591 Version no: 4797</p> <p style="margin: 0;">No. Date Time %BAC Calibration 05/21/08 15:09 .100</p> <p style="margin: 0;">Test Result #001 04/18/14 13:08 .100</p> <p style="margin: 0;">Push No. 3 to erase, or power down to retain</p> </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26
 008850

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1308	34	13A28-13-5-10-KKF24C	.100	21	.101			X		Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15													
Test Result 0001 04/18/14 13:11 .100 Calibration 03/02/06 09:07 .100 No. Date Time 2BAC Version no: 4797 04/18/14 13:11 154 Serial no: 008850 04/18/14 13:11 Version no: 4797 Push No. 3. to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	
													B.O.
													OUT FOR SERVICE
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1311	34	13A28-13-5-10-KKF24C	.100	21		.099			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p>													
<p style="writing-mode: vertical-rl; transform: rotate(180deg);"> Push No. 3 to erase Test Result 0001 04/18/14 13:15 .099 Calibration 03/25/13 14:00 .100 No. Date Time ZINC Version no: 4797 04/18/14 13:16 NS4 Serial no: 003051 Version no: 4797 </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1314	34	13A28-13-5-10-KKF24C	.100	21		.106			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

or power down to retain
 04/18/14 13:19
 HS4 Serial no: 008897
 Version no: 4797
 No. Date Time %DNC
 Calibration 01/18/12 16:13 .100
 Test Result
 0001 04/18/14 13:19 .106
 Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1317	34	13A28-13-5-10-KKF24C	.100	20		.095			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p> <p style="text-align: center;">Test Result 0001 04/18/14 13:20 .095</p> <p style="text-align: center;">Calibration 18/02/08 09:19 .100</p> <p style="text-align: center;">No. Date Time %BAC 04/18/14 13:21 AS4 Serial no: 018624 Version no: 4797</p> <p style="text-align: center;">Push No. 3 to erase, or Power down to retain</p> <p style="text-align: center;">or Power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1320	34	13A28-13-5-10- KKF24C	.100	22		.100					Lee	17046	
Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15														
04/18/14 13:25 604 Serial no: 018646 Version no: 4797 No. Date Time %BAC Calibration 07/29/13 13:32 .100 Test Result 0001 04/18/14 13:24 .100 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1323	34	13A28-13-5-10- KKF24C	.100	21		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Push No. 3 to erase.

 Test Result
0001 04/18/14 13:25 .102

 No. Date Time ZIMC
Calibration
05/21/08 14:35 .100

 04/18/14 13:25
NS4 Serial no: 077488
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1326	34	13A28-13-5-10- KKF24C	.100	21		.103			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

or power down to retain
 04/18/14 13:29
 104 Serial no: 008877
 Version no: 4797
 No. Date Time ZBAC
 Calibration
 05/26/05 10:40 .100
 Test Result
 0001 04/18/14 13:29 .100
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1329	34	13A28-13-5-10-KKF24C	.100	22		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test Result
0001 04/18/14 13:30 .099

No. Date Time ZMNC
Calibration
12/19/08 14:44 .100

04/18/14 13:30
054 Serial no: 018615
Version no: 4797

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1332	34	13A28-13-5-10- KKF24C	.100	22		.100			X		Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

04/18/14 13:32
164 Serial no: 018626
Version no: 4797
No. Date Time %BAC
Calibration 04/29/05 08:59 .101
Test Result
0001 04/18/14 13:32 .100
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
04/18/2014	1335	34	13A28-13-5-10- KKF24C	.100	21							Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			01-28-13												
			Expires on:												
			01-28-15												

Test Result
 0001 04/18/14 13:34 .100
 Calibration
 09/09/13 14:50 .100
 No. Date Time ZMAC
 Version no: 4797
 04/18/14 13:34
 AS4 Serial no: 018696
 04/18/14 13:34
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1338	34	13A28-13-5-10-KKF24C	.100	21		.100					Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

04/18/14 13:43
 654 Serial no: 083030
 Version no: 4797
 No. Date Time ZMAC
 Calibration 01/29/09 10:39 .100
 Test Result
 0001 04/18/14 13:42 .100
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
04/18/2014	1341	34	13A28-13-5-10- KKF24C	.100	21							Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			01-28-13												
			Expires on:												
			01-28-15												

Push No. 3 to erase
 Test Result
 0001 04/18/14 13:47 .100
 No. Date Time ZMNC
 Calibration 08/12/08 14:35 .100
 Version not 4797
 04/18/14 13:47
 PAS Serial no: 018692
 Version not 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1344	34	13A28-13-5-10-KKF24C	.100	21		.102			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13
Expires on:
01-28-15

Power down to return
04/18/14 13:49
MS4 Serial no: 018762
Version no: 4797
No. Date Time ZBNC
Calibration
03/30/09 15:52 .100
Test Result
0001 04/18/14 13:48 .102
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1347	34	13A28-13-5-10- KKF24C	.100	22		.095			X	Lee	17046	will NOT ALLOW ME TO CLEAR memory.
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p>04/18/14 13:52 AS4 Serial no: 018766 Version no: 4797</p> <p>No. Date Time 2DNC Calibration 03/12/12 12:30 .100</p> <p>Test Result 0001 04/12/14 19:03 .000 0002 04/12/14 21:42 .100 0003 04/12/14 21:45 .175 0004 04/12/14 23:31 .154 0005 04/12/14 23:34 .158 0006 04/13/14 01:35 .204 0007 04/13/14 01:38 .190 0008 04/13/14 03:30 .179 0009 04/14/14 21:18 .000 0010 04/18/14 13:51 .095</p> <p>Push No. 3 to erase, or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1350	34	13A28-13-5-10-KKF24C	.100	20		-101				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

04/18/14 14:00
 AS4 Serial no: 018691
 Version no: 4797
 No. Date Time ZMAC
 Calibration 09/08/08 14:01 .100
 Test Result
 0001 04/12/14 22:59 .179
 0002 04/12/14 23:01 .175
 0003 04/13/14 08:49 .152
 0004 04/13/14 08:52 .133
 0005 04/18/14 13:59 .101
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1353	34	13A28-13-5-10-KKF24C	.100	21		.105			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p> <div style="text-align: right; margin-right: 50px;"> <p>04/18/14 14:01 064 Serial no: 017990 Version no: 4797</p> <p>No. Date Time 2BAC Calibration 11/26/03 11:10 .100</p> <p>Test Result 0001 04/13/14 02:39 .175 0002 04/13/14 02:42 .176 0003 04/13/14 12:21 .207 0004 04/13/14 12:23 .263 0005 04/13/14 22:37 .227 0006 04/13/14 22:39 .244 0007 04/18/14 14:00 .105</p> <p>Push No. 3 to erase. or power down to retain.</p> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1356	BY	13A28-13-5-10- KKF24C	.100	22		.101			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13</p> <p>Expires on: 01-28-15</p>														
<p style="writing-mode: vertical-rl; transform: rotate(180deg);">Test Result</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">No. Date Time %NC</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Calibration</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">01/13/09 18:25 .100</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">04/18/14 14:04</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">NC4 Serial no: 018774</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Version no: 4797</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Test Result</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0001 04/15/14 19:58 .000</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0002 04/15/14 19:52 .000</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0003 04/17/14 02:44 .000</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0004 04/17/14 22:46 .092</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0005 04/17/14 22:49 .066</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0006 04/17/14 22:51 .062</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0007 04/18/14 01:59 .160</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0008 04/18/14 02:03 .156</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">0009 04/18/14 14:03 .101</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Tush No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041406
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
04/18/2014	1359	34	13A28-13-5-10- KKF24C	.100	21		.105				X	Lee	17046		
			Prepared by: Douglas C. Talbot												
			01-28-13												
			Expires on:												
			01-28-15												
											04/18/14 14:06 NS4 Serial no: 041406 Version no: 4797				
											No. Date Time YMMC Calibration 03/08/06 13:56 .100				
											Test Result 0001 04/18/14 14:05 .105				
											Push No. 3 to erase or power down to retain				

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	140234		13A28-13-5-10-KKF24C	.100	80		.100			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										
No. Date Time %MC Calibration 02/28/14 11:29 .100 Test Result 0001 04/18/14 14:07 .100 Push No. 3 to erase.													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION: (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014	1405	34	13A28-13-5-10- KKF24C	.100	21		.097			X	Lee	17046	

Prepared by:
Douglas C. Talbot
01-28-13

Expires on:
01-28-15

Test Result
 0001 04/12/14 08:01 .055
 0002 04/12/14 08:05 .051
 0003 04/12/14 08:09 .006
 0004 04/12/14 08:11 .006
 0005 04/12/14 08:12 .104
 0006 04/12/14 08:16 .072
 0007 04/13/14 01:05 .017
 0008 04/13/14 02:04 .166
 0009 04/13/14 02:07 .166
 0010 04/16/14 14:10 .097

No. Date Time %MC
 Calibration
 05/05/09 08:03 .100

04/18/14 14:11
 NS4 Serial no: 041502
 Version no: 4797

or power down to retain

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082654
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/18/2014	1408	34	13A28-13-5-10- KKF24C	.100	20		.100				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test Result
 0001 04/18/14 14:12 .100
 No. Date Time ZINC
 Calibration 03/28/14 11:36 .100
 04/18/14 14:12
 ASA Serial no: 082654
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
04/18/2014			13A28-13-5-10- KKF24C	.100							Lee	17046	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.