





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
05/01/2014	1332	34	13A28-13-5-10- KKF24C	.100	22		.099			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 01-28-13 Expires on: 01-28-15</p>													
<p>Print No. 3. To erase, press down to retain</p>													
<p>Cal Result 06/28/10 16:36 .100</p>													
<p>No. Date Time ZHNC Calibration 06/28/10 16:36 .100</p>													
<p>05/01/14 13:31 154 Serial not response Version not 07/07</p>													
<p>0601 04/25/14 19:22 .141 0602 04/25/14 19:25 .144 0603 04/26/14 03:33 .097 0604 04/26/14 03:36 .091 0605 04/29/14 01:26 .141 0606 04/29/14 01:27 .141 0607 04/29/14 01:31 .266 0608 04/29/14 03:29 .202 0609 04/29/14 03:31 .193 0610 04/30/14 01:41 .096 0611 04/30/14 02:23 .117 0612 05/01/14 01:42 .082 0613 05/01/14 02:15 .155 0614 05/01/14 02:17 .145 0615 05/01/14 13:31 .099</p>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
05/01/2014	1330	34	13A28-13-5-10-KKF24C	.100	21		.102			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result  
 05/01/14 13:41 .102  
 No. Date Time ZPIC  
 Calibration  
 11/30/11 08:55 .100  
 Version no: 6797  
 HSA Serial no: 008559  
 05/01/14 13:41

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.











**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
05/01/2014	1409	34	13A28-13-5-10- KKF24C	.100	21		.098		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test Result  
 0001 05/01/14 14:07 .098  
 Calibration  
 05/21/08 13:16 .100  
 Date Time 2JMC  
 Version No: 4797  
 05/01/14 14:08  
 Serial no: 077505  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

















**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
05/01/2014	1433	34	13A28-13-5-10-KKF24C	.100	20		.098			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test No. 3 10 of 150  
 or power down in 14:30  
 Test Result  
 0001 05/01/14 14:30 .098  
 Cal. Date Time 2000  
 10/26/09 09:50 .106  
 Version no: 4797  
 05/01/14 14:30  
 Test Serial no: 000010  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
05/01/2014	1436	34	13A28-13-5-10- KKF24C	.100	22		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			01-28-13										
			Expires on:										
			01-28-15										

Test No. 3 In Passes  
 Test Result  
 0001 05/01/14 14:32 .100  
 Calibration  
 11/03/11 16:03 .100  
 Date Time 23MC  
 05/01/14 14:32  
 Serial no: 018684  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



























































**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082654
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
05/01/2014	1553	34	13A28-13-5-10- KKF24C	.100	22		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			01-28-13											
			Expires on:											
			01-28-15											

Test No. 5 to error  
 over down to retain  
 Test Result  
 05/01/2014 15:48 .099  
 Calibration  
 05/28/14 11:53 .100  
 Date Time Temp  
 Calibration  
 05/01/14 15:49  
 Test Serial not 082654  
 Precision not 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

