

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008501 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0810 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .092 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

Test Result
 0001 05/21/14 08:12 .092
 Calibration
 03/30/09 15:40 .100
 No. Date Time %BMC
 05/21/14 08:13
 AS4 Serial no: 008501
 Version no: 4797

Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018785 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|--------------------------------------|---------------------------|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 0915 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | | .097 | | | | X | Lee | 17046 | |
| Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | | | | |
| Test Result 0001 05/20/14 12:14 .266 0002 05/20/14 12:17 .306 0003 05/20/14 12:19 .258 0004 05/21/14 08:16 .097 | | | | | | | | | | | | | | |
| No. Date Time XBAC Calibration 07/18/08 13:27 .100 | | | | | | | | | | | | | | |
| 05/21/14 08:17 AS4 Serial no: 018785 Version no: 4797 | | | | | | | | | | | | | | |
| Push No. 3 to erase, or power down to retain | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008546 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0910 | 34 | 14C13-10-5-10 C1307252 | .100 | 22 | | .100 | | | X | Lee | 17046 | |
| <p>Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16</p> | | | | | | | | | | | | | |
| <p>No. Date Time %BAC Calibration 06/28/13 16:36 .100</p> <p>Test Result 0001 05/17/14 04:42 .030 0002 05/18/14 01:23 .164 0003 05/18/14 01:26 .175 0004 05/19/14 03:32 .085 0005 05/19/14 03:35 .086 0006 05/19/14 03:37 .095 0007 05/21/14 08:19 .100</p> <p>Push No. 3 to erase, or power down to retain</p> | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

4

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082659 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|--|--|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | | |
| 05/21/2014 | 0824 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | | | | | | Lee | 17046 | | |
| | | Prepared by: | | | | | | | | | | | | | |
| | | Douglas C. Talbot | | | | | | | | | | | | | |
| | | 03-13-14 | | | | | | | | | | | | | |
| | | Expires on: | | | | | | | | | | | | | |
| | | 03-13-16 | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | <pre> 05/21/14 08:27 ASA Serial no: 082659 Version no: 4797 No. Date Time %BAC Calibration 03/28/14 09:04 .100 Test Result 0001 05/17/14 04:09 .164 0002 05/17/14 04:11 .192 0003 05/17/14 04:14 .157 0004 05/18/14 02:05 .101 0005 05/18/14 02:07 .098 0006 05/18/14 02:10 .032 0007 05/21/14 08:25 .099 </pre> | |
| | | | | | | | | | | | | | | Push No. 3 to erase, or power down to retain | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008559 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 0953 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .103 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |

Test Result
 0001 05/21/14 09:00 .103
 No. Date Time 2BAC
 Calibration 11/30/11 08:55 .100
 Version no: 4797
 AS4 Serial no: 008559
 05/21/14 09:01

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008580 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0900 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .100 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 | | | | | | | | | | |
| | | | Expires on: 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 09:05
AS4 Serial no: 008580
Version no: 4797

No. Date Time %BAC
Calibration 03/06/14 12:39 .100

Test Result
0001 05/20/14 01:19 .135
0002 05/20/14 01:21 .124
0003 05/21/14 01:27 .000
0004 05/21/14 02:44 .000
0005 05/21/14 09:02 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 083032 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP. (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|-----------------------------------|---------------|----------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0905 | 34 | 14C13-10-5-10 | .100 | 21 | | .100 | | | X | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| Prepared by: | | | | | | | | | | | | | |
| Douglas C. Talbot | | | | | | | | | | | | | |
| 03-13-14 | | | | | | | | | | | | | |
| Expires on: | | | | | | | | | | | | | |
| 03-13-16 | | | | | | | | | | | | | |
| <div style="border: 1px solid black; padding: 5px; display: inline-block;"> Test Result 0001 05/21/14 09:05 .100 Calibration 02/17/14 09:01 .100 No. Date Time ZBAC Version no: 4797 AS4 Serial no: 083032 05/21/14 09:06 </div> | | | | | | | | | | | | | |
| Push No. 3 to erase, or power down to retain | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008620 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|-------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0909 | 34 | 14C13-10-5-10 | .100 | 21 | | .093 | | | | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

Push No. 3 to erase.
or Power down to retain

Test Result
 0001 05/19/14 21:18 .094
 0002 05/19/14 21:20 .090
 0003 05/19/14 23:23 .025
 0004 05/19/14 23:25 .024
 0005 05/20/14 01:56 .006
 0006 05/20/14 02:57 .057
 0007 05/21/14 03:01 .067
 0008 05/21/14 09:12 .098

No. Date Time %BAC
 Calibration
 01/30/13 14:34 .100

05/21/14 09:13
 HSA Serial no: 008620
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082957 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|--|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0912 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .101 | | | X | Lee | 17046 | |
| <p>Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16</p> | | | | | | | | | | | | | |
| <p style="text-align: center;">Test Result 0001 05/18/14 02:52 .000 0002 05/18/14 04:20 .125 0003 05/18/14 04:22 .118 0004 05/18/14 04:27 .133 0005 05/19/14 01:49 .123 0006 05/19/14 01:45 .122 0007 05/19/14 02:54 .037 0008 05/21/14 09:12 .101</p> | | | | | | | | | | | | | |
| <p style="text-align: center;">No. Date Time %BAC Calibration 02/17/14 08:52 .100</p> | | | | | | | | | | | | | |
| <p style="text-align: center;">Push No. 9 to erase or remove device from calibration</p> | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

10

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 077505 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 0915 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .098 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

Test Result
 0001 05/21/14 09:16 .098

No. Date Time %BAC
 Calibration 05/21/08 13:16 .100

05/21/14 09:16
 AS4 Serial no: 077505
 Version no: 4797

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 077479 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0910 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | | .098 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

05/21/14 09:19
 AS4 Serial no: 077479
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/17/14 09:10 .100
 Test Result
 0001 05/17/14 04:00 .163
 0002 05/21/14 09:19 .098
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082660 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|--|------|-----------------------------------|---------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0920 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | | .100 | | | X | Lee | 17046 | |
| Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | | | |
| Test Result 0001 05/17/14 04:02 .116 0002 05/17/14 04:05 .110 0003 05/18/14 03:09 .154 0004 05/18/14 03:11 .155 0005 05/21/14 09:20 .100 Calibration 02/17/14 09:16 .100 No. Date Time ZBAC Version no: 4797 H&A Serial no: 082660 05/21/14 09:21 | | | | | | | | | | | | | |
| Push No. 3 to erase, or power down to retain | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008709 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0926 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .101 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 09:30
 H&S4 Serial no: 008709
 Version no: 4797
 No. Date Time ZBNC
 Calibration 02/17/14 09:27 .100
 Test Result
 0001 05/17/14 01:51 .110
 0002 05/17/14 01:59 .112
 0003 05/18/14 04:42 .000
 0004 05/18/14 23:52 .197
 0005 05/18/14 23:58 .207
 0006 05/21/14 09:29 .101

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008714 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0929 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | .104 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 09:32
 NS4 Serial no: 008714
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/11/04 14:05 .100
 Test Result
 0001 05/21/14 09:32 .104
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008888 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|--|------|--------------------------------------|-------------------|---------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0931 | 34 | 14C13-10-5-10 | .100 | 21 | | .101 | | | X | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| Test Result 0001 05/21/14 09:30 .101 Calibration 03/25/13 18:45 .100 No. Date Time %BAC 05/21/14 09:31 AS4 Serial no: 008888 Version no: 4797 Push No. 3 to erase or power down to retain | | | | | | | | | | | | | |

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

17

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008783 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0936 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | .101 | | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

Test Result
 0001 05/18/14 16:25 .259
 0002 05/21/14 09:38 .101
 Calibration
 07/18/08 19:52 .100
 No. Date Time %BAC
 05/21/14 09:39
 NS4 Serial no: 008783
 Version no: 4797
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082657 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0950 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .101 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

```

05/21/14 09:53
AS4 Serial no: 082657
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 09:50 .100

Test Result
0001 05/09/14 01:30 .883
0002 05/09/14 01:33 .878
0003 05/09/14 01:34 .893
0004 05/10/14 00:48 .899
0005 05/10/14 00:50 .897
0006 05/11/14 02:46 .869
0007 05/11/14 02:49 .866
0008 05/12/14 22:15 .812
0009 05/12/14 22:18 .400
0010 05/12/14 22:20 .124
0011 05/17/14 01:53 .824
0012 05/21/14 09:53 .101
    
```

Push No. 3 to erase,
 or enter down to reset.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008810 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|-------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | | | 14C13-10-5-10 | .100 | | | | | | | Lee | 17046 | B. O. |
| | | | C1307252 | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018684 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0955 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .103 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 09:58
 AS4 Serial no: 018684
 Version no: 4797
 No. Date Time %DAC
 Calibration 11/03/11 16:00 .100
 Test Result
 0001 05/21/14 09:58 .103
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008635 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-----------------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 0958 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | .102 | | | | | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 10:01
 AS4 Serial no: 008635
 Version no: 4797
 No. Date Time %BAC
 Calibration
 10/24/03 09:04 .100
 Test Result
 0001 05/21/14 10:01 .102
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

22

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|---------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | LD. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008845 |
|--------------------------------|-----------------|--------------------------------------|---------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1001 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .104 | | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

Push No. 3 to erase.

Test Result
0001 05/21/14 04:22 .097
0002 05/21/14 04:24 .085
0003 05/21/14 10:02 .104

No. Date Time %BAC
Calibration
12/24/03 11:01 .100

05/21/14 18:03
AS4 Serial no: 008845
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082653 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1011 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | | 101 | | | X | | Lee | 17046 | |
| | | Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | | |

Push No. 3 to erase,
or power down to retain

Test Result
 0001 05/17/14 21:39 .063
 0002 05/17/14 21:44 .067
 0003 05/17/14 21:53 .026
 0004 05/18/14 04:16 .153
 0005 05/18/14 04:17 .144
 0006 05/18/14 19:35 .016
 0007 05/18/14 21:44 .059
 0008 05/19/14 05:37 .238
 0009 05/19/14 05:40 .233
 0010 05/21/14 10:12 .101

No. Date Time %BAC
 Calibration
 02/17/14 10:08 .100

05/21/14 10:12
 AS4 Serial no: 082653
 Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008848 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1013 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | | | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

Test Result
 0001 05/21/14 10:16 .104
 Calibration
 01/13/08 09:25 .100
 No. Date Time %BAC
 05/21/14 10:16
 H54 Serial no: 008848
 Version no: 4797
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 077591 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|-------|-----------------------------------|---------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 10:16 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | .102 | | | | X | Lee | 17046 | |
| Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | | | |
| Push No. 3 to erase, or power down to retain Test Result 0001 05/21/14 10:18 .102 Calibration 05/21/08 15:09 .100 No. Date Time ZRAC Version no: 4797 NS4 Serial no: 077591 05/21/14 10:18 | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008850 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP. (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|-------------------|-------|----------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1046 | 34 | 14C13-10-5-10 | .100 | 21 | | .100 | | | X | | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |

05/21/14 10:51
 NS4 Serial no: 0088850
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/02/06 09:07 .100
 Test Result
 0001 05/21/14 10:21 .104
 0002 05/21/14 10:50 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 077573 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|----|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | | NO |
| 05/21/2014 | 1049 | 34 | 14C13-10-5-10 C1307252 | .100 | 22 | | .095 | | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

05/21/14 10:49
NS4 Serial no: 077573
Version no: 4797

No. Date Time %BAC
Calibration
04/09/14 08:36 .000

Test Result
0001 05/19/14 22:59 .000
0002 05/19/14 23:02 .000
0003 05/21/14 10:48 .095

Push No. 3 to erase,
or enter down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

28

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 083051 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|-------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1052 | 34 | 14C13-10-5-10 | .100 | 21 | | .099 | | | X | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 10:52
 A84 Serial no: 083051
 Version no: 4797
 No. Date Time ZBAC
 Calibration 03/25/13 14:00 .100
 Test Result
 0001 05/21/14 10:52 .099
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008897 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | | | 14C13-10-5-10 C1307252 | .100 | | | | | | | Lee | 17046 | NOT IN LOCKER |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018624 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | | | 14C13-10-5-10 C1307252 | .100 | | | | | | | Lee | 17046 | NOT IN LOCKER |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#32

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018646 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|-------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1106 | 34 | 14C13-10-5-10 | .100 | 22 | | .100 | | | X | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |

Test Result

0001 05/21/14 01:11 .073

0002 05/21/14 01:13 .004

0003 05/21/14 11:09 .100

Push No. 3 to erase,
or Power down to retain

No. Date Time %BAC
Calibration 07/29/13 13:32 .100

05/21/14 11:10
AS4 Serial no: 018646
Version no: 4797

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 077488 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|--|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1109 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .102 | | | X | | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | |

No. Date Time %BAC
 Calibration 05/21/08 14:35 .100
 Test Result
 0001 05/21/14 11:09 .102
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 008877 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|--------------------------------------|---------------|---------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1112 | 34 | 14C13-10-5-10 | .100 | 21 | | .103 | | | | Lee | 17046 | |
| | | | C1307252 | | | | | | | | | | |
| Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | | | | |
| <div style="display: flex; justify-content: space-between; align-items: center;"> <div style="border: 1px solid black; padding: 5px; transform: rotate(-90deg); transform-origin: center;"> Push No. 3 to erase, or power down to retain </div> <div style="border: 1px solid black; padding: 5px; transform: rotate(-90deg); transform-origin: center;"> Test Result 0001 05/21/14 11:14 .103 </div> <div style="border: 1px solid black; padding: 5px; transform: rotate(-90deg); transform-origin: center;"> No. Date Time ZBAC Calibration 05/26/05 10:40 .100 </div> <div style="border: 1px solid black; padding: 5px; transform: rotate(-90deg); transform-origin: center;"> 05/21/14 11:16 AS4 Serial no: 008877 Version no: 479? </div> </div> | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018615 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|--|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1115 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .099 | | | | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 11:17
 NS4 Serial no: 018615
 Version no: 4797
 No. Date Time %BAC
 Calibration 12/19/08 14:44 .100
 Test Result
 0001 05/21/14 11:17 .099
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#37

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018626 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|-----------------------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1110 | 34 | 14C13-10-5-10 C1307252 | .100 | 22 | | .102 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | |
| | | | Prepared by: Douglas C. Talbot | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/21/14 11:22
AS4 Serial no: 018626
Version no: 4797

No. Date Time %BNC
Calibration 04/29/05 08:59 .101

Test Result
0001 05/21/14 11:21 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018696 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------------------------------|---|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1121 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .108 | | | X | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 | | | | | | | | | | |
| | | | Expires on: 03-13-16 | | | | | | | | | | |

Push No. 3 to erase,
 or power down to retain

Test Result
 0001 05/21/14 11:23 .108

No. Date Time %BAC
 Calibration
 09/09/13 14:50 .100

05/21/14 11:24
 ASA Serial no: 018696
 Version no: 4797

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#39

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 083030 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|--------------------------------------|---------------------------|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|--|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | | |
| 05/21/2014 | 1150 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .100 | | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | |

05/21/14 11:52
 AS4 Serial no: 083030
 Version no: 4797
 No. Date Time %BAC
 Calibration 01/29/09 10:39 .100
 Test Result
 0001 05/21/14 11:52 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

40

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018692 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1153 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .100 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

05/21/14 11:56
AS4 Serial no: 018692
Version no: 4797

No. Date Time ZBAC
Calibration
08/12/08 14:35 .100

Test Result
0001 05/21/14 11:55 .100

Push No. 3 to erase;
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

42

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018762 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1150 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .102 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

05/21/14 11:58
AS4 Serial no: 018762
Version no: 4797
No. Date Time %BAC
Calibration 03/30/09 15:52 .100
Test Result
0001 05/21/14 11:57 .102
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018766 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|-------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 11:59 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .094 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

05/21/14 12:01
 ASD Serial no: 018766
 Version no: 4797
 No. Date Time %RAC
 Calibration 03/12/12 12:30 .100
 Test Result
 0001 05/21/14 12:00 .094
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

44

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018691 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1202 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .100 | | | X | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 | | | | | | | | | | |
| | | | Expires on: 03-13-16 | | | | | | | | | | |

05/21/14 12:04
 AS4 Serial no: 018691
 Version no: 4797
 No. Date Time %RAC
 Calibration
 09/08/08 14:01 .100
 Test Result
 0001 05/17/14 21:20 .154
 0002 05/17/14 21:22 .154
 0003 05/18/14 00:01 .163
 0004 05/18/14 00:02 .173
 0005 05/18/14 19:42 .131
 0006 05/18/14 19:47 .130
 0007 05/21/14 12:04 .106

Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 017990 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---------------------------|-------|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1205 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .105 | | | X | | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

05/21/14 12:10
 H54 Serial no: 017990
 Version no: 4797
 No. Date Time %MC
 Calibration
 11/26/09 11:10 .100
 Test Result
 0001 05/21/14 12:10 .105
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

46

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 018774 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|----|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | | NO |
| 05/21/2014 | 1200 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .099 | | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

Push No. 3 to erase
 Test Result
 0001 05/21/14 12:13 .099
 Calibration
 01/13/09 10:25 .100
 No. Date Time %BAC
 Calibration
 05/21/14 12:13
 05/21/14 12:13
 Version no: 4797
 05/21/14 12:13
 AS4 Serial no: 018774
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 041486 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|--------------------------------------|---------------------------|------------------|---------|--------------------------------|---------|-------------------------|-----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/21/2014 | 1211 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .106 | | | X | Lee | 17046 | |
| <p>Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16</p> | | | | | | | | | | | | | |
| <p style="text-align: center;"> or power down to retain Test Result 0001 05/21/14 12:15 .106 or power down to retain No. Date Time %DAC Calibration 03/08/06 13:56 .100 AS4 Serial no: 041486 Version no: 4797 05/21/14 12:16 </p> | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 041495 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|--|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1214 | 34 | 14C13-10-5-10 C1307252 | .100 | 20 | .100 | | | | X | | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 | | | | | | | | | | | |
| | | | Expires on: 03-13-16 | | | | | | | | | | | |
| | | | Push No. 3 to erase or power down to retain | | | | | | | | | | | |
| | | | Test Result | | | | | | | | | | | |
| | | | 0001 05/18/14 01:08 .063 | | | | | | | | | | | |
| | | | 0002 05/18/14 01:10 .061 | | | | | | | | | | | |
| | | | 0003 05/18/14 01:57 .020 | | | | | | | | | | | |
| | | | 0004 05/18/14 01:59 .026 | | | | | | | | | | | |
| | | | 0005 05/19/14 03:59 .188 | | | | | | | | | | | |
| | | | 0006 05/19/14 04:01 .216 | | | | | | | | | | | |
| | | | 0007 05/19/14 04:03 .216 | | | | | | | | | | | |
| | | | 0008 05/21/14 12:16 .100 | | | | | | | | | | | |
| | | | No. Date Time XPRC | | | | | | | | | | | |
| | | | Calibration | | | | | | | | | | | |
| | | | 02/20/14 11:28 .100 | | | | | | | | | | | |
| | | | 05/21/14 12:16 | | | | | | | | | | | |
| | | | AS4 Serial no: 041495 | | | | | | | | | | | |
| | | | Version no: 4797 | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 041502 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|---|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1217 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .090 | | | X | | Lee | 17046 | |
| <p>Prepared by: Douglas C. Talbot 03-13-14 Expires on: 03-13-16</p> | | | | | | | | | | | | | | |
| <p>Test Result 0001 05/16/14 19:05 .313 0002 05/17/14 20:52 .101 0003 05/17/14 20:56 .098 0004 05/17/14 23:27 .193 0005 05/17/14 23:29 .178 0006 05/21/14 12:20 .098</p> <p>No. Date Time :DMC Calibration 05/05/09 08:03 .100</p> <p>05/21/14 12:21 AS4 Serial no: 041502 Version no: 4797</p> <p>Push No. 3 to erase, or power down to retain</p> | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. 082654 |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|--------------------------------------|---|--|------------------|-----------------------|--------------------------------|-----------------------|-------------------------|----|--------------------------|----------|----------|-------------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1220 | 34 | 14C13-10-5-10 C1307252 | .100 | 21 | | .100 | | | X | | Lee | 17046 | |
| | | | Prepared by: Douglas C. Talbot 03-13-14 | | | | | | | | | | | |
| | | | Expires on: 03-13-16 | | | | | | | | | | | |
| | | | | <div style="border: 1px solid black; padding: 5px; display: inline-block;"> 05/21/14 12:22 NS4 Serial no: 082654 Version no: 4797 No. Date Time 2BAC Calibration 03/28/14 11:36 .100 Test Result 0001 05/21/14 12:22 .100 Push No. 3 to erase. </div> | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

SGT

| | | | | | |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Officer Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO. <i>008874</i> |
|--------------------------------|-----------------|--------------------------------------|----------------------|--|---------------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---------------------------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/21/2014 | 1223 | 34 | 14C13-10-5-10 C1307252 | .100 | 22 | | | 101 | | | X | Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | Prepared by: | | | | | | | | | | | |
| | | | Douglas C. Talbot | | | | | | | | | | | |
| | | | 03-13-14 | | | | | | | | | | | |
| | | | Expires on: | | | | | | | | | | | |
| | | | 03-13-16 | | | | | | | | | | | |

Push No. 3 to erase or power down to retain

Test Result
0001 05/20/14 02:29 .083
0002 05/20/14 02:31 .080
0003 05/21/14 12:24 .101

No. Date Time %BAC
Calibration
03/28/14 11:40 .100

05/21/14 12:26
AS4 Serial no: 008874
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.