

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
06/19/2015	0747	34	14F27-11-5-10								Lee	17046	
			C1307252										

```

06/19/15 07:48
AS4 Serial no: 077479
Version no: 4797

No. Date Time ZBAC
Calibration
02/17/14 09:10 .100

Test Result
0001 06/14/15 01:42 .216
0002 06/14/15 01:44 .214
0003 06/14/15 03:52 .167
0004 06/14/15 03:55 .169
0005 06/14/15 05:27 .125
0006 06/14/15 05:31 .127
0007 06/19/15 07:47 .099
    
```

Push No. 3 to erase.

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME

South Los Angeles

AREA NO.

530

COORDINATOR NAME

Officer Kirk Lee

I.D. NUMBER

17046

PAS MAKE/MODEL

Intoximeters Inc./Alco-Sensor IV

PAS SERIAL NO.

008709

#13

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
									TEST #1 (Optional)	TEST #2 (Optional)					
06/19/2015	0749	34	14F27-11-5-10	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008709	23	101	X		Lee	17046	
			C1307252												
Prepared by:															
Douglas C. Talbot															
06-27-14															
Expires on:															
06-27-16															
Push No. 3 to erase,															
06/19/15 07:52															
AS4 Serial no: 008709															
Version no: 4797															
or power down to retain															
No. Date Time XBAC Calibration															
02/17/14 09:27 .100															
Test Result															
0001 06/13/15 00:28 .089															
0002 06/13/15 00:31 .090															
0003 06/13/15 03:08 .102															
0004 06/13/15 03:11 .100															
0005 06/14/15 01:14 .019															
0006 06/14/15 01:56 .164															
0007 06/14/15 01:58 .157															
0008 06/14/15 04:13 .145															
0009 06/14/15 04:15 .170															
0010 06/14/15 04:17 .164															
0011 06/16/15 00:09 .013															
0012 06/16/15 04:49 .172															
0013 06/16/15 04:52 .181															
0014 06/18/15 03:00 .043															
0015 06/19/15 07:52 .101															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#14

AREA NAME: **South Los Angeles** AREA NO.: **530** COORDINATOR NAME: **Officer Kirk Lee** I.D. NUMBER: **17046** PAS MAKE/MODEL: **Intoximeters Inc./Alco-Sensor IV** PAS SERIAL NO.: **008714**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/19/2015			14F27-11-5-10		.100							Lee	17046	NOT IN LOG
			C1307252											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008883
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/19/2015	0753	34	14F27-11-5-10	.100	20	20	096	X			Lee	17046		
			C1307252											

or power down to retain

06/19/15 07:52
AS4 Serial no: 0000008
Version no: 4797

No. Date Time XBAC
Calibration
03/25/13 10:45 .100

Test Result
0001 06/19/15 07:51 .098

Push No. 3 to erase,

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 003783
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2015	0755	34	14F27-11-5-10	.100	20				X			Lee	17046	
			C1307252											

or power down to retain

06/19/15 07:58
AS4 Serial no: 008783
Version no: 4797

No. Date Time ZBAC
Calibration
07/18/08 13:52 .100

Test Result
0001 06/13/15 05:09 .000
0002 06/16/15 02:29 .144
0003 06/16/15 02:30 .142
0004 06/19/15 07:57 .101

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#10
 002657

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
06/19/2015	0757	34	14F27-11-5-10	.100	21						X	Lee	17046	
			C1307252											
Prepared by: Douglas C. Talbot														
Expires on: 06-27-14														
06-27-16														
Push No. 3 to erase. Test Result 0001 06/19/15 07:58 .100 Calibration 02/17/14 09:50 .100 No. Date Time ZBAC 06/19/15 07:59 AS4 Serial no: 002657 Version no: 4797 or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					YES
06/19/2015	0759	34	14F27-11-5-10		.100	20					X		Lee	17046		
			C1307252													
			Prepared by:													
			Douglas C. Talbot													
			06-27-14													
			Expires on:													
			06-27-16													

or power down to retain

06/19/15 08:03
AS4 Serial no: 008810
Version no: 4797

No. Date Time %BAC
Calibration
10/30/03 09:53 .100

Test Result
0001 06/19/15 08:02 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	0802-34		14F27-11-5-10	.100	22		103			X		Lee	17046	

06/19/15 08:04
 AS4 Serial no: 018684
 Version no: 4797

 No. Date Time XBAC
 Calibration
 11/03/11 16:03 .100

 Test Result
 0001 06/19/15 08:03 .103

 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

21

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Officer Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
 PAS SERIAL NO: 17046
 I.D. NO: 008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	0805	34	14F27-11-5-10 C1307252		.100	21		.098		X			Lee	17046	

06/19/15 08:08
 AS4 Serial no: 008635
 Version no: 4797

No. Date Time XBAC
 Calibration
 10/24/03 09:04 .100

Test Result
 0001 06/13/15 01:50 .087
 0002 06/13/15 01:52 .085
 0003 06/13/15 04:09 .146
 0004 06/13/15 04:12 .147
 0005 06/14/15 01:14 .000
 0006 06/19/15 08:07 .098

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
06/19/2015	0924	34	14F27-11-5-10	.100			22				Lee	17046	
			C1307252										

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

06/19/15 09:29
AS4 Serial no: 008845
Version no: 4797

No. Date Time ZBAC
Calibration
12/24/03 11:01 .100

Test Result
0001 06/13/15 02:25 .125
0002 06/13/15 02:28 .140
0003 06/14/15 04:09 .206
0004 06/14/15 04:11 .231
0005 06/14/15 04:13 .214
0006 06/14/15 18:42 .143
0007 06/14/15 18:45 .143
0008 06/19/15 04:53 .000
0009 06/19/15 09:27 .098

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 002653
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	VALUE	TEST #1	TEST #2 (Optional)	YES				
06/19/2015	0927	34	14F27-11-5-10		22					X		Lee	17046	
			C1307252											

06/19/15 09:30
MS4 Serial no: 882653
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 10:08 .100

Test Result
0001 06/13/15 11:55 .000
0002 06/19/15 03:18 .143
0003 06/19/15 03:20 .140
0004 06/19/15 09:29 .097

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#24

AREA NAME: South Los Angeles
AREA NO: 530
COORDINATOR NAME: Officer Kirk Lee
ID. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
PAS SERIAL NO.: 009848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	0829	34	14F27-11-5-10		.100	21							Lee	17046	
			C1307252												

06/19/15 09:34
AS4 Serial no: 008048
Version no: 4797
No. Date Time %BAC
Calibration
01/13/03 09:25 .100
Test Result
0001 06/19/15 09:34 .101
Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		P. RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
06/19/2015	0932	34	14F27-11-5-10 C1307252	.100	21	.100	X		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-27-14										
			Expires on:										
			06-27-16										

or power down to retain

06/19/15 09:38
AS4 Serial no: 077591
Version no: 4797

No. Date Time XBAC
Calibration
05/21/08 15:09 .100

Test Result
0001 06/19/15 09:37 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lec	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	0934	34	14F27-11-5-10 C1307252	21		101		X			Lee	17046	

```

06/19/15 09:42
AS4 Serial no: 008850
Version no: 4797

No. Date Time %BAC
Calibration
03/02/06 09:07 .100

Test Result
0001 06/17/15 11:36 .000
0002 06/19/15 09:42 .101

Push No. 3 to erase,
or power down to retain

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)					
06/19/2015	0936	34	14F27-11-5-10	.100	22	.095		X		Lee	17046	
			C1307252									

06/19/15 09:41
AS4 Serial no: 877573
Version no: 4797

No. Date Time ZBAC
Calibration
04/09/14 08:36 .000

Test Result

0001	06/09/15	01:36	.023
0002	06/09/15	01:38	.022
0003	06/09/15	02:51	.155
0004	06/09/15	02:53	.170
0005	06/11/15	03:00	.000
0006	06/11/15	21:08	.226
0007	06/11/15	21:10	.204
0008	06/11/15	21:11	.209
0009	06/11/15	21:13	.194
0010	06/18/15	23:04	.076
0011	06/18/15	23:09	.082
0012	06/19/15	02:30	.185
0013	06/19/15	02:32	.109
0014	06/19/15	02:38	.087
0015	06/19/15	02:39	.000
0016	06/19/15	09:41	.095

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Officer Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
 PAS SERIAL NO.: 003051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
06/19/2015	0943	34	14F27-11-5-10	.100	22				X	Lee	17046	
			C1307252									
Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16												
Test Result 0001 06/19/15 09:44 .099 Calibration 03/25/13 14:00 .100 No. Date Time ZBAC 06/19/15 09:44 A94 Serial no: 003051 Version no: 4797 or power down to retain												
Push No. 3 to erase.												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#30

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.						
South Los Angeles		530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	00 0897						
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
06/19/2015	0945	34	14F27-11-5-10	21	.100	21	.100	X		Lee	17046	
			C1307252									
Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16												
Push No. 3 to erase. or power down to retain 06/19/15 09:49 AS4 Serial no: 008897 Version no: 4797 No. Date Time XBAC Calibration 01/18/12 16:13 .100 Test Result 0001 06/13/15 02:41 .161 0002 06/13/15 02:45 .130 0003 06/13/15 02:47 .148 0004 06/13/15 04:30 .113 0005 06/13/15 04:33 .115 0006 06/13/15 21:41 .109 0007 06/13/15 21:44 .098 0008 06/14/15 01:02 .148 0009 06/14/15 01:06 .130 0010 06/14/15 02:40 .110 0011 06/14/15 02:43 .091 0012 06/19/15 01:57 .128 0013 06/19/15 01:59 .131 0014 06/19/15 09:49 .106												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lec	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

DIR 0624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																																																											
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	YES					NO																																																																																										
06/19/2015	1000	34	14F27-11-5-10	.100	22					X		Lee	17046																																																																																													
			C1307252																																																																																																							
			Prepared by: Douglas C. Talbot																																																																																																							
			06-27-14																																																																																																							
			Expires on: 06-27-16																																																																																																							
			<table border="0"> <tr> <td colspan="13"> <p>Test Result</p> <p>0001 06/09/15 01:48 .200</p> <p>0002 06/09/15 01:58 .198</p> <p>0003 06/09/15 01:59 .000</p> <p>0004 06/17/15 02:05 .030</p> <p>0005 06/19/15 10:08 .093</p> </td> </tr> <tr> <td colspan="13"> <p>Push No. 9 to erase.</p> </td> </tr> <tr> <td colspan="13"> <p>Calibration</p> <p>No. Date Time ZBAC</p> <p>10/07/08 09:19 .100</p> </td> </tr> <tr> <td colspan="13"> <p>Version no: 4797</p> </td> </tr> <tr> <td colspan="13"> <p>AS4 Serial no: 018624</p> </td> </tr> <tr> <td colspan="13"> <p>06/19/15 10:08</p> </td> </tr> <tr> <td colspan="13"> <p>or power down to retain</p> </td> </tr> </table>													<p>Test Result</p> <p>0001 06/09/15 01:48 .200</p> <p>0002 06/09/15 01:58 .198</p> <p>0003 06/09/15 01:59 .000</p> <p>0004 06/17/15 02:05 .030</p> <p>0005 06/19/15 10:08 .093</p>													<p>Push No. 9 to erase.</p>													<p>Calibration</p> <p>No. Date Time ZBAC</p> <p>10/07/08 09:19 .100</p>													<p>Version no: 4797</p>													<p>AS4 Serial no: 018624</p>													<p>06/19/15 10:08</p>													<p>or power down to retain</p>												
<p>Test Result</p> <p>0001 06/09/15 01:48 .200</p> <p>0002 06/09/15 01:58 .198</p> <p>0003 06/09/15 01:59 .000</p> <p>0004 06/17/15 02:05 .030</p> <p>0005 06/19/15 10:08 .093</p>																																																																																																										
<p>Push No. 9 to erase.</p>																																																																																																										
<p>Calibration</p> <p>No. Date Time ZBAC</p> <p>10/07/08 09:19 .100</p>																																																																																																										
<p>Version no: 4797</p>																																																																																																										
<p>AS4 Serial no: 018624</p>																																																																																																										
<p>06/19/15 10:08</p>																																																																																																										
<p>or power down to retain</p>																																																																																																										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICE ACCURACY CHECK/CALIBRATION LOG

#32

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Officer Kirk Lec I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV PAS SERIAL NO.: 018646

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
06/19/2015	1011	31	14F27-11-5-10 C1307252		.100		21						Lee	17046		

or power down to retain

06/19/15 10:25
AS4 Serial no: 018646
Version no: 4797

No. Date Time %BAC
Calibration
07/29/13 13:32 .100

Test Result

Push No. 9 to erase,
or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017498
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME	TEMP ("C)	SIMULATOR SOLUTION (OR DRY GAS)*	LOT NO.	VALUE	PAS TEMP ("C)		PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)		YES	NO				
06/19/2015	1020	34	14F27-11-5-10		.100	2		X				Lee	17046	
			C1307252											

06/19/15 10:24
AS4 Serial no: 077488
Version no: 4797

No. Date Time ZBAC
Calibration
05/21/08 14:35 .100

Test Result
0001 06/19/15 10:24 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

34

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Officer Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
PAS SERIAL NO.: 17046
008677

DATE	TIME (C)	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	1022	34	14F27-11-5-10		.100	21							Lee	17046	

06/19/15 10:30
AS4 Serial no: 008877
Version no: 4797

No. Date Time %BAC
Calibration
05/26/05 10:40 .100

Test Result
0001 06/19/15 10:30 .102

Push No. 3 to erase.

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 019615
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME	TEMP (°C)		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEST #1	TEST #2 (Optional)		TEST #1	TEST #2 (Optional)	YES	NO	YES	NO				
06/19/2015	1024	34		.100	22				X			Lee	17046	
SIMULATOR SOLUTION (OR DRY GAS)** LOT NO. 14F27-11-5-10 CI1307252 Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16														
Push No. 3 to erase. Test Result 0001 06/19/15 10:30 .005 0002 06/19/15 10:30 .098 Calibration 12/19/08 14:44 .100 No. Date Time XBAC AS4 Serial no: 018615 Version no: 4797 or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 093030
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/19/2015		14F27-11-5-10		.100							Lee	17046	NOT IN VEHICLE
		C1307252											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER I7046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018092
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/19/2015	1002-34	14F27-11-5-10		.100	21	098			X		Lee	17046	
		C1307252											
		Prepared by: Douglas C. Talbot											
		Expires on: 06-27-14											
		Push No. 3 to erase.											
		Test Result 0001 06/19/15 10:11 .098											
		Calibration 08/12/08 14:35 .100											
		No. Date Time XBAC AS4 Serial no: 818692 Version no: 4797											
		or Power down to retain											
		06/19/15 10:11											

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#412

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. D19 762
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
06/19/2015	1043	34	14F27-11-5-10	.100	21	.100			X		Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-27-14										
			Expires on:										
			06-27-16										
<p>06/19/15 10:43 AS4 Serial no: 018762 Version no: 4797</p> <p>No. Date Time XBAC Calibration 03/30/09 15:52 .100</p> <p>Test Result 0001 06/19/15 10:42 .100</p> <p>Push No. 3 to erase, or power down to retain</p> <p>or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 019766
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME (C)	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	1045	34	14F27-11-5-10	.100	RI					X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

06/19/15 10:46
 AS4 Serial no: 018766
 Version no: 4797

No. Date Time ZBAC
 Calibration
 09/12/14 13:45 .100

Test Result
 0001 06/19/15 10:45 .097

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 012091
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/19/2015	1047	34	14F27-11-5-10		.100	TEST #1	TEST #2	TEST #1	TEST #2				Lee	17046	
			C1307252							X					

06/19/15 10:49
 AS4 Serial no: 018691
 Version no: 4797

No. Date Time %BAC
 Calibration
 09/08/08 14:01 .100

Test Result
 0001 06/19/15 10:49 .104

Push No. 3 to erase:

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEST #1	TEST #2 (Optional)			TEST #1	TEST #2 (Optional)	YES	NO	YES	NO				
06/19/2015	1049	34		14F27-11-5-10	.100	22		106		X			Lee	17046	
				C1307252											
Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16															

or power down to retain

06/19/15 10:50
 AS4 Serial no: 017990
 Version no: 4797

No. Date Time XBAC
 Calibration
 11/26/03 11:10 .100

Test Result
 0001 06/19/15 10:49 .106

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 010774
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					YES
06/19/2015	1051	34	14F27-11-5-10		.100								Lee	17046		
			C1307252													

or power down to retain

06/19/15 10:53
AS4 Serial no: 010774
Version no: 4797

No. Date Time %BAC
Calibration
01/13/09 10:25 .100

Test Result

0001	06/09/15	22:01	.070
0002	06/09/15	22:03	.074
0003	06/10/15	02:36	.200
0004	06/10/15	02:40	.202
0005	06/10/15	02:51	.000
0006	06/16/15	02:10	.094
0007	06/16/15	02:12	.093
0008	06/17/15	01:32	.061
0009	06/17/15	01:35	.059
0010	06/17/15	02:18	.216
0011	06/17/15	02:22	.208
0012	06/18/15	00:50	.123
0013	06/18/15	00:52	.104
0014	06/19/15	10:51	.102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2	TEST #1 (Optional)	TEST #2 (Optional)				
06/19/2015	1051	34	14F27-11-5-10	.100	21	TEST #1	TEST #2				Lee	17046	
			C1307252						X				

06/19/15 11:04
 AS4 Serial no: 041486
 Version no: 4797

No. Date Time %BAC
 Calibration
 03/08/06 13:56 .100

Test Result
 0001 06/19/15 10:59 .106

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

04/495

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2	TEST #1	TEST #2	TEST #1	TEST #2				
06/19/2015	1053	34	14F27-11-5-10	.100						X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

06/19/15 11:05
AS4 Serial no: 041495
Version no: 4797
No. Date Time ZBAC
Calibration
02/20/14 11:28 .100
Test Result
0001 06/19/15 11:04 .099
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2015	1055	34	14F27-11-5-10	.100	21							Lee	17046	
			C1307252							X				
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

Push No. 3 to erase,
or power down to retain

06/19/15 11:08
AS4 Serial no: 041502
Version no: 4797

No. Date Time %BAC
Calibration
05/05/09 08:03 .100

Test Result
0001 05/29/15 23:11 .128
0002 05/29/15 23:14 .136
0003 05/30/15 05:00 .158
0004 05/30/15 05:02 .169
0005 06/01/15 03:19 .125
0006 06/01/15 03:23 .119
0007 06/05/15 00:08 .171
0008 06/05/15 00:11 .169
0009 06/06/15 01:51 .166
0010 06/06/15 01:55 .148
0011 06/06/15 23:55 .039
0012 06/07/15 00:58 .000
0013 06/07/15 00:58 .000
0014 06/07/15 02:56 .130
0015 06/07/15 02:59 .116
0016 06/08/15 02:28 .162
0017 06/08/15 02:30 .160
0018 06/12/15 19:54 .172
0019 06/12/15 19:57 .168
0020 06/13/15 22:31 .035
0021 06/13/15 23:48 .000
0022 06/13/15 23:48 .000
0023 06/14/15 00:17 .140
0024 06/14/15 00:20 .153
0025 06/15/15 03:04 .132
0026 06/15/15 03:06 .130
0027 06/19/15 11:08 .096

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

DB 2654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/19/2015	1105	34	14F27-11-5-10	.100	21	.098				X		Lee	17046	
			C1307252											

06/19/15 11:10
AS4 Serial no: 082654
Version no: 4797

No. Date Time %BAC
Calibration
03/28/14 11:36 .100

Test Result
0001 06/19/15 11:10 .098

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.