

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008501
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1035	34	14F23-11-5-10 C1307252	.100	21	.096				X	Lee	17046	

Test Result
 0001 08/26/14 02:14 .010
 0002 08/26/14 02:17 .018
 0003 08/28/14 16:39 .096
 Calibration
 03/20/09 15:40 .100
 No. Date Time ZBAC
 08/28/14 18:40
 NCA Serial no: 008501
 Version no: 4797

Push the 3 to green
 or power down to reload

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1038	34	14F23-11-5-10 C1307252	.100	21		.099			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Each No. 5 to ensure
 of proper steps in refilling

Test Result
 0001 08/28/14 10:43 099
 No. Date Time ZBAC
 Calibration 07/18/08 13:27 106
 08/28/14 10:42
 No. Serial no: 018785
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1041	34	14F23-11-5-10	.100	22		.101			X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
 0001 08/23/14 09:30 .000
 0002 08/23/14 09:47 .000
 0003 08/24/14 01:46 .000
 0004 08/25/14 04:27 .017
 0005 08/28/14 10:44 .101

 No. Date Time ZONE
 Calibration 08/28/13 16:36 .100

 08/28/14 10:44
 NS4 Serial no: 008546
 Version no: 4797

 Push No. 0 to cross,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008580
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1055	34	14F23-11-5-10 C1307252	.100	21		1099			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>														
<p>Test Result 0001 08/28/14 10:59 .100</p> <p>Mo. Date Time ZHCO Calibration 03/06/14 12:39 .100</p> <p>08/28/14 10:59 NS4 Serial no: 000380 Version no: 4797</p> <p>Dist. No. 3 to correct or power down to reload</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 093032
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1057	34	14F23-11-5-10	.100	21		.100			X	Lee	17046	
			C1307252										

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Push No. 3 to erase,
or power down to retain.

 Test Result
 0001 08/28/14 00:05 .145
 0002 08/28/14 00:09 .139
 0003 08/28/14 01:09 .090
 0004 08/28/14 10:56 .100

 No. Date Time XMP
 Calibration
 02/17/14 09:01 .100

 08/28/14 10:56
 PAS Serial no: 083032
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1059	34	14F23-11-5-10 C1307252	.100	21		098			X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

08/28/14 11:07
NS4 Serial num 008620
Version num 4097
No. Date Time ZMM
Calibration
01/30/13 14:04 100

Test Result
0001 08/28/14 11:26 .098
0002 08/28/14 13:15 .065
0003 08/28/14 13:17 .092
0004 08/28/14 13:18 .072
0005 08/28/14 13:26 .122
0006 08/28/14 13:22 .076
0007 08/28/14 13:23 .056
0008 08/28/14 13:29 .119
0009 08/28/14 13:53 .139
0010 08/28/14 13:55 .102
0011 08/28/14 13:56 .002
0012 08/28/14 13:57 .116
0013 08/28/14 14:03 .070
0014 08/28/14 14:07 .004
0015 08/28/14 14:07 .002
0016 08/28/14 14:11 .004
0017 08/28/14 14:13 .017
0018 08/28/14 14:16 .102
0019 08/28/14 14:17 .124
0020 08/28/14 14:19 .121
0021 08/28/14 14:37 .069
0022 08/28/14 14:38 .062
0023 08/28/14 15:03 .138
0024 08/28/14 15:05 .069
0025 08/28/14 15:08 .110
0026 08/28/14 15:10 .072
0027 08/28/14 15:13 .094
0028 08/28/14 15:17 .080
0029 08/28/14 15:19 .080
0030 08/28/14 15:45 .101
0031 08/28/14 15:47 .130
0032 08/28/14 15:48 .173
0033 08/28/14 15:52 .119
0034 08/28/14 16:00 .139
0035 08/28/14 16:02 .067
0036 08/28/14 16:04 .111
0037 08/28/14 16:06 .119
0038 08/28/14 16:08 .068
0039 08/28/14 17:04 .173
0040 08/28/14 17:06 .099
0041 08/28/14 17:08 .128
0042 08/28/14 17:10 .061
0043 08/28/14 19:44 .084
0044 08/28/14 19:46 .059
0045 08/28/14 02:02 .095
0046 08/28/14 02:20 .066
0047 08/28/14 04:01 .141
0048 08/28/14 04:07 .138
0049 08/28/14 11:02 .094

Pencil No. 9 - to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1103	34	14F23-11-5-10 C1307252	.100	21	101			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

No. Date Time 2100
 Calibration 02/17/14 08:52.106
 Test Result 0001 08/28/14 11:04.161
 Push No. 3 to erase or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014			14F23-11-5-10 C1307252	.100							Lee	17046	B.D. OUT FOR SERVICE

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
08/28/2014	1105	34	14F23-11-5-10	.100	20		.099				X	Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Just No. 3 to erase
or power down to retain
 No. Date Time %HOC
 Calibration 02/17/14 09:18 .100
 Test Result 0001 08/28/14 11:05 .099

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1107	34	14F23-11-5-10 C1307252	.100	21		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

08/28/14 11:08
 NS4 Serial not 082660
 Version not 4797

 No. Date Time ZONE
 Calibration 02/17/14 09:16 100

 Test Result
 0001 08/24/14 00056 0.00
 0002 08/24/14 00058 1.00
 0003 08/24/14 02147 1.00
 0004 08/24/14 02151 1.00
 0005 08/24/14 23117 0.09
 0006 08/24/14 23116 0.02
 0007 08/24/14 23116 0.02
 0008 08/25/14 02121 0.00
 0009 08/25/14 02122 0.00
 0010 08/28/14 11107 1.00

Push Ho. 3 to erase
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1111	34	14F23-11-5-10 C1307252	.100	22		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

08/28/14 11:13
AS4 Serial num 008709
Version num 4797

No. Date Time ZBAC
Calibration
02/17/14 09:27 .100

Test Result
0001 08/24/14 17:34 .196
0002 08/24/14 17:36 .242
0003 08/24/14 17:38 .289
0004 08/28/14 11:17 .101
0005 08/28/14 11:18 .100

Push No. 3 to erase
or power down to refresh

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008 714
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																								
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO																																																												
08/28/2014	1115	34	14F23-11-5-10 C1307252	.100	21	-104		X		Lee	17046																																																									
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>																																																																				
<p>Test Result</p> <table border="1"> <tr> <th>No.</th> <th>Date</th> <th>Time</th> <th>Z/N/C</th> </tr> <tr> <td>0001</td> <td>08/23/14</td> <td>01:28</td> <td>.146</td> </tr> <tr> <td>0002</td> <td>08/23/14</td> <td>01:31</td> <td>.132</td> </tr> <tr> <td>0003</td> <td>08/23/14</td> <td>02:44</td> <td>.210</td> </tr> <tr> <td>0004</td> <td>08/23/14</td> <td>03:46</td> <td>.216</td> </tr> <tr> <td>0005</td> <td>08/24/14</td> <td>02:13</td> <td>.071</td> </tr> <tr> <td>0006</td> <td>08/24/14</td> <td>02:19</td> <td>.071</td> </tr> <tr> <td>0007</td> <td>08/24/14</td> <td>03:03</td> <td>.091</td> </tr> <tr> <td>0008</td> <td>08/24/14</td> <td>03:03</td> <td>.082</td> </tr> <tr> <td>0009</td> <td>08/24/14</td> <td>21:28</td> <td>.048</td> </tr> <tr> <td>0010</td> <td>08/24/14</td> <td>21:30</td> <td>.094</td> </tr> <tr> <td>0011</td> <td>08/24/14</td> <td>21:32</td> <td>.033</td> </tr> <tr> <td>0012</td> <td>08/24/14</td> <td>22:41</td> <td>.031</td> </tr> <tr> <td>0013</td> <td>08/28/14</td> <td>11:21</td> <td>.104</td> </tr> </table> <p>Push No. 3 to erase, or power down to retain</p>													No.	Date	Time	Z/N/C	0001	08/23/14	01:28	.146	0002	08/23/14	01:31	.132	0003	08/23/14	02:44	.210	0004	08/23/14	03:46	.216	0005	08/24/14	02:13	.071	0006	08/24/14	02:19	.071	0007	08/24/14	03:03	.091	0008	08/24/14	03:03	.082	0009	08/24/14	21:28	.048	0010	08/24/14	21:30	.094	0011	08/24/14	21:32	.033	0012	08/24/14	22:41	.031	0013	08/28/14	11:21	.104
No.	Date	Time	Z/N/C																																																																	
0001	08/23/14	01:28	.146																																																																	
0002	08/23/14	01:31	.132																																																																	
0003	08/23/14	02:44	.210																																																																	
0004	08/23/14	03:46	.216																																																																	
0005	08/24/14	02:13	.071																																																																	
0006	08/24/14	02:19	.071																																																																	
0007	08/24/14	03:03	.091																																																																	
0008	08/24/14	03:03	.082																																																																	
0009	08/24/14	21:28	.048																																																																	
0010	08/24/14	21:30	.094																																																																	
0011	08/24/14	21:32	.033																																																																	
0012	08/24/14	22:41	.031																																																																	
0013	08/28/14	11:21	.104																																																																	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1118	34	14F23-11-5-10 C1307252	.100	22		.100		X			Lee	17046	
														Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16
														Push No. 3 to erase, or power down to retain Test Result 0001 08/28/14 02:14 .164 0002 08/28/14 02:18 .167 0003 08/24/14 01:31 .806 0004 08/24/14 01:35 .806 0005 08/28/14 11:14 .100 No. Date Time 2010 Calibration 08/25/13 10:45 .100 08/28/14 11:15 NS4 Serial no: 60496 Version no: 4297

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 003783
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1120	34	14F23-11-5-10 C1307252	.100	22		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

08/28/14 11:26
ASA Serial: not 000783
Version: not 4797

No. Date Time ZBAC
Calibration 07/18/08 13:52 .100

Test Result
0001 08/25/14 21:56 .203
0002 08/25/14 21:59 .192
0003 08/27/14 01:32 .104
0004 08/27/14 01:39 .106
0005 08/28/14 11:25 .100

Push No. 3 to erase
or power down to retrain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1126	34	14F23-11-5-10 C1307252	.100	21		.100			X	Lee	17046	
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014			14F23-11-5-10	.100								Lee	17046	NOT IN LOCKER
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1134	34	14F23-11-5-10 C1307252	.100	22		101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

08/28/14 11:40
 NS4 Serial no: 018684
 Version no: 4797
 No. Date Time ZBMC
 Calibration
 11/03/11 16:00 .100
 Test Result
 0001 08/28/14 11:39 .100
 Push No. 3 to erase
 or power down to reload

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1138	34	14F23-11-5-10	.100	21		.102			X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Test Result
 AM01 08/28/14 11:44 102
 No. Date Time ZBMC
 Calibration
 10/24/03 09:04 100
 08/28/14 11:43
 (254 Serial) not 008635
 location not 4297

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1140	34	14F23-11-5-10	.100	21		.103			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

08/28/14 11:45
 094 Serial no: 008845
 Version no: 4797
 No. Date Time %BAC
 Calibration 12/24/09 11:01 .105
 Test Result
 0001 08/22/14 23:19 .053
 0002 08/22/14 23:19 .064
 0003 08/24/14 01:56 .229
 0004 08/24/14 01:58 .294
 0005 08/28/14 11:45 .103

Push No. 5 to erase
 or power down to regain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 00BB45
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014			14F23-11-5-10	.100							Lee	17046	NOT IN LOCKER
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1149	34	14F23-11-5-10 C1307252	.100	21	.101			X		Lee	17046		
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16											
			<div style="border: 1px solid black; padding: 5px;"> <p>08/28/14 11:51 [654 serial] not 082653 [4797] not 4797</p> <p>No. Date Time ZBNC Calibration 02/17/14 10:00 100</p> <p>Test Result 0001 08/24/14 07:44 100 0007 08/24/14 07:41 117 0008 08/24/14 07:47 123 0004 08/24/14 04:16 134 0005 08/24/14 04:20 126 0006 08/24/14 19:23 118 0007 08/24/14 19:27 100 0009 08/24/14 21:24 102 0009 08/24/14 21:27 100 0010 08/24/14 23:43 198 0011 08/24/14 23:49 181 0012 08/28/14 11:56 101</p> <p>Push No. 3 to erase, or power down to reclaim.</p> </div>											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1151	34	14F23-11-5-10 C1307252	.100	21		.102				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Test Result
 1001 08/28/14 11:56 .102
 Ho. Date Time ZIMC
 Calibration 01/18/09 09:25 .100
 Version not 4/97
 08/28/14 11:56
 164 Serial not 000000
 Version not 4/97

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1153	34	14F23-11-5-10 C1307252	.100	20	.101				X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>														
<p>Test Result 08/28/14 11:52 .101</p> <p>Calibration 05/21/88 15:09 .100</p> <p>No. Date Time XPNC 05/21/88 15:09 .100</p> <p>08/28/14 11:52 084 Ser 141 No: 077591 Version No: 4297</p> <p>Push No. 3 to erase or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 003850
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1150	34	14F23-11-5-10 C1307252	.100	21		.099			X	Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
08/28/14 15:25 .151
08/27/14 15:28 .152
08/26/14 12:05 .099

Test Result
08/27/14 15:25 .151
08/27/14 15:28 .152
08/26/14 12:05 .099

Push No. 3 to erase
or power down to return

08/28/14 12:06
(54 Serial) not (08/28/14)
Calibration not 4797

08/02/06 09:02 .100

08/28/14 12:06
(54 Serial) not (08/28/14)
Calibration not 4797

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
08/28/2014	1158	34	14F23-11-5-10 C1307252	.100	22		095			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			06-23-14												
			Expires on:												
			06-23-16												
				Push No. 3 to screen or power down to retrain											
				Test Result 0001 08/26/14 02:16 .125 0002 08/26/14 02:17 .126 0003 08/27/14 01:41 .000 0004 08/28/14 11:59 .095											
				No. Date Time ZPNOC Calibration 04/09/14 00:36 .000											
				08/28/14 11:59 not (serial) not 077573 version not 4797											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1144	34	14F23-11-5-10 C1307252	.100	22		.098			X	Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

08/28/14 11:45
084 Serial no: 803051
Version no: 4797

No. Date Time ZIMC
Calibration
08/25/13 14:08 100

Test Result
0001 08/28/14 11:45 100

Push Ho. 5 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008897

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1202	34	14F23-11-5-10	.100	21		-104		X		Lee	17046		
			C1307252											
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16														

Test Result

No. 0001	Date 08/16/14	Time 02:38	ZINC	0.099
0002	08/16/14	02:39	ZINC	0.096
0003	08/16/14	04:08	ZINC	0.092
0004	08/16/14	04:08	ZINC	0.086
0005	08/16/14	04:08	ZINC	0.066
0006	08/19/14	02:44	ZINC	0.145
0007	08/17/14	02:48	ZINC	0.143
0008	08/23/14	03:08	ZINC	0.167
0009	08/23/14	03:10	ZINC	0.149
0010	08/23/14	03:19	ZINC	0.154
0011	08/24/14	06:01	ZINC	0.082
0012	08/24/14	06:08	ZINC	0.108
0013	08/24/14	04:52	ZINC	0.074
0014	08/24/14	02:25	ZINC	0.087
0015	08/24/14	02:48	ZINC	0.099
0016	08/24/14	02:48	ZINC	0.096
0017	08/25/14	02:12	ZINC	0.117
0018	08/25/14	02:18	ZINC	0.109
0019	08/25/14	04:05	ZINC	0.212
0020	08/25/14	04:08	ZINC	0.277
0021	08/28/14	12:18	ZINC	0.104

Push No. 3 to access
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1205	34	14F23-11-5-10 C1307252	.100	22		0.095			X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

10/28/14 12:09
NSA Serial No: 018624
Version: Not 4797
No. Date Time ZMAC
Calibration
10/07/08 09:10 100
Test Result
0001 08/28/14 12:09 .095
Push No. 8 to erase
or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1208	34	14F23-11-5-10 C1307252	.100	22		.099		X		Lee	17046	

08/28/14 12:13
AS4 Serial no: 018646
Version: not 4797

No. Date Time %MC
Calibration
07/29/13 18:39 100

Test Result
0001 08/28/14 12:12 .100

Pack No. 3 to cross
or power down to replace

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1210	34	14F23-11-5-10	.100	21		.102			X	Lee	17046	
			C1307252										
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16													

Test Result
 08/28/14 12:11
 (64 Serial) no: 077488
 Version no: 4797
 No. Date Time 2700
 Calibration
 05/21/08 14:05 .100
 08/28/14 12:10 10/
 Push loc. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*			CALIBRATION REQUIRED		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS				
08/28/2014	1212	34	14F23-11-5-10 C1307252	.100	21	.103				X		Lee	17046		

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Push No. 3 to erase
or power down to refresh

Test Result
0001 08/28/14 12:19 .100

No. Date Time YMAC
Calibration
05/26/05 10:40 .100

08/28/14 12:19
AS4 Serial no: 008877
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
08/28/2014	1214	31	14F23-11-5-10 C1307252	.100	22	.098			X		Lee	17046		

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
0001 08/28/14 12:14 .100
 Calibration
12/19/08 14:44 .100
 No. Date Time 2BAC
 60/28/14 12:10
 154 Serial no: 018615
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1216	34	14F23-11-5-10 C1307252	.100	22		.100			X	Lee	17046	

Prepared by:
 Douglas C. Talbot
 06-23-14
 Expires on:
 06-23-16

08/28/14 12:19
 NS4 Serial no: 018626
 Version no: 4797
 No. Date Time XMC
 Calibration 04/29/08 08:59 .100
 Test Result
 #091 08/28/14 12:19 .100
 Just No. 3 to erase
 or power down to return

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	12:19	34	14F23-11-5-10 C1307252	.100	22		.104				X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>														
<p>Test Result 0091 08/28/14 12:20 .104</p> <p>Calibration 09/09/13 14:50 .100</p> <p>No. Date Time 2100 Calibration 09/09/13 14:50 .100</p> <p>Def. Ion Not 4797 08/28/14 12:21 Def. Serial Not 018696 Def. Ion Not 4797</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
08/28/2014	1221	34	14F23-11-5-10 C1307252	.100	21	-101			X	Lee	17046	
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16												

Push Ho. 5 to power
or power down to return

Test Result
0001 08/28/14 12:22 .100

Calibration
01/29/09 16:39 .100

No. Date Time ZBNC
Version Date 4797
08/28/14 12:22
104 Serial num 083030

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1146	34	14F23-11-5-10 C1307252	.100	22		.099				Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

08/28/14 11:50
 NS4 Serial no: 018692
 Version no: 4797
 No. Date Time ZMTC
 Calibration 08/12/08 14:30 .100
 Test Result
 0001 08/28/14 1146 .099
 Push No. 3 to screen
 or power down to reload

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1223	34	14F23-11-5-10 C1307252	.100	21		.102			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											
<p style="text-align: center;"> Test Result 0.101 08/28/14 12:23 100 No. Date Time %MC Calibration 08/26/09 10:52 100 Version Not Applicable 08/20/14 12:50 NS4 Serial Not 018762 Version Not Applicable Wash No. 0 to 0.000 or lower down to 0.000 </p>														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1229	34	14F23-11-5-10 C1307252	.100	21		.096			X	Lee	17046	
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16													
<pre> Test Result 0001 08/28/14 02:58 .000 0002 08/28/14 03:20 .000 0003 08/28/14 17:32 .000 Push bot. & to screen or power down to retain No. Date Time ZIMC Calibration 08/12/12 12:39 .100 08/28/14 12:39 HSA Serial not 018766 Version not 479P </pre>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014	1232	30	14F23-11-5-10 C1307252	.100	22		.102			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>													
<p>Test Result</p> <p>No. Date Time ZHNC Calibration 09/08/08 14:00 .100</p> <p>08/28/14 12:38 004 Serial not 018691 Version not 4797</p> <p>Test No. 3 to erase, or power down to retain</p> <p>0001 08/28/14 08:51 .045 0002 08/28/14 09:08 .040 0003 08/28/14 12:37 .102</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
08/28/2014	1236	34	14F23-11-5-10 C1307252	.100	22		-.099			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			06-23-14												
			Expires on:												
			06-23-16												

Test No. 3 to pass.
 or power down to retain
 Test Result
 0001 08/26/14 22:20 .100
 0002 08/26/14 22:23 .129
 0003 08/28/14 12:39 .099
 No. Date Time Zone
 Calibration 01/13/09 10:22 .100

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. <i>041486</i>
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
08/28/2014			14F23-11-5-10 C1307252	.100							Lee	17046	<i>B.O.</i>

Prepared by:
 Douglas C. Talbot
 06-23-14
 Expires on:
 06-23-16

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1239	34	14F23-11-5-10 C1307252	.100	20		.100				X	Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
08/28/14 12:39 .100

08/28/14 11:20 .100

No. Date Time SINC
Calibration
08/28/14 11:20 .100

08/28/14 12:39
NS4 Serial no: 041495
Version no: 4797

Push Ho. 3 to erase
or power down to return

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

52

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	041502

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	1241	34	14F23-11-5-10 C1307252	.100	21		09B		X		Lee	17046		
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Push No. 3 to check of eqpt down to retain

- Test Result
- 0001 08/23/14 03:04 .102
- 0002 08/23/14 03:06 .104
- 0003 08/23/14 04:25 .097
- 0004 08/23/14 04:28 .094
- 0005 08/23/14 06:24 .104
- 0006 08/23/14 06:27 .105
- 0007 08/24/14 02:20 .065
- 0008 08/24/14 02:27 .066
- 0009 08/24/14 02:27 .000
- 0010 08/24/14 19:57 .115
- 0011 08/24/14 19:59 .112
- 0012 08/24/14 21:05 .100
- 0013 08/25/14 08:03 .110
- 0014 08/25/14 09:37 .100
- 0015 08/25/14 09:39 .179
- 0016 08/28/14 12:45 .098

No. Date Time ZONE
 Calibration
 05/05/09 09:00 .100

08/28/14 12:45
 AEA Serial no: 041502
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
08/28/2014	12:43	34	14F23-11-5-10 C1307252	.100	22		.099				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											
			Fresh No. 3 to erase or power down to retain											
							Test Result							
							0901 08/25/14 09:22 .132							
							0907 08/25/14 09:23 .127							
							0908 08/26/14 12:46 .099							
							No. Date Time S/MC							
							Calibration							
							09/28/14 11:30 .106							
							Version not done							
							08/28/14 12:46							
							PAS Serial not done							
							082654							
							Version not done							

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

SGT.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
08/28/2014			14F23-11-5-10	.100							Lee	17046	NOT IN LOCKER
			C1307252										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.