

SGT's PAS

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. <i>8874</i>
--------------------------------	-----------------	---	----------------------	---	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
09/12/2014	0831	<i>34</i>	14F23-11-5-10- C1307252	.100	<i>22</i>	<i>097</i>				<i>X</i>		Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p style="text-align: center;"> No. Date Time %BAC Calibration 03/28/14 11:48 .100 Test Result 0001 08/30/14 15:20 .000 0002 08/31/14 16:58 .000 0003 08/31/14 16:59 .000 0004 09/12/14 08:31 .097 </p> <p style="text-align: center;"> Push No. 3 to erase, or power down to retain </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8501
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	08:38	34	14F23-11-5-10- C1307252	.100	20	0.094				X		Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 08:40
 NS4 Serial no: 006501
 Version no: 4797
 No. Date Time %BAC
 Calibration
 09/30/09 15:40 .100
 Test Result
 0001 09/12/14 08:39 .094
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18785
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0841	34	14F23-11-5-10- CI307252	.100	22		.099			X	Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16													
<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p style="text-align: center;"> 09/12/14 08:42 AS4 Serial no: 018795 Version no: 4797 No. Date Time %BAC Calibration 07/18/08 13:27 .100 Test Result 0001 09/12/14 08:42 .099 Push No. 3 to erase, or power down to retain </p> </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8546
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0843	34	14F23-11-5-10- C1307252	.100	22		.100			X		Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p style="text-align: center;"> 09/12/14 08:45 AG4 Serial no: 000546 Version no: 4797 No. Date Time %BAC Calibration 06/28/13 16:36 .100 Test Result 0001 09/06/14 01:16 .127 0002 09/06/14 01:19 .126 0003 09/12/14 08:44 .100 Push No. 3 to erase, or Power down to retain </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82659
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0857	34	14F23-11-5-10- C1307252	.100	21		.097				X	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 08:52
 AS4 Serial no: 082659
 Version no: 4797
 No. Date Time %BAC
 Calibration 09/28/14 09:04 .100
 Test Result
 0001 09/12/14 08:51 .097
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8559
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0900	34	14F23-11-5-10- C1307252	.100	20		101					Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16														
<pre> 09/12/14 09:02 AS4 Serial no: 008559 Version no: 4797 No. Date Time ZBAC Calibration 11/30/11 08:55 .106 Test Result 0001 09/12/14 09:01 .101 Push No. 3 to erase, or Power down to retain </pre>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8580
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0859	34	14F23-11-5-10- C1307252	.100	21		.098			Y	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16										

09/12/14 09:00
 AS4 Serial no: 000580
 Version no: 4797
 No. Date Time ZBAC
 Calibration 03/06/14 12:39 .100
 Test Result
 0001 09/12/14 08:59 .098
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 83032
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0855	34	14F23-11-5-10- C1307252	.100	21	.100						Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											

09/12/14 08:56
 H54 Serial no: 003032
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/17/14 09:01 .100
 Test Result
 0001 09/09/14 22:02 .100
 0002 09/09/14 22:05 .099
 0003 09/10/14 02:39 .065
 0004 09/10/14 02:40 .064
 0005 09/11/14 22:09 .134
 0006 09/11/14 22:11 .128
 0007 09/12/14 08:56 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 81620
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0906	34	14F23-11-5-10- C1307252	.100	21		.098				∅	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

```

09/12/14 09:08
RS4 Serial no: 008620
Version no: 4797

No. Date Time %BAC
Calibration
01/30/13 14:34 .100

Test Result
0001 09/09/14 10:12 .000
0002 09/09/14 20:31 .009
0003 09/10/14 01:02 .179
0004 09/10/14 01:04 .151
0005 09/10/14 01:09 .084
0006 09/10/14 05:17 .095
0007 09/10/14 05:20 .093
0008 09/12/14 09:07 .098

Push No. 3 to erase,
or Power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82957
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0903	34	14F23-11-5-10-C1307252	.100	21		.101				Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16										

09/12/14 09:04
 AG4 Serial no: 082957
 Version no: 4797

No. Date Time XBRAC
 Calibration 02/17/14 08:52 .100

Test Result
 0001 09/12/14 09:04 .101

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 77505 ¹⁰
--------------------------------	-----------------	---	----------------------	---	---------------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	-	-	14F23-11-5-10- C1307252	.100								Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. <i>11</i> 77479
--------------------------------	-----------------	---	----------------------	---	--------------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0906	34	14F23-11-5-10- C1307252	.100	20		.099			X	Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16													
09/12/14 09:07 AS4 Serial no: 077479 Version no: 4737													
No. Date Time %BAC Calibration 02/17/14 09:10 .100													
Test Result 0001 09/05/14 23:03 .167 0002 09/05/14 23:05 .164 0003 09/06/14 23:40 .203 0004 09/06/14 23:42 .210 0005 09/07/14 02:41 .119 0006 09/07/14 02:44 .121 0007 09/07/14 03:21 .121 0008 09/07/14 03:24 .125 0009 09/07/14 20:14 .152 0010 09/07/14 20:16 .159 0011 09/07/14 23:20 .054 0012 09/07/14 23:22 .063 0013 09/08/14 01:40 .126 0014 09/08/14 01:44 .183 0015 09/08/14 01:49 .116 0016 09/12/14 09:07 .099													
Push No. 3 to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82660
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0921	34	14F23-11-5-10- C1307252	.100	20	0.102						Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 09:22
AS4 Serial no: 082660
Version no: 4797
No. Date Time ZBAC
Calibration 02/17/14 09:16 .100
Test Result
0001 09/07/14 00:27 .125
0002 09/07/14 00:29 .115
0003 09/07/14 02:45 .144
0004 09/07/14 02:46 .136
0005 09/07/14 02:52 .043
0006 09/07/14 02:53 .040
0007 09/07/14 19:44 .172
0008 09/07/14 19:47 .174
0009 09/12/14 09:21 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8709
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP. (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0929	34	14F23-11-5-10- C1307252	.100	21		103			∅		Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											

09/12/14 09:31
 HSA Serial no: 008709
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/17/14 09:27 .100
 Test Result 0001 09/12/14 09:30 .103
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8714
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0933	34	14F23-11-5-10- C1307252	.100	21		106				∅	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 09:34
 A94 Serial no: 008714
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/11/04 14:05 .100
 Test Result
 0001 09/11/14 02:17 .149
 0002 09/11/14 02:20 .159
 0003 09/12/14 00:27 .084
 0004 09/12/14 00:30 .091
 0005 09/12/14 09:33 .106
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8888
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION OR DRY GAS**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	09:25	34	14F23-11-5-10- C1307252	.100	22	102						Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p>09/12/14 09:26 AC4 Serial no: 0009988 Version no: 4797</p> <p>No. Date Time %BAC Calibration 03/25/13 10:45 .100</p> <p>Test Result 0001 09/06/14 03:57 .124 0002 09/06/14 04:00 .134 0003 09/06/14 20:58 .099 0004 09/06/14 21:00 .100 0005 09/06/14 21:04 .023 0006 09/06/14 01:04 .000 0007 09/12/14 09:25 .102</p> <p>Push No. 3 to erase or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8783
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0936	34	14F23-11-5-10- C1307252	.100	21		101					Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16														
<pre> 09/12/14 09:36 AS4 Serial no: 000783 Version no: 4797 No. Date Time %BAC Calibration 07/18/08 13:52 .100 Test Result 0001 09/08/14 19:50 .280 0002 09/08/14 19:52 .289 0003 09/09/14 01:54 .216 0004 09/09/14 01:56 .215 0005 09/11/14 01:57 .088 0006 09/11/14 01:59 .090 0007 09/11/14 02:06 .000 0008 09/12/14 09:36 .101 Push No. 3 to erase, or power down to retain </pre>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82657
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0935	34	14F23-11-5-10- C1307252	.100	22		.100			X		Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 09:36
 AS4 Serial no: 082657
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/17/14 09:50 .100
 Test Result 0001 09/12/14 09:36 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Brian Flick	15063	Intoximeters Inc./Alcosensor IV	19 8810

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	—		14F23-11-5-10-C1307252	.100	NOT	IN	LOCKER	—			Flick	15063	
			Prepared by:										
			Douglas C. Talbot										
			on 06-23-14.										
			Expires on 06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18684
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0939	34	14F23-11-5-10- C1307252	.100	22		.102			X	Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>													
<p style="text-align: center;"> No. Date Time %BAC Calibration 11/03/11 16:03 .100 Test Result 0001 09/12/14 09:40 .102 Push No. 3 to erase, or power down to retain </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8635
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	0948	35	14F23-11-5-10- C1307252	.100	21		101			X	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16										
<pre> 09/12/14 09:49 AS4 Serial no: 000635 Version no: 4797 No. Date Time XBAC Calibration 10/24/08 09:04 .100 Test Result 0001 09/12/14 09:48 .101 Push No. 3 to erase, or power down to retain </pre>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8845
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	0959	34	14F23-11-5-10- C1307252	.100	21		.102					Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 10:01
 AS4-Serial no: 000845
 Version no: 4797
 No. Date Time ZBAC
 Calibration 12/24/03 11:01 .100
 Test Result
 0001 09/06/14 04:39 .109
 0002 09/06/14 04:41 .110
 0003 09/06/14 19:48 .328
 0004 09/06/14 19:58 .335
 0005 09/07/14 21:49 .000
 0006 09/07/14 22:22 .124
 0007 09/07/14 22:24 .122
 0008 09/12/14 10:00 .102
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82653
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION OR DRY GAS**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1010	34	14F23-11-5-10- C1307252	.100	21		.099			X		Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 10:11
 H54 Serial no: 862653
 Version no: 4797
 No. Date Time XBAC
 Calibration 02/17/14 10:09 .100
 Test Result
 0001 09/05/14 23:10 .097
 0002 09/05/14 23:13 .077
 0003 09/06/14 03:52 .156
 0004 09/06/14 03:55 .154
 0005 09/07/14 02:31 .171
 0006 09/07/14 02:33 .148
 0007 09/07/14 02:36 .168
 0008 09/07/14 03:30 .201
 0009 09/07/14 03:33 .200
 0010 09/08/14 01:25 .092
 0011 09/08/14 01:28 .095
 0012 09/11/14 23:32 .117
 0013 09/11/14 23:35 .117
 0014 09/12/14 01:40 .119
 0015 09/12/14 01:43 .129
 0016 09/12/14 04:13 .006
 0017 09/12/14 10:11 .999

Push No. 3 to erase
 or Power down to Retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8848
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1016	34	14F23-11-5-10- C1307252	.100	21		101					Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
			No. Date Time %BWC Calibration 01/13/03 09:25 .100 09/12/14 10:17 AS4 Serial no: 008848 Version no: 4797 Test Result 0001 09/12/14 10:17 .101 Push No. 3 to erase, or power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 77591
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1022	34	14F23-11-5-10- C1307252	.100	20		.100					Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p style="text-align: center;"> 09/12/14 10:23 A94 Serial no: 077591 Version no: 4797 No. Date Time %BAC Calibration 06/21/08 15:09 .100 Test Result 0001 09/12/14 10:23 .100 Push No. 3 to erase, or power down to retain </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8850
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1030	34	14F23-11-5-10- C1307252	.100	21		.101					Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p>09/12/14 10:31 AS4 Serial no: 008850 Version no: 4797 No. Date Time ZPAC Calibration 08/02/06 09:07 .100 Test Result 0001 09/12/14 10:31 .101 Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 77573
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1023	34	14F23-11-5-10- C1307252	.100	22	097				X		Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p>Test Result 0001 09/10/14 00:56 .012 0002 09/10/14 01:01 .035 0003 09/10/14 01:33 .105 0004 09/11/14 02:42 .107 0005 09/11/14 02:45 .103 0006 09/12/14 10:23 .097</p> <p>Push No. 3 to erase, or power down to retain</p> <p>No. Date Time ZDHC Calibration 04/09/14 08:36 .080</p> <p>09/12/14 10:24 AS4 Serial no: 077573 Version no: 4797</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 83051
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1027	34	14F23-11-5-10- C1307252	.100	21		.098				⊗	Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16														
<pre> 09/12/14 10:28 AS4 Serial no: 083051 Version no: 4797 No. Date Time ZBAC Calibration 09/25/13 14:06 .100 Test Result 0901 09/12/14 10:127 .098 Push No. 3 to erase, or power down to retain </pre>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

30
 8897

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8897
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1037	34	14F23-11-5-10- C1307252	.100	21		.103				Y	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
			<pre> 09/12/14 10:38 NS4 Serial no: 008897 Version no: 4797 No. Date Time ZBAC Calibration 01/18/12 16:13 .100 Test Result 0001 09/12/14 10:38 .100 Push No. 3 to erase, or power down to retain </pre>											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 31 18624
--------------------------------	-----------------	---	----------------------	---	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1045	34	14F23-11-5-10- C1307252	.100	22		.093			X		Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
														Test Result 0001 09/06/14 02:30 .116 0002 09/07/14 01:26 .065 0003 09/07/14 01:29 .062 0004 09/07/14 01:39 .000 0005 09/07/14 03:16 .000 0006 09/07/14 03:19 .079 0007 09/07/14 03:21 .079 0008 09/12/14 08:11 .015 0009 09/12/14 08:12 .014 0010 09/12/14 02:06 .000 0011 09/12/14 10:46 .093
														No. Date Time %RAC Calibration 10/07/08 09:19 .100
														09/12/14 10:46 NS4 Serial no: 018624 Version no: 4797
			Push No. 3 to erase, or power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18646
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1055	34	14F23-11-5-10- C1307252	.100	23		.097				Ø	Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											
<p>09/12/14 10:56 AS4 Serial no: 018646 Version no: 4797</p> <p>No. Date Time %BAC Calibration 07/29/13 13:32 .100</p> <p>Test Result 0001 09/12/14 10:55 .097</p> <p>Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 77488
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR OR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1053	34	14F23-11-5-10- C1307252	.100	21		.100					Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

09/12/14 10:54
AS4 Serial no: 077488
Version no: 4797
No. Date Time ZBAC
Calibration 05/21/08 14:35 .100
Test Result
0001 09/12/14 10:54 .100
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 8877
--------------------------------	-----------------	---	----------------------	---	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1102	34	14F23-11-5-10- C1307252	.100	21			102				Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											
<p>09/12/14 11:03 AS4 Serial no: 000877 Version no: 4797</p> <p>No. Date Time %BAC Calibration 05/26/05 10:40 .100</p> <p>Test Result 0001 09/12/14 11:02 .102</p> <p>Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18615
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)	PAS ACCURACY CHECK RESULTS*			CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
09/12/2014	1101	34	14F23-11-5-10- C1307252	.100	22	.098						Flick	15063		
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16												
			<pre> 09/12/14 11:02 AS4 Serial no: 018615 Version no: 4797 No. Date Time ZBAC Calibration 12/19/08 14:44 .100 Test Result: 0001 09/12/14 11:02 .098 Push No. 3 to erase, or power down to retain </pre>												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18626
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	1111	34	14F23-11-5-10- C1307252	.100	22		.099				Flick	15063	
			Prepared by:										
			Douglas C. Talbot										
			on 06-23-14.										
			Expires on 06-23-16										

09/12/14 11:13
 HSA Serial no: 018626
 Version no: 4797
 No. Date Time ZBAC
 Calibration 04/29/05 08:59 .101
 Test Result
 0001 09/12/14 11:11 .099
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. <i>38</i> 18696
--------------------------------	-----------------	---	----------------------	---	--------------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/12/2014	1115	34	14F23-11-5-10-C1307252	.100	22		.102			0	Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>													
<p style="text-align: center;"> 09/12/14 11:15 A94 Serial no: 018696 Version no: 4797 No. Date Time ZBAC Calibration 09/09/13 14:50 .100 Test Result 0001 09/12/14 11:15 .100 Push No. 3 to erase, or power down to retain </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 83030
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1117	34	14F23-11-5-10- C1307252	.100	22		.100					Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p style="text-align: center;"> No. Date Time %BAC Calibration 01/29/09 18:39 .100 Test Result 0001 09/12/14 11:17 .100 Push No. 3 to erase, or power down to retain </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18692
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1017	34	14F23-11-5-10- C1307252	.100	25		100				X	Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
			<div style="border: 1px solid black; padding: 5px; text-align: center;"> <p>09/12/14 10:18 AS4 Serial no: 018692 Version no: 4797</p> <p>No. Date Time %BAC Calibration 08/12/08 14:35 .100</p> <p>Test Result 0001 09/12/14 10:17 .100</p> <p>Push No. 3 to erase, or power down to retain</p> </div>											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18766
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1333	34	14F23-11-5-10-	.100	22		.090					Flick	15063	
9-12-14	1349	34	C1307252	.100	25		.098					Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16														
<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px;"> Push No. 3 to erase or power down to retain </div> <div style="border: 1px solid black; padding: 5px;"> Test Result 0001 09/07/14 02:11 .000 0002 09/07/14 04:18 .078 0003 09/07/14 04:21 .075 0004 09/12/14 13:35 .090 0005 09/12/14 13:51 .098 </div> <div style="border: 1px solid black; padding: 5px;"> No. Date Time %BAC Calibration 09/12/14 13:45 .100 </div> <div style="border: 1px solid black; padding: 5px;"> 09/12/14 13:52 AS4 Serial no: 018766 Version no: 4797 </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18691
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
09/12/2014	1339	34	14F23-11-5-10-C1307252	.100	22		.101				✓	Flick	15063		

Prepared by:
Douglas C. Talbot
on 06-23-14.
Expires on 06-23-16

09/12/14 13:40
 ASD Serial no: 018691
 Version no: 4797
 No. Date Time %BAC
 Calibration 09/08/08 14:01 .100
 Test Result
 0001 09/12/14 13:40 .101
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

45
17990

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 17990
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1340	34	14F23-11-5-10- C1307252	.100	22		105					Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
			Test Result: 0001 09/06/14 02:21 .140 0002 09/06/14 02:25 .135 0003 09/07/14 01:57 .043 0004 09/07/14 02:59 .627 0005 09/07/14 21:45 .888 0006 09/07/14 21:48 .096 0007 09/07/14 23:10 .077 0008 09/07/14 23:13 .084 0009 09/12/14 13:40 .105 No. Date Time %BAC Calibration 11/26/03 11:10 .100 09/12/14 13:41 H54 Serial no: 0179990 Version no: 4797 Push No. 3 to erase, or power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

46

18774

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 18774
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1342	34	14F23-11-5-10- C1307252	.100	22		.099			Y		Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			on 06-23-14.											
			Expires on 06-23-16											

09/12/14 13:43
 AS4 Serial no: 618774
 Version no: 4797
 No. Date Time %BAC
 Calibration 01/13/09 18:25 .100
 Test Result
 0001 09/09/14 01:21 .243
 0002 09/09/14 01:24 .243
 0003 09/09/14 23:49 .000
 0004 09/12/14 13:43 .099
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 50 41486
--------------------------------	-----------------	---	----------------------	---	--------------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	—		14F23-11-5-10- C1307252	.100	—	B, D,	—	—				Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 41495
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1345	34	14F23-11-5-10- C1307252	.100	21		.100					Flick	15063	
			Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16											
			Test Result 0001 09/12/14 13:45 .100 Calibration 02/20/14 11:28 .100 No. Date Time XBAC Version no: 4797 09/12/14 13:46 684 Serial no: 041495 09/12/14 13:46											
			Push No. 3 to erase, or power down to retain											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 41502
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1355	34	14F23-11-5-10- C1307252	.100	22		097			X		Flick	15063	
Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16														
No. Date Time %BAC Calibration 05/06/09 08:09 .100 Test Result 0001 09/12/14 13:56 .097 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alcosensor IV	PAS SERIAL NO. 82654
--------------------------------	-----------------	---	----------------------	---	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/12/2014	1356	34	14F23-11-5-10- C1307252	.100	22	.099					Y	Flick	15063	
<p>Prepared by: Douglas C. Talbot on 06-23-14. Expires on 06-23-16</p>														
<p>09/12/14 13:56 AS4 Serial no: 882654 Version no: 4797</p> <p>No. Date Time %BAC Calibration 09/28/14 11:36 .100</p> <p>Test Result 0001 09/12/14 13:56 .099</p> <p>Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.