

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	0918	34	14F23-11-5-10	.100	21		.099			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push No. 3 to erase
or power down to retain

Test Result
0001 09/23/14 13:18 .000
0002 09/23/14 13:21 .000
0003 09/26/14 09:22 .099

Calibration
07/18/08 13:27 .100

09/26/14 09:23
884 Serial no: 0108785
Version 101 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	09:32	34	14F23-11-5-10 C1307252	.100	21		.098		X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Push No. 1 to erase,

Test Result
0001 09/26/14 09:29 .119
0002 09/26/14 09:31 .120
0003 09/26/14 09:40 .098

No. Date Time 2000
Calibration 01/28/13 14:04 .100

09/26/14 09:41
AS4 Serial no: 008620
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	0936	34	14F23-11-5-10	.100	22		.101			X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
0901 09/26/14 09:36 .101
 No. Date Time 2000
 Calibration
02/17/14 08:02 .100
 09/26/14 09:36
 AS4 Serial no: 082957
 Version no: 4797
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	0938	34	14F23-11-5-10 C1307252	.100	21		.098			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Test Result
 0901 09/26/14 09:43 .141
 0902 09/26/14 09:44 .101
 0903 09/26/14 09:57 .098
 No. Date Time 200C
 Calibration 07/17/14 09:10 .106
 09/26/14 09:48
 AS4 Serial no: 077479
 Version no: 4797
 Test No. 0 to Erase,
 or report down to Retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	0941	34	14F23-11-5-10 C1307252	.100	21		.099		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

No. Date Time 2014
 Calibration 02/17/14 00:16 .100
 Version not 4797
 AS4 Serial no: 002660
 09/26/14 09:41
 Test Result
 0001 09/21/14 02:49 .115
 0002 09/21/14 02:52 .110
 0003 09/21/14 19:32 .100
 0004 09/21/14 19:35 .095
 0005 09/26/14 03:26 .191
 0006 09/26/14 03:28 .191
 0007 09/26/14 09:41 .099

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	0944	34	14F23-11-5-10 C1307252	.100	22		.100			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push No. 3 to erase
 the previous data in
 the memory.
 Test Result
 09/26/14 09:51 .100
 No. Date Time ZONE
 Calibration
 07/17/14 09:27 .100
 Version 101 4/07
 09/26/14 09:51
 AS4 Serial 101 09/26/14
 Version 101 4/07

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	0953	34	14F23-11-5-10 C1307252	.100	22		.102			X	Lee	17046	
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16													
Push No. 3 to erase, or power down to retain													
Test Result 0001 09/26/14 10:02 .102													
No. Date Time ZBAC Calibration 10/11/04 14:05 .100													
09/26/14 10:02 AS4 Serial no: 600214 Version: rev 4/97													
Power Down to Retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	0958	34	14F23-11-5-10	.100	22		.099			X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

09/26/14 09:54
654 Serial not present
Version not found

No. Date Time CONC
Calibration
09/25/13 14:45 .100

Test Results
09/20/14 01:37 .080
09/20/14 01:40 .081
09/20/14 22:06 .096
09/21/14 01:45 .090
09/21/14 01:48 .098
09/21/14 28:22 .142
09/26/14 09:58 .099

Push No. 3 to cross
or equipt. down to red pin.

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1003	34	14F23-11-5-10	.100	21		.101			X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Test Result
 0001 09/26/14 10:04 .101
 No. Date Time %MAC
 Calibration 02/17/14 09:00 100
 Version 101 4797
 09/26/14 10:05
 084 Serial no: 082657
 09/26/14 10:05

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
09/26/2014	1007	34	14F23-11-5-10	.100	22	.101			X		Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

09/26/14 10:18
 NS4 Serial no: 018684
 Version no: 4797
 No. Date Time ZDNC
 Calibration 11/08/11 16:00 JPH
 Test Result
 0001 09/26/14 10:59 000
 0002 09/26/14 10:18 101
 Push NO. 3 to erase
 or remove drum to factory

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014			14F23-11-5-10	.100							Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 00B635
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1010	34	14F23-11-5-10 C1307252	.100	21		.101			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>														
<p style="text-align: center;"> Test Result 09/26/14 21:00 .100 09/26/14 21:12 .116 09/26/14 21:01 .147 09/26/14 21:02 .152 09/26/14 21:28 .149 09/26/14 07:23 .143 09/26/14 10:16 .100 09/26/14 10:16 .101 </p> <p style="text-align: center;"> No. Date Time 2010 Calibration 10/24/09 09:04 .100 </p> <p style="text-align: center;"> 09/26/14 10:16 654 Serial no: 00B635 Version no: 4797 </p> <p style="text-align: center;"> Push No. 3 to Erase, or button down to retain. </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1022	39	14F23-11-5-10	.100	21		.100			X	Lee	17046	
			C1307252										
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16													

Test Results
 0801 09/19/14 19:15 .072
 0107 09/19/14 19:18 .073
 0801 09/19/14 19:21 .000
 0809 09/20/14 02:21 .000
 0805 09/20/14 19:32 .201
 0806 09/20/14 19:35 .209
 0807 09/22/14 01:23 .000
 0808 09/26/14 01:23 .007
 0809 09/26/14 01:27 .003
 0810 09/26/14 10:25 .100

No. Date Time 2000
 Calibration
 09/19/14 10:00 .100

09/26/14 10:24
 NS4 Serial not required
 Version not apply

Push No. 9 Cal. Issues
 or Issues shown to Refrain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1025	34	14F23-11-5-10	.100	21		.100			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push Ho. 3 to clear
 or press down to refresh
 Test Result
 0001 09/26/14 10:20 .100
 No. Date Time %BAC
 Calibration 01/13/03 09:25 .100
 Version no: 4797
 09/26/14 10:42
 V04 Serial no: 008848

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1039	34	14F23-11-5-10 C1307252	.100	20	100			X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Push No. 3 In excess
of power down for refash

Test Result
0901 09/26/14 10:48 .100

Calibration
05/21/08 15:09 .100

No. Date Time ZINC
04297

09/26/14 10:48
not serial not 077591
not 4297

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008350
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1042	34	14F23-11-5-10	.100	21		.101			X	Lee	17046	
			C1307252										
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>													
<p>Push No. 2 to erase or power down to 1.50V</p> <p>Test Result 0001 09/26/14 10:52</p> <p>Calibration 03/02/06 09:07</p> <p>No. Date Time 200C Calibration</p> <p>Version no: 4797</p> <p>09/26/14 10:52</p> <p>Lot Serial no: 008056</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1047	34	14F23-11-5-10	.100	22						Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Test Result
 0001 09/19/14 20:50 .100
 0002 09/19/14 20:50 .110
 0003 09/19/14 20:54 .104
 0004 09/20/14 08:40 .107
 0005 09/20/14 08:50 .106
 0006 09/26/14 10:46 .106
 No. Date Time JDMC
 Calibration 04/09/14 08:36 .000
 09/26/14 10:47
 001 Serial not 077573
 Version not 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1051	34	14F23-11-5-10	.100	22		.098			X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Inst. No. 3 to cancel

Test Results
0801 09/26/14 18:14 .092
0802 09/26/14 18:15 .094
1003 09/26/14 18:50 .098

Calibration
09/25/13 14:06 .100

No. Date Time Status
09/26/14 18:51
084 Serial no: 083051
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1104	34	14F23-11-5-10 C1307252	.100	22						Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push On. 3 to erase.

Test Result	No.	Date	Time	YOB	Calibration
0201	09/19/14	20:53			
0802	09/19/14	20:54			
2965	09/20/14	02:20			
0808	09/20/14	02:23			
0807	09/20/14	23:48			
1408	09/20/14	23:51			
0809	09/21/14	00:10			
0800	09/21/14	00:12			
0809	09/21/14	02:35			
0810	09/21/14	02:37			
0811	09/21/14	21:55			
0812	09/21/14	21:59			
0813	09/21/14	22:00			
0814	09/26/14	02:50			
0815	09/26/14	02:52			
0816	09/26/14	02:23			
0817	09/26/14	02:24			
0818	09/26/14	02:49			
0819	09/26/14	11:12			

or cover down to return

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1109	34	14F23-11-5-10 C1307252	.100	22		.097			X	Lee	17046	
		Prepared by:											
		Douglas C. Talbot											
		06-23-14											
		Expires on:											
		06-23-16											
				Test Result		No. Date Time		Calibration					
				0901 09/26/14 11:14 .097		10/07/08 09:19 .100							
				Test Result		No. Date Time		Calibration					

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1113	34	14F23-11-5-10 C1307252	.100	23						Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Printed: 9/26/14 11:18
Test Result
0001 09/26/14 11:18 .098
No. Date Time ZONE
Calibration
09/29/14 10:32 .100
Version
09/26/14 11:18
NSA Serial no: 018646
Version no: 4297

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077499
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1116	34	14F23-11-5-10 C1307252	.100	21		103			X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

09/26/14 11:16
 AS4 Serial no: 077499
 Version no: 4797
 No. Date Time Result
 Calibration 05/21/09 14:35 100
 Test Result 0901 09/26/14 11:16 100
 Calibration OK
 or reader is not to retain

power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1120	34	14F23-11-5-10 C1307252	.100	21		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										
				<div style="border: 1px solid black; padding: 5px; display: inline-block;"> 09/26/14 11:27 PAS Serial no: 008877 Version no: 4797 No. Date Time Yr Calibration 09/26/05 10:40 .100 Test Result 0001 09/26/14 11:27 .101 Test No. 3 to erase or query down to reset </div>									

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
09/26/2014	1135	34	14F23-11-5-10 C1307252	.100	22		.098			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			06-23-14												
			Expires on:												
			06-23-16												

Test Result
0001 09/26/14 11:38 000
 No. Date Time 2000
Calibration
12/19/08 14:44 100
 09/26/14 11:39
084 Serial no: 018615
Version no: 4797
 Test No. 1 11:39:39
of 1500000000 to retain.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1138	34	14F23-11-5-10 C1307252	.100	22		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Test Result
 0001 09/26/14 11:39 .100
 No. Date Time XMMC
 Calibration 04/29/05 08:50 .101
 Version no: 4797
 09/26/14 11:39
 004 Serial no: 018626
 09/26/14 11:39
 Fresh No. 2 to P18626
 or lower down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1149	34	14F23-11-5-10 C1307252	.100	22		.102			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

No. Date Time 2300
 Calibration 09/09/13 14:56
 Test Result
 0001 09/26/14 11:52 .100
 Method: 3
 or pour down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1152	34	14F23-11-5-10 C1307252	.100	21		.099			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										
<p style="text-align: center;"> No. Date Time 284 Calibration 01/29/09 10:30 .100 Test Result 09/26/14 11:52 .099 No. Date Time 284 Version not 4299 09/26/14 11:52 Not Serial not 083030 </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 019692
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1045	34	14F23-11-5-10 C1307252	.100	22		.099			X		Lee	17046	
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16											
			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: 0 auto;"> Push No. 3 An error or power down to return Test Result 0901 09/26/14 10:48 .099 Calibration 08/12/08 14:35 .100 Date Time ZBAC 09/26/14 10:48 109 09/14/14 10:01 019692 109 09/14/14 10:01 4797 </div>											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	1155	34	14F23-11-5-10 C1307252	.100	22		.101			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

09/26/14 12:00
 AS4 Serial no: 018762
 Version no: 4797
 No. Date Time 2000
 Calibration 09/30/09 15:52 1106
 Test Result
 0001 09/26/14 11:00 100
 Print No. 31 of 31
 or scan down to reload

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014			14F23-11-5-10	.100							Lee	17046	B.O.
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
09/26/2014	12:13	34	14F23-11-5-10	.100	22		.099		X		Lee	17046	
			C1307252										
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>													

Test Result
 0001 09/22/14 02:07 101
 0002 09/22/14 02:10 104
 0003 09/23/14 09:57 172
 0004 09/23/14 09:58 112
 0005 09/23/14 04:06 127
 0006 09/23/14 07:30 104
 0007 09/23/14 22:40 102
 0008 09/27/14 15:17 109
 09/26/14 12:15
 001 serial no# 018774
 canister no# 4792
 Calibration
 Date Time Y/M/C
 01/19/09 10:25 100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041436
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014		34	14F23-11-5-10 C1307252	.100								Lee	17046	B.O.
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	1217	34	14F23-11-5-10	.100	21		.09%			X		Lee	17046	
			C1307252											
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16											
														Quality 3 to erase
														Test Record 09/26/14 12:14 098 Calibration 02/28/14 11:20 106 No. Date Time Time Calibration 02/28/14 11:20 106 Version 041495 04797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
09/26/2014	12:19	34	14F23-11-5-10 C1307252	.100	22		.096			X		Lee	17046	
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16											
			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p style="text-align: center;">Test Result</p> <p style="text-align: center;">09/26/14 12:21</p> <p style="text-align: center;">AS4 Serial no: 041502</p> <p style="text-align: center;">Version no: 4797</p> <p style="text-align: center;">No. Date Time 2010</p> <p style="text-align: center;">Calibration 09/26/2014 12:21</p> <p style="text-align: center;">or power down to retain</p> </div>											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

