

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

SGT

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. <i>008874</i>
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
10/08/2014	1720	34	14F23-11-5-10 C1307252	.100	22		.098			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p>													
<p style="text-align: center;"> Push No. 3 to erase, Test Result 0001 10/06/14 17:25 .098 No. Date Time ZMAC Calibration 03/28/14 11:40 .100 10/08/14 17:25 MS4 Serial no: 008874 Version no: 4797 </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008501
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1247	34	14F23-11-5-10 C1307252	.100	21		.097			X	Lee	17046	
Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16													
No. Date Calibration 06/23/14 12:52 .097 06/23/14 03:23 .205 06/23/14 12:52 .097 06/23/14 12:52 .097													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018785
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1426	34	14F23-11-5-10 C1307252	.100	21		.099			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p> <p>Test Result: 091 10/08/14 14:32 .099</p> <p>Calibration: 07/18/08 15:27 .106</p> <p>10/08/14 14:32 AS4 Serial not 018785 Version not 4799</p> <p>power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1430	34	14F23-11-5-10	.100	22		.101				Lee	17046	
			C1307252										
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>													

Bath No. 2 To Erase:
If cover flap to retain
 No. Date Time
 Calibration 06/28/18 14:30 .100
 User: TBM
 10/08/14 14:34
 /64 Serial not stored
 User: TBM
 001 8797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014			14F23-11-5-10	.100							Lee	17046	B.O. OUT for SERVICE
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008580
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1451	34	14F23-11-5-10 C1307252	.100	22		1.099			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>														
<p style="text-align: center;"> No. Date Time 2014 Calibration 03/06/14 10:30 .100 Test Result 10/08/14 14:58 .099 </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#B

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1458	34	14F23-11-5-10	.100	21	.099				X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Push the 1 to erase
 or pairs. Repeat for each...
 10/08/14 15:07
 NS4 Serial 001 008620
 Version 101 (7/07)
 No. Date Time 00:00
 Date 10/08/14 10:00
 .100
 .148
 .153
 .246
 .101
 .098
 .093
 .093
 .093
 .093
 .098
 .098
 .099

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014			14F23-11-5-10 C1307252	.100								Lee	17046	NOT IN LOGGER
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1501	34	14F23-11-5-10 C1307252	.100	22		.079			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Full 100% 3 Accuracy
 or lower down to 0.090
 10/01/09/19/14 08:33 .100
 0807 09/19/14 08:34 .099
 0808 09/19/14 08:35 .100
 0809 10/08/14 15:50 .100
 0810 10/08/14 15:50 .100
 0811 10/08/14 15:50 .100
 0812 10/08/14 15:50 .100
 0813 10/08/14 15:50 .100
 0814 10/08/14 15:50 .100
 0815 10/08/14 15:50 .100
 0816 10/08/14 15:50 .100
 0817 10/08/14 15:50 .100
 0818 10/08/14 15:50 .100
 0819 10/08/14 15:50 .100
 0820 10/08/14 15:50 .100
 0821 10/08/14 15:50 .100
 0822 10/08/14 15:50 .100
 0823 10/08/14 15:50 .100
 0824 10/08/14 15:50 .100
 0825 10/08/14 15:50 .100
 0826 10/08/14 15:50 .100
 0827 10/08/14 15:50 .100
 0828 10/08/14 15:50 .100
 0829 10/08/14 15:50 .100
 0830 10/08/14 15:50 .100
 0831 10/08/14 15:50 .100
 0832 10/08/14 15:50 .100
 0833 10/08/14 15:50 .100
 0834 10/08/14 15:50 .100
 0835 10/08/14 15:50 .100
 0836 10/08/14 15:50 .100
 0837 10/08/14 15:50 .100
 0838 10/08/14 15:50 .100
 0839 10/08/14 15:50 .100
 0840 10/08/14 15:50 .100
 0841 10/08/14 15:50 .100
 0842 10/08/14 15:50 .100
 0843 10/08/14 15:50 .100
 0844 10/08/14 15:50 .100
 0845 10/08/14 15:50 .100
 0846 10/08/14 15:50 .100
 0847 10/08/14 15:50 .100
 0848 10/08/14 15:50 .100
 0849 10/08/14 15:50 .100
 0850 10/08/14 15:50 .100
 0851 10/08/14 15:50 .100
 0852 10/08/14 15:50 .100
 0853 10/08/14 15:50 .100
 0854 10/08/14 15:50 .100
 0855 10/08/14 15:50 .100
 0856 10/08/14 15:50 .100
 0857 10/08/14 15:50 .100
 0858 10/08/14 15:50 .100
 0859 10/08/14 15:50 .100
 0860 10/08/14 15:50 .100
 0861 10/08/14 15:50 .100
 0862 10/08/14 15:50 .100
 0863 10/08/14 15:50 .100
 0864 10/08/14 15:50 .100
 0865 10/08/14 15:50 .100
 0866 10/08/14 15:50 .100
 0867 10/08/14 15:50 .100
 0868 10/08/14 15:50 .100
 0869 10/08/14 15:50 .100
 0870 10/08/14 15:50 .100
 0871 10/08/14 15:50 .100
 0872 10/08/14 15:50 .100
 0873 10/08/14 15:50 .100
 0874 10/08/14 15:50 .100
 0875 10/08/14 15:50 .100
 0876 10/08/14 15:50 .100
 0877 10/08/14 15:50 .100
 0878 10/08/14 15:50 .100
 0879 10/08/14 15:50 .100
 0880 10/08/14 15:50 .100
 0881 10/08/14 15:50 .100
 0882 10/08/14 15:50 .100
 0883 10/08/14 15:50 .100
 0884 10/08/14 15:50 .100
 0885 10/08/14 15:50 .100
 0886 10/08/14 15:50 .100
 0887 10/08/14 15:50 .100
 0888 10/08/14 15:50 .100
 0889 10/08/14 15:50 .100
 0890 10/08/14 15:50 .100
 0891 10/08/14 15:50 .100
 0892 10/08/14 15:50 .100
 0893 10/08/14 15:50 .100
 0894 10/08/14 15:50 .100
 0895 10/08/14 15:50 .100
 0896 10/08/14 15:50 .100
 0897 10/08/14 15:50 .100
 0898 10/08/14 15:50 .100
 0899 10/08/14 15:50 .100
 0900 10/08/14 15:50 .100
 0901 10/08/14 15:50 .100
 0902 10/08/14 15:50 .100
 0903 10/08/14 15:50 .100
 0904 10/08/14 15:50 .100
 0905 10/08/14 15:50 .100
 0906 10/08/14 15:50 .100
 0907 10/08/14 15:50 .100
 0908 10/08/14 15:50 .100
 0909 10/08/14 15:50 .100
 0910 10/08/14 15:50 .100
 0911 10/08/14 15:50 .100
 0912 10/08/14 15:50 .100
 0913 10/08/14 15:50 .100
 0914 10/08/14 15:50 .100
 0915 10/08/14 15:50 .100
 0916 10/08/14 15:50 .100
 0917 10/08/14 15:50 .100
 0918 10/08/14 15:50 .100
 0919 10/08/14 15:50 .100
 0920 10/08/14 15:50 .100
 0921 10/08/14 15:50 .100
 0922 10/08/14 15:50 .100
 0923 10/08/14 15:50 .100
 0924 10/08/14 15:50 .100
 0925 10/08/14 15:50 .100
 0926 10/08/14 15:50 .100
 0927 10/08/14 15:50 .100
 0928 10/08/14 15:50 .100
 0929 10/08/14 15:50 .100
 0930 10/08/14 15:50 .100
 0931 10/08/14 15:50 .100
 0932 10/08/14 15:50 .100
 0933 10/08/14 15:50 .100
 0934 10/08/14 15:50 .100
 0935 10/08/14 15:50 .100
 0936 10/08/14 15:50 .100
 0937 10/08/14 15:50 .100
 0938 10/08/14 15:50 .100
 0939 10/08/14 15:50 .100
 0940 10/08/14 15:50 .100
 0941 10/08/14 15:50 .100
 0942 10/08/14 15:50 .100
 0943 10/08/14 15:50 .100
 0944 10/08/14 15:50 .100
 0945 10/08/14 15:50 .100
 0946 10/08/14 15:50 .100
 0947 10/08/14 15:50 .100
 0948 10/08/14 15:50 .100
 0949 10/08/14 15:50 .100
 0950 10/08/14 15:50 .100
 0951 10/08/14 15:50 .100
 0952 10/08/14 15:50 .100
 0953 10/08/14 15:50 .100
 0954 10/08/14 15:50 .100
 0955 10/08/14 15:50 .100
 0956 10/08/14 15:50 .100
 0957 10/08/14 15:50 .100
 0958 10/08/14 15:50 .100
 0959 10/08/14 15:50 .100
 0960 10/08/14 15:50 .100
 0961 10/08/14 15:50 .100
 0962 10/08/14 15:50 .100
 0963 10/08/14 15:50 .100
 0964 10/08/14 15:50 .100
 0965 10/08/14 15:50 .100
 0966 10/08/14 15:50 .100
 0967 10/08/14 15:50 .100
 0968 10/08/14 15:50 .100
 0969 10/08/14 15:50 .100
 0970 10/08/14 15:50 .100
 0971 10/08/14 15:50 .100
 0972 10/08/14 15:50 .100
 0973 10/08/14 15:50 .100
 0974 10/08/14 15:50 .100
 0975 10/08/14 15:50 .100
 0976 10/08/14 15:50 .100
 0977 10/08/14 15:50 .100
 0978 10/08/14 15:50 .100
 0979 10/08/14 15:50 .100
 0980 10/08/14 15:50 .100
 0981 10/08/14 15:50 .100
 0982 10/08/14 15:50 .100
 0983 10/08/14 15:50 .100
 0984 10/08/14 15:50 .100
 0985 10/08/14 15:50 .100
 0986 10/08/14 15:50 .100
 0987 10/08/14 15:50 .100
 0988 10/08/14 15:50 .100
 0989 10/08/14 15:50 .100
 0990 10/08/14 15:50 .100
 0991 10/08/14 15:50 .100
 0992 10/08/14 15:50 .100
 0993 10/08/14 15:50 .100
 0994 10/08/14 15:50 .100
 0995 10/08/14 15:50 .100
 0996 10/08/14 15:50 .100
 0997 10/08/14 15:50 .100
 0998 10/08/14 15:50 .100
 0999 10/08/14 15:50 .100
 1000 10/08/14 15:50 .100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1504	34	14F23-11-5-10 C1307252	.100	21		.099			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Test Result
 02/17/14 09:10 .100
 Calibration
 No. Date Time 2010
 02/17/14 09:10 .100
 version no: 4797
 02/17/14 19:06
 02/17/14 no: 077479
 02/17/14 no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1507	34	14F23-11-5-10	.100	21		.100				X	Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

No. Date Time 2014
 Calibration
 02/17/14 09:10:10
 Test Result
 0.091 10/08/2014 15:07:34
 14F23-11-5-10
 C1307252

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014		34	14F23-11-5-10 C1307252	.100	23		100			X		Lee	17046	
														Prepared by:
														Douglas C. Talbot
														06-23-14
														Expires on:
														06-23-16

Inst. No. 8 to PHAS-01
 or power down to retain
 data
 Date: 10/08/2014 21:20:00
 Time: 10/08/2014 21:20:00
 Serial: 008709
 Version: 4.000
 No. Tests: 1
 Date Time: 10/08/2014 21:20:00
 Calibration: 10/08/2014 21:20:00
 No. Tests: 1
 Date Time: 10/08/2014 21:20:00
 Serial: 008709
 Version: 4.000

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1555	34	14F23-11-5-10	.100	22		.102			X		Lee	17046	
			C1307252											
			Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16											

Dish No. 9 to 19 are

Test Result
 1001 10/06/14 08:06 .048
 1002 10/06/14 20:17 .061
 1003 10/06/14 28:10 .064
 1004 10/05/14 04:35 .121
 1005 10/05/14 04:38 .116
 1006 10/05/14 04:44 .000
 1007 10/05/14 16:05 .102

No. Date Time Job
 Calibration
 02/11/04 14:00 .100

10/08/14 16:55
 (04 serial) not (0000)
 Version not 475

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1600	34	14F23-11-5-10 C1307252	.100	22		.099			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p> <p style="text-align: center;">Push No. 3 to erase:</p> <pre> Test Result 0001 10/06/14 01:56 .135 0002 10/06/14 01:59 .137 0003 10/06/14 02:06 .142 0004 10/06/14 02:07 .129 0005 10/08/14 15:56 .099 </pre> <p style="text-align: center;">No. Date Time ZBAC Calibration 09/25/13 18:45 .100</p> <p style="text-align: center;">10/08/14 15:57 MS4 Serial no: 000888 Version no: 4797</p> <p style="text-align: center;">or press down to refresh</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1603	34	14F23-11-5-10 C1307252	.100	21		.100		X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p> <p style="text-align: center;">Push No. 3 to erase</p> <pre> Test Result 0001 10/03/14 02:11 .123 0002 10/03/14 02:19 .121 0003 10/05/14 01:23 .180 0004 10/05/14 01:25 .178 0005 10/05/14 03:17 .076 0006 10/05/14 03:18 .075 0007 10/07/14 02:28 .069 0008 10/07/14 02:30 .071 0009 10/08/14 02:29 .078 0010 10/08/14 02:31 .080 0011 10/08/14 16:09 .100 </pre> <p style="text-align: center;">No. Date Time 28AC Calibration 07/18/08 13:52 .100</p> <p style="text-align: center;">10/08/14 16:09 AS4 Serial no: 008783 Version no: 4797</p> <p style="text-align: center;">or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014			14F23-11-5-10	.100								Lee	17046	NOT IN LOWER
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1606	34	14F23-11-5-10 C1307252	.100	21		.100			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p>													
<p style="text-align: center;"> Push No. 3 to erase. Test Result 0001 10/08/14 16:11 .100 Calibration 10/30/03 09:53 .100 No. Date Time ZBAC Version no: 4797 10/08/14 16:12 NS4 Serial no: 008810 10/08/14 16:12 Version no: 4797 OF POWER DOWN TO RETAIN </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1608	34	14F23-11-5-10 C1307252	.100	22		.101			X		Lee	17046	

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

10/08/14 16:12
AS4 Serial no: 018684
Version no: 4797

No. Date Time ZBAC
Calibration
11/03/11 16:03 .100

Test Result
0001 10/08/14 16:11 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1809	34	14F23-11-5-10	.100	21		.102			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

```

10/08/14 16:18
#84 Serial no: 008635
Version no: 4797

No. Date Time %MAC
Calibration
10/24/03 09:04 .100

Test Result
0001 10/03/14 01:56 .053
0002 10/06/14 23:47 .051
0003 10/06/14 23:51 .058
0004 10/08/14 16:17 .102
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	161234	34	14F23-11-5-10	.100	22						Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push No. 3 to erase.
 or power down to retain.

10/08/14 16:17
 AS4 Serial no: 008845
 Version no: 4797

No. Date Time %BNC
 Calibration 12/24/03 11:01 .100

Test Result
 0001 10/04/14 04:04 .116
 0002 10/04/14 04:07 .113
 0003 10/04/14 22:29 .075
 0004 10/04/14 22:31 .074
 0005 10/04/14 22:39 .000
 0006 10/05/14 02:29 .137
 0007 10/05/14 02:32 .148
 0008 10/08/14 16:16 .104

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1614	34	14F23-11-5-10	.100	21		.101				X	Lee	17046	
			C1307252											

Prepared by:
 Douglas C. Talbot
 06-23-14
 Expires on:
 06-23-16

10/08/14 16:16
 H54 Serial no: 082653
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/17/14 10:08 .100
 Test Result
 0001 10/05/14 04:17 .129
 0002 10/05/14 04:19 .127
 0003 10/05/14 05:01 .112
 0004 10/05/14 05:04 .090
 0005 10/05/14 05:06 .109
 0006 10/05/14 05:09 .091
 0007 10/08/14 16:16 .101

Push No. 3 to erase

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1020	34	14F23-11-5-10	.100							Lee	17046	NOT IN LOCKER
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1621	34	14F23-11-5-10 C1307252	.100	20		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Push No. 3 to erase,
or power down to retain
 Test Result
0001 10/08/14 16:25 .100
 No. Date Time %BAC
Calibration
05/21/08 15:09 .100
 Version no: 4797
 AS4 Serial no: 077591
10/08/14 16:27
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008350
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1624	34	14F23-11-5-10	.100	21	.102			X		Lee	17046		
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Push No. 3 to erase
 Test Result
 0001 10/08/14 16:35 .102
 No. Date Time ZMNC
 Calibration 09/02/06 09:07 .100
 10/08/14 16:36
 AS4 Serial no: 000850
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	LD. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1626	34	14F23-11-5-10	.100	22		.096				X	Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											
				Push No. 3 to erase, or power down to retain	0001 10/02/14 28:05 .258	0002 10/02/14 20:09 .257	0003 10/03/14 02:49 .162	0004 10/03/14 02:46 .170	0005 10/08/14 16:27 .096	No. Date Time %BAC Calibration 04/09/14 08:36 .000	10/08/14 16:28 AS4 Serial no: 077573 Version no: 4797	or power down to retain		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	11:29	34	14F23-11-5-10	.100	21		1099					Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

No. Date Time %BAC
 Calibration 03/25/13 14:00 .100
 Test Result
 0001 10/08/14 16:30 .099

10/08/14 16:31
 AS4 Serial no: 083051
 Version no: 4797

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1636	34	14F23-11-5-10	.100	22		.096			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Push No. 3 to erase,
or power down to retain
 Test Result
 0001 10/08/14 16:40 .096
 No. Date Time ZBAC
 Calibration
 10/07/08 09:19 .100
 10/08/14 16:41
 ASD Serial no: 018624
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1639	34	14F23-11-5-10	.100	22		.099			X		Lee	17046	
			C1307252											
			Prepared by: Douglas C. Talbot 06-23-14											
			Expires on: 06-23-16											

No. Date Time %BAC
 Calibration 07/29/13 13:32 .100
 Test Result
 0001 10/08/14 16:45 .099
 Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1642	34	14F23-11-5-10	.100	21		.103			X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-23-14
Expires on:
06-23-16

Push No. 3 to erase,
or power down to retain

Test Result
0001 10/08/14 16:42 .100

No. Date Time ZBAC
Calibration
05/21/08 14:35 .100

Version no: 4797
AS4 Serial no: 077488
10/08/14 16:43

OR power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1644	34	14F23-11-5-10 C1307252	.100	21		.104				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

10/08/14 16:52
 AS4 Serial no: 008877
 Version no: 4797

No. Date Time ZBAC
 Calibration 05/26/05 10:40 .100

Test Result
 0001 10/08/14 16:52 .104

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1046	34	14F23-11-5-10 C1307252	.100	22		.096			X		Lee	17046	
Prepared by: Douglas C. Talbot 06-23-14														
Expires on: 06-23-16														
10/08/14 16:51 AS4 Serial no: 018615 Version no: 4797 No. Date Time ZBNC Calibration 12/19/08 14:44 .100 Test Result 0001 10/08/14 16:56 .098 Push No. 3 to erase														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
10/08/2014	1652	34	14F23-11-5-10 C1307252	.100	22	.100			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p> <p style="text-align: center;"> 10/08/14 16:55 AS4 Serial no: 018626 Version no: 4797 No. Date Time ZBAC Calibration 04/29/05 08:59 .101 Test Result 0001 10/08/14 16:54 .100 Push No. 3 to erase, or power down to retain </p>												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 019696
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1654	34	14F23-11-5-10	.100	22		104			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

10/08/14 16:57
 AS4 Serial no: 018696
 Version no: 4797
 No. Date Time Z/DAC
 Calibration 09/09/13 14:50 .100
 Test Result
 0001 10/08/14 16:57 .104
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	1656	34	14F23-11-5-10	.100	22	-	.100			X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

Test Result
 0001 10/04/14 23:34 .034
 0002 10/04/14 23:36 .031
 0003 10/05/14 00:57 .128
 0004 10/05/14 01:00 .127
 0005 10/06/14 00:50 .198
 0006 10/06/14 00:59 .212
 0007 10/06/14 03:43 .027
 0008 10/08/14 16:57 .100

No. Date Time %BAC
 Calibration
 01/29/09 18:39 .100

10/08/14 16:58
 AS4 Serial no: 083030
 Version no: 4797

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1059	34	14F23-11-5-10	.100	22	.100			X		Lee	17046		
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											
														10/08/14 17:04 AS4 Serial no: 018692 Version no: 4797 No. Date Time %BAC Calibration 08/12/08 14:05 .100 Test Result 0001 10/08/14 17:03 .100 Push No. 3 to erase, or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
10/08/2014	1701	34	14F23-11-5-10 C1307252	.100	21		.102			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			06-23-14												
			Expires on:												
			06-23-16												

10/08/14 17:05
 AS4 Serial no: 018762
 Version no: 4797
 No. Date Time ZBAC
 Calibration 03/30/09 15:52 .100
 Test Result
 0001 10/08/14 17:05 .102
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014			14F23-11-5-10	.100							Lee	17046	B.O.
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1704	34	14F23-11-5-10 C1307252	.100	22		.102			X		Lee	17046	
			Prepared by: Douglas C. Talbot 06-23-14											
			Expires on: 06-23-16											

10/08/14 17:11
 AS4 Serial no: 018691
 Version no: 479P
 No. Date Time ZBAC
 Calibration 09/08/08 14:01 .100
 Test Result
 0001 10/04/14 08:15 .144
 0002 10/04/14 08:18 .143
 0003 10/08/14 17:10 .102
 Push No. 3 to erase
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1700	34	14F23-11-5-10 C1307252	.100	22							Lee	17046	
<div style="display: flex; justify-content: space-between;"> <div style="width: 30%;"> <p>Prepared by: Douglas C. Talbot 06-23-14 Expires on: 06-23-16</p> </div> <div style="width: 30%; text-align: center;"> <p>Test Result 0001 10/08/14 00:05 .012 0002 10/08/14 02:22 .096 0003 10/08/14 02:24 .098 0004 10/08/14 17:10 .107</p> </div> <div style="width: 30%; text-align: center;"> <p>No. Date Time ZHAC Calibration 11/26/03 11:10 .100</p> </div> </div> <p style="text-align: center; margin-top: 10px;">Push No. 3 to erase.</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014	8:17:09	34	I4F23-11-5-10 C1307252	.100	22		.099			X	Lee	17046	
<div style="display: flex; justify-content: space-between;"> <div style="width: 30%;"> <p>Prepared by: Douglas C. Talbot 06-23-14</p> <p>Expires on: 06-23-16</p> </div> <div style="width: 30%; border: 1px solid black; padding: 5px;"> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Push No. 3 to erase, or power down to retain</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">Test Result 0001 10/06/14 19:57 .149 0002 10/06/14 20:01 .167 0003 10/06/14 17:13 .099</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">No. Date Time %BAC Calibration 01/13/09 10:25 .100</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">10/08/14 17:14 MS4 Serial no: 018774 Version no: 4797</p> <p style="writing-mode: vertical-rl; transform: rotate(180deg);">or power down to retain</p> </div> </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/08/2014			14F23-11-5-10	.100							Lee	17046	B.O.
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-23-14										
			Expires on:										
			06-23-16										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1714	34	14F23-11-5-10	.100	21		.096		X		Lee	17046		
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Push No. 3 to erase,
or Power down to retain

Test Result
 0001 10/03/14 23:45 .000
 0002 10/03/14 23:49 .000
 0003 10/04/14 04:14 .048
 0004 10/04/14 04:16 .048
 0005 10/04/14 23:41 .000
 0006 10/04/14 23:43 .000
 0007 10/05/14 05:43 .144
 0008 10/05/14 05:45 .145
 0009 10/08/14 17:19 .096

10/08/14 17:20
 AS4 Serial no: 041502
 Version no: 4797
 No. Date Time XMAC
 Calibration 05/05/09 08:03 .100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082654
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/08/2014	1717	34	14F23-11-5-10	.100	22						Lee	17046		
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-23-14											
			Expires on:											
			06-23-16											

Test Result

0001 10/03/14 20:10 .122

0002 10/03/14 20:20 .122

0003 10/03/14 22:04 .201

0004 10/03/14 22:07 .200

0005 10/04/14 02:43 .121

0006 10/04/14 02:46 .113

0007 10/05/14 03:13 .157

0008 10/05/14 03:17 .149

0009 10/05/14 23:04 .105

0010 10/05/14 23:07 .118

0011 10/06/14 02:50 .000

0012 10/06/14 02:52 .005

0013 10/06/14 17:20 .098

No. Date Time %BAC

Calibration

08/28/14 11:36 .100

10/08/14 17:26

AS4 Serial no: 082654

Version no: 4797

Push No. 3 to erase,

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.