STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

SGT

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008874

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	1012	34	C659	.106	20		.103			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

```
12/02/15 10:12
AS4 Serial no: 008874
Version    no: 4797

No.   Date      Time   %BAC
Calibration
      03/28/14 11:40  .100

Test Result
0001 11/25/15 18:40  .000
0002 11/25/15 20:31  .114
0003 11/26/15 02:01  .124
0004 11/26/15 02:03  .118
0005 12/02/15 10:11  .103

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008501

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0625	34	C659	.106	21		.093			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 06:41
ASA Serial no: 008501
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/30/09	15:40	.100

Test Result

0001	11/25/15	22:22	.165
0002	11/25/15	22:24	.173
0003	12/02/15	06:26	.098

Push No. 3 to erase,
or Power down to retain

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	013785

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0644	34	C659	.106	19		.104			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 06:45
ASA Serial no: 013785
Version no: 4797

No.	Date	Time	%BAC
Calibration	07/18/08	13:27	.100

Test Result
0001 12/02/15 06:44 .104

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008546

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0646	34	C659	.100	20		.105			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

```
or power down to retain

12/02/15 06:48
ASA Serial no: 008546
Version   no: 4797

No.  Date   Time  %BAC
Calibration
     06/28/13 16:36 .100

Test Result
0001 12/02/15 06:47 .105

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082659

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0650	34	C659	.106	20		.104			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 06:50
AS4 Serial no: 082659
Version no: 4797

No. Date Time %BAC
Calibration
03/28/14 09:04 .100

Test Result
0001 12/02/15 06:50 .104

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME.	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0652	34	C659	.106	20		.106			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 06:54
ASA Serial no: 008559
Version no: 4797

No. Date Time %BAC
Calibration
11/30/11 08:55 .100

Test Result
0001 12/02/15 06:53 .106

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008580

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0654	34	C659	.106	19		.105			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

12/02/15 06:54
AS4 Serial no# 008580
Version no# 4797

No.	Date	Time	%BAC
Calibration	03/06/14	12:39	.100

Test Result

0001	11/24/15	02:13	.143
0002	11/24/15	02:16	.146
0003	11/24/15	02:25	.063
0004	11/26/15	02:14	.131
0005	11/26/15	02:16	.131
0006	11/26/15	04:20	.162
0007	11/26/15	04:22	.146
0008	12/02/15	06:53	.105

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	083032

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	065834		C659	.106	20		.105			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

```
Push No. 3 to erase,
or power down to retain

12/02/15 07:00
A54 Serial no: 083032
Version  no: 4797

No.  Date      Time  %BAC
Calibration
     02/17/14 09:01  .100

Test Result
0001 11/25/15 22:51  .000
0002 11/26/15 21:35  .165
0003 11/26/15 21:39  .175
0004 12/02/15 06:59  .105

Push No. 3 to erase,
or power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	009620

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0700	34	C659	.106	20		.103			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 07:06
ASA Serial no: 008620
Version no: 4797

No.	Date	Time	%BAC
Calibration	01/30/13	14:34	.100

Test	Result			
0001	11/26/15	01:48	.195
0002	11/26/15	01:50	.183
0003	12/02/15	00:48	.042
0004	12/02/15	00:50	.039
0005	12/02/15	07:05	.103

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082957

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0705	34	C659	.106	20		.106			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,
or power down to retain

12/02/15 07:06
AS4 Serial no: 082957
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 08:52 .100

Test Result
0001 12/02/15 07:06 .106

Push No. 3 to erase,

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0709	34	C659	.106	20		.103			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,

Test Result
0001 12/02/15 07:06 .103

No. Date Time %BAC
Calibration
09/18/14 07:21 .099

12/02/15 07:06
ASA Serial no: 077505
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077479

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0710	34	C659	.100	20		.103			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 07:11
ASA Serial no: 077479
Version no: 4797
No. Date Time %BAC
Calibration
02/17/14 09:10 .100
Test Result
0001 12/02/15 07:10 .103
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0712	34	C659	.106	20		.105			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

Test Result
0001 11/28/15 01:06 .125
0002 11/28/15 01:09 .108
0003 11/29/15 02:11 .098
0004 11/29/15 02:13 .101
0005 11/29/15 04:05 .131
0006 11/29/15 04:08 .123
0007 12/02/15 07:13 .105

No. Date Time %BAC
Calibration
02/17/14 09:27 .100

12/02/15 07:14
ASA Serial no: 008709
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0734	34	C659	.106	20		.102			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

Test Result
0001 11/28/15 22:47 .046
0002 12/02/15 03:38 .102
0003 12/02/15 03:48 .102
0004 12/02/15 07:34 .102

No. Date Time %BAC
Calibration
03/11/04 14:05 .100

12/02/15 07:35
AS4 Serial no: 008714
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	009888

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0736	34	C659	.106	21		.100			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Push No. 3 to erase.

Test Result
0001 11/29/15 01:40 .174
0002 11/29/15 01:42 .190
0003 12/02/15 07:34 .100

No. Date Time %BAC
Calibration
03/25/13 10:45 .100

12/02/15 07:35
ASA Serial no: 009888
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	005763

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0746	34	C659	.106	39							Lee	17046	OUT of SERVICE
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/02/2015	0740	34	C659	.106	20		.102			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Push No. 3 to erase.

Test Result
0001 12/02/15 07:39 .102

No. Date Time %BAC
Calibration
02/17/14 09:50 .100

12/02/15 07:39
ASA Serial no: 082657
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008810

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0742	34	C659	.106	19		.102			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

Test Result
0001 12/02/15 07:42 .102

No. Date Time %BAC
Calibration
10/30/03 09:53 .100

12/02/15 07:43
AS4 Serial no: 008810
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018684

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0744	34	C659	.106	21		.107			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

Test Result
0001 11/26/15 17:05 .294
0002 11/26/15 17:09 .301
0003 12/02/15 07:44 .107

No. Date Time %BAC
Calibration
11/03/11 16:03 .100

12/02/15 07:44
454 Serial no: 018684
Version no: 4757

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0809	34	C659	.106	20		.104			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,
or power down to retain

Test Result
0001 12/02/15 08:11 .104

No. Date Time %BAC
Calibration
10/24/03 09:04 .100

12/02/15 08:11
ASA Serial no: 008635
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008845

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0812	34	C659	.106	20		.110			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 06:12
ASA Serial no: 008845
Version no: 4797

No. Date Time %BAC
Calibration
12/24/03 11:01 .100

Test Result
0001 12/02/15 06:12 .110

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0814	34	C659	.106	20		.105			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase
or power down to retain

12/02/15 08:16
ASA Serial no: 082653
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 10:00 .100

Test Result
0001 12/02/15 08:15 .105

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008948

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0816	34	C659	.106	20		.107			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 08:16
ASA Serial no: 008948
Version no: 4797
No. Date Time %BAC
Calibration
01/13/03 09:25 .100
Test Result
0001 12/02/15 08:16 .107

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077591

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0820	34	C659	.106	19							Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,
or Power down to retain

12/02/15 08:21
ASA Serial no: 077591
Version no: 4797

No. Date Time %BAC
Calibration
05/21/08 15:09 .100

Test Result
0001 12/02/15 08:20 .104

Push No. 3 to erase,

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008850

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/02/2015	1150	34	C659	.106	33		.100			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase
or power down to retain

Test Result
0001 12/02/15 11:51 .100

12/02/15 11:52
ASA Serial no: 008850
Version no: 4797

No. Date Time %BAC
Calibration
03/02/06 09:07 .100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0837	34	C659	.106	.25		.099			X		Lee	17046	CALIBRATED PREVIOUS CALIBRATION WAS .080
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Test Result
0001 12/02/15 08:26 .099
0002 12/02/15 08:38 .099

No. Date Time %BAC
Calibration
12/02/15 08:37 .100

12/02/15 08:40
AS4 Serial no:-077573
Version no: 4797

Push No. 3 to erase,
or power down to retain

12/02/15 08:26 .099

Test Result
0001 12/02/15 08:26 .099

No. Date Time %BAC
Calibration
04/09/14 08:36 .080

12/02/15 08:27
AS4 Serial no: 077573
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	083051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0830	34	C659	.106	20		.101			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 08:30
ASA Serial no: 083051
Version no: 4797

No.	Date	Time	%BAC
Calibration	03/25/13	14:00	.100

Test	Result		
0001	11/27/15	13:08	.147
0002	12/02/15	08:28	.101

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008397

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0833	34	C659	.106	.20		.109			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,

Test Result
0001 11/29/15 01:55 .056
0002 11/29/15 01:59 .056
0003 12/02/15 08:34 .109

No. Date Time %BAC
Calibration
01/18/12 16:13 .100

12/02/15 08:34
AS4 Serial no: 008997
Version no: 4797

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0843	34	C659	.106	22		.100			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

```
Push No. 3 to erase
or Power down to retain

12/02/15 08:43
ASA Serial no: 018624
Version no: 4797

No.  Date     Time  %BAC
Calibration
     10/07/08 09:19 .100

Test Result
0001 11/26/15 03:08 .137
0002 11/26/15 03:12 .132
0003 12/01/15 03:29 .175
0004 12/01/15 03:32 .168
0005 12/02/15 08:43 .100
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018646

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0845	34	C659	.100	22		.102			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,

Test Result
0001 11/28/15 01:22 .028
0002 11/28/15 01:23 .032
0003 11/29/15 02:30 .095
0004 11/29/15 02:33 .092
0005 12/02/15 08:45 .102

No. Date Time %BAC
Calibration
07/29/13 13:32 .100

12/02/15 08:45
ASA Serial no: 018646
Version no: 4797

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	077488

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0947	34	C659	.106	20		.107			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Push No. 3 to erase
or power down to retain

Test Result
0001 12/02/15 08:45 .107

No. Date Time %BAC
Calibration
05/21/08 14:35 .100

12/02/15 08:46
AS4 Serial no: 077488
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0949	34	C659	.106	20		.106			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase

12/02/15 08:51
AS4 Serial no: 008877
Version no: 4797

No.	Date	Time	%BAC
Calibration	05/26/05	10:40	.100
Test Result			
0001	11/29/15	04:50	.107
0002	11/29/15	04:52	.108
0003	12/01/15	03:00	.173
0004	12/01/15	03:02	.168
0005	12/02/15	08:50	.106

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018615

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0851	34	C659	.106	20		.104			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Push No. 3 to erase, or Power down to retain

Test Result
0001 12/02/15 09:52 .104

No. Date Time %DAC
Calibration
12/19/08 14:44 .100

12/02/15 09:52
NS4 Serial no: 018615
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018626

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0853	34	C659	.106	21		.103			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 08:53
ASA Serial no: 018626
Version no: 4797
No. Date Time %BAC
Calibration
04/29/05 08:59 .101
Test Result
0001 12/02/15 08:52 .103
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018696

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0928	34	C659	.106	27		.100			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,
or power down to retain

Test Result
0001 12/02/15 08:57 .119
0002 12/02/15 09:27 .100

No. Date Time %BAC
Calibration
12/02/15 09:24 .100

12/02/15 09:28
AS4 Serial no: 018696
Version no: 4797

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018692

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0859	34	C659	.106	21		.103			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 09:01
ASA Serial no: 018692
Version no: 4797
No. Date Time %BAC
Calibration
08/12/08 14:35 .100
Test Result
0001 12/02/15 09:00 .103
Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018762

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0902	34	C659	.106	20		.105			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase,
or Power down to retain

Test Result
0001 12/02/15 09:01 .105

No. Date Time %BAC
Calibration
03/30/09 15:52 .100

12/02/15 09:02
ASA Serial no: 018762
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0922	34	C659	.106	20							Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 09:22
ASA Serial no: 018766
Version no: 4797
No. Date Time %BAC
Calibration
09/12/14 12:45 .100
Test Result
0001 12/02/15 09:21 .103
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0926	34	C659	.106	22		.107			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

12/02/15 09:26
AS4 Serial no: 018691
Version no: 4797

No.	Date	Time	%BAC
Calibration	09/08/08	14:01	.100

Test	Result		
0001	11/26/15	02:16	.106
0002	11/26/15	02:18	.093
0003	11/28/15	01:57	.057
0004	11/28/15	01:59	.064
0005	11/28/15	02:26	.088
0006	11/28/15	02:29	.106
0007	11/30/15	02:25	.047
0008	11/30/15	03:12	.176
0009	11/30/15	03:15	.179
0010	12/01/15	01:15	.230
0011	12/01/15	22:11	.168
0012	12/01/15	22:14	.163
0013	12/02/15	03:14	.100
0014	12/02/15	03:15	.096
0015	12/02/15	09:26	.107

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	017990

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0930	34	C659	.106	20		.109			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 09:33
ASA Serial no: 017990
Version no: 4797

No. Date Time %BAC
Calibration
11/26/03 11:10 .100

Test Result
0001 12/02/15 09:30 .108

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	018774

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0940	34	C659	.106	21		.104			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

Push No. 3 to erase.

12/02/15 09:40
ASA Serial no: 018774
Version no: 4797

No.	Date	Time	%BAC
Calibration	01/13/09	10:25	.100

Test Result			
0001	11/24/15	00:19	.181
0002	11/24/15	00:23	.170
0003	11/24/15	00:26	.036
0004	11/25/15	02:46	.095
0005	11/25/15	02:49	.007
0006	11/27/15	00:10	.162
0007	11/27/15	00:12	.168
0008	11/27/15	02:20	.018
0009	11/27/15	02:22	.017
0010	12/02/15	09:40	.104

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	041486

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0954	34	C659	.106	21		.107			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

Push No. 3 to erase.

Test Result
0001 12/02/15 09:56 .107

No. Date Time %BAC
Calibration
03/08/06 13:56 .100

12/02/15 09:57
AS4 Serial no: 041486
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	041495

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	0950	34	C659	.106	20							Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 09:58
ASA Serial no: 041495
Version no: 4792

No. Date Time %BAC
Calibration
02/20/14 11:28 .100

Test Result
0001 12/02/15 09:57 .101

Push No. 3 to erase,
or power down to retain

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	041502

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	1000	34	C659	.106	20		.100			X		Lee	17046	
			Prepared by:											
			LASO Crime Lab											
			Expires on:											
			08/21/2016											

12/02/15 10:08
AS4 Serial no: 041502
Version no: 4797

No. Date Time %BAC
Calibration
05/05/09 08:03 .100

Test Result
0001 12/02/15 09:59 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
12/02/2015	1010	34	C659	.106	20		.101			X		Lee	17046	

Prepared by:
LASO Crime Lab

Expires on:
08/21/2016

```
12/02/15 10:09
AS4 Serial no: 082654
Version nos: 4797

No.  Date     Time  %BAC
Calibration
     03/28/14 11:36 .100

Test Result
0001 11/27/15 21:55 .047
0002 11/28/15 00:14 .152
0003 11/28/15 00:18 .000
0004 11/28/15 02:33 .120
0005 11/28/15 02:35 .119
0006 11/28/15 19:22 .173
0007 11/28/15 19:28 .171
0008 12/02/15 10:09 .101

Push No. 3 to erase,
or Power down to retain
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.