

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/04/2016 08:21

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3321
Check sum	26423	Download date/time:	08/04/2016	08:36		Record ID	8R46
Record imported (Y/N)				N	Source Diagnostic		

Instrument serial No: 68-012340 Accuracy check date/time: 08/04/2016 08:22

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 004

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	59220	Download date/time:	08/04/2016	08:36		Record ID	8QW2
Record imported (Y/N)				N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/04/2016 08:25

First calibration check: .119 Interferent detected: No
Second calibration check: .120 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 004

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 33717	Download date/time: 08/04/2016	08:36	Record ID 8QX1
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 08/09/2016 19:34

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49037	Download date/time: 09/01/2016	08:40	Record ID 8S87
Record imported (Y/N) N	Source Exception		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/11/2016 08:18

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3295	
Check sum	59656	Download date/time:	09/01/2016	08:40		Record ID	8S36	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 08/11/2016 08:20

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	32509	Download date/time:	09/01/2016	08:40		Record ID	8RU1	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/11/2016 08:24

First calibration check: .118 Interferent detected: No
Second calibration check: .120 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 018

Operator A BERGERON 10821

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 26403	Download date/time: 09/01/2016	08:40	Record ID 8RV6
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 08/18/2016 11:03

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3293
Check sum 38972	Download date/time: 09/01/2016	08:40	Record ID 8S49
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/18/2016 11:05

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 39326	Download date/time: 09/01/2016	08:40	Record ID 8RW1
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 08/18/2016 11:11

First calibration check: .111 Interferent detected: No
Second calibration check: .114 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.04
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 014

Operator A BERGERON 10821 _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 65473	Download date/time: 09/01/2016	08:40	Record ID 8RX8
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/18/2016 11:20

First calibration check: .121 Interferent detected: No
Second calibration check: .121 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.09
Simulator Ser No. DR3707

COLD SIM HOSE LAST CAL CHECK

Number of people tested since last periodic determination of accuracy: 000

Operator A BERGERON 10821

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 27048	Download date/time: 09/01/2016	08:40	Record ID 8RY3
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340 Accuracy check date/time: 08/25/2016 07:15

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom	Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom	Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram	Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3285
Check sum 18880	Download date/time: 09/01/2016	08:40	Record ID 8S50	
Record imported (Y/N) N		Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/25/2016 07:18

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 018

Operator _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 15856	Download date/time: 09/01/2016	08:40	Record ID 8RZ8
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340 Accuracy check date/time: 08/25/2016 07:20

First calibration check: .115 Interferent detected: No
Second calibration check: .115 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 018

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 63500	Download date/time: 09/01/2016	08:40	Record ID 8S01
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 08/25/2016 07:25

First calibration check: .117 Interferent detected: No
Second calibration check: .120 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.07
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 15091	Download date/time: 09/01/2016	08:40	Record ID 8S12
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 08/28/2016 00:35

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3274
Check sum 52013	Download date/time: 09/01/2016	08:40	Record ID 8S63
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.