STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

SGT

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008874

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1254	34	15E30-15-5-10-C14L	.100	22		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result
0001 12/31/16 01:38 .067
0002 12/31/16 01:39 .065
0003 12/31/16 02:40 .000
0004 12/31/16 12:39 .144
0005 12/31/16 12:41 .139
0006 01/06/17 13:01 .105

No. Date Time %BAC
Calibration
01/22/16 13:23 .100

01/06/17 13:02
AS4 Serial no: 008874
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	0943	34	15E30-15-5-10-C14L	.100	22		.099			X		Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 5-30-15											
			Expiration date:											
			5-30-17											

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120188
UOM	BAC
Last Check	01/06/2017
Time:	09:42 AM
Cal Standard	.100
Result	.099
Last Cal	10/21/2016
Time:	08:16 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the [illegible] 110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	0945	34	15E30-15-5-10-C14L	.100	21		.103			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120194
UOM	BAC
Last Check	01/06/2017
Time:	09:47 AM
Cal Standard	.100
Result	.103
Last Cal	11/21/2016
Time:	10:51 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the … 110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	0949	34	15E30-15-5-10-C14L	.100	22		.099			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120197
UOM	BAC
Last Check	01/06/2017
Time:	09:50 AM
Cal Standard	.100
Result	.099
Last Cal	10/21/2016
Time:	10:22 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the ac

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

0).

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	0957	34	15E30-15-5-10-C14L	.100	22		.099			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120195
UOM	BAC
Last Check	01/06/2017
Time:	09:58 AM
Cal Standard	.100
Result	.099
Last Cal	10/21/2016
Time:	09:53 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the ac).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1001	34	15E30-15-5-10-C14L	.100	20		.097			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result
0001 01/06/17 10:08 .097

No. Date Time %BAC
Calibration
01/15/16 10:40 .100

01/06/17 10:08
AS4 Serial no: 008559
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1004	34	15E30-15-5-10-C14L	.100	21		.101			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Lifeloc Technologies	Calibration / Check Printout
FC20	8.7.24
Serial No.	16040161
UOM	BAC
Last Check	01/06/2017
Time:	10:05 AM
Cal Standard	.100
Result	.101
Last Cal	10/21/2016
Time:	10:20 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the act

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/06/2017	1007	34	15E30-15-5-10-C14L	.100	21		.103			X		Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 5-30-15											
			Expiration date:											
			5-30-17											

Lifeloc Technologies	Calibration / Check Printout	FC20	Serial No.	UOM	Last Check	Time:	Cal Standard	Result	Last Cal	Time:	Cal Standard
		8.7.24	16040140	BAC	01/06/2017	10:08 AM	.100	.103	11/21/2016	12:08 PM	.100

* If the reading is within plus or minus (±) 0.010 percent of the [illegible] .0).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008620

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1018	34	15E30-15-5-10-C14L	.100	20		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase.

01/06/17 10:21 .101
Test Result
0001 01/06/17 10:22 .100
No. Date Time %BAC
Calibration
01/15/16 11:10
AS4 Serial no: 008620
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082 957

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1022	34	15E30-15-5-10-C14L	.100	21		.095			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 10:23
AS4 Serial no: 082957
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 08:52 .100

Test Result
0001 01/06/17 10:23 .095

Push No. 3 to erase.

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1024	34	15E30-15-5-10-C14L	.100	20		.104			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or Power down to retain

Test Result
0001 01/06/17 10:25 .104

No. Date Time %BAC
Calibration
01/22/16 10:42 .100

01/06/17 10:25
AS4 Serial no: 077505
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077479

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1044	34	15E30-15-5-10-C14L	.100	20		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 10:46
ASA Serial no: 077479
Version no: 4797

No. Date Time %BAC
Calibration
01/22/16 11:01 .100

Test Result
0001 01/06/17 10:46 .101

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1046	34	15E30-15-5-10-C14L	.100	20		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result Date Time %BAC
0001 12/30/16 03:57 .108
0002 12/30/16 03:58 .109
0003 12/31/16 22:27 .061
0004 12/31/16 22:28 .063
0005 01/01/17 02:40 .112
0006 01/01/17 02:42 .119
0007 01/06/17 10:48 .101

Calibration
02/11/16 10:14 .100

01/06/17 10:49
AS4 Serial no: 082653
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1048	34	15E30-15-5-10-C14L	.100	21		.102			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase.
Test Result
0001 01/06/17 10:57 .102
No. Date Time %BAC
Calibration
03/17/16 09:40 .100
01/06/17 10:58
AS4 Serial no: 008709
Version no: 4797
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1053	34	15E30-15-5-10-C14L	.100	20		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 11:02
AS4 Serial no: 008714
Version no: 4797

No.	Date	Time	%BAC
Calibration	01/15/16	12:37	.100

Test Result

0001	12/30/16	00:58	.143
0002	12/30/16	01:00	.111
0003	12/30/16	01:02	.140
0004	12/30/16	01:04	.130
0005	12/31/16	02:35	.062
0006	12/31/16	02:38	.060
0007	12/31/16	03:21	.257
0008	12/31/16	03:24	.241
0009	12/31/16	21:33	.162
0010	12/31/16	21:35	.166
0011	01/02/17	03:20	.134
0012	01/02/17	03:22	.123
0013	01/02/17	03:29	.087
0014	01/06/17	11:01	.105

Push No. 3 to erase,
or power down to retain

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008888

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1056	34	15E30-15-5-10-C14L	.100	21		.102			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 10:57
ASA Serial no: 008888
Version no: 4797

No. Date Time %BAC
Calibration
03/17/16 09:44 .100

Test Result
0001 01/06/17 10:56 .102

Push No. 3 to erase
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008783

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1058	34	15E30-15-5-10-C14L	.100	20		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 11:06
AS4 Serial no: 008783
Version no: 4797

No.	Date	Time	%BAC
Calibration	02/11/16	10:10	.100

Test Result			
0001	01/03/17	20:06	.176
0002	01/03/17	20:08	.173
0003	01/06/17	11:05	.103

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1100	34	15E30-15-5-10-C14L	.100	21		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase.

Test Result
0001 01/06/17 11:03 .103

No. Date Time %BAC
Calibration
01/22/16 11:19 .100

01/06/17 11:03
034 Serial no: 082657
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008810

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1102	34	15E30-15-5-10-C14L	.100	20		.104			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase
or power down to retain

Test Result
0001 01/06/17 11:10 .104

No. Date Time %BAC
Calibration
01/22/16 12:34 .100

01/06/17 11:11
ASA Serial no: 008810
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018684

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1104	34	15E30-15-5-10-C14L	.100	20		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or Power down to retain

Test Result
0001 12/31/16 04:23 .041
0002 01/02/17 04:57 .111
0003 01/02/17 04:58 .109
0004 01/06/17 11:10 .103

No. Date Time %BAC
Calibration
01/12/16 14:36 .093

01/06/17 11:10
ASA Serial no: 018684
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1106	34	15E30-15-5-10-C14L	.100	20		.106			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 11:21
ASA Serial no: 008635
Version no: 4797

No. Date Time %BAC
Calibration
01/22/16 12:39 .100

Test Result
0001 12/30/16 01:09 .081
0002 12/30/16 01:12 .089
0003 12/30/16 01:21 .067
0004 12/31/16 03:53 .089
0005 12/31/16 03:54 .095
0006 12/31/16 19:29 .163
0007 12/31/16 19:31 .167
0008 01/01/17 01:56 .092
0009 01/01/17 01:59 .093
0010 01/01/17 02:28 .085
0011 01/01/17 03:42 .214
0012 01/01/17 03:44 .169
0013 01/01/17 03:45 .193
0014 01/06/17 11:20 .106

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008845

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1129	34	15E30-15-5-10-C14L	.100	21		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase
or Power down to retain

Test Result
0001 01/02/17 02:59 .191
0002 01/02/17 03:03 .196
0003 01/06/17 11:36 .101

No. Date Time %BAC
Calibration
12/01/16 10:43 .100

01/06/17 11:35
ASA Serial no: 008845
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1131	34	15E30-15-5-10-C14L	.100	23		.100			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120177
UOM	BAC
Last Check	01/06/2017
Time:	11:32 AM
Cal Standard	.100
Result	.100
Last Cal	10/21/2016
Time:	1:48 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the act ... 10).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1143	34	15E30-15-5-10-C14L	.100	20		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 11:51
AS4 Serial no: 008848
Version no: 4797

No. Date Time %BAC
Calibration
02/11/16 10:33 .100

Test Result
0001 01/06/17 11:51 .105

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077591

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1145	34	15E30-15-5-10-C14L	.100	20		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result
0001 01/06/17 11:50 .103

No. Date Time %BAC
Calibration
02/11/16 10:27 .100

01/06/17 11:50
AS4 Serial no: 077591
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008850

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1117	34	15E30-15-5-10-C14L	.100	21		.094			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or Power down to retain

Test Result
0001 01/06/17 11:52 .094
No. Date Time XBAC
Calibration
05/24/16 16:01 .104
Version nos: 47327
ASA Serial nos: 008850
01/06/17 11:52

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1149	34	15E30-15-5-10-C14L	.100	21		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 11:51
ASA Serial nos: 077573
Version nos: 4797

No. Date Time %BAC
Calibration
10/21/16 14:23 .100

Test Result
0001 01/06/17 11:51 .101

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	083051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	0954	34	15E30-15-5-10-C14L	.100	20		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 09:55
ASA Serial no: 083051
Version no: 4797
No. Date Time %BAC
Calibration
05/24/16 16:24 .100
Test Result
0001 01/06/17 09:54 .103
Push No. 3 to erase,

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008897

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1151	34	15E30-15-5-10-C14L	.100	20		.099			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or power down to retain

Test Result
0001 01/06/17 11:58 .099

No. Date Time %BAC
Calibration
01/18/12 16:13 .100

01/06/17 11:59
ASA Serial no: 008897
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1153	30	15E30-15-5-10-C14L	.100	22		.100			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:00
RS4 Serial no: 018624
Version no: 4797

No. Date Time %BAC
Calibration
02/01/16 11:02 .100

Test Result
0001 01/06/17 11:59 .100

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018646

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1156	34	15E30-15-5-10-C14L	.100	20		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:05
ASA Serial no: 018646
Version no: 4797

No.	Date	Time	%BAC
Calibration	01/22/16	12:59	.100

Test Result

0001	12/30/16	02:41	.005
0002	12/31/16	02:44	.026
0003	12/31/16	03:26	.090
0004	12/31/16	03:29	.092
0005	12/31/16	03:37	.103
0006	12/31/16	03:39	.132
0007	01/01/17	02:19	.073
0008	01/01/17	02:39	.154
0009	01/01/17	02:41	.152
0010	01/01/17	03:35	.079
0011	01/01/17	03:39	.083
0012	01/06/17	12:04	.105

Push No. 3 to erase

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077488

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1159	30	15E30-15-5-10-C14L	.100	21		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or Power down to retain

Test Result
0001 01/06/17 12:01 .101

No. Date Time %BAC
Calibration
02/11/16 10:52 .100

01/06/17 12:02
ASA Serial no: 077488
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1201	34	15E30-15-5-10-C14L	.100	22		.104			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result
0001 12/30/16 19:35 .157
0002 12/30/16 19:37 .161
0003 12/30/16 22:14 .079
0004 12/30/16 22:16 .077
0005 12/31/16 03:49 .181
0006 12/31/16 03:52 .197
0007 01/06/17 12:05 .104

No. Date Time %BAC
Calibration
02/11/16 11:15 .100

01/06/17 12:06
ASA Serial no: 008877
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1204	34	15E30-15-5-10-C14L	.100	23		.106			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Lifeloc Technologies	Calibration / Check Printout	FC20	Serial No.	UOM	Last Check	Time:	Cal Standard	Result	Last Cal	Time:	Cal Standard
		8.7.24	16040163	BAC	01/06/2017	12:02 PM	.100	.106	10/27/2016	1:23 PM	.100

* If the reading is within plus or minus (±) 0.010 percent of the .110).

** If using dry gas for accuracy checks, note the canister serial n

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018626

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1207	34	15E30-15-5-10-C14L	.100	20		.096			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or power down to retain

Test Result
0001 01/06/17 12:09 .096

No. Date Time %BAC
Calibration
07/15/16 15:21 .100

01/06/17 12:10
NS4 Serial no: 018626
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	01 8696

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1210	34	15E30-15-5-10-C14L	.100	21		.100			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:14
ASA Serial no: 018696
Version no: 4797
No. Date Time %BAC
Calibration
04/12/16 10:21 .100

Test Result
0001 01/06/17 12:13 .100

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	083030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1212	34	15E30-15-5-10-C14L	.100	21		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:13
ASA Serial no: 083030
Version no: 4797

No. Date Time %BAC
Calibration
02/11/16 11:51 .100

Test Result
0001 01/06/17 12:12 .105

Push No. 3 to erase,
or Power down to retain

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018692

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1214	34	15E30-15-5-10-C14L	.100	21		.101			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:21
ASA Serial no: 018692
Version no: 4797

No. Date Time %BAC
Calibration
02/01/16 11:30 .100

Test Result
0001 01/06/17 12:20 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018762

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1216	34	15E30-15-5-10-C14L	.100	21		.100			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:22
AS4 Serial no: 018762
Version no: 4797

No.	Date	Time	%BAC
	Calibration		
	03/17/16	15:50	.100

Test	Result		
0001	12/30/16	21:02	.218
0002	12/30/16	21:04	.216
0003	12/31/16	20:35	.203
0004	12/31/16	20:37	.216
0005	12/31/16	20:39	.000
0006	01/01/17	02:22	.129
0007	01/01/17	02:25	.127
0008	01/06/17	12:22	.100

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	019766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1222	34	15E30-15-5-10-C14L	.100	21		.095			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:26
ASA Serial no: 019766
Version no: 4797

No. Date Time %BAC
Calibration
01/22/16 13:16 .100

Test Result
0001 01/06/17 12:26 .095

Push No. 3 to erase,
or power down to retain

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	018691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1233	34	15E30-15-5-10-C14L	.100	21		.103			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:44
ASA Serial no: 018691
Version no: 4797

No. Date Time %BAC
Calibration
01/22/16 13:18 .100

Test Result
0001 01/06/17 12:43 .103

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	017990

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1235	34	15E30-15-5-10-C14L	.100	21		.099			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase,
or Power down to retain

Test Result
0001 01/04/17 02:46 .155
0002 01/04/17 02:49 .153
0003 01/06/17 12:43 .099

No. Date Time %BAC
Calibration
11/26/03 11:10 .100

01/06/17 12:44
454 Serial no: 017990
Version no: 4797

or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)** TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)** LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)** VALUE	PAS TEMP (°C) TEST #1	PAS TEMP (°C) TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS* TEST #1	PAS ACCURACY CHECK RESULTS* TEST #2 (Optional)	CALIBRATION REQUIRED YES	CALIBRATION REQUIRED NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1238	34	15E30-15-5-10-C14L	.100	22		.099			X		Lee	17046	
			19002											

Solution prepared by:
Douglas C. Talbot
Date: 5-30-15
Expiration date:
5-30-17

Cal Standard	.100	
Time:	10:32 AM	
Last Cal	10/21/2016	
Result	.099	
Cal Standard	.100	
Time:	12:37 PM	
Last Check	01/06/2017	
UOM	BAC	
Serial No.	16120189	
FC20	8.7.24	

Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of th

** If using dry gas for accuracy checks, note the canister serial

0.110).

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	041486

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1240	34	15E30-15-5-10-C14L	.100	21		.098			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase

Test Result
0001 01/06/17 12:48 .098

No. Date Time %BAC
Calibration
03/08/06 13:56 .100

01/06/17 12:48
ASA Serial no: 041486
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	041495

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1242	34	15E30-15-5-10-C14L	.100	20		.105			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase, or power down to retain

Test Result
0001 01/06/17 12:41 .105

No. Date Time %BAC
Calibration
01/22/16 14:31 .100

01/06/17 12:42
ASY Serial no: 041495
Version no: 4797

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	041502

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1244	34	15E30-15-5-10-C14L	.100	21		.097			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

01/06/17 12:53
ASA Serial no: 041502
Version no: 4797
No. Date Time %BAC
Calibration
01/15/16 14:20 .100
Test Result
0001 01/06/17 12:52 .097
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

		SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED					
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
01/06/2017	1252-31		15E30-15-5-10-C14L	.100	20		.096			X		K. Lee	17046	
			19002											
			Solution prepared by:											
			Douglas C. Talbot											
			Date: 05-30-15											
			Expiration date:											
			5-30-17											

Push No. 3 to erase.

Test Result
0001 12/30/16 02:14 .054
0002 12/30/16 02:16 .061
0003 01/05/17 00:42 .000
0004 01/06/17 00:37 .057
0005 01/06/17 00:40 .059
0006 01/06/17 12:55 .098

No. Date Time %BAC
Calibration
12/19/16 11:26 .100

01/06/17 12:56
HS4 Serial no: 082654
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.