

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/02/2017 14:48

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3184	
Check sum	18957	Download date/time:	02/02/2017	15:02		Record ID	9442	
Record imported (Y/N)				N	Source Diagnostic			

Instrument serial No: 68-012335 Accuracy check date/time: 02/02/2017 14:49

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .126

Solution lot#: IRS070116

Soln temp C: 34.02

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	21912	Download date/time:	02/02/2017	15:02		Record ID	93Z7	
Record imported (Y/N)				N	Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/02/2017 14:53

First calibration check: .123 Interferent detected: No
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.08
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 65195	Download date/time: 02/02/2017	15:02	Record ID 9402
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012335 Accuracy check date/time: 02/09/2017 08:57

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3197
Check sum 31996	Download date/time: 03/02/2017	12:41	Record ID 9574
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/09/2017 09:00

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.08
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 59170	Download date/time: 03/02/2017	12:41	Record ID 9503
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012335 Accuracy check date/time: 02/09/2017 09:02

First calibration check: .121 Interferent detected: No
Second calibration check: .124 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.04
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 37453	Download date/time: 03/02/2017	12:41	Record ID 9510
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/13/2017 02:35

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3193
Check sum	64257	Download date/time:	03/02/2017 12:41	Record ID	9589		
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012335 Accuracy check date/time: 02/16/2017 10:14

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3183
Check sum	36548	Download date/time:	03/02/2017 12:41	Record ID	9596		
Record imported (Y/N) N				Source Diagnostic			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/16/2017 10:15

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.04
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49623	Download date/time: 03/02/2017	12:41	Record ID 9527
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012335 Accuracy check date/time: 02/16/2017 10:20

First calibration check: .124 Interferent detected: No
Second calibration check: .126 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.07
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 25707	Download date/time: 03/02/2017	12:41	Record ID 9534
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/23/2017 10:37

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3184	
Check sum	39731	Download date/time:	03/02/2017	12:41		Record ID	95A7	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012335 Accuracy check date/time: 02/23/2017 10:38

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.07
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	14032	Download date/time:	03/02/2017	12:41		Record ID	9543	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012335 Accuracy check date/time: 02/23/2017 10:41

First calibration check: .121 Interferent detected: No
Second calibration check: .123 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.06
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 18480	Download date/time: 03/02/2017	12:41	Record ID 9550
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.