

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/09/2017 14:01

First calibration check: .121 Interferent detected: No  
Second calibration check: .124 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.01  
Simulator Ser No. DR3707

PLACED INTO SERVICE - BOOKING

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 12881	Download date/time: 03/02/2017	13:24	Record ID 9635
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/09/2017 14:57

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3431
Check sum 48742	Download date/time: 03/02/2017	13:24	Record ID 96B7
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/09/2017 14:59

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3477	
Check sum	20405	Download date/time:	03/02/2017	13:24		Record ID	96C4	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012337 Accuracy check date/time: 02/10/2017 23:44

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Unstable reference

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	17825	Download date/time:	03/02/2017	13:24		Record ID	96G2	
Record imported (Y/N)				N				Source Exception

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/12/2017 17:15

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 34138	Download date/time: 03/02/2017	13:24	Record ID 96H1
Record imported (Y/N) N	Source	Exception	

---

Instrument serial No: 68-012337 Accuracy check date/time: 02/16/2017 08:49

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3547
Check sum 44005	Download date/time: 03/02/2017	13:24	Record ID 96D5				
Record imported (Y/N) N	Source	Diagnostic					

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/16/2017 08:51

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 016

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 25516	Download date/time: 03/02/2017	13:24	Record ID 9648
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012337 Accuracy check date/time: 02/16/2017 08:58

First calibration check: .121 Interferent detected: No  
Second calibration check: .121 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 33.97  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 016

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49440	Download date/time: 03/02/2017	13:24	Record ID 9655
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/23/2017 09:15

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3555	
Check sum	58468	Download date/time:	03/02/2017	13:24		Record ID	96E2	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012337 Accuracy check date/time: 02/23/2017 09:17

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 006

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	18511	Download date/time:	03/02/2017	13:24		Record ID	9660	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 02/23/2017 09:19

First calibration check: .120 Interferent detected: No  
Second calibration check: .120 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.04  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 006

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49000	Download date/time: 03/02/2017	13:24	Record ID 9677
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.