

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/13/2017 08:52

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 003

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 65440	Download date/time: 05/04/2017	08:39	Record ID 98V5
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 04/13/2017 08:56

First calibration check: .127 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.01  
Simulator Ser No. DR3707

PLACED INTO SERVICE - BOOKING

Number of people tested since last periodic determination of accuracy: 003

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 63360	Download date/time: 05/04/2017	08:39	Record ID 98W6
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/13/2017 09:24

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass		
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass		
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3273		
Check sum	56369	Download date/time:	05/04/2017	08:39		Record ID	9944		
Record imported (Y/N)				N				Source	Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 04/15/2017 19:20

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards			
Prom		Real Time clock		Motor		Printer			
Ram		Auto calibration		Range		Digital voltmeter			
Check sum	37140	Download date/time:	05/04/2017	08:39		Record ID	9995		
Record imported (Y/N)				N				Source	Exception

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/21/2017 08:50

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3466	
Check sum	13060	Download date/time:	05/04/2017	08:39		Record ID	9959	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 04/21/2017 08:53

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 017

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	63494	Download date/time:	05/04/2017	08:39		Record ID	98X9	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/21/2017 08:55

First calibration check: .115 Interferent detected: No  
Second calibration check: .117 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.01  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 017

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 48381	Download date/time: 05/04/2017	08:39	Record ID 98Y0
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012340 Accuracy check date/time: 04/21/2017 09:01

First calibration check: .120 Interferent detected: No  
Second calibration check: .120 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.02  
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SOLN

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 44121	Download date/time: 05/04/2017	08:39	Record ID 98Z1
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/22/2017 03:02

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49676	Download date/time: 05/04/2017	08:39	Record ID 99A2
Record imported (Y/N) N	Source Exception		

Instrument serial No: 68-012340 Accuracy check date/time: 04/27/2017 08:00

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3466
Check sum 40051	Download date/time: 05/04/2017	08:39	Record ID 9960
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/27/2017 08:02

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 025

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 54150	Download date/time: 05/04/2017	08:39	Record ID 9900
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 04/27/2017 08:07

First calibration check: .120 Interferent detected: No  
Second calibration check: .121 Interferent detected: No

Concentration (%W/V) .126 Solution lot#: IRS070116 Soln temp C: 34.00  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 025

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 48054	Download date/time: 05/04/2017	08:39	Record ID 9915
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.