

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 06/01/2017 09:16

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3444
Check sum	64078	Download date/time:	06/01/2017 09:30			Record ID	9BG7
	Record imported (Y/N)	N	Source	Diagnostic			

Instrument serial No: 68-012337 Accuracy check date/time: 06/01/2017 09:18

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.05  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards		
Prom	Real Time clock	Motor	Printer		
Ram	Auto calibration	Range	Digital voltmeter		
Check sum	22833	Download date/time:	06/01/2017 09:29	Record ID	9BB6
	Record imported (Y/N)	N	Source	Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012337 Accuracy check date/time: 06/01/2017 09:33

First calibration check: .123 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.02  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 29030	Download date/time: 06/01/2017	09:45	Record ID 9BH0
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.