

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321 Accuracy check date/time: 09/01/2017 16:40

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3552
Check sum	10889	Download date/time:	10/05/2017 11:24			Record ID	9LA2
		Record imported (Y/N)	N			Source Diagnostic	

Instrument serial No: 68-012321 Accuracy check date/time: 09/07/2017 07:45

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3340
Check sum	22091	Download date/time:	10/05/2017 11:24			Record ID	9LB7
		Record imported (Y/N)	N			Source Diagnostic	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321 Accuracy check date/time: 09/12/2017 23:51

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3857
Check sum 46589	Download date/time: 10/05/2017 11:24		Record ID 9LC4
Record imported (Y/N) N	Source Diagnostic		

Instrument serial No: 68-012321 Accuracy check date/time: 09/13/2017 01:01

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3859
Check sum 42012	Download date/time: 10/05/2017 11:24		Record ID 9LD9
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012321 Accuracy check date/time: 09/14/2017 09:34

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3911	
Check sum	58759	Download date/time:	10/05/2017	11:24		Record ID	9LE6	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012321 Accuracy check date/time: 09/21/2017 10:31

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards		
Prom	Pass	Real Time clock		Motor	Pass	Printer		
Ram	Pass	Auto calibration	Pass	Range	Fail	Digital voltmeter	???	
Check sum	63128	Download date/time:	10/05/2017	11:24		Record ID	9LF1	
Record imported (Y/N)				N				Source Diagnostic

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.