

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

SGT

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	--------------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	1120	34	14F27-11-5-10	.100	21	21	.099			X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											
		07/15/15 11:30 AS4 Serial no: 008874 Version no: 4797 No. Date Time XBAC Calibration 03/28/14 11:40 .100 Test Result 0001 07/15/15 11:29 .099 Push No. 3 to erase, or power down to retain 01 POWER DOWN COMPLETE												

* If the reading is within plus or minus (±) 0.010 percent of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008501
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0820	34		14F27-11-5-10		80		.097		X		Lee	17046	
				C1307252										

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

07/15/15 09:23
AS4 Serial no: 008501
Version no: 4797

No. Date Time ZBAC
Calibration
03/30/09 15:40 .100

Test Result
0001 07/15/15 09:22 .097

Push No. 9 to erase.
or power down to retain.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#2

PAS SERIAL NO.
 018785

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Officer Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2015	0923	34		14F27-11-5-10	.100	21		.102		X			Lee	17046	
				C1307252											

Prepared by:
 Douglas C. Talbot

Expires on:
 06-27-16

```

Push No. 3 to erase,
or power down to retain

-----
Test Result
0001 07/15/15 09:25 .102

-----
Calibration
07/18/08 13:27 .100

-----
No. Date Time ZDAC
07/15/15 09:26
AS4 Serial no: 018785
Version no: 4797
  
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPT 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
					TEST #1	TEST #2	TEST #1	(Optional)	YES	NO					YES
07/15/2015	0926	34	14F27-11-5-10								X	Lee	17046		
			C1307252												

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

07/15/15 09:28
AS4 Serial no: 008546
Version no: 4797

No. Date Time %BAC
Calibration
06/28/13 16:36 .100

Test Result
0001 07/03/15 01:24 .030
0002 07/03/15 04:59 .055
0003 07/03/15 05:03 .053
0004 07/04/15 02:21 .000
0005 07/04/15 03:34 .000
0006 07/04/15 03:38 .000
0007 07/04/15 03:38 .000
0008 07/05/15 01:12 .104
0009 07/05/15 01:15 .144
0010 07/05/15 01:18 .142
0011 07/06/15 01:26 .211
0012 07/06/15 01:35 .163
0013 07/15/15 09:28 .101

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 1002659
--------------------------------	-----------------	--------------------------------------	----------------------	--	---------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)							
07/15/2015	0928	34	14F27-11-5-10		.100	21				X			Lee	17046		
			CI307252													
<p>Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16</p>																
<p>Push No. 3 to erase.</p>																
<p>No. Date Time ZBAC Calibration 03/28/14 09:04 .100</p>																
<p>Test Result 0001 07/11/15 03:08 .112 0002 07/11/15 03:11 .138 0003 07/11/15 03:13 .129 0004 07/11/15 04:34 .146 0005 07/11/15 04:38 .139 0006 07/15/15 09:30 .098</p>																
<p>07/15/15 09:30 AS4 Serial no: 082659 Version no: 4797</p>																
<p>or power down to retain</p>																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Officer Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
 PAS SERIAL NO.: 008559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0930	34	14F27-11-5-10		.100	21		.102		X		Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

or power down to retain

07/15/15 09:36
 AS4 Serial no: 008559
 Version no: 4797

No. Date Time %BAC
 Calibration
 11/30/11 08:55 .100

Test Result
 0001 07/15/15 09:35 .102

Push No. 3 to erase

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 000580
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)	LOT NO.	VALUE	PAS TEMP (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
						TEST #1	TEST #2	TEST #1	TEST #2	YES	NO					YES
07/15/2015	09:32	34		14F27-11-5-10	.100	21						X	Lee	17046		
				C1307252												

07/15/15 09:35
AS4 Serial no: 000580
Version no: 4797

No. Date Time %BAC
Calibration
03/06/14 12:39 .100

Test Result
0001 07/15/15 09:35 .100

Push No. 3 to erase,
or power down to retain

or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 000020
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0925	34	14F27-11-5-10	.100	21	.100				X	Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-27-14										
			Expires on:										
			06-27-16										

07/15/15 09:40
AS4 Serial no: 000620
Version no: 4797

No. Date Time ZBAC
Calibration
01/30/13 14:34 .100

Test Result
0001 07/15/15 09:40 .100

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 002957
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0937	34		14E27-11-5-10	.100	22		.101		X			Lee	17046	
				C1307252											

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

or power down to retain

```

07/15/15 09:37
AS4 Serial no: 082957
Version no: 4797

No. Date Time %DAC
Calibration
02/17/14 08:52 .100
    
```

```

Test Result
0001 07/15/15 09:37 .101
    
```

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0739	34	14F27-11-5-10	.100	22							Lee	17046	
			C1307252											

Prepared by:
 Douglas C. Talbot
 06-27-14
 Expires on:
 06-27-16

07/15/15 09:39
 AS4 Serial no: 077505
 Version no: 4797

No. Date Time ZBAC
 Calibration
 09/18/14 07:21 .099

Test Result
 0001 07/06/15 14:14 .086
 0002 07/06/15 14:15 .084
 0003 07/15/15 09:39 .099

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

PAS SERIAL NO. 077479

PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV

I.D. NUMBER 17046

COORDINATOR NAME Officer Kirk Lee

AREA NO. 530

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)					
07/15/2015	0943	34	14F27-11-5-10	.100	RD		091	X		Lee	17046	
			C1307252									

or power down to retain

07/15/15 09:44
AS4 Serial no: 077479
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 09:10 .100

Test Result
0001 07/10/15 19:46 .000
0002 07/10/15 19:47 .000
0003 07/11/15 22:52 .241
0004 07/11/15 22:54 .273
0005 07/11/15 22:56 .262
0006 07/15/15 09:43 .097

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Officer Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
PAS SERIAL NO.: 008709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0945	34	14F27-11-5-10	.100	22						X	Lee	17046	
			C1307252											
			Prepared by: Douglas C. Talbot 06-27-14											
			Expires on: 06-27-16											

07/15/15 09:49
AS4 Serial no: 008709
Version no: 4797

No. Date Time %BAC
Calibration
02/17/14 09:27 .100

Test Result
0001 07/10/15 02:18 .092
0002 07/10/15 02:20 .082
0003 07/10/15 02:25 .076
0004 07/14/15 01:15 .000
0005 07/14/15 03:19 .127
0006 07/14/15 03:21 .123
0007 07/15/15 09:49 .101

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Kirk Lee	17046	Intoximeters Inc./Alco-Sensor IV	008714

DATE	TIME (C)	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)	PAS ACCURACY CHECK/RESULTS		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
			LOT NO.	VALUE		TEST #1	TEST #2						YES	NO
07/15/2015	0955	34	14F27-11-5-10	.100	22	21		X		Lee	17046			
			C1307252											

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

Test Result
0001 07/09/15 20:06 .084
0002 07/09/15 20:08 .068
0003 07/09/15 20:10 .081
0004 07/10/15 01:37 .069
0005 07/10/15 01:38 .068
0006 07/10/15 02:12 .149
0007 07/10/15 02:14 .159
0008 07/11/15 22:29 .000
0009 07/12/15 01:24 .180
0010 07/12/15 01:27 .193
0011 07/15/15 09:58 .101

No. Date Time ZBAC
Calibration
03/11/04 14:05 .100

07/15/15 09:59
AS4 Serial no: 008714
Version no: 4797

or power down to retain

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16
008888

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME (C)	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	0959	34	14F27-11-5-10	.100	21							Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

```

07/15/15 09:57
AS4 Serial no: 0088888
Version no: 4797

No. Date Time ZBAC
Calibration
03/25/13 10:45 .100

Test Result
0001 07/15/15 09:56 .099

Push No. 3 to erase,

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 002783
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2015	1001	34	14F27-11-5-10	.100	20						X	Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

or power down to retain

07/15/15 10:04
AS4 Serial no: 008783
Version no: 4797

No. Date Time %BAC
Calibration
07/18/08 13:52 .100

Test Result
0001 07/11/15 01:02 .109
0002 07/11/15 01:04 .122
0003 07/15/15 10:04 .100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

18

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Officer Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
 PAS SERIAL NO.: 002657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
07/15/2015	1003	34	14F27-11-5-10	.100		.100			X		Lee	17046	
			C1307252										

Prepared by:
 Douglas C. Talbot
 06-27-14
 Expires on:
 06-27-16

Push No. 3 to erase,
 or power down to retain

07/15/15 10:05
 AS4 Serial no: 002657
 Version no: 4797

No. Date Time %BAC
 Calibration
 02/17/14 09:50 .100

Test Result

0001	07/11/15	03:03	.094
0002	07/11/15	03:05	.091
0003	07/11/15	03:10	.000
0004	07/11/15	15:05	.092
0005	07/11/15	15:07	.095
0006	07/11/15	23:09	.154
0007	07/11/15	23:12	.153
0008	07/12/15	01:56	.008
0009	07/12/15	01:57	.008
0010	07/12/15	02:15	.068
0011	07/12/15	02:17	.116
0012	07/12/15	02:20	.109
0013	07/12/15	21:58	.256
0014	07/12/15	22:00	.263
0015	07/13/15	00:23	.107
0016	07/13/15	00:25	.100
0017	07/13/15	03:36	.228
0018	07/13/15	03:38	.226
0019	07/15/15	10:04	.100

Push No. 3 to erase,

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
07/15/2015	1006	34		14F27-11-5-10	.100	30					X	Lee	17046	
				C1307252										
				Prepared by: Douglas C. Talbot										
				06-27-14										
				Expires on: 06-27-16										

or power down to retain

07/15/15 10:10
AS4 Serial no: 008810
Version no: 4797

No. Date Time ZBAC
Calibration
10/30/03 09:53 .100

Test Result
0001 07/15/15 10:09 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

South Los Angeles

#20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Officer Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV PAS SERIAL NO.: 018684

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	PAS TEMP (°C)			PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2015	1011	34			20		.100	.099				Lee	17046		

07/15/15 10:20
AS4 Serial no: 018684
Version: no: 4797

No. Date Time ZBAC
Calibration
11/03/11 16:03 .100

Test Result
0001 07/06/15 14:44 .098
0002 07/15/15 10:13 .099

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21
008635

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Officer Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV
PAS SERIAL NO.: 17046

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)					
07/15/2015	1020	34	14F27-11-5-10	.100	22				X		Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-27-14										
			Expires on:										
			06-27-16										

or power down to retain

07/15/15 10:24
AS4 Serial no: 008635
Version no: 4797

No. Date Time ZBAC
Calibration
10/24/03 09:04 .100

Test Result

0001	06/20/15	02:15	.161
0002	06/20/15	02:18	.156
0003	06/22/15	00:39	.000
0004	06/22/15	00:41	.000
0005	06/22/15	02:35	.163
0006	06/22/15	02:37	.176
0007	06/26/15	21:27	.081
0008	06/26/15	21:30	.080
0009	06/27/15	02:40	.081
0010	06/27/15	02:42	.082
0011	06/27/15	23:33	.161
0012	06/27/15	23:35	.153
0013	06/28/15	01:27	.000
0014	06/28/15	01:28	.000
0015	06/28/15	01:53	.030
0016	06/28/15	01:54	.031
0017	06/28/15	03:26	.076
0018	06/28/15	03:28	.074
0019	06/28/15	04:05	.171
0020	06/29/15	10:37	.099
0021	07/15/15	10:23	.095

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

22

AREA NAME
South Los Angeles

AREA NO.
530

COORDINATOR NAME
Officer Kirk Lee

I.D. NUMBER
17046

PAS MAKE/MODEL
Intoximeters Inc./Alco-Sensor IV

PAS SERIAL NO.

008045

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)					
07/15/2015	1023	34	14F27-11-5-10	.100	21	.102		X		Lee	17046	
			C1307252									
			Prepared by:									
			Douglas C. Talbot									
			06-27-14									
			Expires on:									
			06-27-16									

or power down to retain

07/15/15 10:26
AS4 Serial no: 008045
Version no: 4797

No. Date Time ZBAC
Calibration
12/24/03 11:01 .100

Test Result
0001 07/15/15 10:25 .102

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Officer Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV PAS SERIAL NO.: 002653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	TEMP (°C)		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
07/15/2015	1026	34	14F27-11-5-10		.100	81					X		Lee	17046		
			C1307252													

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

Push No. 3 to erase,
or power down to retain

```

07/15/15 10:28
AS4 Serial no: 002653
Version no: 4797

No.   Date   Time  ZBAC
Calibration
02/17/14 10:00 .100

Test Result
0001 07/11/15 23:53 .235
0002 07/11/15 23:56 .214
0003 07/12/15 03:35 .156
0004 07/12/15 03:38 .199
0005 07/12/15 03:42 .200
0006 07/15/15 10:28 .098
    
```

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24
 008848

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 17046
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2	YES	NO				
07/15/2015	1029	34	14F27-11-5-10	.100	21							Lee	17046	
			C1307252											

07/15/15 10:34			
AS4 Serial no: 008848			
Version no: 4797			
No.	Date	Time	XBAC
Calibration			
	01/13/09	09:25	.100
Test Result			
0001	07/15/15	10:33	.104

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)					
07/15/2015	1033	34	14F27-11-5-10		.100	20				X		Lee	17046	
			C1307252											

Push No. 3 to erase,
or power down to retain

07/15/15 10:36
AS4 Serial no: 077591
Version no: 4797

No. Date Time ZBAC
Calibration
05/21/08 15:09 .100

Test Result
0001 07/15/15 10:34 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	1046	34	14F27-11-5-10	.100	22						X	Lee	17046	
			C1307252											
			Prepared by:											
			Douglas C. Talbot											
			06-27-14											
			Expires on:											
			06-27-16											

07/15/15 10:51
 AS4 Serial no: 018646
 Version no: 4797

No. Date Time %BAC
 Calibration
 07/29/13 13:32 .100

Test Result
 0001 07/15/15 10:51 .098

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#36
018615

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	10:55	34	14F27-11-5-10		22					X		Lee	17046	
			C1307252											

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

07/15/15 10:57
AS4 Serial no: 018615
Version no: 4797

No.	Date	Time	XBAC
	12/19/08	14:44	.100

Test Result

0001	07/11/15	21:09	.048
0002	07/15/15	10:56	.100

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPT 051

#37
 O/B 6/26

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Officer Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV PAS SERIAL NO: 17046

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)		YES	NO				
07/15/2015	1057	34	14F27-11-5-10	.100	21		.100	X			Lee	17046	
			C1307252										
			Prepared by:										
			Douglas C. Talbot										
			06-27-14										
			Expires on:										
			06-27-16										
Push No. 3 to erase. 07/15/15 10:57 AS4 Serial no: 018626 Version no: 4797 No. Date Time ZBAC Calibration 04/29/05 08:59 .101 Test Result 0001 07/11/15 01:12 .000 0002 07/11/15 01:14 .000 0003 07/11/15 01:34 .154 0004 07/11/15 01:37 .151 0005 07/11/15 04:32 .083 0006 07/11/15 04:35 .093 0007 07/15/15 10:56 .100 or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g, with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2015	1059	34	14F27-11-5-10		.100	21				X			Lee	17046	
			C1307252												

or power down to retain

07/15/15 10:05
 AS4 Serial no: 018696
 Version no: 4797

No.	Date	Time	ZBAC
	04/30/15	09:29	.100

Test Result
 0001 07/15/15 10:04 .106

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 17046
--------------------------------	-----------------	--------------------------------------	----------------------	--	-------------------------

083030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
07/15/2015	1101	34	14F27-11-5-10	.100	21	.100			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 06-27-14 Expires on: 06-27-16</p>													
<p>07/15/15 11:01 AS4 Serial no: 083030 Version no: 4797</p> <p>No. Date Time XPAC Calibration 01/29/09 10:39 .100</p> <p>Test Result 0001 07/15/15 11:01 .100</p> <p>Push No. 3 to erase, or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/15/2015	1103	34	14F27-11-5-10		.100	22		.099		X			Lee	17046	
			C1307252												
			Prepared by:												
			Douglas C. Talbot												
			06-27-14												
			Expires on:												
			06-27-16												

```

07/15/15 11:05
AS4 Serial no: 018692
Version no: 4797

No. Date Time ZBAC
Calibration
00/12/00 14:35 .100

Test Result
0001 07/15/15 11:05 .099

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	--------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/15/2015	1105	34	14F27-11-5-10	21		.100		X			Lee	17046	
			C1307252										

Prepared by:
Douglas C. Talbot
06-27-14
Expires on:
06-27-16

Push No. 9 to erase,
or power down to retain

07/15/15 11:07
AS4 Serial no: 018762
Version no: 4797

No. Date Time XBAC
Calibration
03/30/09 15:52 .100

Test Result
0001 07/06/15 20:50 .176
0002 07/06/15 20:53 .170
0003 07/07/15 20:32 .187
0004 07/07/15 20:34 .185
0005 07/15/15 11:06 .100

Push No. 9 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI.051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	--------------------------------------	----------------------	--	----------------

018766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED	PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (C)	LOT NO.		TEST #1	TEST #2	TEST #1 (Optional)	TEST #2 (Optional)		YES	NO			
07/15/2015	1100	39	14F27-11-5-10		.100	21					X		Lee	17046	

07/15/15 11:10
AS4 Serial no: 018766
Version no: 4797

No.	Date	Time	ZBAC
Calibration	09/12/14	13:45	.100
Test Result	0001	07/15/15 11:10	.100

Push No. 3 to erase.

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#44
 012691

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Officer Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters Inc./Alco-Sensor IV | PAS SERIAL NO.:

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PASSES RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)		YES	NO				
07/15/2015	1110	34	14F27-11-5-10	C1307252	.100	22				X		Lee	17046	

or power down to retain

07/15/15 11:13
 ASA Serial no: 018691
 Version no: 4797

No. Date Time ZBAC
 Calibration
 09/00/00 14:01 .100

Test Result
 0001 07/15/15 11:13 .105

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

