

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) 01/051

# 1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8501
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2052	34	13B06-16-5-10- KKF24C	.100	19		.098			X		Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="width: 30%;"> <p>11/14/13 20:53 AS4 Serial no: 008501 Version no: 4797</p> </div> <div style="width: 30%;"> <p>No. Date Time %BAC Calibration 02/30/09 15:40 .100</p> </div> <div style="width: 35%;"> <p>Test Result 0001 09/27/13 11:10 .097 0002 09/28/13 15:48 .167 0003 09/28/13 15:53 .167 0004 09/29/13 18:52 .197 0005 09/29/13 18:56 .168 0006 09/29/13 18:58 .187 0007 10/03/13 11:22 .097 0008 10/15/13 13:55 .096 0009 10/23/13 13:42 .098 0010 10/31/13 13:37 .098 0011 11/09/13 17:07 .081 0012 11/10/13 18:35 .095 0013 11/10/13 18:38 .099 0014 11/14/13 20:53 .098</p> </div> </div>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18785
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2054	34	13B06-16-5-10- KKF24C	.100	19		.100			X	Valadez	16071	Time Change
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										
					<pre> 11/14/13 20:55 AS4 Serial no: 018785 Version no: 4797 No. Date Time ZBAC Calibration 07/18/08 13:27 .100 Test Result 0001 09/18/13 13:02 .101 0002 09/27/13 11:10 .101 0003 10/03/13 11:26 .100 0004 10/11/13 11:53 .101 0005 10/15/13 13:52 .099 0006 10/23/13 13:39 .100 0007 10/31/13 13:37 .100 0008 11/14/13 20:55 .100 </pre>								

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8546
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																																																																																																		
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																																																																																																																																						
11/14/2013	2058	34	13B06-16-5-10- KKF24C	.100	21		102			X		Valadez	16071	Time Change																																																																																																																																		
Prepared by: Douglas C. Talbot 02-06-2013																																																																																																																																																
Expires on: 02-06-2015																																																																																																																																																
<table border="0" style="width: 100%;"> <tr> <td style="width: 20%;">11/14/13 20:59</td> <td style="width: 20%;">AS4 Serial no: 008546</td> <td style="width: 20%;">Version no: 4797</td> <td style="width: 20%;">No. Date Time %BAC</td> <td style="width: 20%;">Calibration</td> </tr> <tr> <td></td> <td></td> <td></td> <td></td> <td>06/28/13 16:36 .100</td> </tr> <tr> <td>Test Result</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0001</td> <td>09/19/13</td> <td>02:25</td> <td>.000</td> <td></td> </tr> <tr> <td>0002</td> <td>09/20/13</td> <td>02:07</td> <td>.065</td> <td></td> </tr> <tr> <td>0003</td> <td>09/20/13</td> <td>02:10</td> <td>.060</td> <td></td> </tr> <tr> <td>0004</td> <td>09/21/13</td> <td>14:51</td> <td>.101</td> <td></td> </tr> <tr> <td>0005</td> <td>09/21/13</td> <td>14:54</td> <td>.111</td> <td></td> </tr> <tr> <td>0006</td> <td>09/27/13</td> <td>11:13</td> <td>.103</td> <td></td> </tr> <tr> <td>0007</td> <td>10/02/13</td> <td>02:09</td> <td>.097</td> <td></td> </tr> <tr> <td>0008</td> <td>10/02/13</td> <td>02:13</td> <td>.096</td> <td></td> </tr> <tr> <td>0009</td> <td>10/03/13</td> <td>11:30</td> <td>.102</td> <td></td> </tr> <tr> <td>0010</td> <td>10/03/13</td> <td>21:21</td> <td>.000</td> <td></td> </tr> <tr> <td>0011</td> <td>10/09/13</td> <td>02:34</td> <td>.113</td> <td></td> </tr> <tr> <td>0012</td> <td>10/09/13</td> <td>02:36</td> <td>.110</td> <td></td> </tr> <tr> <td>0013</td> <td>10/11/13</td> <td>11:48</td> <td>.101</td> <td></td> </tr> <tr> <td>0014</td> <td>10/15/13</td> <td>13:52</td> <td>.100</td> <td></td> </tr> <tr> <td>0015</td> <td>10/16/13</td> <td>02:48</td> <td>.160</td> <td></td> </tr> <tr> <td>0016</td> <td>10/16/13</td> <td>02:50</td> <td>.155</td> <td></td> </tr> <tr> <td>0017</td> <td>10/23/13</td> <td>19:57</td> <td>.099</td> <td></td> </tr> <tr> <td>0018</td> <td>10/29/13</td> <td>03:48</td> <td>.147</td> <td></td> </tr> <tr> <td>0019</td> <td>10/29/13</td> <td>03:50</td> <td>.138</td> <td></td> </tr> <tr> <td>0020</td> <td>10/31/13</td> <td>13:42</td> <td>.101</td> <td></td> </tr> <tr> <td>0021</td> <td>11/13/13</td> <td>00:28</td> <td>.060</td> <td></td> </tr> <tr> <td>0022</td> <td>11/13/13</td> <td>00:29</td> <td>.064</td> <td></td> </tr> <tr> <td>0023</td> <td>11/14/13</td> <td>20:59</td> <td>.102</td> <td></td> </tr> </table>															11/14/13 20:59	AS4 Serial no: 008546	Version no: 4797	No. Date Time %BAC	Calibration					06/28/13 16:36 .100	Test Result					0001	09/19/13	02:25	.000		0002	09/20/13	02:07	.065		0003	09/20/13	02:10	.060		0004	09/21/13	14:51	.101		0005	09/21/13	14:54	.111		0006	09/27/13	11:13	.103		0007	10/02/13	02:09	.097		0008	10/02/13	02:13	.096		0009	10/03/13	11:30	.102		0010	10/03/13	21:21	.000		0011	10/09/13	02:34	.113		0012	10/09/13	02:36	.110		0013	10/11/13	11:48	.101		0014	10/15/13	13:52	.100		0015	10/16/13	02:48	.160		0016	10/16/13	02:50	.155		0017	10/23/13	19:57	.099		0018	10/29/13	03:48	.147		0019	10/29/13	03:50	.138		0020	10/31/13	13:42	.101		0021	11/13/13	00:28	.060		0022	11/13/13	00:29	.064		0023	11/14/13	20:59	.102	
11/14/13 20:59	AS4 Serial no: 008546	Version no: 4797	No. Date Time %BAC	Calibration																																																																																																																																												
				06/28/13 16:36 .100																																																																																																																																												
Test Result																																																																																																																																																
0001	09/19/13	02:25	.000																																																																																																																																													
0002	09/20/13	02:07	.065																																																																																																																																													
0003	09/20/13	02:10	.060																																																																																																																																													
0004	09/21/13	14:51	.101																																																																																																																																													
0005	09/21/13	14:54	.111																																																																																																																																													
0006	09/27/13	11:13	.103																																																																																																																																													
0007	10/02/13	02:09	.097																																																																																																																																													
0008	10/02/13	02:13	.096																																																																																																																																													
0009	10/03/13	11:30	.102																																																																																																																																													
0010	10/03/13	21:21	.000																																																																																																																																													
0011	10/09/13	02:34	.113																																																																																																																																													
0012	10/09/13	02:36	.110																																																																																																																																													
0013	10/11/13	11:48	.101																																																																																																																																													
0014	10/15/13	13:52	.100																																																																																																																																													
0015	10/16/13	02:48	.160																																																																																																																																													
0016	10/16/13	02:50	.155																																																																																																																																													
0017	10/23/13	19:57	.099																																																																																																																																													
0018	10/29/13	03:48	.147																																																																																																																																													
0019	10/29/13	03:50	.138																																																																																																																																													
0020	10/31/13	13:42	.101																																																																																																																																													
0021	11/13/13	00:28	.060																																																																																																																																													
0022	11/13/13	00:29	.064																																																																																																																																													
0023	11/14/13	20:59	.102																																																																																																																																													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82659
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2056	34	13B06-16-5-10- KKF24C	.100	19		.105				X	Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											
No. Date Time %BAC Calibration 10/22/08 17:37 .100 Test Result 0001 10/15/13 13:45 .103 0002 10/23/13 13:34 .105 0003 10/31/13 13:37 .105 0004 11/09/13 03:52 .199 0005 11/09/13 03:56 .185 0006 11/14/13 02:46 .000 0007 11/14/13 20:56 .105														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8559
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2169	34	13B06-16-5-10- KKF24C	.100	20		.104				X	Valadez	16071	Time Change
		Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015												
		<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px;"> 11/14/13 21:10 AS4 Serial no: 008559 Version no: 4797 </div> <div style="border: 1px solid black; padding: 5px;"> No. Date Time XRAC Calibration 11/30/11 08:55 .100 </div> <div style="border: 1px solid black; padding: 5px;"> Test Result  0001 09/27/13 11:20 .103  0002 10/03/13 11:40 .103  0003 10/11/13 11:57 .102  0004 10/15/13 13:55 .100  0005 10/31/13 14:15 .106  0006 11/10/13 18:09 .106  0007 11/14/13 21:10 .104 </div> </div>												

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8580
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2106	34	13B06-16-5-10- KKF24C	.100	20		.102			X	Valadez	16071	Time Change
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/14/13 21:07  
 AS4 Serial no: 008580  
 Version no: 4797

No.	Date	Time	%BAC
Calibration	12/01/06	08:14	.099
Test Result			
0001	09/20/13	01:03	.218
0002	09/20/13	01:06	.215
0003	09/24/13	00:09	.169
0004	09/24/13	00:12	.181
0005	09/24/13	01:52	.140
0006	09/25/13	02:21	.000
0007	09/27/13	11:16	.103
0008	10/01/13	02:02	.000
0009	10/03/13	11:37	.103
0010	10/03/13	20:12	.000
0011	10/08/13	01:08	.168
0012	10/08/13	01:11	.165
0013	10/11/13	11:55	.103
0014	10/15/13	02:12	.040
0015	10/15/13	02:44	.037
0016	10/15/13	03:54	.021
0017	10/15/13	13:48	.101
0018	10/16/13	02:42	.084
0019	10/16/13	02:45	.083
0020	10/16/13	02:51	.044
0021	10/17/13	01:56	.000
0022	10/18/13	00:42	.017
0023	10/18/13	02:15	.209
0024	10/18/13	02:18	.179
0025	10/18/13	02:20	.208
0026	10/23/13	13:37	.103
0027	10/29/13	00:03	.109
0028	10/29/13	00:06	.103
0029	10/29/13	23:36	.063
0030	10/31/13	01:35	.000
0031	10/31/13	02:09	.167
0032	10/31/13	02:12	.195
0033	10/31/13	13:45	.104
0034	11/01/13	00:40	.000
0035	11/01/13	01:40	.100
0036	11/01/13	01:44	.172
0037	11/01/13	02:16	.161
0038	11/01/13	02:19	.154
0039	11/05/13	04:15	.000
0040	11/12/13	02:26	.027
0041	11/14/13	21:06	.102

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8303
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS						
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO										
11/14/2013	2059	34	13B06-16-5-10- KKF24C	.100	20		101			X		Valadez	16071	Time Change						
<table border="0" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%; border: none;">                     Prepared by:                      Douglas C. Talbot                      02-06-2013                      Expires on:                      02-06-2015                 </td> <td style="width: 20%; border: none; vertical-align: top;">                     11/14/13 21:00                      AS4 Serial no: 083032                      Version no: 4797                 </td> <td style="width: 20%; border: none; vertical-align: top;">                     No. Date %BAC                      Calibration 12/01/11 07:38 .061                 </td> <td style="width: 30%; border: none; vertical-align: top;">                     Test Result                      0001 09/27/13 11:08 .100                      0002 10/03/13 11:32 .097                      0003 10/05/13 01:44 .147                      0004 10/05/13 01:46 .144                      0005 10/05/13 04:51 .036                      0006 10/05/13 22:33 .238                      0007 10/05/13 22:35 .246                      0008 10/06/13 00:21 .080                      0009 10/06/13 00:24 .076                      0010 10/06/13 00:31 .044                      0011 10/06/13 02:54 .135                      0012 10/06/13 02:56 .119                      0013 10/11/13 00:12 .080                      0014 10/11/13 00:15 .082                      0015 10/11/13 01:11 .181                      0016 10/11/13 01:14 .175                      0017 10/11/13 11:48 .101                      0018 10/11/13 20:21 .229                      0019 10/11/13 20:23 .233                      0020 10/12/13 02:39 .175                      0021 10/12/13 02:41 .164                      0022 10/12/13 23:20 .000                      0023 10/13/13 00:15 .007                      0024 10/13/13 00:18 .006                      0025 10/13/13 03:05 .133                 </td> </tr> <tr> <td style="border: none; vertical-align: top;">                     0026 10/13/13 03:08 .139                      0027 10/14/13 01:05 .155                      0028 10/14/13 01:08 .145                      0029 10/15/13 13:36 .097                      0030 10/19/13 03:31 .145                      0031 10/19/13 03:35 .122                      0032 10/20/13 04:16 .105                      0033 10/20/13 04:18 .101                      0034 10/23/13 13:26 .098                      0035 10/25/13 02:16 .048                      0036 10/25/13 02:19 .051                      0037 10/26/13 01:32 .091                      0038 10/26/13 01:34 .086                      0039 10/26/13 01:40 .087                      0040 10/26/13 01:42 .128                      0041 10/26/13 01:44 .000                      0042 10/27/13 23:56 .127                      0043 10/27/13 23:59 .148                      0044 10/31/13 13:36 .101                      0045 11/02/13 21:30 .045                      0046 11/02/13 21:33 .047                      0047 11/03/13 03:13 .131                      0048 11/03/13 03:15 .132                      0049 11/04/13 03:07 .170                      0050 11/04/13 03:10 .164                      0051 11/09/13 03:06 .015                      0052 11/09/13 03:08 .014                      0053 11/10/13 00:45 .039                      0054 11/10/13 01:57 .117                      0055 11/10/13 01:59 .113                      0056 11/10/13 05:24 .010                      0057 11/10/13 21:09 .012                      0058 11/14/13 20:59 .101                 </td> <td style="border: none;"></td> </tr> </table>															Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015	11/14/13 21:00 AS4 Serial no: 083032 Version no: 4797	No. Date %BAC Calibration 12/01/11 07:38 .061	Test Result 0001 09/27/13 11:08 .100 0002 10/03/13 11:32 .097 0003 10/05/13 01:44 .147 0004 10/05/13 01:46 .144 0005 10/05/13 04:51 .036 0006 10/05/13 22:33 .238 0007 10/05/13 22:35 .246 0008 10/06/13 00:21 .080 0009 10/06/13 00:24 .076 0010 10/06/13 00:31 .044 0011 10/06/13 02:54 .135 0012 10/06/13 02:56 .119 0013 10/11/13 00:12 .080 0014 10/11/13 00:15 .082 0015 10/11/13 01:11 .181 0016 10/11/13 01:14 .175 0017 10/11/13 11:48 .101 0018 10/11/13 20:21 .229 0019 10/11/13 20:23 .233 0020 10/12/13 02:39 .175 0021 10/12/13 02:41 .164 0022 10/12/13 23:20 .000 0023 10/13/13 00:15 .007 0024 10/13/13 00:18 .006 0025 10/13/13 03:05 .133	0026 10/13/13 03:08 .139 0027 10/14/13 01:05 .155 0028 10/14/13 01:08 .145 0029 10/15/13 13:36 .097 0030 10/19/13 03:31 .145 0031 10/19/13 03:35 .122 0032 10/20/13 04:16 .105 0033 10/20/13 04:18 .101 0034 10/23/13 13:26 .098 0035 10/25/13 02:16 .048 0036 10/25/13 02:19 .051 0037 10/26/13 01:32 .091 0038 10/26/13 01:34 .086 0039 10/26/13 01:40 .087 0040 10/26/13 01:42 .128 0041 10/26/13 01:44 .000 0042 10/27/13 23:56 .127 0043 10/27/13 23:59 .148 0044 10/31/13 13:36 .101 0045 11/02/13 21:30 .045 0046 11/02/13 21:33 .047 0047 11/03/13 03:13 .131 0048 11/03/13 03:15 .132 0049 11/04/13 03:07 .170 0050 11/04/13 03:10 .164 0051 11/09/13 03:06 .015 0052 11/09/13 03:08 .014 0053 11/10/13 00:45 .039 0054 11/10/13 01:57 .117 0055 11/10/13 01:59 .113 0056 11/10/13 05:24 .010 0057 11/10/13 21:09 .012 0058 11/14/13 20:59 .101	
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015	11/14/13 21:00 AS4 Serial no: 083032 Version no: 4797	No. Date %BAC Calibration 12/01/11 07:38 .061	Test Result 0001 09/27/13 11:08 .100 0002 10/03/13 11:32 .097 0003 10/05/13 01:44 .147 0004 10/05/13 01:46 .144 0005 10/05/13 04:51 .036 0006 10/05/13 22:33 .238 0007 10/05/13 22:35 .246 0008 10/06/13 00:21 .080 0009 10/06/13 00:24 .076 0010 10/06/13 00:31 .044 0011 10/06/13 02:54 .135 0012 10/06/13 02:56 .119 0013 10/11/13 00:12 .080 0014 10/11/13 00:15 .082 0015 10/11/13 01:11 .181 0016 10/11/13 01:14 .175 0017 10/11/13 11:48 .101 0018 10/11/13 20:21 .229 0019 10/11/13 20:23 .233 0020 10/12/13 02:39 .175 0021 10/12/13 02:41 .164 0022 10/12/13 23:20 .000 0023 10/13/13 00:15 .007 0024 10/13/13 00:18 .006 0025 10/13/13 03:05 .133																	
0026 10/13/13 03:08 .139 0027 10/14/13 01:05 .155 0028 10/14/13 01:08 .145 0029 10/15/13 13:36 .097 0030 10/19/13 03:31 .145 0031 10/19/13 03:35 .122 0032 10/20/13 04:16 .105 0033 10/20/13 04:18 .101 0034 10/23/13 13:26 .098 0035 10/25/13 02:16 .048 0036 10/25/13 02:19 .051 0037 10/26/13 01:32 .091 0038 10/26/13 01:34 .086 0039 10/26/13 01:40 .087 0040 10/26/13 01:42 .128 0041 10/26/13 01:44 .000 0042 10/27/13 23:56 .127 0043 10/27/13 23:59 .148 0044 10/31/13 13:36 .101 0045 11/02/13 21:30 .045 0046 11/02/13 21:33 .047 0047 11/03/13 03:13 .131 0048 11/03/13 03:15 .132 0049 11/04/13 03:07 .170 0050 11/04/13 03:10 .164 0051 11/09/13 03:06 .015 0052 11/09/13 03:08 .014 0053 11/10/13 00:45 .039 0054 11/10/13 01:57 .117 0055 11/10/13 01:59 .113 0056 11/10/13 05:24 .010 0057 11/10/13 21:09 .012 0058 11/14/13 20:59 .101																				

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8620
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2119	34	13B06-16-5-10- KKF24C	.100	20		.102			X		Valadez	16071	Time Change

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77505
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2114	34	13B06-16-5-10- KKF24C	.100	20		.102			X		Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015														
11/14/13 21:17 AS4 Serial no: 077505 Version no: 4797 Mo. Date Time %BAC Calibration 05/21/08 13:16 .100 Test Result 0001 09/27/13 15:01 .095 0002 10/03/13 11:52 .100 0003 10/11/13 12:26 .099 0004 10/18/13 13:33 .095 0005 10/23/13 13:25 .099 0006 10/31/13 13:45 .100 0007 11/14/13 21:16 .102														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77479
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	21:21	34	13B06-16-5-10- KKF24C	.100	19	1097			X		Valadez	16071	Time Change

11/14/13 21:22  
 AS4 Serial no: 077479  
 Version no: 4797

No. Date Time %BAC  
 Calibration 03/07/13 09:48 .099

Test Result	No.	Date	Time	%BAC
.196	0048	10/13/13	00:25	.196
.125	0049	10/13/13	00:31	.125
.009	0050	10/13/13	02:28	.009
.081	0051	10/13/13	04:20	.081
.080	0052	10/13/13	04:22	.080
.210	0053	10/14/13	00:46	.210
.207	0054	10/14/13	00:50	.207
.095	0055	10/15/13	13:38	.095
.193	0056	10/19/13	00:15	.193
.197	0057	10/19/13	00:17	.197
.153	0058	10/20/13	00:22	.153
.150	0059	10/20/13	00:25	.150
.129	0060	10/20/13	02:10	.129
.122	0061	10/20/13	02:13	.122
.146	0062	10/21/13	00:22	.146
.142	0063	10/21/13	00:24	.142
.158	0064	10/21/13	02:40	.158
.153	0065	10/21/13	02:44	.153
.021	0066	10/21/13	02:49	.021
.097	0067	10/23/13	13:26	.097
.155	0068	10/25/13	01:24	.155
.147	0069	10/25/13	01:27	.147
.148	0070	10/25/13	01:31	.148
.149	0071	10/26/13	00:38	.149
.141	0072	10/26/13	00:41	.141
.137	0073	10/26/13	02:05	.137
.199	0074	10/26/13	02:08	.199
.188	0075	10/26/13	04:16	.188
.018	0076	10/27/13	00:34	.018
.000	0077	10/27/13	01:33	.000
.180	0078	10/27/13	01:36	.180
.174	0079	10/27/13	03:51	.174
.108	0080	10/27/13	03:54	.108
.126	0081	10/27/13	04:00	.126
.123	0082	10/31/13	13:52	.123
.098	0083	11/01/13	21:52	.098
.119	0084	11/01/13	21:55	.119
.120	0085	11/02/13	22:42	.120
.178	0086	11/02/13	22:44	.178
.178	0087	11/04/13	03:15	.178
.168	0088	11/04/13	03:19	.168
.119	0089	11/04/13	03:21	.119
.225	0090	11/09/13	01:54	.225
.145	0091	11/09/13	01:57	.145
.157	0092	11/09/13	02:34	.157
.217	0093	11/09/13	02:36	.217
.247	0094	11/09/13	02:39	.247
.201	0095	11/09/13	23:23	.201
.056	0096	11/10/13	01:00	.056
.148	0097	11/10/13	01:03	.148
.160	0098	11/10/13	01:05	.160
.159	0099	11/10/13	03:29	.159
.127	0100	11/10/13	03:32	.127
.138	0101	11/11/13	00:18	.138
.272	0102	11/11/13	00:22	.272
.269	0103	11/11/13	03:59	.269
.000	0104	11/11/13	04:00	.000
.000	0105	11/11/13	04:07	.000
.015	0106	11/14/13	21:22	.015
.097				.097

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82660
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2103	34	13B06-16-5-10- KKF24C	.100	19		.096			X		Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

No.	Date	Time	%BAC
11/14/13 21:24			
AS4 Serial no: 082660			
Version no: 4797			
Calibration 11/18/10 06:21 .081			
Test Result			
0001	10/15/13	13:34	.095
0002	10/19/13	03:14	.160
0003	10/19/13	03:17	.158
0004	10/19/13	21:28	.068
0005	10/19/13	21:30	.086
0006	10/20/13	00:09	.111
0007	10/20/13	00:11	.124
0008	10/20/13	00:19	.079
0009	10/20/13	03:03	.051
0010	10/20/13	03:04	.052
0011	10/23/13	13:22	.097
0012	10/26/13	02:01	.113
0013	10/26/13	02:03	.110
0014	10/27/13	22:51	.115
0015	10/27/13	22:52	.117
0016	10/28/13	03:02	.140
0017	10/28/13	03:04	.140
0018	10/31/13	14:01	.099
0019	11/02/13	22:37	.106
0020	11/02/13	22:40	.102
0021	11/02/13	22:43	.000
0022	11/02/13	22:45	.000
0023	11/03/13	02:49	.149
0024	11/03/13	02:52	.157
0025	11/03/13	02:56	.105
0026	11/08/13	03:16	.132
0027	11/08/13	03:18	.129
0028	11/09/13	03:26	.167
0029	11/09/13	03:29	.155
0030	11/09/13	05:02	.130
0031	11/09/13	05:04	.155
0032	11/09/13	05:06	.138
0033	11/09/13	05:21	.127
0034	11/09/13	05:22	.138
0035	11/09/13	05:26	.130
0036	11/09/13	05:27	.127
0037	11/10/13	03:32	.113
0038	11/10/13	03:34	.122
0039	11/11/13	01:59	.027
0040	11/11/13	02:01	.029
0041	11/11/13	04:02	.133
0042	11/11/13	04:05	.152
0043	11/14/13	21:24	.096

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8888
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2130	34	13B06-16-5-10- KKF24C	.100	21	.100			X		Valadez	16071	Time Change

11/14/13 21:31  
 ASA Serial no: 008888  
 Version no: 4797

No. Date Time %BAC  
 Calibration 03/25/13 10:45 .100

Test Result  
 0001 09/13/13 23:26 .000  
 0002 09/14/13 02:03 .186  
 0003 09/14/13 02:07 .189  
 0004 09/14/13 19:55 .206  
 0005 09/14/13 19:58 .197  
 0006 09/15/13 20:04 .262  
 0007 09/15/13 20:08 .258  
 0008 09/16/13 01:45 .174  
 0009 09/16/13 01:49 .183  
 0010 09/16/13 22:22 .056  
 0011 09/16/13 22:24 .058  
 0012 09/16/13 23:01 .019  
 0013 09/18/13 12:38 .102  
 0014 09/21/13 23:08 .165  
 0015 09/21/13 23:10 .162  
 0016 09/23/13 01:03 .177  
 0017 09/23/13 01:06 .175  
 0018 09/26/13 21:47 .106  
 0019 09/26/13 21:50 .112  
 0020 09/27/13 00:20 .192  
 0021 09/27/13 00:22 .200  
 0022 09/27/13 02:27 .059  
 0023 09/27/13 02:29 .060  
 0024 09/27/13 02:55 .027  
 0025 09/27/13 02:56 .030  
 0026 09/27/13 11:25 .099  
 0027 09/27/13 23:30 .210  
 0028 09/27/13 23:34 .213  
 0029 09/28/13 22:36 .086  
 0030 09/28/13 22:39 .081  
 0031 09/29/13 02:25 .000  
 0032 09/29/13 02:27 .000  
 0033 09/29/13 02:33 .065  
 0034 09/29/13 03:07 .012  
 0035 09/29/13 03:09 .012  
 0036 09/29/13 03:33 .080  
 0037 09/29/13 03:36 .092  
 0038 10/03/13 11:54 .101

0039 10/11/13 12:31 .102  
 0040 10/12/13 02:13 .194  
 0041 10/12/13 02:18 .192  
 0042 10/12/13 22:50 .072  
 0043 10/12/13 22:52 .068  
 0044 10/13/13 04:23 .157  
 0045 10/13/13 04:26 .156  
 0046 10/14/13 00:50 .064  
 0047 10/14/13 00:53 .061  
 0048 10/14/13 02:11 .037  
 0049 10/14/13 03:34 .149  
 0050 10/14/13 03:37 .157  
 0051 10/15/13 13:24 .099  
 0052 10/19/13 00:32 .226  
 0053 10/19/13 00:34 .207  
 0054 10/19/13 01:24 .157  
 0055 10/19/13 01:28 .143  
 0056 10/19/13 05:02 .116  
 0057 10/19/13 05:06 .112  
 0058 10/20/13 01:31 .182  
 0059 10/20/13 01:34 .183  
 0060 10/20/13 04:31 .051  
 0061 10/20/13 04:34 .050  
 0062 10/21/13 01:19 .000  
 0063 10/21/13 02:42 .083  
 0064 10/21/13 02:46 .084  
 0065 10/23/13 13:09 .101  
 0066 10/25/13 22:14 .058  
 0067 10/25/13 22:30 .009  
 0068 10/26/13 03:53 .125  
 0069 10/26/13 03:56 .122  
 0070 10/27/13 04:43 .108  
 0071 10/27/13 04:46 .093  
 0072 10/31/13 14:29 .102  
 0073 11/01/13 21:08 .000  
 0074 11/01/13 21:11 .000  
 0075 11/01/13 23:47 .057  
 0076 11/02/13 01:31 .053  
 0077 11/03/13 01:32 .128  
 0078 11/03/13 01:35 .133  
 0079 11/09/13 23:26 .184  
 0080 11/09/13 23:28 .201  
 0081 11/09/13 23:32 .181  
 0082 11/10/13 03:38 .057  
 0083 11/10/13 03:39 .052  
 0084 11/10/13 04:32 .118  
 0085 11/10/13 04:34 .141  
 0086 11/10/13 04:37 .132  
 0087 11/10/13 04:40 .086  
 0088 11/10/13 04:41 .092  
 0089 11/14/13 21:30 .100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8783
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2146	34	13B06-16-5-10- KKF24C	.100	19		103			X		Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/14/13 21:47	AS4 Serial no: 000783	Version no: 4797	No. Date Time %BAC	Calibration 02/18/08 13:52 .100	Test Result	0001 09/12/13 01:43 .074	0002 09/12/13 01:45 .068	0003 09/12/13 02:32 .086	0004 09/12/13 02:34 .091	0005 09/17/13 22:30 .051	0006 09/17/13 22:33 .053	0007 09/18/13 12:56 .102	0008 09/19/13 00:53 .019	0009 09/19/13 02:39 .049	0010 09/20/13 18:46 .196	0011 09/20/13 18:47 .000	0012 09/20/13 18:47 .000	0013 09/23/13 20:42 .122	0014 09/23/13 20:45 .127	0015 09/24/13 01:24 .000	0016 09/24/13 01:26 .057	0017 09/24/13 01:27 .048	0018 09/27/13 11:38 .100	0019 10/01/13 21:24 .071	0020 10/01/13 21:26 .068	0021 10/03/13 02:19 .075	0022 10/03/13 02:22 .063	0023 10/03/13 12:06 .102	0024 10/08/13 00:44 .065	0025 10/08/13 00:46 .063	0026 10/08/13 01:49 .143	0027 10/08/13 01:51 .147	0028 10/11/13 12:44 .101	0029 10/15/13 13:33 .100	0030 10/17/13 05:13 .110	0031 10/17/13 05:16 .131	0032 10/17/13 05:19 .124	0033 10/23/13 13:19 .102	0034 10/29/13 00:57 .128	0035 10/29/13 01:00 .124	0036 10/29/13 21:51 .072	0037 10/29/13 21:53 .072	0038 10/31/13 14:42 .103	0039 11/05/13 03:41 .055	0040 11/05/13 03:43 .056	0041 11/05/13 03:46 .048	0042 11/06/13 03:10 .098	0043 11/06/13 03:13 .091	0044 11/06/13 03:17 .079	0045 11/06/13 03:19 .020	0046 11/06/13 03:21 .047	0047 11/11/13 23:44 .125	0048 11/11/13 23:46 .126	0049 11/11/13 23:50 .050	0050 11/12/13 03:18 .075	0051 11/12/13 03:20 .076	0052 11/14/13 21:47 .103
----------------	-----------------------	------------------	--------------------	---------------------------------	-------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------	--------------------------

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82657
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2143	34	13B06-16-5-10- KKF24C	.100	20	.099			X		Valadez	16071	Time Change	
			Prepared by: Douglas C. Talbot 02-06-2013											
			Expires on: 02-06-2015											
			11/14/13 21:44 #54 Serial no: 082657 Version no: 4797 No. Date Time %BAC Calibration 10/07/11 12:22 .081 Test Result 0001 10/15/13 13:24 .095 0002 10/23/13 13:12 .097 0003 10/31/13 14:36 .098 0004 11/14/13 21:43 .099											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8810
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2152	34	13B06-16-5-10- KKF24C	.100	19		.099			X		Valadez	16071	Time Change
			Prepared by: Douglas C. Talbot											
			Expires on: 02-06-2015											

11/14/13 21:53  
AS4 Serial no: 008810  
Version no: 4797

No. Date Time %BAC  
Calibration 10/30/08 09:53 .100

Test Result  
0001 09/10/13 22:40 .000  
0002 09/11/13 00:00 .089  
0003 09/11/13 00:10 .079  
0004 09/18/13 02:07 .111  
0005 09/18/13 02:09 .118  
0006 09/18/13 12:58 .099  
0007 09/18/13 20:28 .000  
0008 09/18/13 21:12 .180  
0009 09/18/13 21:15 .185  
0010 09/23/13 22:33 .036  
0011 09/24/13 01:55 .138  
0012 09/24/13 01:58 .124  
0013 09/24/13 01:59 .123  
0014 09/24/13 20:29 .148  
0015 09/24/13 20:30 .142  
0016 09/24/13 20:35 .000  
0017 09/25/13 02:09 .125  
0018 09/25/13 02:12 .115  
0019 09/25/13 02:18 .059  
0020 09/25/13 02:22 .058  
0021 09/27/13 11:47 .096  
0022 10/03/13 00:31 .029  
0023 10/03/13 01:48 .060  
0024 10/03/13 12:08 .097  
0025 10/03/13 23:25 .000  
0026 10/04/13 02:08 .142  
0027 10/04/13 02:12 .141  
0028 10/07/13 23:23 .000  
0029 10/08/13 01:14 .154  
0030 10/08/13 01:16 .165  
0031 10/11/13 12:53 .096  
0032 10/14/13 22:51 .055  
0033 10/15/13 13:29 .096  
0034 10/16/13 00:15 .159  
0035 10/16/13 00:18 .173  
0036 10/17/13 00:36 .074  
0037 10/17/13 00:38 .079  
0038 10/22/13 05:54 .000  
0039 10/22/13 23:58 .084  
0040 10/23/13 00:00 .000  
0041 10/23/13 13:13 .097  
0042 10/24/13 01:52 .120  
0043 10/24/13 01:53 .105  
0044 10/31/13 14:46 .099  
0045 11/12/13 00:23 .048  
0046 11/12/13 00:25 .047  
0047 11/12/13 02:09 .014  
0048 11/12/13 02:59 .043  
0049 11/13/13 01:29 .000  
0050 11/13/13 03:10 .049  
0051 11/13/13 03:45 .050  
0052 11/13/13 04:06 .079  
0053 11/13/13 04:08 .082  
0054 11/14/13 21:53 .099

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18684
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2155	34	13B06-16-5-10- KKF24C	.100	21		.105			X		Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														

11/14/13 21:57  
AS4 Serial no: 018684  
Version no: 4797

No. Date Time ZRAC  
Calibration 11/03/11 16:03 .100

Test Result
0001 09/11/13 08:21 .072
0002 09/11/13 08:23 .050
0003 09/12/13 02:39 .019
0004 09/17/13 23:21 .031
0005 09/18/13 01:59 .023
0006 09/18/13 03:38 .000
0007 09/19/13 02:16 .094
0008 09/19/13 02:19 .094
0009 09/19/13 21:06 .096
0010 09/27/13 11:47 .101
0011 10/03/13 12:10 .101
0012 10/11/13 12:59 .102
0013 10/15/13 13:25 .102
0014 10/16/13 06:39 .122
0015 10/16/13 08:42 .113
0016 10/16/13 22:54 .075
0017 10/22/13 01:39 .030
0018 10/23/13 19:55 .100
0019 10/29/13 02:03 .127
0020 10/29/13 02:06 .109
0021 10/29/13 23:40 .400
0022 10/29/13 23:41 .190
0023 10/29/13 23:44 .045
0024 10/31/13 14:48 .105
0025 11/14/13 21:56 .105

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8635
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2200	34	13B06-16-5-10- KKF24C	.100	27		.100				Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013				Expires on: 02-06-2015		11/14/13 22:07 AS4 Serial no: 008635 Version no: 4797		No. Date Time %RAC Calibration 10/24/03 09:04 .100		Test Result 0001 09/17/13 02:25 .111 0002 09/17/13 02:27 .111 0003 09/17/13 02:37 .076 0004 09/18/13 01:35 .157 0005 09/18/13 01:38 .157 0006 09/18/13 13:06 .103 0007 09/19/13 01:02 .000 0008 09/19/13 01:14 .000 0009 09/20/13 01:55 .138 0010 09/20/13 01:58 .130 0011 09/20/13 04:36 .000 0012 09/23/13 23:21 .218 0013 09/23/13 23:29 .195 0014 09/23/13 23:26 .218 0015 09/24/13 03:03 .179 0016 09/24/13 03:05 .184 0017 09/25/13 21:46 .000 0018 09/25/13 21:47 .336 0019 09/25/13 21:49 .353 0020 09/25/13 21:53 .937 0021 09/27/13 11:53 .162 0022 10/02/13 02:41 .134 0023 10/02/13 02:43 .139 0024 10/03/13 00:57 .051 0025 10/03/13 12:25 .102 0026 10/04/13 02:43 .119 0027 10/04/13 02:45 .112 0028 10/04/13 02:51 .120			
0029	10/06/13	13:55	.066										
0030	10/06/13	13:56	.047										
0031	10/06/13	16:38	.065										
0032	10/06/13	17:53	.063										
0033	10/08/13	01:36	.095										
0034	10/08/13	01:39	.094										
0035	10/08/13	03:37	.170										
0036	10/08/13	03:39	.166										
0037	10/08/13	03:44	.000										
0038	10/08/13	03:45	.000										
0039	10/11/13	13:07	.103										
0040	10/15/13	01:29	.089										
0041	10/15/13	01:31	.093										
0042	10/15/13	13:29	.099										
0043	10/16/13	01:59	.129										
0044	10/16/13	02:01	.117										
0045	10/17/13	02:39	.201										
0046	10/17/13	02:41	.204										
0047	10/23/13	01:55	.000										
0048	10/23/13	02:43	.219										
0049	10/23/13	02:46	.168										
0050	10/23/13	02:49	.192										
0051	10/23/13	19:59	.091										
0052	10/23/13	21:51	.132										
0053	10/23/13	21:54	.142										
0054	10/23/13	23:48	.114										
0055	10/23/13	23:50	.109										
0056	10/29/13	02:25	.000										
0057	10/30/13	02:03	.213										
0058	10/30/13	02:05	.214										
0059	10/31/13	15:02	.101										
0060	11/05/13	03:39	.000										
0061	11/06/13	04:35	.060										
0062	11/06/13	04:37	.064										
0063	11/07/13	02:33	.240										
0064	11/07/13	02:35	.248										
0065	11/11/13	21:05	.133										
0066	11/11/13	21:09	.115										
0067	11/12/13	03:49	.152										
0068	11/12/13	03:52	.158										
0069	11/14/13	22:07	.100										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8845
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2046	34	13B06-16-5-10- KKF24C	.100	20		.106			X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													

No.	Date	Time	%BAC
11/14/13 20:47			
AS4 Serial no: 008845			
Version no: 4797			
Calibration 12/24/03 11:01 .100			
Test Result			
0001	09/12/13	20:10	.000
0002	09/18/13	02:41	.326
0003	09/18/13	02:43	.344
0004	09/18/13	20:06	.098
0005	09/24/13	02:39	.042
0006	09/24/13	20:51	.114
0007	09/24/13	20:53	.118
0008	09/24/13	23:24	.117
0009	09/24/13	23:26	.108
0010	09/25/13	02:46	.210
0011	09/27/13	11:50	.103
0012	10/03/13	01:10	.000
0013	10/03/13	02:13	.063
0014	10/03/13	02:14	.063
0015	10/03/13	12:22	.104
0016	10/08/13	23:04	.000
0017	10/09/13	02:42	.000
0018	10/11/13	13:10	.105
0019	10/15/13	13:24	.105
0020	10/15/13	23:17	.260
0021	10/15/13	23:19	.252
0022	10/22/13	22:10	.211
0023	10/22/13	22:12	.203
0024	10/23/13	13:11	.104
0025	10/23/13	21:58	.155
0026	10/23/13	22:00	.169
0027	10/30/13	00:08	.085
0028	10/30/13	00:10	.093
0029	10/31/13	21:14	.091
0030	10/31/13	21:18	.094
0031	10/31/13	21:22	.030
0032	11/01/13	14:04	.106
0033	11/05/13	21:20	.271
0034	11/05/13	21:22	.293
0035	11/12/13	01:20	.143
0036	11/12/13	01:22	.137
0037	11/12/13	01:27	.071
0038	11/14/13	00:15	.000
0039	11/14/13	20:46	.106

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82653
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2:58	34	13B06-16-5-10- KKF24C	.100	20		.096			X		Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/14/13 21:59  
AS4 Serial no: 082653  
Version no: 4797

No. Date Time %BAC  
Calibration  
12/01/11 07:44 .081

Test Result  
0001 10/15/13 12:55 .098  
0002 10/23/13 13:01 .097  
0003 10/31/13 14:51 .098  
0004 11/14/13 21:59 .096

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8848
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2210	34	13B06-16-5-10-KKF24C	.100	20		102			X	Valadez	16071	Time Change
Prepared by:													
Douglas C. Talbot													
02-06-2013													
Expires on:													
02-06-2015													

```

11/14/13 22:11
AS4 Serial no: 008848
Version no: 4797

No. Date Time %BAC
Calibration
01/12/03 09:25 .100

Test Result
0001 09/18/13 13:07 .104
0002 09/27/13 11:52 .103
0003 10/03/13 12:25 .102
0004 10/11/13 13:13 .103
0005 10/15/13 13:02 .102
0006 10/23/13 13:07 .102
0007 10/31/13 15:01 .104
0008 11/14/13 22:10 .102
    
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77591
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2209	34	13B06-16-5-10- KKF24C	.100	19		.099				X	Valadez	16071	Time Change

Prepared by:  
Douglas C. Talbot  
02-06-2013  
Expires on:  
02-06-2015

11/14/13 22:10  
AS4 Serial no: 077591  
Version no: 4797  
No. Date Time %BAC  
Calibration 05/21/08 15:09 .100  
Test Result  
0001 09/18/13 13:06 .103  
0002 09/27/13 11:51 .101  
0003 10/03/13 12:23 .101  
0004 10/11/13 13:12 .102  
0005 10/15/13 12:55 .101  
0006 10/23/13 13:03 .097  
0007 10/31/13 14:59 .103  
0008 11/02/13 10:03 .000  
0009 11/03/13 11:07 .000  
0010 11/14/13 22:09 .099

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8850
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2213	34	13B06-16-5-10- KKF24C	.100	20		.03			X	Valadez	16071	Time Change
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

```

11/14/13 22:14
AS4 Serial no: 008850
Version no: 4797

No. Date Time %BAC
Calibration
03/02/06 09:07 .100

Test Result
0001 09/18/13 20:40 .100
0002 09/19/13 23:26 .134
0003 09/19/13 23:29 .148
0004 09/20/13 04:02 .137
0005 09/20/13 04:05 .152
0006 09/24/13 01:49 .103
0007 09/24/13 01:51 .099
0008 09/24/13 19:50 .000
0009 09/25/13 02:18 .112
0010 09/25/13 02:21 .112
0011 09/25/13 20:20 .088
0012 09/25/13 20:21 .017
0013 09/26/13 02:00 .207
0014 09/26/13 02:02 .195
0015 09/27/13 12:55 .100
0016 10/04/13 02:47 .113
0017 10/04/13 02:49 .109
0018 10/08/13 01:12 .086
0019 10/09/13 23:43 .040
0020 10/11/13 13:49 .101
0021 10/15/13 01:33 .142
0022 10/15/13 01:36 .137
0023 10/15/13 12:54 .100
0024 10/16/13 22:57 .021
0025 10/17/13 02:09 .000
0026 10/18/13 01:01 .000
0027 10/18/13 02:11 .131
0028 10/18/13 02:14 .129
0029 10/18/13 02:18 .084
0030 10/23/13 01:44 .000
0031 10/23/13 22:57 .157
0032 10/23/13 23:00 .183
0033 10/23/13 23:01 .183
0034 10/30/13 01:29 .190
0035 10/30/13 01:31 .178
0036 10/31/13 15:01 .103
0037 11/01/13 03:03 .000
0038 11/05/13 00:03 .000
0039 11/05/13 23:19 .000
0040 11/14/13 22:13 .103
    
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77573
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2631	34	13B06-16-5-10- KKF24C	.100	21		.099			✓		Valadez	16071	Time Change

No.	Date	Time	%BAC
0001	09/10/13	02:23	.000
0002	09/11/13	02:49	.000
0003	09/11/13	02:51	.000
0004	09/16/13	20:30	.245
0005	09/16/13	20:40	.225
0006	09/17/13	01:57	.061
0007	09/17/13	03:05	.000
0008	09/17/13	22:51	.101
0009	09/17/13	22:54	.106
0010	09/18/13	13:05	.099
0011	09/18/13	22:37	.194
0012	09/18/13	22:39	.203
0013	09/24/13	02:12	.117
0014	09/24/13	02:15	.125
0015	09/25/13	20:16	.000
0016	09/25/13	22:28	.112
0017	09/25/13	22:30	.120
0018	09/27/13	12:53	.099
0019	09/30/13	22:23	.000
0020	10/01/13	01:24	.110
0021	10/01/13	01:27	.144
0022	10/01/13	01:29	.133
0023	10/02/13	21:50	.260
0024	10/02/13	21:52	.256
0025	10/03/13	02:42	.153
0026	10/03/13	02:44	.124
0027	10/03/13	02:47	.126
0028	10/03/13	12:22	.100
0029	10/04/13	01:16	.018
0030	10/04/13	02:30	.141
0031	10/04/13	02:33	.138
0032	10/10/13	03:17	.090
0033	10/10/13	03:20	.094
0034	10/10/13	03:22	.000
0035	10/11/13	14:34	.096
0036	10/13/13	00:05	.092
0037	10/13/13	00:06	.196
0038	10/13/13	00:07	.028
0039	10/13/13	00:08	.034
0040	10/13/13	00:09	.086
0041	10/13/13	00:11	.210
0042	10/13/13	00:12	.163
0043	10/13/13	00:13	.094
0044	10/13/13	02:46	.037

0045	10/13/13	02:49	.123
0046	10/13/13	02:50	.102
0047	10/13/13	02:52	.157
0048	10/13/13	02:53	.208
0049	10/13/13	03:48	.092
0050	10/13/13	04:50	.069
0051	10/13/13	10:53	.031
0052	10/14/13	23:09	.063
0053	10/14/13	23:11	.064
0054	10/15/13	00:34	.000
0055	10/15/13	01:43	.128
0056	10/15/13	01:46	.124
0057	10/15/13	12:40	.099
0058	10/15/13	19:54	.076
0059	10/15/13	19:56	.073
0060	10/16/13	20:37	.000
0061	10/18/13	02:11	.129
0062	10/18/13	02:13	.128
0063	10/22/13	02:37	.093
0064	10/22/13	02:39	.085
0065	10/24/13	02:50	.000
0066	10/24/13	02:52	.000
0067	10/29/13	01:05	.152
0068	10/29/13	01:08	.145
0069	10/30/13	01:26	.000
0070	10/30/13	21:34	.096
0071	10/30/13	21:36	.081
0072	10/30/13	21:38	.098
0073	10/30/13	21:46	.014
0074	10/31/13	01:24	.073
0075	10/31/13	01:26	.073
0076	10/31/13	03:39	.062
0077	10/31/13	03:42	.062
0078	10/31/13	12:51	.100
0079	10/31/13	14:59	.098
0080	11/01/13	02:31	.000
0081	11/01/13	03:43	.027
0082	11/01/13	03:46	.026
0083	11/04/13	21:47	.000
0084	11/05/13	03:02	.068
0085	11/05/13	03:04	.069
0086	11/05/13	03:27	.071
0087	11/06/13	02:19	.161
0088	11/06/13	02:21	.147
0089	11/06/13	22:51	.051
0090	11/07/13	03:17	.084
0091	11/07/13	03:19	.087
0092	11/07/13	03:21	.024
0093	11/11/13	20:28	.028
0094	11/13/13	20:27	.191
0095	11/13/13	20:30	.189
0096	11/14/13	03:03	.148
0097	11/14/13	03:06	.154
0098	11/14/13	20:31	.099

\* If th  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83051
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2209	34	13B06-16-5-10- KKF24C	.100	20		.100				X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015														

11/14/13 22:10  
AS4 Serial no: 003051  
Version no: 4797

No. Date Time %BAC  
Calibration 09/25/13 14:00 .100

Test Result  
0001 09/18/13 13:06 .099  
0002 09/21/13 13:46 .177  
0003 09/21/13 13:47 .206  
0004 09/27/13 12:52 .100  
0005 10/03/13 12:22 .100  
0006 10/10/13 14:06 .080  
0007 10/10/13 14:10 .105  
0008 10/11/13 12:18 .099  
0009 10/15/13 12:45 .099  
0010 10/23/13 12:52 .099  
0011 10/31/13 15:00 .099  
0012 11/14/13 22:09 .100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8897
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2224	34	13B06-16-5-10- KKF24C	.100	19		.109			X		Valadez	16071	Time Change

No.	Date	Time	%BAC
0001	09/12/13	02:20	.206
0002	09/12/13	02:24	.191
0003	09/13/13	00:22	.170
0004	09/13/13	00:24	.166
0005	09/14/13	03:01	.125
0006	09/14/13	03:05	.118
0007	09/14/13	21:50	.052
0008	09/15/13	01:29	.241
0009	09/15/13	01:32	.262
0010	09/15/13	01:34	.255
0011	09/18/13	13:20	.107
0012	09/21/13	04:34	.000
0013	09/22/13	03:48	.126
0014	09/22/13	03:51	.130
0015	09/23/13	00:49	.209
0016	09/23/13	00:51	.085
0017	09/23/13	00:55	.191
0018	09/26/13	22:44	.000
0019	09/27/13	04:16	.051
0020	09/27/13	04:18	.050
0021	09/27/13	13:04	.107
0022	09/28/13	00:57	.077
0023	09/28/13	00:59	.085
0024	09/28/13	01:00	.089
0025	09/28/13	02:55	.063
0026	09/28/13	02:57	.060
0027	09/28/13	03:24	.196
0028	09/28/13	03:26	.188
0029	09/28/13	03:28	.191
0030	09/29/13	02:33	.231
0031	09/29/13	02:37	.230
0032	09/30/13	01:22	.000
0033	09/30/13	01:23	.000
0034	09/30/13	03:00	.119
0035	09/30/13	03:02	.119
0036	10/03/13	12:35	.107
0037	10/05/13	00:37	.168
0038	10/05/13	00:40	.210
0039	10/05/13	00:43	.214
0040	10/05/13	03:45	.070
0041	10/06/13	02:26	.103
0042	10/06/13	02:28	.102
0043	10/06/13	04:17	.000
0044	10/11/13	14:46	.104
0045	10/12/13	02:03	.148
0046	10/12/13	02:06	.152
0047	10/12/13	04:14	.089
0048	10/12/13	04:18	.099
0049	10/13/13	01:21	.153
0050	10/13/13	01:24	.164
0051	10/13/13	03:53	.054
0052	10/13/13	03:54	.067
0053	10/14/13	01:16	.000

0054	10/14/13	03:50	.000
0055	10/15/13	12:56	.105
0056	10/19/13	02:51	.121
0057	10/19/13	02:53	.119
0058	10/19/13	02:54	.140
0059	10/20/13	03:46	.170
0060	10/20/13	03:48	.180
0061	10/20/13	03:50	.153
0062	10/21/13	02:11	.039
0063	10/21/13	02:14	.038
0064	10/21/13	03:36	.189
0065	10/21/13	03:39	.172
0066	10/23/13	13:01	.107
0067	10/25/13	00:19	.109
0068	10/25/13	00:22	.104
0069	10/26/13	03:43	.130
0070	10/26/13	03:46	.118
0071	10/26/13	21:33	.097
0072	10/26/13	21:36	.139
0073	10/26/13	21:37	.137
0074	10/26/13	21:40	.108
0075	10/26/13	21:49	.000
0076	10/27/13	02:44	.093
0077	10/27/13	02:47	.091
0078	10/27/13	02:51	.093
0079	10/27/13	02:57	.065
0080	10/28/13	03:03	.113
0081	10/28/13	03:06	.116
0082	10/31/13	15:13	.106
0083	11/01/13	23:24	.051
0084	11/01/13	23:29	.046
0085	11/02/13	00:52	.182
0086	11/02/13	00:56	.170
0087	11/02/13	03:47	.032
0088	11/02/13	04:41	.096
0089	11/02/13	04:45	.099
0090	11/02/13	04:52	.028
0091	11/02/13	22:23	.185
0092	11/02/13	22:27	.178
0093	11/03/13	01:32	.000
0094	11/03/13	02:43	.173
0095	11/03/13	02:45	.164
0096	11/03/13	03:53	.134
0097	11/03/13	03:56	.134
0098	11/04/13	03:48	.252
0099	11/04/13	03:50	.243
0100	11/08/13	03:38	.076
0101	11/08/13	03:41	.070
0102	11/09/13	02:17	.104
0103	11/09/13	02:21	.100
0104	11/09/13	22:10	.098
0105	11/09/13	22:13	.108
0106	11/10/13	00:36	.226
0107	11/10/13	00:30	.225
0108	11/10/13	02:34	.023
0109	11/10/13	03:57	.119
0110	11/10/13	04:00	.116
0111	11/10/13	04:07	.082
0112	11/10/13	04:10	.107
0113	11/10/13	06:52	.122
0114	11/10/13	06:56	.118
0115	11/14/13	22:24	.109

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18624
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013			13B06-16-5-10- KKF24C	.100								Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

B.O.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18646
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013		34	13B06-16-5-10- KKF24C	.100								Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

*NOT IN LOCKER*

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77488
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2216	34	13B06-16-5-10- KKF24C	.100	20		105			X		Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="width: 30%;"> <p>11/14/13 22:17  AS4 Serial no: 077488  Version no: 4797</p> </div> <div style="width: 30%;"> <p>No. Date Time ZBAC  Calibration  05/21/08 14:35 .100</p> </div> <div style="width: 30%;"> <p>Test Result  0001 09/18/13 13:15 .102  0002 09/27/13 12:55 .102  0003 10/03/13 12:32 .102  0004 10/11/13 14:36 .102  0005 10/11/13 14:37 .101  0006 10/15/13 13:13 .104  0007 10/23/13 12:45 .104  0008 10/31/13 15:03 .102  0009 11/14/13 22:17 .105</p> </div> </div>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8877
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS															
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																			
11/14/2013	2230	34	13B06-16-5-10- KKF24C	.100	20		.106			X		Valadez	16071	Time Change															
Prepared by: Douglas C. Talbot 02-06-2013																													
Expires on: 02-06-2015																													
<table border="1"> <tr> <td>11/14/13 22:31</td> <td>AS4 Serial no: 008877</td> <td>Version no: 4797</td> <td>No. Date Time %BAC</td> <td>Calibration</td> <td>05/26/05 10:40 .100</td> <td>Test Result</td> <td>0001 09/18/13 13:28 .104</td> <td>0002 09/27/13 13:06 .102</td> <td>0003 10/03/13 12:44 .106</td> <td>0004 10/11/13 14:49 .103</td> <td>0005 10/15/13 13:21 .105</td> <td>0006 10/23/13 12:53 .105</td> <td>0007 10/31/13 15:15 .104</td> <td>0008 11/14/13 22:30 .106</td> </tr> </table>															11/14/13 22:31	AS4 Serial no: 008877	Version no: 4797	No. Date Time %BAC	Calibration	05/26/05 10:40 .100	Test Result	0001 09/18/13 13:28 .104	0002 09/27/13 13:06 .102	0003 10/03/13 12:44 .106	0004 10/11/13 14:49 .103	0005 10/15/13 13:21 .105	0006 10/23/13 12:53 .105	0007 10/31/13 15:15 .104	0008 11/14/13 22:30 .106
11/14/13 22:31	AS4 Serial no: 008877	Version no: 4797	No. Date Time %BAC	Calibration	05/26/05 10:40 .100	Test Result	0001 09/18/13 13:28 .104	0002 09/27/13 13:06 .102	0003 10/03/13 12:44 .106	0004 10/11/13 14:49 .103	0005 10/15/13 13:21 .105	0006 10/23/13 12:53 .105	0007 10/31/13 15:15 .104	0008 11/14/13 22:30 .106															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18615
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2233	34	13B06-16-5-10- KKF24C	.100	21		.099			X		Valadez	16071	Time Change
			Prepared by: Douglas C. Talbot 02-06-2013											
			Expires on: 02-06-2015											

11/14/13 22:34  
 AS4 Serial no: 018615  
 Version no: 4797

No. Date Time %BAC  
 Calibration 12/19/08 14:44 .100

Test Result  
 0001 09/14/13 01:42 .073  
 0002 09/14/13 02:02 .000  
 0003 09/14/13 23:53 .059  
 0004 09/15/13 01:15 .135  
 0005 09/15/13 01:17 .138  
 0006 09/15/13 01:54 .142  
 0007 09/15/13 01:57 .145  
 0008 09/15/13 03:39 .152  
 0009 09/15/13 03:41 .141  
 0010 09/15/13 04:20 .000  
 0011 09/15/13 05:26 .169  
 0012 09/15/13 05:28 .158  
 0013 09/18/13 13:28 .098  
 0014 09/21/13 03:41 .044  
 0015 09/21/13 04:12 .199  
 0016 09/21/13 04:14 .195  
 0017 09/23/13 01:05 .179  
 0018 09/23/13 01:07 .184  
 0019 09/23/13 02:40 .148  
 0020 09/23/13 02:41 .145  
 0021 09/27/13 13:05 .098  
 0022 09/27/13 20:03 .061  
 0023 09/27/13 22:52 .183  
 0024 09/27/13 22:55 .179  
 0025 09/27/13 22:58 .088  
 0026 09/28/13 04:12 .021  
 0027 09/28/13 04:14 .020  
 0028 09/29/13 21:19 .067  
 0029 09/29/13 21:21 .063  
 0030 09/30/13 02:09 .114  
 0031 09/30/13 02:12 .115  
 0032 10/03/13 12:45 .099  
 0033 10/06/13 06:53 .223  
 0034 10/06/13 06:55 .229  
 0035 10/06/13 16:26 .179  
 0036 10/06/13 16:27 .207  
 0037 10/06/13 16:31 .180  
 0038 10/11/13 14:47 .100  
 0039 10/15/13 13:17 .101  
 0040 10/23/13 12:46 .100  
 0041 10/31/13 15:15 .100  
 0042 11/14/13 22:33 .099

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18626
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2233	34	13B06-16-5-10- KKF24C	.100	21		.100			X		Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015														
<pre> 11/14/13 22:35 AS4 Serial no: 018626 Version no: 4797  No. Date Time %BAC Calibration 04/29/05 08:59 .101  Test Result 0001 09/18/13 13:27 .100 0002 09/27/13 13:03 .101 0003 10/09/13 12:45 .100 0004 10/11/13 14:45 .101 0005 10/15/13 13:12 .103 0006 10/23/13 12:33 .101 0007 10/31/13 15:12 .102 0008 11/14/13 22:34 .100                     </pre>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18696
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013			13B06-16-5-10- KKF24C	.100								Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

B.C.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83030
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2236	34	13B06-16-5-10- KKF24C	.100	20		.099			X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015													
11/14/13 22:37 #54 Serial no: 083030 Version no: 4797 No. Date Time %BAC Calibration 01/29/09 10:39 .100 Test Result 0001 09/18/13 13:33 .101 0002 09/27/13 13:04 .102 0003 10/03/13 12:47 .101 0004 10/11/13 14:47 .101 0005 10/15/13 13:02 .103 0006 10/23/13 12:28 .102 0007 10/31/13 15:12 .101 0008 11/14/13 22:36 .099													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18692
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2242	34	13B06-16-5-10- KKF24C	.100	20		.099				X	Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

```

11/14/13 22:44
AS4 Serial no: 018692
Version no: 4797

No. Date Time %BAC
Calibration
08/12/08 14:35 .100

Test Result
0001 09/18/13 13:40 .098
0002 09/27/13 11:36 .100
0003 10/03/13 12:52 .100
0004 10/11/13 11:35 .096
0005 10/15/13 13:04 .101
0006 10/23/13 12:30 .102
0007 10/31/13 15:17 .100
0008 11/14/13 22:43 .099
    
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18762
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																																																																																							
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																																																																																																																						
11/14/2013	2241	34	13B06-16-5-10-KKF24C	.100	20		.100			X	Valadez	16071	Time Change																																																																																																																							
Prepared by: Douglas C. Talbot 02-06-2013																																																																																																																																				
Expires on: 02-06-2015																																																																																																																																				
<table border="0" style="width: 100%;"> <tr> <td style="width: 20%;">11/14/13 22:42</td> <td style="width: 20%;">AS4 Serial no: 018762</td> <td style="width: 20%;">Version no: 4797</td> <td style="width: 20%;">No. Date Time %BAC</td> <td style="width: 20%;">Calibration</td> <td style="width: 20%;">03/30/09 15:52 .100</td> </tr> <tr> <td>Test Result</td> <td>0001</td> <td>09/18/13</td> <td>20:09</td> <td>.101</td> <td>0002</td> <td>09/20/13</td> <td>02:47</td> <td>.131</td> <td>0003</td> <td>09/20/13</td> <td>02:49</td> <td>.110</td> <td>0004</td> <td>09/20/13</td> <td>02:52</td> <td>.123</td> <td>0005</td> <td>09/27/13</td> <td>13:05</td> <td>.100</td> <td>0006</td> <td>10/03/13</td> <td>12:48</td> <td>.100</td> <td>0007</td> <td>10/03/13</td> <td>12:50</td> <td>.101</td> <td>0008</td> <td>10/08/13</td> <td>00:07</td> <td>.000</td> <td>0009</td> <td>10/08/13</td> <td>00:59</td> <td>.060</td> <td>0010</td> <td>10/08/13</td> <td>01:02</td> <td>.070</td> <td>0011</td> <td>10/08/13</td> <td>21:26</td> <td>.030</td> <td>0012</td> <td>10/09/13</td> <td>01:09</td> <td>.024</td> <td>0013</td> <td>10/11/13</td> <td>14:48</td> <td>.100</td> <td>0014</td> <td>10/15/13</td> <td>12:37</td> <td>.100</td> <td>0015</td> <td>10/16/13</td> <td>00:22</td> <td>.174</td> <td>0016</td> <td>10/16/13</td> <td>00:24</td> <td>.166</td> <td>0017</td> <td>10/23/13</td> <td>12:23</td> <td>.102</td> <td>0018</td> <td>10/24/13</td> <td>01:01</td> <td>.000</td> <td>0019</td> <td>10/24/13</td> <td>01:02</td> <td>.177</td> <td>0020</td> <td>10/31/13</td> <td>02:17</td> <td>.112</td> <td>0021</td> <td>10/31/13</td> <td>02:20</td> <td>.106</td> <td>0022</td> <td>10/31/13</td> <td>15:14</td> <td>.101</td> <td>0023</td> <td>11/07/13</td> <td>03:12</td> <td>.173</td> <td>0024</td> <td>11/07/13</td> <td>03:15</td> <td>.178</td> <td>0025</td> <td>11/12/13</td> <td>00:11</td> <td>.120</td> <td>0026</td> <td>11/12/13</td> <td>00:13</td> <td>.124</td> <td>0027</td> <td>11/12/13</td> <td>00:18</td> <td>.000</td> <td>0028</td> <td>11/14/13</td> <td>22:42</td> <td>.100</td> </tr> </table>														11/14/13 22:42	AS4 Serial no: 018762	Version no: 4797	No. Date Time %BAC	Calibration	03/30/09 15:52 .100	Test Result	0001	09/18/13	20:09	.101	0002	09/20/13	02:47	.131	0003	09/20/13	02:49	.110	0004	09/20/13	02:52	.123	0005	09/27/13	13:05	.100	0006	10/03/13	12:48	.100	0007	10/03/13	12:50	.101	0008	10/08/13	00:07	.000	0009	10/08/13	00:59	.060	0010	10/08/13	01:02	.070	0011	10/08/13	21:26	.030	0012	10/09/13	01:09	.024	0013	10/11/13	14:48	.100	0014	10/15/13	12:37	.100	0015	10/16/13	00:22	.174	0016	10/16/13	00:24	.166	0017	10/23/13	12:23	.102	0018	10/24/13	01:01	.000	0019	10/24/13	01:02	.177	0020	10/31/13	02:17	.112	0021	10/31/13	02:20	.106	0022	10/31/13	15:14	.101	0023	11/07/13	03:12	.173	0024	11/07/13	03:15	.178	0025	11/12/13	00:11	.120	0026	11/12/13	00:13	.124	0027	11/12/13	00:18	.000	0028	11/14/13	22:42	.100
11/14/13 22:42	AS4 Serial no: 018762	Version no: 4797	No. Date Time %BAC	Calibration	03/30/09 15:52 .100																																																																																																																															
Test Result	0001	09/18/13	20:09	.101	0002	09/20/13	02:47	.131	0003	09/20/13	02:49	.110	0004	09/20/13	02:52	.123	0005	09/27/13	13:05	.100	0006	10/03/13	12:48	.100	0007	10/03/13	12:50	.101	0008	10/08/13	00:07	.000	0009	10/08/13	00:59	.060	0010	10/08/13	01:02	.070	0011	10/08/13	21:26	.030	0012	10/09/13	01:09	.024	0013	10/11/13	14:48	.100	0014	10/15/13	12:37	.100	0015	10/16/13	00:22	.174	0016	10/16/13	00:24	.166	0017	10/23/13	12:23	.102	0018	10/24/13	01:01	.000	0019	10/24/13	01:02	.177	0020	10/31/13	02:17	.112	0021	10/31/13	02:20	.106	0022	10/31/13	15:14	.101	0023	11/07/13	03:12	.173	0024	11/07/13	03:15	.178	0025	11/12/13	00:11	.120	0026	11/12/13	00:13	.124	0027	11/12/13	00:18	.000	0028	11/14/13	22:42	.100																				

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#43  
18766

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18766
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2248	34	13B06-16-5-10- KKF24C	.100	20		892			X		Valadez	16071	Time Change
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/14/13 22:50  
AS4 Serial no: 018766  
Version no: 4797

No. Date Time %BAC  
Calibration 03/12/12 12:30 .100

Test Result  
0001 09/18/13 13:44 .095  
0002 09/21/13 08:32 .167  
0003 09/21/13 08:35 .169  
0004 09/21/13 08:37 .170  
0005 09/27/13 13:11 .096  
0006 09/28/13 03:38 .145  
0007 09/28/13 03:41 .159  
0008 09/29/13 01:20 .064  
0009 09/29/13 03:05 .116  
0010 09/29/13 03:06 .124  
0011 09/29/13 03:15 .000  
0012 09/30/13 02:06 .196  
0013 09/30/13 02:08 .197  
0014 10/03/13 13:12 .095  
0015 10/12/13 11:30 .175  
0016 10/12/13 11:32 .176  
0017 10/15/13 12:39 .095  
0018 10/23/13 12:26 .097  
0019 10/26/13 07:56 .144  
0020 10/26/13 07:59 .140  
0021 10/27/13 08:00 .172  
0022 10/27/13 08:03 .164  
0023 10/31/13 15:19 .096  
0024 11/14/13 22:49 .092

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18691
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2256	39	13B06-16-5-10-	.100	21		105			X	Valadez	16071	Time Change
			KKF24C										

11/14/13 22:57  
AS4 Serial no: 018691  
Version no: 4797

No. Date Time XMAC  
Calibration 09/08/08 14:01 .100

Test Result

0045 10/20/13 05:16 .136  
0046 10/21/13 02:16 .155  
0047 10/21/13 02:20 .131  
0048 10/23/13 12:17 .105  
0049 10/25/13 01:27 .070  
0050 10/25/13 01:30 .068  
0051 10/25/13 02:21 .042  
0052 10/25/13 02:49 .032  
0053 10/26/13 01:50 .213  
0054 10/26/13 01:55 .159  
0055 10/26/13 04:50 .172  
0056 10/26/13 04:52 .182  
0057 10/27/13 00:08 .058  
0058 10/27/13 00:57 .031  
0059 10/27/13 01:15 .123  
0060 10/27/13 01:18 .121  
0061 10/27/13 03:12 .000  
0062 10/28/13 02:23 .077  
0063 10/28/13 02:25 .000  
0064 10/28/13 03:56 .154  
0065 10/28/13 03:57 .134  
0066 10/31/13 15:25 .104  
0067 11/02/13 03:53 .125  
0068 11/02/13 03:54 .117  
0069 11/03/13 01:40 .049  
0070 11/03/13 02:01 .125  
0071 11/03/13 02:03 .126  
0072 11/03/13 02:39 .132  
0073 11/03/13 02:42 .138  
0074 11/03/13 03:56 .157  
0075 11/03/13 03:59 .149  
0076 11/04/13 01:20 .120  
0077 11/04/13 01:23 .119  
0078 11/04/13 02:12 .170  
0079 11/04/13 02:15 .155  
0080 11/04/13 01:53 .077  
0081 11/08/13 01:55 .075  
0082 11/08/13 04:45 .132  
0083 11/08/13 04:47 .126  
0084 11/09/13 00:36 .201  
0085 11/09/13 00:38 .207  
0086 11/09/13 03:02 .177  
0087 11/09/13 03:07 .188  
0088 11/09/13 07:13 .164  
0089 11/09/13 07:16 .190  
0090 11/09/13 22:33 .193  
0091 11/09/13 22:35 .195  
0092 11/10/13 03:14 .140  
0093 11/10/13 03:17 .128  
0094 11/11/13 04:02 .092  
0095 11/11/13 04:04 .093  
0096 11/14/13 22:56 .105

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 17990
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013			13B06-16-5-10- KKF24C	.100							Valadez	16071	Time Change
			Prepared by: Douglas C. Talbot										
			02-06-2013										
			Expires on: 02-06-2015										

B.O.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18774
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2254	34	13B06-16-5-10- KKF24C	.100	21		101				X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														

11/14/13 22:57  
 AS4 Serial no: 018774  
 Version no: 4797

No. Date Time %BAC  
 Calibration 01/13/09 10:25 .100

Test Result
0001 09/10/13 23:55 .161
0002 09/10/13 23:57 .141
0003 09/11/13 02:19 .196
0004 09/11/13 02:21 .195
0005 09/12/13 02:32 .000
0006 09/17/13 02:23 .067
0007 09/17/13 02:25 .072
0008 09/18/13 01:36 .167
0009 09/18/13 01:38 .201
0010 09/18/13 13:49 .100
0011 09/18/13 19:43 .057
0012 09/18/13 19:45 .052
0013 09/24/13 01:06 .156
0014 09/24/13 01:09 .154
0015 09/24/13 05:50 .000
0016 09/25/13 00:01 .024
0017 09/25/13 00:02 .023
0018 09/27/13 13:13 .102
0019 10/02/13 20:45 .128
0020 10/02/13 20:48 .141
0021 10/03/13 13:24 .100
0022 10/09/13 01:27 .220
0023 10/09/13 01:30 .219
0024 10/11/13 14:53 .100
0025 10/15/13 12:03 .099
0026 10/16/13 02:00 .000
0027 10/22/13 02:13 .135
0028 10/22/13 02:16 .128
0029 10/22/13 02:19 .089
0030 10/22/13 21:34 .000
0031 10/23/13 12:09 .102
0032 10/31/13 02:45 .091
0033 10/31/13 02:48 .087
0034 10/31/13 15:22 .101
0035 11/05/13 01:15 .000
0036 11/11/13 22:01 .106
0037 11/11/13 22:04 .105
0038 11/13/13 21:43 .039
0039 11/13/13 21:46 .039
0040 11/14/13 22:56 .101

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41486
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2300	34	13B06-16-5-10- KKF24C	.100	20		.107			X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px;">           11/14/13 23:01            AS4 Serial no: 041486            Version no: 4797         </div> <div style="border: 1px solid black; padding: 5px;">           No. Date Time %BAC            Calibration            03/08/06 13:56 .100         </div> <div style="border: 1px solid black; padding: 5px;">           Test Result            0001 09/18/13 13:52 .106            0002 09/27/13 13:15 .105            0003 10/03/13 13:26 .106            0004 10/11/13 14:56 .105            0005 10/15/13 12:02 .105            0006 10/23/13 12:08 .106            0007 10/31/13 15:24 .106            0008 11/14/13 23:01 .107         </div> </div>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41495
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2256	37	13B06-16-5-10- KKF24C	.100	19		.097			X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
<div style="border: 1px solid black; padding: 5px;"> <p>11/14/13 22:57 AS4 Serial no: 041495 Version no: 4797</p> <p>No. Date Time ZBAC Calibration 09/29/10 10:45 .098</p> <p>Test Result 0001 09/18/13 13:40 .095 0002 09/27/13 13:14 .095 0003 10/03/13 13:22 .096 0004 10/11/13 14:52 .095 0005 10/15/13 11:54 .095 0006 10/23/13 12:01 .096 0007 10/31/13 15:19 .096 0008 11/14/13 22:56 .097</p> </div>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41502
--------------------------------	-----------------	---	----------------------	--	-------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/14/2013	2:30	34	13B06-16-5-10- KKF24C	.100	20		.100				X	Valadez	16071	Time Change

Pre  
Exp

11/14/13 23:05  
AS4 Serial no: 041502  
Version no: 4797

No. Date Time ZPAC  
Calibration 05/05/09 08:03 .100

Test Result

- 0001 09/14/13 02:15 .083
- 0002 09/14/13 02:18 .086
- 0003 09/14/13 22:18 .000
- 0004 09/14/13 23:58 .000
- 0005 09/14/13 23:58 .000
- 0006 09/14/13 23:59 .000
- 0007 09/15/13 00:59 .039
- 0008 09/15/13 01:02 .041
- 0009 09/18/13 13:54 .099
- 0010 09/22/13 01:33 .000
- 0011 09/22/13 03:53 .000
- 0012 09/22/13 03:54 .000
- 0013 09/27/13 13:28 .098
- 0014 10/03/13 13:28 .098
- 0015 10/04/13 19:27 .295
- 0016 10/04/13 19:31 .295
- 0017 10/05/13 01:18 .000
- 0018 10/05/13 03:18 .136
- 0019 10/05/13 03:20 .140
- 0020 10/05/13 22:01 .122
- 0021 10/05/13 22:04 .122
- 0022 10/06/13 01:06 .171
- 0023 10/06/13 01:08 .177
- 0024 10/06/13 03:06 .124
- 0025 10/06/13 03:09 .124
- 0026 10/07/13 03:45 .161
- 0027 10/07/13 03:49 .125
- 0028 10/07/13 03:51 .155
- 0029 10/11/13 15:02 .097
- 0030 10/13/13 00:24 .000
- 0031 10/13/13 01:40 .208
- 0032 10/13/13 01:42 .186
- 0033 10/13/13 01:44 .198
- 0034 10/18/13 13:39 .096
- 0035 10/18/13 23:26 .107
- 0036 10/18/13 23:28 .107
- 0037 10/19/13 03:14 .029
- 0038 10/19/13 03:47 .046
- 0039 10/19/13 03:49 .043
- 0040 10/20/13 02:28 .080
- 0041 10/20/13 02:30 .077
- 0042 10/20/13 21:06 .094
- 0043 10/20/13 21:08 .092
- 0044 10/23/13 12:05 .099
- 0045 10/24/13 23:00 .158
- 0046 10/24/13 23:01 .169
- 0047 10/24/13 23:04 .165
- 0048 10/26/13 04:30 .165
- 0049 10/26/13 04:33 .166
- 0050 10/26/13 22:00 .000
- 0051 10/26/13 23:59 .102
- 0052 10/27/13 00:02 .113
- 0053 10/27/13 02:52 .197
- 0054 10/27/13 02:54 .166
- 0055 10/27/13 02:56 .204
- 0056 10/27/13 21:10 .000
- 0057 10/28/13 04:50 .136
- 0058 10/28/13 04:53 .132
- 0059 10/31/13 15:26 .101
- 0060 11/02/13 03:08 .209
- 0061 11/02/13 03:12 .199
- 0062 11/14/13 23:05 .100

\* If the reading is within plus or minus (±) 0.05 percent of the true value (acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



S.G.T.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8874
--------------------------------	-----------------	---	----------------------	--	------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/14/2013	2048	34	13B06-16-5-10- KKF24C	.100	20		.102			X	Valadez	16071	Time Change
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015													
<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px;">           11/14/13 20:49            AS4 Serial no: 008874            Version no: 4797         </div> <div style="border: 1px solid black; padding: 5px;">           No. Date Time %RAC            Calibration 06/25/13 10:32 .107         </div> <div style="border: 1px solid black; padding: 5px;">           Test Result            0001 09/27/13 11:00 .103            0002 10/03/13 11:19 .104            0003 10/11/13 11:43 .102            0004 10/15/13 11:55 .102            0005 10/23/13 12:02 .103            0006 10/31/13 13:35 .104            0007 11/14/13 20:48 .102         </div> </div>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.