

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8501
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1243	34	13B06-16-5-10-KKF24C	.100	20		.093			X		Flick	15063	
Prepared by:														
Douglas C. Talbot														
02-06-2013														
Expires on:														
02-06-2015														
11/21/13 12:45 AS4 Serial no: 008501 Version no: 4797 No. Date Time %BAC Calibration 03/30/09 15:40 .100 Test Result 0001 11/21/13 12:44 .093 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18785
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1254	34	13B06-16-5-10-KKF24C	.100	20		.098				X	Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 12:55
 AS4 Serial no: 018785
 Version no: 4797
 No. Date Time %BAC
 Calibration 07/18/08 13:27 .100
 Test Result
 0001 11/21/13 12:51 .000
 0002 11/21/13 12:54 .098
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8546
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1302	34	13B06-16-5-10-KKF24C	.100	20		.099			X		Flick	15063	
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px; width: 20%;"> Push No. 3 to erase, or power down to retain </div> <div style="border: 1px solid black; padding: 5px; width: 40%;"> <pre> Test Result 0001 11/15/13 01:07 .119 0002 11/15/13 01:09 .119 0003 11/15/13 01:11 .115 0004 11/15/13 06:19 .000 0005 11/15/13 06:20 .000 0006 11/18/13 22:47 .182 0007 11/18/13 22:49 .187 0008 11/18/13 23:27 .182 0009 11/18/13 23:29 .184 0010 11/21/13 13:03 .099 </pre> </div> <div style="border: 1px solid black; padding: 5px; width: 20%;"> <pre> No. Date Time 2BAC Calibration 06/28/13 16:36 .100 </pre> </div> <div style="border: 1px solid black; padding: 5px; width: 20%;"> <pre> 11/21/13 13:03 AS4 Serial no: 008546 Version no: 4797 </pre> </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

41

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82659
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1304	34	13B06-16-5-10-KKF24C	.100	20		105				Flick	15063	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										
											11/21/13 13:06		
											MS4 Serial no: 082659		
											Version no: 4797		
											No. Date Time %BAC		
											Calibration		
											10/22/08 17:37 .100		
											Test Result		
											0001 11/21/13 13:06 .105		
											Push No. 3 to erase,		
											or Power down to retain		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8559
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1908	54	13B06-16-5-10- KKF24C	.100	26		103				X	Flick	15063	
		Prepared by: Douglas C. Talbot 02-06-2013												
		Expires on: 02-06-2015												
		<div style="border: 1px solid black; padding: 5px; display: inline-block;"> 11/21/13 13:10 MS4 Serial no: 008559 Version no: 4797 No. Date Time %BAC Calibration 11/30/11 08:55 .106 Test Result 0001 11/21/13 13:09 .103 Push No. 3 to erase, or power down to retain </div>												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8580
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1911	34	13B06-16-5-10-KKF24C	.100	20		104			X	Flick	15063	
Prepared by:													
Douglas C. Talbot													
02-06-2013													
Expires on:													
02-06-2015													
<div style="text-align: right;"> 11/21/13 13:12 RS4 Serial no: 000500 Version no: 4797 No. Date Time %BAC Calibration 12/01/06 08:14 .099 Test Result 0001 11/19/13 02:53 .019 0002 11/21/13 13:12 .104 </div>													
<div style="text-align: center;"> Push No. 3 to erase. US OF POWER down to retain </div>													

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83032
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1313	34	13B06-16-5-10-KKF24C	.100	20		.100			X		Flick	15063	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>														
<p style="text-align: right;"> 11/21/13 13:14 AS4 Serial no: 083032 Version no: 4797 No. Date Time ZBAC Calibration 12/01/11 07:30 .081 Test Result 0001 11/16/13 01:42 .100 0002 11/16/13 01:44 .102 0003 11/16/13 04:10 .102 0004 11/16/13 04:20 .092 0005 11/16/13 04:25 .027 0006 11/16/13 04:26 .027 0007 11/21/13 13:13 .100 </p> <p style="text-align: center;"> Push No. 3 to erase, or Power down to retain </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8620
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1317	54	13B06-16-5-10- KKF24C	.100	20		100				Flick	15063	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>													
<pre> 11/21/13 13:18 AS4 Serial no: 008620 Version no: 4797 No. Date Time %BAC Calibration 01/30/13 14:34 .100 Test Result 0001 11/18/13 19:13 .000 0002 11/18/13 21:04 .025 0003 11/19/13 21:06 .000 0004 11/21/13 02:14 .104 0005 11/21/13 02:16 .104 0006 11/21/13 13:17 .100 Push No. 3 to erase, or power down to retain </pre>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82957
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1318	34	13B06-16-5-10-KKF24C	.100	21		.098					Flick	15063	

Prepared by:
Douglas C. Talbot
02-06-2013
Expires on:
02-06-2015

11/21/13 13:19
AS4 Serial no: 882957
Version no: 4797
No. Date Time %BAC
Calibration 01/24/13 06:06 .082
Test Result 0001 11/21/13 13:19 .098
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77505
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1821	84	13B06-16-5-10- KKF24C	.100	20		0097			X	Flick	15063	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>													
<pre> 11/21/13 13:22 AS4 Serial no: 077505 Version no: 4797 No. Date Time ZBAC Calibration 05/21/08 13:16 .100 Test Result 0001 11/21/13 13:21 .097 Push No. 3 to erase, or power down to retain </pre>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77479
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1323	34	13B06-16-5-10-KKF24C	.100	20		0.095			X	Flick	15063	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>													
<p>Test Result</p> <p>0001 11/16/13 02:29 .203 0002 11/16/13 02:32 .181 0003 11/16/13 20:07 .263 0004 11/16/13 20:10 .274 0005 11/17/13 00:40 .057 0006 11/17/13 02:22 .175 0007 11/17/13 02:25 .154 0008 11/17/13 02:27 .136 0009 11/21/13 13:24 .095</p> <p>No. Date Time %BAC Calibration 03/07/13 09:48 .099</p> <p>11/21/13 13:24 AS4 Serial no: 077479 Version no: 4797</p> <p>Push No. 3 to erase, or Power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1329	34	13B06-16-5-10- KKF24C	.100	20		.094			X	Lee, Kirk	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/21/13 13:30
 AS4 Serial no: 082660
 Version no: 4797

 No. Date Time ZBAC
 Calibration 11/18/10 06:21 .081

 Test Result
 0001 11/18/13 08:18 .181
 0002 11/18/13 08:22 .178
 0003 11/21/13 13:29 .094

 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1333	34	13B06-16-5-10-	.100	21		.010			X		Lee, Kirk	17046	MANUAL BUTTON ACCIDENTLY PUSHED PRERE TO BLOWING.
	1334	34	KKF24C		21		.102			X				
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="border: 1px solid black; padding: 5px;"> Push No. 3 to erase, or Power down to retain </div> <div style="font-family: monospace;"> Version no: 4797 No. Date Time %BAC Calibration 95/20/04 11:06 .096 Test Result 0001 11/16/13 02:44 .096 0002 11/16/13 02:56 .097 0003 11/16/13 02:58 .101 0004 11/16/13 21:24 .041 0005 11/17/13 02:27 .145 0006 11/17/13 02:28 .147 0007 11/17/13 02:39 .105 0008 11/17/13 02:35 .106 0009 11/21/13 13:34 .010 0010 11/21/13 13:36 .102 </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1346	34	13B06-16-5-10- KKF24C	.100	20		.107			X	Lee, Kirk	17046	
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
Test Result 0001 11/21/13 13:47 .107 Calibration 03/11/04 14:05 .100 No. Date Time %BAC 11/21/13 13:48 AS4 Serial no: 008714 Version no: 4797 Push No. 3 to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1341	34	13B06-16-5-10-KKF24C	.100	21	.101				X	Lee, Kirk	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/21/13 13:43
AS4 Serial no: 008888
Version no: 4797

No. Date Time %BAC
Calibration 03/25/13 18:45 .100

Test Result
0001 11/16/13 03:23 .167
0002 11/16/13 03:24 .158
0003 11/17/13 01:41 .104
0004 11/17/13 01:46 .106
0005 11/17/13 04:13 .118
0006 11/18/13 01:27 .000
0007 11/21/13 13:42 .101

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	135	34	13B06-16-5-10- KKF24C	.100	19		.104			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

```

11/21/13 13:53
AS4 Serial no: 008783
Version no: 4797

No. Date Time %BAC
Calibration
02/18/08 13:52 .100

Test Result
0001 11/19/13 08:24 .050
0002 11/19/13 08:26 .048
0003 11/19/13 01:16 .021
0004 11/19/13 01:17 .019
0005 11/21/13 13:52 .104

Push No. 3 to erase,
or power down to retain
  
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1356	34	13B06-16-5-10- KKF24C	.100	20		.097			X		Lee, Kirk	17046	
			Prepared by: Douglas C. Talbot											
			Expires on: 02-06-2015											

11/21/13 13:56
 HS4 Serial no: 082657
 version no: 4797
 No. Date Time %BAC
 Calibration 10/07/11 12:22 .081
 Test Result
 0001 11/19/13 11:30 .080
 0002 11/19/13 15:26 .080
 0003 11/21/13 13:56 .097
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1357	34	13B06-16-5-10- KKF24C	.100	19		.098			X		Lee, Kirk	17046	
			Prepared by: Douglas C. Talbot 02-06-2013											
			Expires on: 02-06-2015											

11/21/13 13:58
 AS4 Serial no: 008810
 Version no: 4797
 No. Date Time ZBAC
 Calibration
 10/30/03 09:53 .100
 Test Result
 0001 11/21/13 13:58 .098
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																												
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																											
11/21/2013	1359	34	13B06-16-5-10-KKF24C	.100	21		.102			X	Lee, Kirk	17046																													
Prepared by: Douglas C. Talbot 02-06-2013																																									
Expires on: 02-06-2015																																									
<table border="1"> <tr> <td>11/21/13 14:01</td> <td>AS4 Serial no: 018684</td> <td>Version no: 4797</td> <td>No. Date Time %BAC</td> </tr> <tr> <td colspan="3"></td> <td>Calibration 11/03/11 16:03 .100</td> </tr> <tr> <td colspan="4" style="text-align: center;">Test Result</td> </tr> <tr> <td>0001</td> <td>11/20/13</td> <td>02:02</td> <td>.155</td> </tr> <tr> <td>0002</td> <td>11/20/13</td> <td>02:04</td> <td>.215</td> </tr> <tr> <td>0003</td> <td>11/20/13</td> <td>02:07</td> <td>.220</td> </tr> <tr> <td>0004</td> <td>11/21/13</td> <td>14:01</td> <td>.102</td> </tr> </table> <p style="text-align: center;">Push No. 3 to erase, or power down to retain</p>														11/21/13 14:01	AS4 Serial no: 018684	Version no: 4797	No. Date Time %BAC				Calibration 11/03/11 16:03 .100	Test Result				0001	11/20/13	02:02	.155	0002	11/20/13	02:04	.215	0003	11/20/13	02:07	.220	0004	11/21/13	14:01	.102
11/21/13 14:01	AS4 Serial no: 018684	Version no: 4797	No. Date Time %BAC																																						
			Calibration 11/03/11 16:03 .100																																						
Test Result																																									
0001	11/20/13	02:02	.155																																						
0002	11/20/13	02:04	.215																																						
0003	11/20/13	02:07	.220																																						
0004	11/21/13	14:01	.102																																						

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1403	34	13B06-16-5-10- KKF24C	.100	20		.098			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 14:04
 AS4 Serial no: 008635
 Version no: 4797

No. Date Time %BAC
 Calibration 10/24/03 09:04 .100

Test Result
 0001 11/19/13 01:49 .122
 0002 11/19/13 01:51 .106
 0003 11/21/13 14:04 .098

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1409	34	13B06-16-5-10- KKF24C	.100	20		.102			X	Lee, Kirk	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/21/13 14:10
 AS4 Serial no: 008845
 Version no: 4797
 No. Date Time %BAC
 Calibration 12/24/03 11:01 .100
 Test Result
 0001 11/14/13 21:16 .000
 0002 11/19/13 22:52 .000
 0003 11/21/13 14:10 .102
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1411	34	13B06-16-5-10- KKF24C	.100	20		.093			X	Lee, Kirk	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/21/13 14:12
 AS4 Serial no: 082653
 Version no: 4797
 No. Date Time %BAC
 Calibration 12/01/11 07:44 .081
 Test Result
 0001 11/21/13 14:12 .093
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 003348
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1407	34	13B06-16-5-10-KKF24C	.100	20		.097			X	Lee, Kirk	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

```

11/21/13 14:08
AS4 Serial no: 008848
Version no: 4797

No. Date Time %BAC
Calibration 01/13/03 09:25 .100

Test Result
0001 11/17/13 08:46 .113
0002 11/17/13 08:48 .115
0003 11/21/13 14:08 .097

Push No. 3 to erase,
or power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1415	34	13B06-16-5-10- KKF24C	.100	19		.096			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 14:16
AS4 Serial no: 077591
Version no: 4797
No. Date Time %BAC
Calibration
05/21/08 15:09 .100
Test Result
0001 11/21/13 14:15 .096
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer Brian Flick	15063	Intoximeters Inc./Alco-Sensor IV	008850

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1417	34	13B06-16-5-10- KKF24C	.100	20		.098			X		Lee, Kirk	17046	

Prepared by:
 Douglas C. Talbot
 02-06-2013
 Expires on:
 02-06-2015

```

11/21/13 14:19
AS4 Serial no: 008850
Version no: 4797
No. Date Time %PAC
Calibration 03/02/06 09:07 .100
Test Result
0001 11/21/13 14:18 .098

Push No. 3 to erase,
or power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1419	34	13B06-16-5-10-KKF24C	.100	21		.095			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 14:20
 #64 Serial no: 077573
 Version no: 4797
 No. Date Time %BAC
 Calibration 05/21/08 14:43 .100
 Test Result
 0001 11/18/13 22:01 .000
 0002 11/19/13 20:22 .145
 0003 11/19/13 20:25 .154
 0004 11/19/13 22:15 .191
 0005 11/19/13 22:17 .188
 0006 11/21/13 00:03 .130
 0007 11/21/13 00:05 .132
 0008 11/21/13 00:15 .073
 0009 11/21/13 14:20 .095
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	14:22	34	13B06-16-5-10- KKF24C	.100	21		.097			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 14:23
 AS4 Serial no: 083051
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/25/13 14:00 .100
 Test Result 0001 11/21/13 14:22 .097
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	---	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1425	34	13B06-16-5-10- KKF24C	.100	20		.106			X		Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

```

11/21/13 14:26
AS4 Serial no: 008897
Version no: 4797

No. Date Time %MAC
Calibration
01/18/12 16:13 .100

Test Result
0001 11/16/13 02:28 .053
0002 11/16/13 02:30 .050
0003 11/17/13 02:27 .080
0004 11/17/13 02:30 .070
0005 11/17/13 02:41 .000
0006 11/21/13 14:25 .106

Push No. 3 to erase,
or power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18624
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013			13B06-16-5-10- KKF24C	.100			B.O					Lee, Kirk	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18646
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																															
11/21/2013	1442	34	13B06-16-5-10- KKF24C	.100	20		.099			X	Flick	15063	<i>[Signature]</i>																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																
<table border="0" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> </tr> <tr> <td>Prepared by:</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>Douglas C. Talbot</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>02-06-2013</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>Expires on:</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>02-06-2015</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> </table>																												Prepared by:														Douglas C. Talbot														02-06-2013														Expires on:														02-06-2015																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																									
Prepared by:																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
Douglas C. Talbot																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
02-06-2013																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
Expires on:																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
02-06-2015																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
<table border="0" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 30%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> <td style="width: 10%;"></td> </tr> <tr> <td>Test Result</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0001 09/18/13 20:07 .097</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0002 09/19/13 01:24 .006</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0003 09/20/13 01:28 .062</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0004 09/20/13 01:30 .061</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0005 09/20/13 02:48 .098</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0006 09/20/13 02:50 .096</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0007 09/20/13 19:52 .071</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0008 09/20/13 19:53 .143</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0009 09/24/13 22:30 .040</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0010 09/24/13 22:32 .127</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0011 09/24/13 22:32 .060</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0012 09/24/13 22:33 .023</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0013 09/24/13 22:34 .032</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0014 09/24/13 22:35 .042</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0015 09/24/13 22:37 .094</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0016 09/24/13 23:05 .040</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0017 09/24/13 23:06 .030</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0018 09/24/13 23:08 .045</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0019 09/24/13 23:10 .020</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0020 09/24/13 23:10 .049</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0021 09/24/13 23:11 .045</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0022 09/24/13 23:14 .177</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0023 10/02/13 01:09 .036</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0024 10/02/13 02:09 .033</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0025 10/02/13 03:15 .183</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0026 10/02/13 03:19 .153</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0027 10/02/13 03:22 .177</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0028 10/03/13 12:38 .099</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0029 10/11/13 14:42 .098</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0030 10/15/13 02:36 .046</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0031 10/18/13 13:39 .093</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0032 10/22/13 00:47 .000</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0033 10/30/13 01:45 .061</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0034 10/31/13 15:09 .098</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0035 11/05/13 02:36 .084</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0036 11/05/13 02:40 .091</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0037 11/05/13 02:46 .058</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0038 11/12/13 04:53 .120</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0039 11/12/13 04:55 .137</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0040 11/19/13 08:11 .187</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0041 11/19/13 08:15 .176</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0042 11/20/13 02:36 .073</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0043 11/20/13 02:39 .057</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0044 11/20/13 02:40 .064</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0045 11/21/13 02:53 .037</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0046 11/21/13 15:41 .099</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> </table>																												Test Result														0001 09/18/13 20:07 .097														0002 09/19/13 01:24 .006														0003 09/20/13 01:28 .062														0004 09/20/13 01:30 .061														0005 09/20/13 02:48 .098														0006 09/20/13 02:50 .096														0007 09/20/13 19:52 .071														0008 09/20/13 19:53 .143														0009 09/24/13 22:30 .040														0010 09/24/13 22:32 .127														0011 09/24/13 22:32 .060														0012 09/24/13 22:33 .023														0013 09/24/13 22:34 .032														0014 09/24/13 22:35 .042														0015 09/24/13 22:37 .094														0016 09/24/13 23:05 .040														0017 09/24/13 23:06 .030														0018 09/24/13 23:08 .045														0019 09/24/13 23:10 .020														0020 09/24/13 23:10 .049														0021 09/24/13 23:11 .045														0022 09/24/13 23:14 .177														0023 10/02/13 01:09 .036														0024 10/02/13 02:09 .033														0025 10/02/13 03:15 .183														0026 10/02/13 03:19 .153														0027 10/02/13 03:22 .177														0028 10/03/13 12:38 .099														0029 10/11/13 14:42 .098														0030 10/15/13 02:36 .046														0031 10/18/13 13:39 .093														0032 10/22/13 00:47 .000														0033 10/30/13 01:45 .061														0034 10/31/13 15:09 .098														0035 11/05/13 02:36 .084														0036 11/05/13 02:40 .091														0037 11/05/13 02:46 .058														0038 11/12/13 04:53 .120														0039 11/12/13 04:55 .137														0040 11/19/13 08:11 .187														0041 11/19/13 08:15 .176														0042 11/20/13 02:36 .073														0043 11/20/13 02:39 .057														0044 11/20/13 02:40 .064														0045 11/21/13 02:53 .037														0046 11/21/13 15:41 .099													
Test Result																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0001 09/18/13 20:07 .097																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0002 09/19/13 01:24 .006																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0003 09/20/13 01:28 .062																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0004 09/20/13 01:30 .061																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0005 09/20/13 02:48 .098																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0006 09/20/13 02:50 .096																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0007 09/20/13 19:52 .071																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0008 09/20/13 19:53 .143																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0009 09/24/13 22:30 .040																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0010 09/24/13 22:32 .127																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0011 09/24/13 22:32 .060																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0012 09/24/13 22:33 .023																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0013 09/24/13 22:34 .032																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0014 09/24/13 22:35 .042																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0015 09/24/13 22:37 .094																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0016 09/24/13 23:05 .040																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0017 09/24/13 23:06 .030																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0018 09/24/13 23:08 .045																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0019 09/24/13 23:10 .020																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0020 09/24/13 23:10 .049																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0021 09/24/13 23:11 .045																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0022 09/24/13 23:14 .177																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0023 10/02/13 01:09 .036																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0024 10/02/13 02:09 .033																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0025 10/02/13 03:15 .183																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0026 10/02/13 03:19 .153																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0027 10/02/13 03:22 .177																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0028 10/03/13 12:38 .099																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0029 10/11/13 14:42 .098																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0030 10/15/13 02:36 .046																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0031 10/18/13 13:39 .093																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0032 10/22/13 00:47 .000																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0033 10/30/13 01:45 .061																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0034 10/31/13 15:09 .098																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0035 11/05/13 02:36 .084																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0036 11/05/13 02:40 .091																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0037 11/05/13 02:46 .058																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0038 11/12/13 04:53 .120																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0039 11/12/13 04:55 .137																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0040 11/19/13 08:11 .187																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0041 11/19/13 08:15 .176																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0042 11/20/13 02:36 .073																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0043 11/20/13 02:39 .057																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0044 11/20/13 02:40 .064																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0045 11/21/13 02:53 .037																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													
0046 11/21/13 15:41 .099																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																																													

Push No. 3 to erase, or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77488
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1439	34	13B06-16-5-10- KKF24C	.100	20	.102				X	Flick	15063	
			Prepared by: Douglas C. Talbot 02-06-2013										
			Expires on: 02-06-2015										

11/21/13 14:40
AS4 Serial no: 077488
Version no: 4797
No. Date Time %BAC
Calibration
05/21/08 14:35 .100
Test Result
0001 11/21/13 14:39 .102
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8877
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1440	34	13B06-16-5-10- KKF24C	.100	20		103			X	Flick	15063	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

No. Date Time ZBAC
 Calibration 05/26/05 10:40 .100
 Test Result
 0001 11/21/13 14:41 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18615
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	144234		13B06-16-5-10- KKF24C	.100	21		.098				Flick	15063	
			Prepared by: Douglas C. Talbot										
			Expires on: 02-06-2015										

11/21/13 14:43
 AS4 Serial no: 018615
 Version no: 4797
 No. Date Time %BAC
 Calibration 12/19/08 14:44 .100
 Test Result 0001 11/21/13 14:43 .098
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18626
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1442	34	13B06-16-5-10-KKF24C	.100	21		098				Y	Flick	15063	
Prepared by:														
Douglas C. Talbot														
02-06-2013														
Expires on:														
02-06-2015														
<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p>11/21/13 14:43 AS4 Serial no: 018626 Version no: 4797</p> <p>No. Date Time 2BAC Calibration 04/29/05 08:59 .101</p> <p>Test Result 0001 11/21/13 14:43 .098</p> <p>Push No. 3 to erase, or power down to retain</p> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 12696
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013			13B06-16-5-10-KKF24C	.100								Flick	15063	

B.O.

Prepared by:
Douglas C. Talbot
02-06-2013
Expires on:
02-06-2015

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83230
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1444	34	13B06-16-5-10- KKF24C	.100	20		1099				Flick	15063	
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015													
11/21/13 14:45 HSA Serial no: 083030 Version no: 4797 No. Date Time %BAC Calibration 01/29/09 10:39 .100 Test Result 0001 11/21/13 14:44 .099 Push No. 3 to erase, or Power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18692
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1258	34	13B06-16-5-10-KKF24C	.100	21		.099				✓	Flick	15063	

Prepared by:
Douglas C. Talbot
02-06-2013
Expires on:
02-06-2015

11/21/13 12:59
 AS4 Serial no: 018692
 Version no: 4797
 No. Date Time %BAC
 Calibration 08/12/08 14:35 .100
 Test Result 0001 11/21/13 12:59 .099
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18762
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1447	34	13B06-16-5-10- KKF24C	.100	20		.100			0	Flick	15063	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

11/21/13 14:48
 HS4 Serial no: 018762
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/30/09 15:52 .100
 Test Result
 0001 11/18/13 23:06 .086
 0002 11/18/13 23:08 .084
 0003 11/21/13 14:48 .100
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18766
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1452	34	13B06-16-5-10-KKF24C	.100	21		090					Flick	15063	
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Push No. 3 to erase, or power down to retain </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 11/16/13 08:17 .154 0002 11/16/13 08:21 .145 0003 11/21/13 14:53 .090 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time ZBAC Calibration 09/12/12 12:30 .100 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> 11/21/13 14:53 AS4 Serial no: 018766 Version no: 4797 </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18691
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	12:51	34	13B06-16-5-10- KKF24C	.100	21		.100			0		Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

No. Date Time %BAC
 Calibration 09/08/09 14:01 .100

 Test Result
 0001 11/17/13 01:56 .181
 0002 11/17/13 01:58 .142
 0003 11/17/13 02:00 .175
 0004 11/17/13 20:44 .048
 0005 11/17/13 20:46 .046
 0006 11/21/13 14:51 .100

 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 17990
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013			13B06-16-5-10- KKF24C	.100			80					Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18774
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	1454	34	13B06-16-5-10-KKF24C	.100	21		.097			X	Flick	15063	
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015													
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Push No. 3 to erase, or power down to retain </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 11/14/13 23:51 .000 0002 11/20/13 23:31 .184 0003 11/20/13 23:34 .140 0004 11/20/13 23:37 .145 0005 11/21/13 14:55 .097 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time %BAC Calibration 01/13/09 10:25 .100 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> 11/21/13 14:55 AS4 Serial no: 018774 Version no: 4797 </div> </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41486
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1456	34	13B06-16-5-10- KKF24C	.100	21		4102				X	Flick	15063	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>														
<p>11/21/13 14:57 ASA Serial no: 041486 Version no: 4797</p> <p>No. Date Time %BAC Calibration 03/08/06 13:56 .100</p> <p>Test Result 0001 11/21/13 14:56 .102</p> <p>Push No. 3 to erase, or power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41502
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
11/21/2013	14:59	34	13B06-16-5-10-KKF24C	.100	20		0.097			X	Flick	15063	
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
Test Result 0001 11/16/13 20:52 .000 0002 11/17/13 00:35 .274 0003 11/17/13 00:37 .259 0004 11/17/13 02:47 .052 0005 11/17/13 02:50 .000 0006 11/17/13 02:50 .000 0007 11/17/13 03:27 .090 0008 11/17/13 03:29 .093 0009 11/17/13 18:59 .222 0010 11/17/13 19:02 .213 0011 11/21/13 15:00 .097 No. Date Time ?BAC Calibration 05/05/09 09:03 .100 11/21/13 15:00 AS4 Serial no: 041502 Version no: 4797													
Push No. 3 to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82654
--------------------------------	-----------------	---	----------------------	--	-------------------------

53

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
11/21/2013	1502	34	13B06-16-5-10-KKF24C	.100	21		094			Y		Flick	15063	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

11/21/13 15:03
 AS4 Serial no: 082654
 Version no: 4797
 No. Date Time ZBAC
 Calibration 04/19/10 09:27 .081
 Test Result 0001 11/21/13 15:02 .094
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

507's PAS

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. <i>8874</i>
--------------------------------	-----------------	---	----------------------	--	-------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
11/21/2013	<i>1237</i>	<i>34</i>	13B06-16-5-10- KKF24C	.100	<i>21</i>		<i>.099</i>			<input checked="" type="checkbox"/>		Flick	15063		
			Prepared by:												
			Douglas C. Talbot												
			02-06-2013												
			Expires on:												
			02-06-2015												

11/21/13 12:39
AS4 Serial no: 008874
Version no: 4797

No. Date Time %RAC
Calibration
06/25/13 10:32 .107

Test Result
0001 11/21/13 12:38 .099

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.