

Sgt.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8874
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2013	1134	34	13B06-16-5-10-KKF24C	.100	19		.102			X		Valadez	16071	
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
No. Date Time %BAC Calibration 06/25/13 10:32 .107 Test Result 0001 12/04/13 11:35 .102														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8501
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)	YES
12/04/2013	1139	34	13B06-16-5-10- KKF24C	.100	18		.097			X		Valadez	16071	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>														
<pre> 12/04/13 11:40 NS4 Serial no: 008501 Version no: 4797 No. Date Time XBAC Calibration 03/30/09 15:40 .100 Test Result 0001 11/30/13 21:41 .122 0002 11/30/13 21:43 .123 0003 12/01/13 17:13 .085 0004 12/01/13 17:16 .081 0005 12/04/13 11:40 .097 </pre>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.11%).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18785
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																								
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																																							
12/04/2013	1254	34	13B06-16-5-10- KKF24C	.100	25		.103			X	Valadez	16071																																									
Prepared by: Douglas C. Talbot 02-06-2013																																																					
Expires on: 02-06-2015																																																					
<table border="1"> <tr> <td>Test Result</td> <td>No.</td> <td>Date</td> <td>Time</td> <td>2BAC</td> </tr> <tr> <td>0001 11/28/13 06:39 .086</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0002 11/28/13 06:41 .092</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0003 12/03/13 08:59 .103</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0004 12/03/13 09:00 .102</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>0005 12/04/13 12:54 .103</td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td colspan="5">Calibration</td> </tr> <tr> <td colspan="5">07/18/08 13:27 .100</td> </tr> </table>														Test Result	No.	Date	Time	2BAC	0001 11/28/13 06:39 .086					0002 11/28/13 06:41 .092					0003 12/03/13 08:59 .103					0004 12/03/13 09:00 .102					0005 12/04/13 12:54 .103					Calibration					07/18/08 13:27 .100				
Test Result	No.	Date	Time	2BAC																																																	
0001 11/28/13 06:39 .086																																																					
0002 11/28/13 06:41 .092																																																					
0003 12/03/13 08:59 .103																																																					
0004 12/03/13 09:00 .102																																																					
0005 12/04/13 12:54 .103																																																					
Calibration																																																					
07/18/08 13:27 .100																																																					
<table border="1"> <tr> <td>12/04/13 12:55</td> <td>RS4 Serial no:</td> <td>018785</td> </tr> <tr> <td>Version</td> <td>no:</td> <td>4797</td> </tr> </table>														12/04/13 12:55	RS4 Serial no:	018785	Version	no:	4797																																		
12/04/13 12:55	RS4 Serial no:	018785																																																			
Version	no:	4797																																																			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82657
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2013	1249	34	13B06-16-5-10- KKF24C	.100	19		.098				X	Valadez	16071	
Prepared by: Douglas C. Talbot 02-06-2013														
Expires on: 02-06-2015														
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> 12/04/13 12:50 NS4 Serial no: 082657 Version no: 4797 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time ZBAC Calibration 10/07/11 12:22 .081 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 12/04/13 12:50 .098 </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18684
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013	130234	34	13B06-16-5-10- KKF24C	.100	20		.104			X	Valadez	16071	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

12/04/13 13:02
 NS4 Serial no: 018684
 Version no: 4797
 No. Date Time %BHC
 Calibration 11/03/11 16:03 .100
 Test Result
 0001 12/04/13 13:02 .104

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82653
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
12/04/2013	1451	34	13B06-16-5-10- KKF24C	.100	19		.096				X	Valadez	16071	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

12/04/13 14:52
 HS4 Serial no: 082653
 Version no: 4797
 No. Date Time ZBAC
 Calibration 12/01/11 07:44 .081
 Test Result
 0001 12/04/13 14:52 .096

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8848
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013	1456	34	13B06-16-5-10- KKF24C	.100	19		.103			X	Valadez	16071	
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 11/30/13 02:33 .213 0002 11/30/13 02:37 .205 0003 12/04/13 14:57 .100 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time %BAC Calibration 01/13/03 09:25 .100 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> 12/04/13 14:58 #54 Serial no: 008848 Version no: 4797 </div> </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 8850
--------------------------------	-----------------	---	----------------------	--	------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																							
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO																											
12/04/2013	1502	34	13B06-16-5-10-KKF24C	.100	19		.102			X		Valadez	16071																								
Prepared by: Douglas C. Talbot 02-06-2013																																					
Expires on: 02-06-2015																																					
<table border="0" style="width: 100%; text-align: right;"> <tr> <td>12/04/13 15:03</td> <td>RS4 Serial no: 008850</td> <td>Version no: 4797</td> <td>No. Date Time ZBAC</td> <td>Calibration</td> <td>03/02/06 09:07 .100</td> </tr> <tr> <td>Test Result</td> <td>0001 11/29/13 02:57 .056</td> <td>0002 11/29/13 03:31 .049</td> <td>0003 11/29/13 05:09 .136</td> <td>0004 11/29/13 05:33 .175</td> <td>0005 11/29/13 05:56 .191</td> </tr> <tr> <td></td> <td>0006 11/29/13 06:02 .120</td> <td>0007 11/29/13 22:01 .000</td> <td>0008 11/29/13 22:02 .000</td> <td>0009 12/01/13 01:29 .131</td> <td>0010 12/03/13 00:35 .087</td> </tr> <tr> <td></td> <td>0011 12/03/13 01:33 .053</td> <td>0012 12/03/13 01:54 .000</td> <td>0013 12/04/13 15:03 .102</td> <td></td> <td></td> </tr> </table>														12/04/13 15:03	RS4 Serial no: 008850	Version no: 4797	No. Date Time ZBAC	Calibration	03/02/06 09:07 .100	Test Result	0001 11/29/13 02:57 .056	0002 11/29/13 03:31 .049	0003 11/29/13 05:09 .136	0004 11/29/13 05:33 .175	0005 11/29/13 05:56 .191		0006 11/29/13 06:02 .120	0007 11/29/13 22:01 .000	0008 11/29/13 22:02 .000	0009 12/01/13 01:29 .131	0010 12/03/13 00:35 .087		0011 12/03/13 01:33 .053	0012 12/03/13 01:54 .000	0013 12/04/13 15:03 .102		
12/04/13 15:03	RS4 Serial no: 008850	Version no: 4797	No. Date Time ZBAC	Calibration	03/02/06 09:07 .100																																
Test Result	0001 11/29/13 02:57 .056	0002 11/29/13 03:31 .049	0003 11/29/13 05:09 .136	0004 11/29/13 05:33 .175	0005 11/29/13 05:56 .191																																
	0006 11/29/13 06:02 .120	0007 11/29/13 22:01 .000	0008 11/29/13 22:02 .000	0009 12/01/13 01:29 .131	0010 12/03/13 00:35 .087																																
	0011 12/03/13 01:33 .053	0012 12/03/13 01:54 .000	0013 12/04/13 15:03 .102																																		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 83051
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
12/04/2013	1506	34	13B06-16-5-10-KKF24C	.100	20		.100				K	Valadez	16071	
			Prepared by:											
			Douglas C. Talbot											
			02-06-2013											
			Expires on:											
			02-06-2015											

12/04/13 15:07
 MS4 Serial no: 083051
 Version no: 4797
 No. Date Time ZBAC
 Calibration 03/25/13 14:00 .100
 Test Result 0001 12/04/13 15:06 .100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. L8624
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013			13B06-16-5-10- KKF24C	.100							Valadez	16071	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 77488
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013	1527	34	13B06-16-5-10- KKF24C	.100	20		.102			X	Valadez	16071	
Prepared by: Douglas C. Talbot 02-06-2013 Expires on: 02-06-2015													
Test Result 0901 12/04/13 15:27 .102 No. Date Time %BAC Calibration 05/21/08 14:35 .100 Version no: 4797 454 Serial no: 077488 12/04/13 15:28													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18626
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013	1535	34	13B06-16-5-10- KKF24C	.100	21		.102			X	Valadez	16071	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										
										12/04/13 15:37 PAS4 Serial no: 018626 Version no: 4797 No. Date Time %BAC Calibration 04/29/05 08:59 .101 Test Result 0001 12/04/13 15:35 .102			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18696
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013			13B06-16-5-10- KKF24C	.100							Valadez	16071	
			Prepared by:										
			Douglas C. Talbot										
			02-06-2013										
			Expires on:										
			02-06-2015										

B.O.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 18766
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/04/2013	15:48	34	13B06-16-5-10- KKF24C	.100	20		.097				X	Valadez	16071		
			Prepared by: Douglas C. Talbot 02-06-2013												
			Expires on: 02-06-2015												
			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> 12/04/13 15:50 AS4 Serial no: 018766 Version no: 4797 No. Date Time 2BAC Calibration 03/12/12 12:30 .100 Test Result 0001 12/04/13 15:49 .097 </div>												

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 41502
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/04/2013	1601	34	13B06-16-5-10- KKF24C	.100	20		.097			X		Valadez	16071	
<p>Prepared by: Douglas C. Talbot 02-06-2013</p> <p>Expires on: 02-06-2015</p>														
<p>12/04/13 16:02 AS4 Serial no: 041502 Version no: 4797</p> <p>No. Date Time %BAC Calibration 05/05/09 08:03 .100</p> <p>Test Result 0001 11/30/13 01:23 .000 0002 11/30/13 02:34 .144 0003 11/30/13 02:38 .140 0004 11/30/13 03:33 .113 0005 11/30/13 03:32 .118 0006 11/30/13 21:50 .203 0007 12/02/13 02:40 .000 0008 12/02/13 02:41 .000 0009 12/02/13 03:49 .181 0010 12/02/13 03:53 .183 0011 12/04/13 16:02 .097</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer Brian Flick	I.D. NUMBER 15063	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 82654
--------------------------------	-----------------	---	----------------------	--	-------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/04/2013	1605	34	13B06-16-5-10- KKF24C	.100	21		.093			X	Valadez	16071	
Prepared by: Douglas C. Talbot 02-06-2013													
Expires on: 02-06-2015													
12/04/13 16:05 H54 Serial no: 082654 Version no: 4797 No. Date Time %IAC Calibration 04/19/10 09:27 .081 Test Result 0001 12/04/13 16:05 .093													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.