

5675

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1449	34	13B06-16-5-10- KKF24C	.100	21		103					Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

12/19/13 14:50
AS4 Serial no: 008874
Version no: 4797

No. Date Time ZRAC
Calibration
06/25/13 10:32 .107

Test Result
0001 12/19/13 01:17 .847
0002 12/19/13 01:18 .845
0003 12/19/13 02:58 .135
0004 12/19/13 03:00 .138
0005 12/19/13 14:49 .103

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

1

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	008501

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1217	34	13B06-16-5-10-KKF24C	.100	20		.093			X		Lee	17046	

Prepared by:
 Douglas C. Talbot
 02-06-13
 Expires on:
 02-06-15

12/19/13 12:17
 #54 Serial no: 008501
 Version no: 4797

No. Date Time %BAC
 Calibration 03/30/09 15:40 .100

Test Result
 0001 12/13/13 21:20 .023
 0002 12/13/13 21:55 .149
 0003 12/13/13 21:57 .142
 0004 12/19/13 12:16 .093

Push No. 3 to erase or
 power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018 785
--------------------------------	-----------------	------------------------------------	----------------------	--	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	12:21	34	13B06-16-5-10- KKF24C	.100	20		.097			X		Lee	17046	
			Prepared by: Douglas C. Talbot 02-06-13											
			Expires on: 02-06-15											

12/19/13 12:20
AS4 Serial no: 018785
Version no: 4797

No. Date Time %BAC
Calibration
07/18/08 13:27 .100

Test Result
0001 12/19/13 12:19 .097

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008546
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1224	34	13B06-16-5-10- KKF24C	.100	21		.102			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 12:23
MS4 Serial no: 008546
Version no: 4797

No. Date Time %BAC
Calibration
06/29/13 16:36 .100
Test Result
0001 12/15/13 02:21 .168
0002 12/15/13 02:22 .169
0003 12/19/13 12:22 .102

Push No. 3 to erase
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082659
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1227	34	13B06-16-5-10-KKF24C	.100	20		.104			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

Push No. 3 to erase,
or power down to retain

Test Result
0001 12/19/13 12:25 .104

No. Date Time %RQA
Calibration
18/22/08 17:37 .100

12/19/13 12:26
AS4 Serial no: 082659
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1229	34	13B06-16-5-10- KKF24C	.100	20		.103			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

12/19/13 12:29
 NS4 Serial no: M08559
 Version no: 4797
 No. Date Time ZBAC
 Calibration 11/30/11 08:55 .100
 Test Result
 0001 12/19/13 12:29 .103
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008580
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1232	34	13B06-16-5-10- KKF24C	.100	20	103			X		Lee	17046	
			Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15										

Push No. 3 to erase,
or power down to retain

Test Result
 0001 12/16/13 22:57 .107
 0002 12/16/13 23:01 .095
 0003 12/17/13 02:18 .127
 0004 12/17/13 02:21 .117
 0005 12/18/13 03:25 .027
 0006 12/19/13 00:57 .142
 0007 12/19/13 01:01 .114
 0008 12/19/13 01:03 .118
 0009 12/19/13 12:31 .103

No. Date Time ZBAC
 Calibration
 12/01/06 08:14 .099
 12/19/13 12:32
 H54 Serial no: 008580
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	083032

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1235	34	13B06-16-5-10- KKF24C	.100	20		.100		X			Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

```

12/19/13 12:33
AS4 Serial no: 083032
Version no: 4797
  
```

```

No. Date Time %BAC
Calibration
12/01/11 07:38 .081
  
```

```

Test Result
0001 12/13/13 23:42 .101
0002 12/13/13 23:44 .090
0003 12/15/13 03:24 .095
0004 12/15/13 03:27 .093
0005 12/19/13 12:32 .100
  
```

```

Push No. 3 to erase,
or power down to retain
  
```

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1237	34	13B06-16-5-10- KKF24C	.100	20		101			X	Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

12/19/13 12:39
HS4 Serial no: 008620
Version no: 4797

No. Date Time %BAC
Calibration
01/30/13 14:34 .100

Test Result
0001 12/19/13 12:39 .101

Push No. 3 to erase
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1241	34	13B06-16-5-10- KKF24C	.100	21		.097			X	Lee	17046	
			Prepared by: Douglas C. Talbot										
			02-06-13										
			Expires on: 02-06-15										

12/19/13 12:43
 #54 Serial no: 082957
 Version no: 4797
 No. Date Time %BAC
 Calibration 01/24/13 06:06 .092
 Test Result
 0601 12/19/13 12:39 .097
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1247	34	13B06-16-5-10- KKF24C	.100	20		.099			X		Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

12/19/13 12:45
AS4 Serial no: 077505
Version no: 4797

No. Date Time %BAC
Calibration 05/21/08 13:16 .100

Test Result
0001 12/19/13 12:45 .099

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013			13B06-16-5-10- KKF24C	.100							Lee	17046	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 0B2660
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1253	34	13B06-16-5-10-	.100	20		.097				X	Lee	17046	
			KKF24C											
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

```

12/19/13 12:52
AS4 Serial no: 0B2660
Version no: 4797
No. Date Time %BAC
Calibration
11/18/10 06:21 .081

Test Result
0001 12/14/13 01:53 .059
0002 12/14/13 01:56 .053
0003 12/14/13 02:25 .158
0004 12/14/13 02:27 .151
0005 12/14/13 02:49 .152
0006 12/14/13 02:51 .160
0007 12/19/13 12:51 .097

Push No. 3 to erase,
or Power down to retain

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1256	34	13B06-16-5-10- KKF24C	.100	20		106			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Push No. 3 to erase:
 No. Date Time %BAC
 Calibration 05/28/04 11:06 .096
 Test Result
 0001 12/14/13 23:04 .182
 0002 12/14/13 23:08 .187
 0003 12/19/13 12:55 .106

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

14

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS AGCURITY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1300	34	13B06-16-5-10-KKF24C	.100	21	.105				X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

12/19/13 13:01
 AS4 Serial no: 008214
 Version no: 4797
 No. Date Time ZMAC
 Calibration 03/11/04 14:05 .100
 Test Result 0091 12/19/13 13:00 .105
 Push No. 3 to erase,
 or Power down to retain
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#16

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1303	34	13B06-16-5-10- KKF24C	.100	21	.102			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Push No. 3 to erase
or Power down to retain

```

Test Result
0001 12/14/13 00:48 .172
0002 12/14/13 00:52 .180
0003 12/14/13 03:51 .130
0004 12/14/13 03:54 .123
0005 12/15/13 02:25 .000
0006 12/15/13 02:52 .127
0007 12/15/13 02:55 .134
0008 12/19/13 13:00 .102

```

```

No. Date Time ZBAC
Calibration 03/25/13 10:45 .100
12/19/13 13:00
MS4 Serial no: 008888
Version no: 479?

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1306	34	13B06-16-5-10- KKF24C	.100	20	.102				X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

```

12/19/13 13:05
AS4 Serial no: 008783
Version no: 4797

```

```

No. Date Time %BAC
Calibration
02/18/08 13:52 .100

```

```

Test Result
0001 12/17/13 00:26 .076
0002 12/17/13 00:29 .073
0003 12/19/13 01:40 .128
0004 12/19/13 01:42 .129
0005 12/19/13 13:05 .102

```

Push No. 3 to erase, or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

18

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1308	34	13B06-16-5-10- KKF24C	.100	20		.097				X	Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

12/19/13 13:09
 AS4 Serial no: 082657
 Version no: 4797
 No. Date Time ZBAC
 Calibration 18/07/11 12:22 .081
 Test Result
 0001 12/19/13 13:08 .097
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1311	34	13B06-16-5-10- KKF24C	.100	20		.098				X	Lee	17046		

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

12/19/13 13:10
#54 Serial no: 008810
Version no: 4797
No. Date Time ZRAC
Calibration 10/30/09 09:53 .100
Test Result
0001 12/17/13 01:18 .077
0002 12/17/13 01:21 .078
0003 12/17/13 02:07 .000
0004 12/17/13 18:22 .000
0005 12/18/13 19:33 .158
0006 12/18/13 19:35 .160
0007 12/19/13 13:10 .098
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1315	34	13B06-16-5-10- KKF24C	.100	20	.103				X	Lee	17046	CHANGED BATTERY
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

12/19/13 13:16
 PAS4 Serial no: 010884
 Version no: 4797

No. Date Time ZBAC
 Calibration 11/03/11 16:03 .100

Test Result
 0001 12/16/13 13:15 .000
 0002 12/16/13 14:29 .000
 0003 12/16/13 16:25 .000
 0004 12/17/13 12:33 .000
 0005 12/17/13 18:57 .000
 0006 12/18/13 14:45 .000
 0007 12/19/13 13:15 .103

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1319	34	13B06-16-5-10- KKF24C	.100	21	.102			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

12/19/13 13:19
 #54 Serial no: 008635
 Version no: 4797

No. Date Time %MHC
 Calibration
 10/24/03 09:04 .100

Test Result
 0001 12/18/13 03:39 .093
 0002 12/18/13 03:42 .081
 0003 12/19/13 01:40 .116
 0004 12/19/13 01:43 .114
 0005 12/19/13 13:10 .102

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008845
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1323	34	13B06-16-5-10-KKF24C	.100	20		.106			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Test Result
 0001 12/17/13 22:37 .000
 0002 12/18/13 08:41 .217
 0003 12/18/13 08:43 .000
 0004 12/18/13 08:44 .201
 0005 12/19/13 13:22 .106
 Calibration
 12/24/09 11:01 .100
 No. Date Time ZBAC
 12/19/13 13:22
 AS4 Serial no: 0008845
 Version no: 4797
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1326	34	13B06-16-5-10-KKF24C	.100	20		.098			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

12/19/13 13:24
MS4 Serial no: 082653
Version no: 4797

No. Date Time %RAC
Calibration 12/01/11 07:44 .081

Test Result
0001 12/19/13 13:23 .098

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1333	34	13B06-16-5-10- KKF24C	.100	20		.101			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 13:33
 AS4 Serial no: 008848
 Version no: 4797
 No. Date Time %RAC
 Calibration 01/13/03 09:25 .100
 Test Result
 0001 12/15/13 21:25 .211
 0002 12/15/13 21:27 .208
 0003 12/16/13 02:17 .130
 0004 12/16/13 02:19 .128
 0005 12/19/13 13:32 .101
 Push No. 3 to erase, or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1336	34	13B06-16-5-10- KKF24C	.100	19		.101				X	Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

12/19/13 13:35
 AS4 Serial no: 077591
 Version no: 4797
 No. Date Time ZBAC
 Calibration 05/21/08 15:09 .100
 Test Result
 0001 12/14/13 12:58 .000
 0002 12/15/13 18:16 .155
 0003 12/15/13 18:19 .153
 0004 12/19/13 13:34 .101
 Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1350	34	13B06-16-5-10- KKF24C	.100	21		.098		X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Test Result
 0001 12/16/13 21:55 .000
 0002 12/17/13 21:58 .105
 0003 12/17/13 22:00 .111
 0004 12/18/13 02:19 .140
 0005 12/18/13 02:21 .149
 0006 12/18/13 21:46 .000
 0007 12/19/13 13:50 .098
 Calibration
 03/02/06 09:07 .100
 Ho. Date Time XPAC
 Calibration
 03/02/06 09:07 .100
 12/19/13 13:51
 H54 Serial no: 008850
 Version no: 4797
 or Power down to retain

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1352	34	13B06-16-5-10-KKF24C	.100	21		.098			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 13:51
AS4 Serial no: 077573
Version no: 4797

No. Date Time ZHAC
Calibration 05/21/08 14:43 .100

Test Result
0001 12/18/13 02:08 .104
0002 12/18/13 02:10 .103
0003 12/19/13 13:50 .098

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1355	34	13B06-16-5-10- KKF24C	.100	2		.098			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

12/19/13 13:54
 NS4 Serial no: 083051
 Version no: 4797
 No. Date Time %BAC
 Calibration 08/25/13 14:00 .100
 Test Result 0001 12/19/13 13:53 .098
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1358	34	13B06-16-5-10- KKF24C	.100	21	.104				X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 13:58
AS4 Serial no: 008897
Version no: 4797

No. Date Time %BAC
Calibration 01/18/12 16:13 .102

Test Result
0001 12/14/13 02:17 .079
0002 12/14/13 02:20 .093
0003 12/14/13 02:23 .074
0004 12/14/13 02:25 .076
0005 12/14/13 04:33 .177
0006 12/14/13 04:36 .173
0007 12/15/13 00:59 .175
0008 12/15/13 01:02 .178
0009 12/19/13 13:58 .104

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Officer K. Lee	17046	Intoximeters Inc./Alco-Sensor IV	018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013			13B06-16-5-10- KKF24C	.100								Lee	17046	

OUT FOR SERVICE

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1400	34	13B06-16-5-10- KKF24C	.100	22		.099			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 14:06
 AS4 Serial no: 018646
 Version no: 4797
 No. Date Time %BAC
 Calibration 07/29/13 13:32 .100
 Test Result
 0001 12/19/13 13:59 .099
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1404	34	13B06-16-5-10- KKF24C	.100	21	.101			X		Lee	17046		

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

12/19/13 14:03
AS4 Serial no: 077488
Version no: 4797
Mo. Date Time ZBAC
Calibration
05/21/08 14:35 .100
Test Result
0001 12/19/13 14:02 .101
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1407	34	13B06-16-5-10- KKF24C	.100	21		.103			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 14:07
 AS4 Serial no: 008877
 Version no: 4797
 No. Date Time %BAC
 Calibration 05/26/05 10:40 .100
 Test Result
 0001 12/19/13 14:07 .103
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1410	34	13B06-16-5-10- KKF24C	.100	22	.098			X		Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15													
<div style="float: right; text-align: center;"> No. Date Time ZBAC Calibration 12/19/08 14:44 .100 Test Result 0001 12/14/13 12:36 .000 0002 12/19/13 14:09 .098 Push No. 3 to erase. </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1414	34	13B06-16-5-10- KKF24C	.100	22		.100			X		Lee	17046	
			Prepared by: Douglas C. Talbot 02-06-13											
			Expires on: 02-06-15											

12/19/13 14:12
HS4 Serial no: 018626
Version no: 4797
No. Date Time ZBAC
Calibration
04/29/05 08:59 .101

Test Result
0001 12/19/13 14:11 .100

Push No. 3 to erase,
or power down to retain

or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013			I3B06-16-5-10- KKF24C	.100							Lee	17046	
			Prepared by: Douglas C. Talbot										
			02-06-13										
			Expires on: 02-06-15										

OUT OF REPAIR

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1417	34	13B06-16-5-10- KKF24C	.100	21		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 14:15
AS4 Serial no: 083030
Version no: 4797
No. Date Time ZTMC
Calibration 01/29/09 10:39 .100
Test Result
0001 12/19/13 14:14 .100
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1419	34	13B06-16-5-10- KKF24C	.100	21		.100			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

Push No. 3 to erase.

Test Result
 0001 12/19/13 15:05 .000
 0002 12/19/13 14:18 .100

No. Date Time ZROC
 Calibration
 08/12/08 14:35 .100

12/19/13 14:18
 ASD Serial no: 018692
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1421	34	13B06-16-5-10- KKF24C	.100	21		103			X		Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15														

12/19/13 14:21
 AS4 Serial no: 018762
 Version no: 4797
 No. Date Time %BAC
 Calibration 02/30/09 15:52 .100
 Test Result
 0001 12/16/13 14:31 .000
 0002 12/16/13 16:27 .037
 0003 12/16/13 20:00 .000
 0004 12/17/13 12:36 .000
 0005 12/17/13 15:41 .000
 0006 12/18/13 12:36 .000
 0007 12/19/13 14:28 .103
 Push No. 3 to erase,
 or power down to retain
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	1423	34	13B06-16-5-10- KKF24C	.100	21		.096				X	Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

Push No. 3 to erase,
or power down to retain
 Test Result
 0091 12/19/13 14:22 .096
 No. Date Time %PAC
 Calibration
 03/12/12 12:36 .100
 12/19/13 14:22
 AS4 Serial no: 018766
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1428	34	13B06-16-5-10- KKF24C	.100	22		.105					Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

12/19/13 14:28
 NS4 Serial no: 018691
 Version no: 4797

No. Date Time ZMAC
 Calibration
 09/08/09 14:01 .100

Test Result
 0001 12/14/13 08:45 .019
 0002 12/14/13 08:46 .018
 0003 12/14/13 08:50 .155
 0004 12/14/13 08:53 .160
 0005 12/14/13 22:41 .006
 0006 12/14/13 22:43 .006
 0007 12/15/13 01:55 .189
 0008 12/15/13 01:56 .184
 0009 12/15/13 04:39 .125
 0010 12/15/13 04:42 .109
 0011 12/16/13 00:51 .188
 0012 12/16/13 00:53 .192
 0013 12/19/13 14:27 .105

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013			13B06-16-5-10-	.100							Lee	17046	
			KKF24C										
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										OUT FOR SERVICE
			Expires on:										
			02-06-15										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/19/2013	1431	34	13B06-16-5-10- KKF24C	.100	21	.101				X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Test Result
 0001 12/18/13 02:15 .068
 0002 12/18/13 02:18 .079
 0003 12/18/13 02:20 .000
 0004 12/18/13 03:08 .129
 0005 12/18/13 03:12 .127
 0006 12/19/13 01:55 .027
 0007 12/19/13 01:56 .004
 0008 12/19/13 14:29 .101
 No. Date Time ZMAC
 Calibration 01/13/09 10:25 .100
 12/19/13 14:29
 ASD Serial no: 018774
 Version no: 4797
 Push No. 3 to erase
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1433	34	13B06-16-5-10- KKF24C	.100	21		.106			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

No. Date Time %BAC
 Calibration 03/08/06 13:56 .100
 Test Result 0001 12/19/13 14:31 .106
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 091495
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	14:36	34	13B06-16-5-10-	.100	20		.094			X		Lee	17046	
			KKF24C											
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

Test Result
0001 12/19/13 14:33 .094

Calibration
08/29/10 10:45 .098

No. Date Time %BAC
454 Serial no: 841495
Version no: 4797

12/19/13 14:34

Push No. 3 to erase
or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
12/19/2013	14:39	34	13B06-16-5-10- KKF24C	.100	21		.096			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

Push No. 3 to erase or power down to retain

```

Test Result
0001 12/14/13 00:09 .096
0002 12/14/13 00:10 .082
0003 12/14/13 02:06 .111
0004 12/14/13 02:06 .116
0005 12/15/13 02:15 .227
0006 12/16/13 02:58 .113
0007 12/16/13 03:01 .114
0008 12/19/13 14:37 .090

No. Date Time %BAC
Calibration 05/05/09 09:03 .100

12/19/13 14:37
H34 Serial no: 041502
Version no: 4797
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. OB 2654
--------------------------------	-----------------	------------------------------------	----------------------	--	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/19/2013	1441	34	13B06-16-5-10- KKF24C	.100	21							Lee	17046	"TEST"
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

B.O.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.