

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083032
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0851	34	13B06-16-5-10- KKF24C	.100	16		.099			X		Lee	17046	NOT IN LOCKER <i>KL</i>
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

Test Result: 0001 01/03/14 08:48 .099
 No. Date Time %RNC
 Calibration 12/01/11 07:38 .081
 Version not 4797
 01/03/14 08:49
 PC4 Serial no: 083002
 0001 01/03/14 08:48 .099
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0742	34	13B06-16-5-10- KKF24C	.100	20		.100			X		Lee	17046	NOT TO LOCKER
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

01/03/14 07:44
 NS4 Serial no: 008620
 Version no: 4797
 No. Date Time ZBMC
 Calibration 01/30/13 14:34 .100
 Test Result
 0001 01/03/14 07:43 .100
 Push No. 3 to erase,
 or power down to retain
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
01/03/2014	0713	34	13B06-16-5-10- KKF24C	.100	20		.104			X		Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												

01/03/14 07:11
654 Serial no: 077505
Version no: 4797
No. Date Time ZBAC
Calibration
05/21/08 13:16 .100
Test Result
0001 01/03/14 07:10 .104
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0716	34	13B06-16-5-10- KKF24C	.100	17		.097			X	Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

Test Result
 0001 12/28/13 00:32 .042
 0002 12/28/13 03:27 .123
 0003 12/28/13 03:29 .118
 0004 12/29/13 01:46 .090
 0005 01/03/14 01:16 .041
 0006 01/03/14 03:40 .079
 0007 01/03/14 03:43 .080
 0008 01/03/14 07:15 .097

Push No. 3 to erase,
or power down to retain

Calibration
 No. Date Time %BNC
 03/07/13 09:48 .099

01/03/14 07:16
 As4 Serial no: 077479
 Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082660
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0726	34	13B06-16-5-10- KKF24C	.100	19	.097			X		Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15													
Test Result 0001 12/28/13 02:41 .045 0002 12/28/13 02:43 .041 0003 12/28/13 04:26 .128 0004 12/28/13 04:29 .129 0005 01/03/14 07:24 .097 Calibration 11/19/10 06:21 .081													
Push No. 3 to erase.													

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
01/03/2014	0729	34	13B06-16-5-10- KKF24C	.100	20		.106				X	Lee	17046		
			Prepared by:												
			Douglas C. Talbot												
			02-06-13												
			Expires on:												
			02-06-15												
										No. Date Time ZBAC Calibration 05/20/04 11:06 .096					
										Test Result 0001 12/27/13 23:35 .103 0002 12/27/13 23:37 .090 0003 12/29/13 01:29 .193 0004 12/29/13 01:27 .108 0005 12/29/13 03:51 .056 0006 12/29/13 03:53 .055 0007 12/30/13 02:55 .176 0008 12/30/13 02:56 .182 0009 01/03/14 07:29 .106					
										Push No. 3 to erase or power down to retain					
										01/03/14 07:29 NS4 Serial no: 008709 Version no: 4797					
										or power down to retain					

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0732	34	13B06-16-5-10- KKF24C	.100	19		.107			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

01/03/14 07:34
MSA Serial no: 008714
Version no: 4797
No. Date Time %RAC
Calibration
03/11/04 14:05 .100
Test Result
0001 01/03/14 08:32 .000
0002 01/03/14 08:55 .000
0003 01/03/14 02:18 .071
0004 01/03/14 02:19 .073
0005 01/03/14 02:42 .093
0006 01/03/14 02:44 .144
0007 01/03/14 02:47 .160
0008 01/03/14 07:33 .107
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0735	34	13B06-16-5-10-KKF24C	.100	21		.102			X	Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

No. Date Time %BAC
Calibration
03/25/13 10:45 .100
Test Result
0001 01/03/14 07:33 .102
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0738	34	13B06-16-5-10- KKF24C	.100	19	.102				X		Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15														
<div style="display: flex; justify-content: space-between;"> <div style="writing-mode: vertical-rl; transform: rotate(180deg);">Push No. 3 to erase.</div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> Test Result 0001 12/31/13 23:56 .092 0002 12/31/13 23:58 .095 0003 01/01/14 01:53 .187 0004 01/01/14 01:57 .182 0005 01/03/14 02:38 .102 </div> <div style="writing-mode: vertical-rl; transform: rotate(180deg);"> No. Date Time ZBAC Calibration 07/18/08 13:52 .100 </div> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 002657
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0745	34	13B06-16-5-10- KKF24C	.100	20	.097			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

01/03/14 07:45
 HSA Serial No: 002657
 Version no: 4797
 No. Date Time ZBAC
 Calibration 10/07/11 12:22 .081
 Test Result
 0001 01/03/14 07:45 .097
 Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0748	34	13B06-16-5-10- KKF24C	.100	19		.098			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

or power down to retain
 10/03/14 07:48
 604 Serial no: 008810
 Version no: 4797
 No. Date Time %BAC
 Calibration
 10/30/08 09:53 .100
 Test Result
 0001 01/03/14 07:48 .098
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0759	34	13B06-16-5-10-	.100	43							Lee	17046	TEMP'ND GO'
1/3/2014	0827	34	KKF24C	.100	22		.100			X		LEE	17046	REPLACED BATTERY.
<p>Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15</p>														
<p>Push No. 3 to erase, or Power down to retain</p>														
<p>Test Result 0001 12/31/13 01:08 .069 0002 12/31/13 01:11 .083 0003 12/31/13 03:54 .000 0004 12/31/13 21:24 .228 0005 12/31/13 21:26 .208 0006 01/01/14 01:43 .189 0007 01/01/14 01:45 .182 0008 01/03/14 08:27 .106</p>														
<p>No. Date Time %BAC Calibration 10/24/03 09:04 .100</p>														
<p>01/03/14 08:29 454 Serial no: 008635 Version no: 4797</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO.
--------------------------------	-----------------	------------------------------------	----------------------	--	----------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0802	39	13B06-16-5-10-KKF24C	.100	20	.105			X		Lee	17046	

Prepared by:
 Douglas C. Talbot
 02-06-13
 Expires on:
 02-06-15

Test Result
 0001 12/31/13 20:00 .000
 0002 01/02/14 00:32 .000
 0003 01/02/14 02:39 .137
 0004 01/02/14 02:42 .129
 0005 01/03/14 00:02 .105

Calibration
 12/24/03 11:01 .100

Version no: 4799

01/03/14 08:03
 NS4 Serial no: 000845

Push No. 3 to erase.

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0806	34	13B06-16-5-10- KKF24C	.100	20		.097			X	Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13													
Expires on: 02-06-15													
<p>Test Result 0001 01/03/14 08:05 .097</p> <p>No. Date Time ZBAC Calibration 12/01/11 07:44 .081</p> <p>Version no: 4797</p> <p>01/03/14 08:05 No. Serial no: 082653 Version no: 4797</p>													
<p>Push No. 3 to erase, or Power down to retain</p>													
<p>or Power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0810	34	13B06-16-5-10-KKF24C	.100	19		.103			X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15</p>													
<p>Test Result 0001 12/28/13 01:21 .045 0002 12/28/13 02:18 .000 0003 12/29/13 01:38 .179 0004 12/29/13 01:40 .228 0005 12/29/13 01:42 .227 0006 01/03/14 03:26 .058 0007 01/03/14 08:10 .103</p>													
<p>Calibration 01/13/03 09:25 .100</p>													
<p>No. Date Time XBAC Version no: 4797</p>													
<p>01/03/14 08:11 454 Serial no: 008848 Version no: 4797</p>													
<p>Push No. 3 to erase, or power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0833	34	13B06-16-5-10- KKF24C	.100	19	.099			X		Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										
										Test Result 0001 12/29/13 20:22 .060 0002 01/03/14 08:32 .099 Calibration 05/21/08 15:09 .100 No. Date Time 2DNC Version no: 4797 AS4 Serial no: N77591 01/03/14 08:32			

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014			13B06-16-5-10- KKF24C	.100								Lee	17046	NOT IN LOCKER
			Prepared by: Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0837	34	13B06-16-5-10- KKF24C	.100	21		.099			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15</p>														
<p>Test Result</p> <p>0001 12/31/13 01:46 .143 0002 12/31/13 22:00 .000 0003 01/01/14 01:49 .000 0004 01/01/14 19:17 .110 0005 01/01/14 19:21 .114 0006 01/01/14 19:29 .000 0007 01/02/14 08:13 .018 0008 01/02/14 08:15 .017 0009 01/03/14 08:35 .099</p> <p>Push No. 3 to erase, or power down to retain</p>														
<p>No. Date Time %RAC Calibration 05/21/08 14:43 .100</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS			
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO							
01/03/2014	0840	34	13B06-16-5-10- KKF24C	.100	21		.098			X		Lee	17046				
			Prepared by:														
			Douglas C. Talbot														
			02-06-13														
			Expires on:														
			02-06-15														
										Test Result 0001 01/03/14 08:40 .098		Calibration 03/25/13 14:00 .100		No. Date Time ZBAC 01/03/14 08:40 694 Serial no: 083051 Version no: 4797			
										Push No. 3 to erase, or power down to retain							

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0644	34	13B06-16-5-10- KKF24C	.100	20		.106				X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

01/03/14 08:48
 ASA Serial no: 008897
 Version no: 4797
 No. Date Time ZHNC
 Calibration 01/18/12 16:13 .100
 Test Result
 0001 12/28/13 01:55 .000
 0002 12/28/13 02:41 .131
 0003 12/28/13 02:43 .138
 0004 01/03/14 08:45 .106
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
01/03/2014			13B06-16-5-10- KKF24C	.100							Lee	17046	

B. D.

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018646
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0856	34	13B06-16-5-10-KKF24C	.100	22		.100			X	Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

01/03/14 08:55
NS4 Serial no: 018646
Version no: 4797
No. Date Time ZHAC
Calibration 07/29/13 18:32 .100
Test Result
0061 12/31/13 22:28 .091
0062 12/31/13 23:26 .082
0063 01/03/14 08:55 .100
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0859	34	13B06-16-5-10-KKF24C	.100	20	.103					X	Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											
												Calibration 05/21/08 14:35 .100		
												Test Result 0601 01/03/14 08:58 .103		
												Push No. 3 to erase, or power down to retain		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#34

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
01/03/2014	0902 34	34	13B06-16-5-10-KKF24C	.100	20		.103				X	Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15														
Test Result 0001 01/03/14 09:02 .103 No. Date Time %IMC Calibration 05/26/05 10:40 .106 Version no: 4797 01/03/14 09:03 NS4 Serial no: 008877 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018615
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0905	34	13B06-16-5-10- KKF24C	.100	21	.100			X		Lee	17046		
			Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15											
											01/03/14 09:05 654 Serial no: 018615 Version no: 4797			
											No. Date Time 2BAC Calibration 12/19/08 14:44 .100			
											Test Result 0001 12/29/13 01:28 .046 0002 12/29/13 01:29 .046 0003 01/03/14 09:05 .100			
											Push No. 3 to erase, or power down to retain			

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0908	34	13B06-16-5-10- KKF24C	.100	21	.100			X		Lee	17046		
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

01/03/14 09:06
NS4 Serial no: 018626
Version no: 4797
No. Date Time ZBAC
Calibration 04/29/05 08:59 .101
Test Result
0001 01/03/14 09:05 .100
Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018696
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014			13B06-16-5-10- KKF24C	.100								Lee	17046	

B.D.

Prepared by:
 Douglas C. Talbot
 02-06-13
 Expires on:
 02-06-15

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0911	34	13B06-16-5-10- KKF24C	.100	21	.100				X	Lee	17046	
<p>Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15</p>													
<p style="text-align: center;"> Test Result 0001 01/03/14 09:00 .100 Calibration 01/29/09 10:39 .100 No. Date Time ZHMC Calibration 01/29/09 10:39 .100 Version no: 4797 01/03/14 09:10 ASA Serial no: 083030 01/03/14 09:10 </p>													
<p style="text-align: center;"> Push No. 3 to erase, or power down to retain </p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0917	34	13B06-16-5-10-KKF24C	.100	21		.099		X		Lee	17046	REPLACED BATTERY
			Prepared by: Douglas C. Talbot 02-06-13										
			Expires on: 02-06-15										
<p style="text-align: center;">Test Result 0601 01/03/14 09:19 .099</p> <p style="text-align: center;">Calibration 08/12/08 14:35 .100</p> <p style="text-align: center;">No. Date Time XMAC Version no: 4297</p> <p style="text-align: center;">01/03/14 09:20 NS4 Serial no: 018692</p> <p style="text-align: center;">Push No. 3 to erase, or Power down to retain</p>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0922	34	13B06-16-5-10-KKF24C	.100	20	.103				X	Lee	17046	
Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15													
<div style="float: right; text-align: right;"> 01/03/14 09:22 ACS Serial no: 018762 Version no: 4797 No. Date Time ZBMC Calibration 03/30/09 15:52 .100 Test Result 0001 12/31/13 20:03 .097 0002 12/31/13 20:05 .091 0003 01/01/14 02:48 .000 0004 01/03/14 09:22 .103 Push No. 3 to erase, or power down to retain </div>													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0925	34	13B06-16-5-10- KKF24C	.100	20		095		Y	X		Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

Test Result
0001 01/03/14 09:23 .095
Calibration
03/12/12 12:30 .100
Version No: 479P
No. Date Time %HAC
01/03/14 09:24
NS4 Serial no: 018766
01/03/14 09:24

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#49

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018691
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0928	34	13B06-16-5-10- KKF24C	.100	20		102			X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

Push No. 3 to erase,
or power down to retain

Test Result
0001 12/29/13 01:10 .171
0002 12/29/13 01:13 .169
0003 12/29/13 22:00 .286
0004 12/29/13 22:07 .242
0005 12/29/13 22:13 .249
0006 12/29/13 01:00 .198
0007 12/29/13 01:02 .165
0008 12/29/13 01:03 .163
0009 01/03/14 03:46 .000
0010 01/03/14 03:48 .000
0011 01/03/14 09:28 .102

No. Date Time ZMVC
Calibration
09/08/08 14:01 .100
01/03/14 09:29
054 Serial no: 018691
Version no: 4797

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014			13B06-16-5-10- KKF24C	.100							Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

B.O.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 018774
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0931	39	13B06-16-5-10- KKF24C	.100	21		.10			X	Lee	17046	
			Prepared by:										
			Douglas C. Talbot										
			02-06-13										
			Expires on:										
			02-06-15										

Push No. 3 to erase
or power down to retain

Test Result
0001 12/31/13 01:44 .000
0002 12/31/13 23:47 .166
0003 12/31/13 23:49 .164
0004 01/01/14 03:20 .095
0005 01/01/14 03:23 .090
0006 01/01/14 03:56 .173
0007 01/01/14 03:59 .162
0008 01/01/14 04:02 .044
0009 01/01/14 21:05 .104
0010 01/01/14 21:08 .090
0011 01/01/14 21:15 .000
0012 01/03/14 09:30 .101

No. Date Time %MC
Calibration
01/13/09 10:25 .100
01/03/14 09:30
654 Serial no: 010774
Version no: 4797

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0937	34	13B06-16-5-10-KKF24C	.100	21	.106				X		Lee	17046	
			Prepared by:											
			Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

01/03/14 09:36
 NS4 Serial no: 041486
 Version no: 4797
 No. Date Time ZBAC
 Calibration 03/02/06 13:56 .100
 Test Result
 0001 01/03/14 09:36 .106
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0940	34	13B06-16-5-10- KKF24C	.100	19		.095			X		Lee	17046	
			Prepared by: Douglas C. Talbot											
			02-06-13											
			Expires on:											
			02-06-15											

01/03/14 09:09
 PAS Serial no: 041495
 Version no: 4797
 No. Date Time ZMAC
 Calibration 09/29/10 10:45 .098
 Test Result
 0001 01/03/14 09:09 .095
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 041502
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
01/03/2014	0943	34	13B06-16-5-10- KKF24C	.100	20	.097			X		Lee	17046	

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

Test Result
0001 12/28/13 06:47 .209
0002 12/28/13 06:50 .187
0003 12/28/13 19:49 .104
0004 12/28/13 19:52 .124
0005 12/28/13 19:56 .114
0006 12/29/13 19:27 .150
0007 12/29/13 19:29 .144
0008 12/29/13 21:46 .043
0009 01/03/14 01:56 .006
0010 01/03/14 09:47 .097

No. Date Time XPRC
Calibration
05/05/09 08:03 .100

01/03/14 09:43
NSA Serial no: 041502
Version no: 4797

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 082654
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014			13B06-16-5-10-	.100								Lee	17046	
			KKF24C											

B.O.

Prepared by:
Douglas C. Talbot
02-06-13
Expires on:
02-06-15

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

567

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Officer K. Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters Inc./Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	------------------------------------	----------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)			PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
01/03/2014	0948	34	13B06-16-5-10- KKF24C	.100	20		.104			X		Lee	17046	
<p>Prepared by: Douglas C. Talbot 02-06-13 Expires on: 02-06-15</p> <p style="text-align: center;">Test No. 3 to erase or power down to retain</p> <p style="text-align: center;"> Test Result 0001 12/31/13 20:29 .093 0002 12/31/13 20:31 .092 0003 01/03/14 09:48 .104 </p> <p style="text-align: center;"> No. Date Time %BAC Calibration 06/25/13 18:32 .107 </p> <p style="text-align: center;"> 01/03/14 09:48 (54 Serial not 008874 Version not 4797 </p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.