

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#1  
16120188

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1114	34	17AP41036	.100	23	.103					M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-11-17  
Expiration date:  
1-11-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:16 AM	10/21/2016	.103	.100	11:17 AM	10/24/2017	BAC	16120188	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/24/2017	1116	34	17AP41036	.100	22		.107					M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-11-17  
Expiration date:  
1-11-19

FC20	8.7.24	Serial No.	16120194	UOM	BAC	Last Check	10/24/2017	Time:	11:18 AM	Cal Standard	.100	Result	.107	Time:	08:35 AM	Last Cal	04/25/2017	Cal Standard	.100
------	--------	------------	----------	-----	-----	------------	------------	-------	----------	--------------	------	--------	------	-------	----------	----------	------------	--------------	------

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/24/2017	1120	34	17AP41036	.100	22		.100					M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120197
UOM	BAC
Last Check	10/24/2017
Time:	10:26 AM
Cal Standard	.100
Result	.100
Last Cal	10/21/2016
Time:	10:22 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1122	34	17AP41036	.100	22		107				M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-11-17  
Expiration date:  
1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120195
UOM	BAC
Last Check	10/24/2017
Time:	11:22 AM
Cal Standard	.100
Result	.107
Last Cal	04/25/2017
Time:	08:47 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/24/2017	0914	34	17AP41036	.100	20		101				X	M. Tippett	20137	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 1-11-17											
			Expiration date:											
			1-11-19											

10/24/17 09:24  
 NS4 Serial no: 008559  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/15/16 10:40 .100  
 Test Result  
 0001 10/24/17 09:23 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
10/24/2017	1124	74	17AP41036	.100	22		.105				X	M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040161
UOM	BAC
Last Check	10/24/2017
Time:	11:29 AM
Cal Standard	.100
Result	.105
Last Cal	04/25/2017
Time:	08:56 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040140
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	11/26	34	17AP41036	.100	22		.107				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

**Lifeloc Technologies**

Calibration / Check Printout

FC20	8.7.24
Serial No.	16040140
UOM	BAC
Last Check	10/24/2017
Time:	11:23 AM
Cal Standard	.100
Result	.107
Last Cal	04/25/2017
Time:	08:58 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.









**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0922	34	17AP41036	.100	20		1107				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

Push No. 3 to erase,  
or power down to retain

Test Result  
 0001 10/14/17 20:51 .155  
 0002 10/14/17 20:54 .153  
 0003 10/15/17 00:42 .101  
 0004 10/15/17 00:45 .111  
 0005 10/16/17 00:52 .141  
 0006 10/15/17 00:53 .140  
 0007 10/16/17 00:14 .097  
 0008 10/16/17 00:17 .099  
 0009 10/16/17 00:20 .036  
 0010 10/22/17 02:04 .065  
 0011 10/22/17 02:03 .065  
 0012 10/22/17 02:06 .065  
 0013 10/22/17 02:07 .000  
 0014 10/22/17 04:46 .045  
 0015 10/24/17 09:27 .107

10/24/17 09:28  
 NS4 Serial) no: 002653  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 02/11/16 10:14 .100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																								
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																							
10/24/2017	09:24	34	17AP41036	.100	21		.107			Y	M. Tippett	20137																									
			Solution prepared by: Cesar A. Mejia Date: 1-11-17 Expiration date: 1-11-19																																		
			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p style="text-align: center;">Push No. 3 to erase or power down to retain</p> <p style="text-align: center;">Test Result</p> <table border="1"> <tr><td>0001</td><td>10/12/17</td><td>22:59</td><td>.152</td></tr> <tr><td>0002</td><td>10/13/17</td><td>23:03</td><td>.134</td></tr> <tr><td>0003</td><td>10/13/17</td><td>23:06</td><td>.143</td></tr> <tr><td>0004</td><td>10/15/17</td><td>08:49</td><td>.166</td></tr> <tr><td>0005</td><td>10/15/17</td><td>08:54</td><td>.171</td></tr> <tr><td>0006</td><td>10/24/17</td><td>09:35</td><td>.107</td></tr> </table> <p style="text-align: center;">No. Date Time %BAC Calibration 03/17/16 09:40 .100</p> <p style="text-align: center;">10/24/17 09:35 AS4 Serial no: 008709 Version no: 4797</p> </div>											0001	10/12/17	22:59	.152	0002	10/13/17	23:03	.134	0003	10/13/17	23:06	.143	0004	10/15/17	08:49	.166	0005	10/15/17	08:54	.171	0006	10/24/17	09:35	.107
0001	10/12/17	22:59	.152																																		
0002	10/13/17	23:03	.134																																		
0003	10/13/17	23:06	.143																																		
0004	10/15/17	08:49	.166																																		
0005	10/15/17	08:54	.171																																		
0006	10/24/17	09:35	.107																																		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0928	34	17AP41036	.100	21		.107				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

Push No. 3 to erase  
 or power down to retain  
 Test Result  
 0001 10/24/17 09:29 .107  
 No. Date Time %BAC  
 Calibration 03/17/16 09:44 .100  
 10/24/17 09:30  
 AS4 Serial no: 008888  
 Version no: 4797

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0932	34	17AP41036	.100	21		0.103			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 09:41  
 PAS Serial no: 082657  
 Version no: 4797  
 No. Date Time PRNC  
 Calibration 05/21/17 08:10 .100  
 Test Result  
 0001 10/24/17 09:40 .100  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#20

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018684
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0936	34	17AP41036	.100	20		20.108			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 09:46  
 NS4 Serial no: 018684  
 Version no: 4797  
 No. Date Time ZTMC  
 Calibration 01/12/16 14:36 .099  
 Test Result  
 0001 10/24/17 09:46 .108  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1128	74	17AP41036	.100	22		.100			0	M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040162
UOM	BAC
Last Check	10/24/2017
Time:	11:22 AM
Cal Standard	.100
Result	.100
Last Cal	08/14/2017
Time:	10:14 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS																																																																												
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO																																																																											
10/24/2017	1130	34	17AP41036	.100	22		.105				M. Tippet	20137																																																																													
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td colspan="14">Solution prepared by: Cesar A. Mejia</td> </tr> <tr> <td colspan="14">Date: 1-11-17</td> </tr> <tr> <td colspan="14">Expiration date: 1-11-19</td> </tr> <tr> <td colspan="14" style="text-align: center;"> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">FC20</td> <td style="width: 10%;">8.7.24</td> <td style="width: 10%;">Serial No.</td> <td style="width: 10%;">16120177</td> <td style="width: 10%;">UOM</td> <td style="width: 10%;">BAC</td> <td style="width: 10%;">Last Check</td> <td style="width: 10%;">10/24/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">11:33 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> <td style="width: 10%;">Result</td> <td style="width: 10%;">.105</td> <td style="width: 10%;">Last Cal</td> <td style="width: 10%;">04/25/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">09:50 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> </tr> </table> </td> </tr> </table>														Solution prepared by: Cesar A. Mejia														Date: 1-11-17														Expiration date: 1-11-19														<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">FC20</td> <td style="width: 10%;">8.7.24</td> <td style="width: 10%;">Serial No.</td> <td style="width: 10%;">16120177</td> <td style="width: 10%;">UOM</td> <td style="width: 10%;">BAC</td> <td style="width: 10%;">Last Check</td> <td style="width: 10%;">10/24/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">11:33 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> <td style="width: 10%;">Result</td> <td style="width: 10%;">.105</td> <td style="width: 10%;">Last Cal</td> <td style="width: 10%;">04/25/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">09:50 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> </tr> </table>														FC20	8.7.24	Serial No.	16120177	UOM	BAC	Last Check	10/24/2017	Time:	11:33 AM	Cal Standard	.100	Result	.105	Last Cal	04/25/2017	Time:	09:50 AM	Cal Standard	.100
Solution prepared by: Cesar A. Mejia																																																																																									
Date: 1-11-17																																																																																									
Expiration date: 1-11-19																																																																																									
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="width: 10%;">FC20</td> <td style="width: 10%;">8.7.24</td> <td style="width: 10%;">Serial No.</td> <td style="width: 10%;">16120177</td> <td style="width: 10%;">UOM</td> <td style="width: 10%;">BAC</td> <td style="width: 10%;">Last Check</td> <td style="width: 10%;">10/24/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">11:33 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> <td style="width: 10%;">Result</td> <td style="width: 10%;">.105</td> <td style="width: 10%;">Last Cal</td> <td style="width: 10%;">04/25/2017</td> <td style="width: 10%;">Time:</td> <td style="width: 10%;">09:50 AM</td> <td style="width: 10%;">Cal Standard</td> <td style="width: 10%;">.100</td> </tr> </table>														FC20	8.7.24	Serial No.	16120177	UOM	BAC	Last Check	10/24/2017	Time:	11:33 AM	Cal Standard	.100	Result	.105	Last Cal	04/25/2017	Time:	09:50 AM	Cal Standard	.100																																																								
FC20	8.7.24	Serial No.	16120177	UOM	BAC	Last Check	10/24/2017	Time:	11:33 AM	Cal Standard	.100	Result	.105	Last Cal	04/25/2017	Time:	09:50 AM	Cal Standard	.100																																																																						

**Lifeloc Technologies**

Calibration / Check Printout

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0947	34	17AP41036	.100	20		.103				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 09:55  
 HSA Serial no: 077591  
 Version no: 4797  
 No. Date Time ZMAC  
 Calibration 05/21/17 08:35 .100  
 Test Result  
 0001 10/24/17 09:58 .100  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008850
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	0948 34		17AP41036	.100	21		.108				M. Tippet	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 10:06  
 NS4 Serial) no: 008850  
 Version no: 4797  
 No. Date Time Y2MC  
 Calibration 02/06/17 14:07 .100  
 Test Result  
 0001 10/24/17 10:06 .100  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.









**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1132	34	17AP41036	.100	22		.102				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	10:41 AM	08/14/2017	.102	.100	11:25 AM	10/24/2017	BAC	16080071	8.7.24

Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 160 40163
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	-----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1134	34	17AP41036	.100	23		.097			X	M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040163
UOM	BAC
Last Check	10/24/2017
Time:	11:32 AM
Cal Standard	.100
Result	.097
Last Cal	06/07/2017
Time:	11:02 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.











**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
10/24/2017	1032	34	17AP41036	.100	21		104				X	M. Tippet	20137	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 1-11-17											
			Expiration date:											
			1-11-19											

Push No. 5 to erase,  
or power down to retain

```

10/24/17 10:41
604 Serial no: 018766
Version no: 4797

No. Date Time XBAC
Calibration
01/22/16 13:16 .100

Test Result
0001 10/14/17 22:10 .040
0002 10/14/17 22:14 .175
0003 10/14/17 22:17 .169
0004 10/22/17 01:09 .193
0005 10/22/17 01:12 .185
0006 10/22/17 02:50 .000
0007 10/24/17 10:41 .104
    
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1036	34	17AP41036	.100	22	.107					M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 10:47  
 NS4 Serial no: 017990  
 Version no: 4797  
 No. Date Time ZDNC  
 Calibration 11/26/08 11:10 .100  
 Test Result  
 0001 10/17/17 04:19 .289  
 0002 10/17/17 04:21 .285  
 0003 10/10/17 03:28 .085  
 0004 10/10/17 03:31 .099  
 0005 10/24/17 10:47 .107  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#57

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	104034		17AP41036	.100	70		26.101		X		M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-11-17										
			Expiration date:										
			1-11-19										

10/24/17 10:48  
 NS4 Serial no: 041495  
 Version no: 4797  
 No. Date Time XDAC  
 Calibration 08/17/17 13:39 .108  
 Test Result  
 0001 10/24/17 10:47 .101  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
10/24/2017	1138	34	17AP41036	.100	22		.097				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-11-17

Expiration date:

1-11-19

<b>Lifeloc Technologies</b>	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080013
UOM	BAC
Last Check	10/24/2017
Time:	11:44 AM
Cal Standard	.100
Result	.097
Last Cal	08/14/2017
Time:	09:43 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



