

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	092834		17AP41035	.100	21		103				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard

Time:

Last Cal

Result

Cal Standard

Time:

Last Check

UOM

Serial No.

FC20

Calibration / Check Printout

**Lifeloc Technologies**

.100

08:16 AM

10/21/2016

.103

.100

09:32 AM

12/08/2017

BAC

16120188

8.7.24

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0930	34	17AP41035	.100	20		.107				M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia

Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard

Time: 08:35 AM

Last Cal 04/25/2017

Result .107

Cal Standard .100

Time: 09:33 AM

Last Check 12/08/2017

UOM BAC

Serial No. 16120194

FC20 8.7.24

Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0932	34	17AP41035	.100	20		.103				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout		<b>Lifeloc Technologies</b>
.100	10:22 AM	10/21/2016	.103	.100	09:38 AM	12/08/2017	BAC	16120197	8.7.24			

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	0934	34	17AP41035	.100	20		108					M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:47 AM	04/25/2017	.108	.100	09:38 AM	12/08/2017	BAC	16120195	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	083034		17AP41035	.100	19		.097			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 09:52  
 NS4 Serial no: 008559  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/15/16 10:40 .100  
 Test Result  
 0001 12/08/17 09:39 .086  
 0002 12/08/17 09:52 .097  
 Push No. 3 to green,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0936	34	17AP41035	.100	20		.107			X	M. Tippett	20137	

Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19		Cal Standard	Time: 08:56 AM	Last Cal 04/25/2017	Result .107	Cal Standard .100	Time: 10:42 AM	Last Check 12/08/2017	UOM BAC	Serial No. 16040161	FC20 8.7.24	Calibration / Check Printout	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040140
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0940	34	17AP41035	.100	21		.107			X	M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	08:58 AM	04/25/2017	.107	.100	09:40 AM	12/08/2017	BAC	16040140	8.7.24

**Lifeloc Technologies**  
Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	084034	34	17AP41035	.100	19		.104			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 09:56  
 HSA SERIAL NO: 000620  
 VERSION: 101: 4797  
 No. of Calibration: 01/15/16 11:10 AM  
 Test Result: 0001 12/08/17 09:55 .104  
 Push No. 0 to erase, or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0846	34	17AP41035	.100	20		.102				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 08:51  
 NS4 Serial no: 082957  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 02/17/14 08:52 .108  
 Test Result  
 0001 12/08/17 08:50 .102  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	084834		17AP41035	.100	19		.109				M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 09:51  
 ASA Serial no: 077505  
 Version: 4797  
 No. Date Time ZBAC  
 Calibration 01/22/16 10:42 .100  
 Test Result 0001 12/08/17 09:50 .109  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0852	34	17AP41035	.100	19		107			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 09:58  
 HQ4 Serial no: 082653  
 Version: 101 4797  
 No. Date Time ZBAC  
 Calibration 02/11/16 10:14 .100  
 Test Result  
 0001 11/05/17 01:29 .184  
 0002 11/05/17 01:31 .182  
 0003 11/12/17 23:40 .122  
 0004 11/12/17 23:42 .121  
 0005 11/14/17 13:46 .108  
 0006 11/19/17 20:25 .110  
 0007 11/19/17 20:29 .110  
 0008 11/19/17 23:10 .141  
 0009 11/19/17 23:13 .148  
 0010 11/21/17 11:59 .108  
 0011 11/25/17 00:29 .162  
 0012 11/25/17 00:32 .159  
 0013 11/25/17 00:41 .099  
 0014 11/25/17 00:43 .095  
 0015 12/01/17 11:30 .103  
 0016 12/03/17 04:40 .000  
 0017 12/03/17 23:25 .056  
 0018 12/08/17 03:31 .097  
 0019 12/08/17 03:38 .092  
 0020 12/08/17 09:57 .107  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017			17AP41035	.100	20		.107				M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19										
			<div style="border: 1px solid black; padding: 5px; width: fit-content; margin: auto;"> <p style="text-align: center;">12/08/17 10:07 494 Serial no: 008709 Version no: 4797</p> <p style="text-align: center;">No. Date Time READ Calibration 03/17/16 09:40 .100</p> <p style="text-align: center;">Test Result</p> <p>0001 12/01/17 21:19 .142 0002 12/01/17 21:21 .147 0003 12/08/17 02:22 .051 0004 12/08/17 06:21 .060 0005 12/08/17 06:24 .046 0006 12/04/17 04:47 .000 0007 12/04/17 04:48 .000 0008 12/08/17 10:06 .107</p> <p style="text-align: center;">Push No. 3 to erase, or Power down to retain</p> </div>										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008714
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO	
12/08/2017	0852	34	17AP41035	.100	19		.106				X		M. Tippett	20137	
			Solution prepared by:												
			Cesar A. Mejia												
			Date: 1-09-17												
			Expiration date:												
			1-09-19												

12/08/17 10:19  
 AS4 Serial no: 008714  
 Version: not 4797  
 No. Date Time Zoned  
 Calibration 05/14/17 09:50 .100  
 Test Result 0091 12/08/17 10:09 .106  
 Dash No. 2 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0902	34	17AP41035	.100	18		.106			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 10:12  
 MS4 Serial no: 0082103  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 05/14/17 10:02 .106  
 Test Result  
 0001 12/08/17 10:12 .106  
 Push No. 18 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0904	34	17AP41035	.100	18		.105			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 10:16  
 NS4 Serial no: 008810  
 Version no: 4799  
 No. Date Time ZBAC  
 Calibration 05/21/17 08:20 .100  
 Test Result  
 0001 12/08/17 10:15 .105  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017			17AP41035	.100	21		103					M. Tippett	20137	

Solution prepared by: Cesar A. Mejia		Cal Standard	Time: 10:14 AM	Last Cal 08/14/2017	Result .103	Cal Standard .100	Time: 09:42 AM	Last Check 12/08/2017	UOM BAC	Serial No. 16040162	FC20 8.7.24	<p style="text-align: center;"><b>Lifeloc Technologies</b> Calibration / Check Printout</p>	
Date: 1-09-17													
Expiration date: 1-09-19													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	094234	34	17AP41035	.100	21		.106			X		M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard

.100

Time:

09:50 AM

Last Cal

04/25/2017

Result

.106

Cal Standard

.100

Time:

09:45 AM

Last Check

12/08/2017

UOM

BAC

Serial No.

16120177

FC20

8.7.24

Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.











**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 083051
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	1020	34	17AP41035	.100	20		.109			X		M. Tippett	20137	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 1-09-17											
			Expiration date:											
			1-09-19											

12/08/17 11:26  
 NS4 Serial No: 083051  
 Version: Not 4797  
 No. Date Time %BAC  
 Calibration 05/24/16 16:24 .100  
 Test Result  
 0001 12/08/17 11:25 .109  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
12/08/2017	1024	34	17AP41035	.100	19		.109				X	M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19											
			<div style="border: 1px solid black; padding: 5px; text-align: center;">                     12/08/17 11:37                      091 Serial no: 008897                      Version no: 4797                      No. Jplc Time ZHAC                      Calibration                      01/12/12 16:13 .100                      Test Result                      0001 12/08/17 11:37 .109                      Push No. 3 to erase                      or power down to retain                 </div>											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	1028	34	17AP41035	.100	20		107			X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 11:35  
 684 Serial no: 018624  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 02/01/16 11:02 .100  
 Test Result 0001 12/08/17 11:34 .107  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	0944	34	17AP41035	.100	20		.104			X		M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard

Time:

Last Cal

Result

Cal Standard

Time:

Last Check

UOM

Serial No.

FC20

Calibration / Check Printout

**Lifeloc Technologies**

.100

10:41 AM

08/14/2017

.104

.100

09:47 AM

12/08/2017

BAC

16080071

8.7.24

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.







**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0448	34	17AP41035	.100	21		.099				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

FC20	8.7.24	Serial No.	16040163	UOM	BAC	Last Check	12/08/2017	Time:	09:50 AM	Cal Standard	.100	Result	.099	Last Cal	06/07/2017	Time:	11:02 AM	Cal Standard	.100
------	--------	------------	----------	-----	-----	------------	------------	-------	----------	--------------	------	--------	------	----------	------------	-------	----------	--------------	------

Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	1030	34	17AP41035	.100	20		.107			X	M. Tippet	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 1-09-17										
			Expiration date:										
			1-09-19										

12/08/17 11:33  
 154 Serial no: 018626  
 Division no: 4597  
 No. Date Time ZMAC  
 Calibration 07/15/16 15:21 .100  
 Test Result 0001 12/08/17 11:33 .107  
 Push No. 3 to erase  
 or Power Down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.









**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018 766
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	1040	34	17AP41035	.100	19		.103				X	M. Tippett	20137	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 1-09-17											
			Expiration date:											
			1-09-19											

12/08/17 11:45  
 NS4 Serial no: 018766  
 Version no: 4737  
 No. Date Time MBMC  
 Calibration 01/22/16 13:16 .100  
 Test Result  
 0001 12/07/17 08:13 .107  
 0002 12/07/17 08:17 .119  
 0003 12/07/17 01:08 .156  
 0004 12/07/17 01:11 .165  
 0005 12/07/17 04:14 .120  
 0006 12/07/17 04:17 .133  
 0007 12/07/17 04:34 .090  
 0008 12/08/17 05:06 .009  
 0009 12/08/17 11:45 .103  
 Push No. 3 to erase  
 or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.







**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
12/08/2017	0950	34	17AP41035	.100	2		.100				X	M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	.100
Time:	12:17 PM
Last Cal	09/01/2017
Result	.100
Cal Standard	.100
Time:	10:47 AM
Last Check	12/08/2017
UOM	BAC
Serial No.	16120189
FC20	8.7.24
Calibration / Check Printout	
<b>Lifeloc Technologies</b>	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
12/08/2017	0954	34	17AP41035	.100	21		101				M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

FC20	8.7.24	Serial No.	16080013	UOM	BAC	Last Check	12/08/2017	Time:	09:58 AM	Cal Standard	.100	Result	.101	Last Cal	08/14/2017	Time:	09:43 AM	Cal Standard	.100
------	--------	------------	----------	-----	-----	------------	------------	-------	----------	--------------	------	--------	------	----------	------------	-------	----------	--------------	------

Calibration / Check Printout

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



