

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/04/2018 09:40

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3493	
Check sum	47615	Download date/time:	01/04/2018	10:03		Record ID	9RA6	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 01/04/2018 09:42

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V) .133

Solution lot#: IRS042717

Soln temp C: 34.01

Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	21325	Download date/time:	01/04/2018	10:03		Record ID	9R55	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/04/2018 09:45

First calibration check: .127 Interferent detected: No
Second calibration check: .134 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.08
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 38394	Download date/time: 01/04/2018	10:03	Record ID 9R66
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 01/11/2018 09:25

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3490
Check sum 46327	Download date/time: 02/01/2018	10:52	Record ID 9RJ3
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/11/2018 09:26

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.08
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22464	Download date/time: 02/01/2018	10:52	Record ID 9RB7
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340 Accuracy check date/time: 01/11/2018 09:28

First calibration check: .134 Interferent detected: No
Second calibration check: .138 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 44120	Download date/time: 02/01/2018	10:52	Record ID 9RC0
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/17/2018 13:23

First calibration check:
Second calibration check:

Interferent detected:
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3503	
Check sum	53191	Download date/time:	02/01/2018	10:52		Record ID	9RK6	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 01/17/2018 13:24

First calibration check: Not performed
Second calibration check: Not performed

Interferent detected: No
Interferent detected: No

Concentration (%W/V) .133

Solution lot#: IRS042717

Soln temp C: 34.05
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	23454	Download date/time:	02/01/2018	10:52		Record ID	9RD5	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/17/2018 13:26

First calibration check: .140 Interferent detected: No
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 53775	Download date/time: 02/01/2018	10:52	Record ID 9RE8
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 01/25/2018 09:27

First calibration check: Interferent detected:
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

Diagnostics check

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3495
Check sum 65099	Download date/time: 02/01/2018	10:52	Record ID 9RL1
Record imported (Y/N) N	Source Diagnostic		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/25/2018 09:28

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49817	Download date/time: 02/01/2018	10:52	Record ID 9RF9
Record imported (Y/N) N	Source Calibration		

Instrument serial No: 68-012340 Accuracy check date/time: 01/25/2018 09:30

First calibration check: .136 Interferent detected: No
Second calibration check: .137 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 14739	Download date/time: 02/01/2018	10:52	Record ID 9RG2
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/31/2018 13:21

First calibration check: Interferent detected:
Second calibration check: Interferent detected:
Concentration (%W/V) Solution lot#: Soln temp C:
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator _____

Diagnostics check

Eeprom Pass Serial No. Pass Temp Pass Internal standards Pass
Prom Pass Real Time clock Pass Motor Pass Printer Pass
Ram Pass Auto calibration Pass Range Pass Digital voltmeter 3515
Check sum 29453 Download date/time: 02/01/2018 10:52 Record ID 9RM4
Record imported (Y/N) N Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 01/31/2018 13:24

First calibration check: Not performed Interferent detected: No
Second calibration check: Not performed Interferent detected: No
Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.02
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 _____

Diagnostics check

Eeprom Serial No. Temp Internal standards
Prom Real Time clock Motor Printer
Ram Auto calibration Range Digital voltmeter
Check sum 45500 Download date/time: 02/01/2018 10:52 Record ID 9RH9
Record imported (Y/N) N Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department
Forensic Science Services Division**

1400 Canal Ave
Long Beach, Ca 90813
PH: (562) 570-7680 Fax: (562) 570-7689

Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record

Instrument serial No: 68-012340 Accuracy check date/time: 01/31/2018 13:29

First calibration check: .136 Interferent detected: No
Second calibration check: .139 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 33.91
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

Diagnostics check

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 55740	Download date/time: 02/01/2018	10:52	Record ID 9RI2
Record imported (Y/N) N	Source Calibration		

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.