

#1

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16120188

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/16/2018	1520	34	17AP41035	.100	21	.092		X			M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.092	.100	1:36 PM	02/16/2018	BAC	16120188	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#2

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
02/16/2018	1322	34	17AP41035	.100	20									M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:50 AM	01/29/2018	.093	.100	1:37 PM	02/16/2018	BAC	16120194	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
02/16/2018	1524	74	17AP41035		.100	72		.096			Y		M. Tippett	20137	

Solution prepared by														
Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														
Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout			
.100		10:22 AM	10/21/2016	.096	.100	1:44 PM	02/16/2018	BAC	16120197	8.7.24	Lifeloc Technologies			

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

*#4*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	1328 34		17AP41035	.100	20								M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:47 AM	04/25/2017	.101	.100	1:41 PM	02/16/2018	BAC	16120195	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051 #5

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 008559

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0838	34	17AP41035	.100	19		096		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time %BAC AS4 Serial no: 008559 Version no: 4797 02/16/18 08:41 Calibration 01/15/16 10:40 .100 Test Result 0001 02/16/18 08:41 .096 Push No. 3 to erase OR Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

*HC*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16040161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1528 34		17AP41035	.100	19		.099					M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard		Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
	.100	08:56 AM	04/25/2017	.099	.100	2:46 PM	02/16/2018	BAC	16040161	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 7

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Lifeloc/FC20  
PAS SERIAL NO.: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEST #1	TEST #2	YES	NO	TEST #1	TEST #2				
02/16/2018	1:30	34	17AP41035	.100	20									M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:58 AM	04/25/2017	.100	.100	1:41 PM	02/16/2018	BAC	16040140	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*78*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: *008620*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0844	74	17AP41035	.100	19						X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

02/16/18 08:45  
 AS4 Serial no: 008620  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/15/16 11:10 .100  
 Test Result 0001 02/16/18 08:45 .101  
 Push No. 3 to erase,  
 or Power down to retain



49

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 082957

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0921	34	17AP41035	.100	19				X			M. Tippet	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19</p>														
<p>02/16/18 09:22      A54 Serial no: 082957      Version no: 4797 No. Date Time ZBAC      Calibration      02/17/14 08:52 .100 Test Result      0001 02/16/18 09:22 .098 Push No. 3 to erase. or Power down to retain</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO.: 077505

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	0923	34	17AP41035	.100	19			098					M. Tippet	20137	
Solution prepared by: Cesar A. Mejia															
Date: 1-09-17															
Expiration date: 1-09-19															
<pre> 02/16/18 09:24 AS4 Serial no: 077505 Version no: 4797 No. Date Time %BAC 01/18/18 10:56 .100 Test Result 0001 02/16/18 09:23 .098 Push No. 3 to erase, OR Power down to retain </pre>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#11

CHP 202K (Rev. 8-08) OPI 051

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077479

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				

02/16/2018 0930 17AP41035 .100

M. Tippett

20137

Device not in lockers for testing

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 082653

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	0927	57	17AP41035	.100	19		103			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time XBAC Version no: 4797 AS4 Serial no: 082653 Calibration 02/11/16 10:14 .100 Test Result 0001 02/10/18 04:44 .078 0002 02/16/18 09:28 .103 Push No. 3 to erase, or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#115

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008709

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/16/2018	0930	74	17AP41035	.100	19		103				X		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													
02/16/18 09:31 NS4 Serial no: 008709 Version no: 4797 No. Date Time XBAC Calibration 03/17/16 09:40 .100 Test Result 0001 02/16/18 09:31 .103 Push No. 3 to erase or Power down to retain													

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#14

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	0934	34	17AP41035	.100	19	*.102					X		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time %BAC Version no: 4797 NS4 Serial no: 008714 02/16/18 09:35 Calibration 05/14/17 09:50 .100 Test Result 0001 02/16/18 09:35 .102 Push No. 3 to erase, OR Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008888

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	0933	34	17AP41035	.100	19						M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
Push No. 3 to erase, OR Power down to retain.													
02/16/18 09:34 AS4 Serial no: 008888 Version no: 4797 No. Date Time %BAC Calibration 03/17/16 09:44 .100 Test Result 0001 02/13/18 20:18 .153 0002 02/13/18 20:20 .152 0003 02/15/18 01:03 .135 0004 02/16/18 09:34 .104													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008783

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	093734	17AP41035		.100	18		102		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time %BAC Calibration 05/14/17 10:07 .100 Test Result 0001 02/16/18 09:38 .102 Push No. 3 to erase, or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf



4/18

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082657
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	0940	34	17AP41035		.100	19					X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Test Result 0001 02/11/18 00:39 .121 0002 02/11/18 00:40 .116 0003 02/11/18 00:42 .106 0004 02/11/18 02:55 .178 0005 02/11/18 02:59 .188 0006 02/11/18 23:17 .043 0007 02/16/18 09:41 .098 No. Date Time %BAC Calibration 05/21/17 08:18 .100 AS4 Serial no: 082657 Version no: 4797															
Push No. 3 to erase, OR Power down to retain															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

#19

AREA NAME South Los Angeles AREA NO. 530 COORDINATOR NAME Kirk Lee I.D. NUMBER 17046 PAS MAKE/MODEL Intoximeters/Alco-Sensor IV PAS SERIAL NO. 008810

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0945	34	17AP41035	.100	18		.100			X		M. Tippett	20137	

02/16/18 09:46  
 AS4 Serial no: 008810  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 05/21/17 08:28 .100  
 Test Result 0001 02/16/18 09:45 .100  
 Push No. 3 to erase,  
 or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

#20

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Kirk Lee

I.D. NUMBER  
 17046

PAS MAKE/MODEL  
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
 018684

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0948	18	17AP41035	.100								M. Tippett	20137	
<p>Solution prepared by:            Cesar A. Mejia            Date: 1-09-17            Expiration date:            1-09-19</p>														
<p>02/16/18 10:49            NS4 Serial no: 018684            Version no: 4797            No. Date Time %BMC            Calibration 01/12/16 14:36 .093            Test Result 0001 02/16/18 10:49 .105            Push No. 3 to erase,            or Power down to retain</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Cnp202k\_1108.pdf

#21

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008635

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	0957	34	17AP41035	.100	18		103				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
Push No. 3 to erase, OR Power down to retain													
Test Result 0001 02/04/18 23:46 .111 0002 02/04/18 23:49 .109 0003 02/05/18 04:34 .238 0004 02/05/18 04:39 .238 0005 02/08/18 14:11 .105 0006 02/11/18 03:30 .142 0007 02/11/18 03:33 .142 0008 02/16/18 10:58 .103													
No. Date Time %BAC Calibration 05/21/17 08:23 .100													
Version no: 4797 NS4 Serial no: 008635													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16040162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1343	34	17AP41035	.100	20		.092			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date: 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	10:14 AM	08/14/2017	.092	.100	1:44 PM	02/16/2018	BAC	16040162	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Lifeloc/FC20  
PAS SERIAL NO.: 16120177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1346	34	17AP41035	.100	21		.100					M. Tippet	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
09:50 AM	04/25/2017	.100	.100	1:46 PM	02/16/2018	BAC	16120177	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008848

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0957	39	17AP41035	.100	18	104		X			M. Tippet	20137	

02/16/18 09:52  
 NS4 Serial no: 008848  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 05/21/17 08:30 .100  
 Test Result  
 0001 02/08/18 13:05 .106  
 0002 02/16/18 09:51 .104  
 Push No. 3 to erase,  
 OR Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 077591

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.			TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1530	34	17AP41035		.100	27							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
02/16/18 13:30 NS4 Serial no: 077591 Version no: 4797 No. Date Time ZBNC Calibration 05/21/17 08:35 .100 Test Result 0001 02/10/18 06:09 .181 0002 02/10/18 06:13 .174 0003 02/11/18 07:39 .027 0004 02/16/18 13:30 .097 Push No. 3 to erase, OR Power down to retain															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#26

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/16/2018	1340	34	17AP41035	.100	20		.097			X		M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard		Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.		FC20	
	.100		1:54 PM		12/19/2017		.097		.100		2:40 PM		02/16/2018		BAC		16080065		8.7.24

Lifeloc Technologies

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0952	34	17AP41035	.100	19		096		X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
No. Date Time %BAC Calibration 01/18/18 11:23 .100 Test Result 0001 02/16/18 09:52 .096 Push No. 3 to erase, or Power down to retain														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 083051

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	0955	34	17AP41035	.100	19								M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19</p> <pre> 02/16/18 09:56 AS4 Serial no: 083051 Version no: 4797 No. Date Time %BAC Calibration 12/28/17 13:29 .100 Test Result 0001 02/16/18 09:55 .095 Push No. 3 to erase, or Power down to retain                 </pre>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	0958	34	17AP41035	.100	18								M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
No. Date Time %BAC Calibration 01/18/12 16:13 .100 Test Result 0001 02/16/18 09:59 .103 Push No. 3 to erase, or power down to retain															
02/16/18 09:59 AS4 Serial no: 008897 Version no: 4797															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 018624

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1 (Optional)		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	1002	34	17APA1035		.100	20				X			M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
<pre> 02/16/18 10:03 AS4 Serial no: 018624 Version no: 4797 No. Date Time %BAC Calibration 02/01/16 11:02 .100 Test Result 0001 02/16/18 10:02 .100 Push No. 3 to erase, OR Power down to retain           </pre>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 32

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16080071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1352	34	17AP41035	.100	20	.097					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.097	.100	1:53 PM	02/16/2018	BAC	16080071	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*#33*

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Kirk Lee

I.D. NUMBER  
 17046

PAS MAKE/MODEL  
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
*077488*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	1062	24	17APA1035	.100	19			X		M. Tippet	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
02/16/18 10:03 NS4 Serial no: 077488 Version no: 4797 No. Date Time %BAC Calibration 02/11/16 10:52 .100 Test Result 0001 02/16/18 10:02 .104 Push No. 3 to erase, or power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

# 34

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Kirk Lee

I.D. NUMBER  
17046

PAS MAKE/MODEL  
Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
008877

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
02/16/2018	1258	34	17AP41035	.100	19									M. Tippet	20137	
<div style="text-align: center;"> <p>Solution prepared by:  Cesar A. Mejia  Date: 1-09-17  Expiration date:  1-09-19</p> </div> <hr/> <div style="text-align: center;"> <p>02/16/18 12:59  A84 Serial no: 008877  Version no: 4797  No. Date Time XBAC  Calibration 05/28/17 08:49 .100  Test Result 0001 02/16/18 12:59 .095  Push No. 3 to erase,  OR Power down to retain</p> </div>																

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions  
Chp202K\_1108.pdf



#36

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16040163

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1352	51	17AP41035	.100	22		.092					M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	11:02 AM	06/07/2017	.092	.100	1:52 PM	02/16/2018	BAC	16040163	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 018626

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1259 34	17AP41035	.100	19		103		X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

02/16/18 13:00  
 AS4 Serial no: 018626  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 07/15/16 15:21 .100  
 Test Result 0001 02/16/18 12:59 .103  
 Push No. 3 to erase,  
 or Power down to retain

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#58

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
02/16/2018	1459	57	17AP41035	.100	20	097			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date: 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/ COMMENTS
.100	2:05 PM	12/19/2017	.097	.100	3:00 PM	02/16/2018	BAC	16080072	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

# 39

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 083030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
02/16/2018	1301	34	17AP41035	.100	19	.096		X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time Z/DAC Calibration 12/28/17 13:51 .100 Test Result 0001 02/16/18 13:01 .096 AS4 Serial no: 083030 Version no: 4797 Push No. 3 to erase OR Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions      Chp202K\_1108.pdf

*#440*

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: *018692*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	<i>1005 34</i> <i>1257 38</i>		17AP41035	.100	13		<i>111</i>				M. Tippett	20137	<i>1st TEST 720</i> <i>CO10</i>

02/16/18 12:55  
 AS4 Serial no: 018692  
 Version no: 4797  
 No. Date Time XBMC  
 Calibration 02/01/16 11:30 .100  
 Test Result  
 0001 02/16/18 10:06 .111  
 0002 02/16/18 12:54 .102  
 Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

442

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1304	34		17AP41035	.100	18							M. Tippett	20137	
			Solution prepared by:												
			Cesar A. Mejia												
			Date: 1-09-17												
			Expiration date:												
			1-09-19												
			Test Result 0001 02/11/18 01:33 .019 0002 02/11/18 03:38 .067 0003 02/12/18 02:16 .132 0004 02/12/18 02:20 .117 0005 02/16/18 13:05 .095 Push No. 3 to erase, or Power down to retain												
			No. Date Time %BAC Calibration 12/28/17 14:01 .100 Version no: 4797 AS4 Serial no: 018762 02/16/18 13:05												

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI051

*#43*

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO.: *018766*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
02/16/2018	1304	34	17AP41035		.100	19							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19  Push No. 3 to erase, or Power down to retain Test Result: 0001 02/16/18 03:27 .172 Calibration: 01/22/16 13:16 .100 No. Date Time %BAC Version no: 4797 AS4 Serial no: 018766 02/16/18 13:07															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Cnp202k\_1108.pdf

#44

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO.: 018691

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	1309 34	17AP41035	.100	19		.097							
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time XBAC    Version    AS4 Serial no: 018691 Calibration    05/28/17 08:26 .100    Test Result    0001 02/16/18 13:09 .097 Push No. 3 to erase, or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions    CHP202K\_1108.pdf



STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO.: 017990

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP ("C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP ("C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
02/16/2018	1311	54	17AP41035	.100	19		.102				M. Tippett	20137	

02/16/18 13:12  
AS4 Serial) no: 017990  
Version no: 4797  
No. Date Time XBAC  
Calibration 11/26/03 11:10 .100  
Test Result 0001 02/16/18 13:11 .102  
Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions  
Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16120189

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1452	34	17AP41035	21		.091		YES	NO		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia													
Date: 1-09-17													
Expiration date: 1-09-19													

Cal Standard		Last Cal		Cal Standard		Last Check		UOM		Serial No.		FC20		Calibration / Check Printout
Result	Time:	Result	Time:	Result	Time:	Result	Time:							
.100	12:17 PM	.091	09/01/2017	.100	2:52 PM		02/16/2018	BAC		16120189		8.7.24		

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*450*

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	<i>041486</i>

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	<i>1312</i>	<i>39</i>	17AP41035	.100	<i>19</i>		<i>102</i>		<i>X</i>			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
No. Date Time KBAC Version no: 4797 AS4 Serial no: 041486 02/16/18 13:13 Calibration 03/08/06 13:56 .100 Test Result 0001 02/16/18 13:12 .102 Push No. 3 to erase, OR Power down to retain														

\* If the reading is within plus or minus (+) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*457*

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	<i>041495</i>

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
02/16/2018	1312	34	17AP41035		.100	19							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
02/16/18 13:15 AS4 Serial no: 041495 Version no: 4797 No. Date Time %BNC Calibration 12/28/17 14:10 .100 Test Result 0001 02/16/18 13:14 .097 Push No. 3 to erase, OR Power down to retain															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

#52

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/F/C20      PAS SERIAL NO.: 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP ('C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP ('C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	1404	34	17AP41035	.100	21		.097					M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.097	.100	2:05 PM	02/16/2018	BAC	16080013	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#53

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Infoximeters/Alco-Sensor IV	PAS SERIAL NO. <i>082654</i>
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
02/16/2018	131734	17AP41035		.100	19			✓				M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
02/16/18 13:18 AS4 Serial no: 082654 Version no: 4797 No. Date Time XBAC Calibration 12/19/16 11:26 .100 Test Result 0001 02/13/18 04:30 .000 0002 02/16/18 13:17 .103 Push No. 3 to erase OR Power down to retain														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*SGT'S PAS*

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008874

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS			
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO		
02/16/2018	1320	34	17AP41035	.100	19		098					X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
Push No. 3 to erase, or Power down to retain																
02/16/18 13:21 AS4 Serial no: 008874 Version no: 4797 No. Date Time %BAC Calibration 10/24/17 11:04 .100 Test Result 0001 02/16/18 13:21 .098																

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf