

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/07/2017 09:28

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3295	
Check sum	24411	Download date/time:	12/07/2017	09:51		Record ID	9P04	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012333 Accuracy check date/time: 12/07/2017 09:30

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	45082	Download date/time:	12/07/2017	09:51		Record ID	9OV5	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/07/2017 09:36

First calibration check: .121 Interferent detected: No  
Second calibration check: .123 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.09  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 32828	Download date/time: 12/07/2017	09:53	Record ID 9P20
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012333 Accuracy check date/time: 12/07/2017 09:38

First calibration check: .124 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.05  
Simulator Ser No. DR3707

LAST CAL CHECK - COLD SIM HOSE

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 20797	Download date/time: 12/07/2017	09:55	Record ID 9P37
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/14/2017 09:05

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3249	
Check sum	26170	Download date/time:	01/04/2018	09:23		Record ID	9QS8	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012333 Accuracy check date/time: 12/14/2017 09:07

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 001

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	36419	Download date/time:	01/04/2018	09:23		Record ID	9QK9	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/14/2017 09:14

First calibration check: .126 Interferent detected: No  
Second calibration check: .127 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 33.93  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 47337	Download date/time: 01/04/2018	09:23	Record ID 9QL6
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012333 Accuracy check date/time: 12/21/2017 06:27

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3237
Check sum 72920	Download date/time: 01/04/2018	09:23	Record ID 9QT1
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/21/2017 06:29

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 33.93  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 001

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 52420	Download date/time: 01/04/2018	09:23	Record ID 9QM9
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012333 Accuracy check date/time: 12/21/2017 06:32

First calibration check: .124 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.00  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 21311	Download date/time: 01/04/2018	09:23	Record ID 9QN6
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/28/2017 09:37

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3231	
Check sum	23695	Download date/time:	01/04/2018	09:23		Record ID	9QU8	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012333 Accuracy check date/time: 12/28/2017 09:38

First calibration check: Not performed  
Second calibration check: Not performed

Interferent detected: No  
Interferent detected: No

Concentration (%W/V) .133

Solution lot#: IRS042717

Soln temp C: 34.00  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	12719	Download date/time:	01/04/2018	09:23		Record ID	9QO9	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012333 Accuracy check date/time: 12/28/2017 09:42

First calibration check: .124 Interferent detected: No  
Second calibration check: .125 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 13471	Download date/time: 01/04/2018	09:23	Record ID 9QP1
Record imported (Y/N) N	Source Calibration		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.