

#3

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/19/2018	092034		17AP41035	.100	19		.096				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
				Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
				.100	10:22 AM	10/21/2016	.096	.100	09:21 AM	03/19/2018	BAC	16120197	8.7.24	
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/19/2018	0922-34		17AP41035	.100	21	.103					M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	08:47 AM	04/25/2017	.103	.100	09:20 AM	03/19/2018	BAC	16120195	8.7.24	
										Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
03/19/2018	1014	34	17AP41035	.100	20					X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

03/19/18 09:18
 AS4 Serial no: 008559
 Version no: 4797
 No. Date Time ZDHC
 Calibration 01/15/16 10:40 .100
 Test Result 0001 03/19/18 09:18 .096
 Push No. 3 to erase
 or Power down to retain

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RESULTS				
03/19/2018	0924 24	17AP41035	.100	21		n/01		X				M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.101	.100	09:25 AM	03/19/2018	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

47

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2018	092534		17AP41035	.100	21		.102		X		M. Tippett	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
				Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
				.100	08:58 AM	04/25/2017	.102	.100	09:23 AM	03/19/2018	BAC	16040140	8.7.24	
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/19/2018	1029 34	17AP41035		.100	20		.101			X	M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
Push No. 3 to erase. OR Power down to retain													
03/19/18 09:30 NS4 Serial no: 082653 Version no: 4797 No. Date Time KBAC Calibration 02/11/16 10:14 .100 Test Result 0001 03/11/18 02:14 .275 0002 03/11/18 02:18 .279 0003 03/16/18 02:20 .172 0004 03/16/18 02:23 .170 0005 03/19/18 09:29 .101													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OP1 051

#118

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee ID. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
03/19/2018	1047	34	17AP41035		.100	20								M. Tippett	20137	ADJUSTED FOR DST	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																	
No. Date Time ZBAC Calibration 05/21/17 08:18 .100 Test Result 0001 03/19/18 10:47 .096 Push No. 3 to erase, or power down to retain																	
03/19/18 10:48 NS4 Serial no: 082657 Version no: 4797																	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
03/19/2018	0924	34	17AP41035	.100	21		.096		NO		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
				Cal Standard									
					Time:								
						Last Cal							
				.100									
					Time:								
						Last Cal							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							
				.100									
					Time:								
						Last Check							
				.096									
					Time:								
						Last Check							

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REPEAT RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	TEST #1	TEST #2			
03/19/2018	0925	34	17AP41035		.100	22		102						M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.102	.100	09:25 AM	03/19/2018	BAC	16120177	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#26

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/19/2018	092834		17AP41035		.100	21		.102				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
			Cal Standard		.100										
			Time:			1:54 PM									
			Last Cal			12/19/2017									
			Result		.102										
			Cal Standard		.100										
			Time:			09:28 AM									
			Last Check			03/19/2018									
			UOM			BAC									
			Serial No.			16080065									
			FC20			8.7.24									
Calibration / Check Printout															
Lifeloc Technologies															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

36

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	YES	NO				
03/19/2018	0930	34	17AD41035	.100	22		.094				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														
				Cal Standard										
				.100										
					Time:									
					11:02 AM									
						Last Cal								
						06/07/2017								
							Result							
							.094							
								Cal Standard						
							.100							
									Time:					
									09:30 AM					
										Last Check				
										03/19/2018				
											UOM			
											BAC			
												Serial No.		
												16040163		
													FC20	
													8.7.24	
														Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Lifeloc Technologies

#42

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: *018762*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									TEST #1	TEST #2 (Optional)	YES	NO				
03/19/2018	1254	34	17AP41035		.100	21				<i>.095</i>			X	M. Tippett	20137	<i>ADJUSTED TIME FOR DST</i>
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
Test Result 0001 03/11/18 01:35 .014 0002 03/11/18 01:38 .011 0003 03/12/18 00:56 .084 0004 03/12/18 00:59 .078 0005 03/12/18 01:41 .279 0006 03/12/18 01:44 .276 0007 03/19/18 11:55 .095 No. Date Time %BAC Calibration 12/28/17 14:01 .100 MS4 Serial no: 018762 Version no: 4797																
Push No. 3 to erase. OR Power down to retain																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#46

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
03/19/2018	0935	34	17AD41035	.100	21		.094		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
				Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
				.100	12:17 PM	09/01/2017	.094	.100	09:35 AM	03/19/2018	BAC	16120189	8.7.24
Calibration / Check Printout													
Lifeloc Technologies													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/19/2018	0938	34	17AP41035	.100	21		.098	.100		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
				Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
				.100	12:59 PM	01/29/2018	.098	.100	09:38 AM	03/19/2018	BAC	16080013	8.7.24	
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/19/2018	1313	34	17AP41035		.100	21			.100				X	M. Tippett	20137	ADJUSTED TIME FOR DST
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
Push No. 3 to erase, or Power down to retain																
Test Result 0001 03/13/18 00:36 .145 0002 03/13/18 00:38 .177 0003 03/13/18 00:41 .172 0004 03/14/18 02:59 .159 0005 03/14/18 03:02 .163 0006 03/19/18 12:13 .100																
No. Date Time XBAC Calibration 12/19/16 11:26 .100																
Version no: 4797 NS4 Serial no: 082654 03/19/18 12:14																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

SGT's PAS

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: *008874*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS/ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #2 (Optional)	TEST #1 (Optional)		YES	NO				
03/19/2018	1317	34	17AP41035		.100	20			.096		NO	X	M. Tippett	20137	<i>ADJUSTED TIME FOR DST</i>
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Test Result 0001 03/11/18 03:46 .087 0002 03/11/18 03:50 .088 0003 03/13/18 19:00 .000 0004 03/16/18 23:20 .194 0005 03/16/18 23:22 .201 0006 03/18/18 18:26 .109 0007 03/18/18 18:30 .113 0008 03/19/18 12:17 .096															
No. Date Time ZBAC Calibration 10/24/17 11:04 .100															
03/19/18 12:18 AS4 Serial no: 008874 Version no: 4797															
Push No. 3 to erase, or Power down to retain															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.