



#2

AREA NAME  
South Los Angeles

AREA NO. 530  
COORDINATOR NAME Kirk Lee

I.D. NUMBER 17046  
PAS MAKE/MODEL Lifeloc/FC20

PAS SERIAL NO. 16120194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1	TEST #2 (Optional)	YES	NO				
03/30/2018	12:04 PM	34	17AP41035	.100	22								M. Tippett	20137	
Solution prepared by:															
Cesar A. Mejia															
Date: 1-09-17															
Expiration date: 1-09-19															
Cal Standard															
Time: 08:50 AM															
Last Cal 01/29/2018															
Result .098															
Cal Standard .100															
Time: 12:04 PM															
Last Check 03/30/2018															
UOM BAC															
Serial No. 16120194															
FC20 8.7.24															
Calibration / Check Printout															
Lifeloc Technologies															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#5

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	12:11	34	17AP41035	100	22	101			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19												
Cal Standard												
					Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.
					10:22 AM	10/21/2016	.101	.100	12:12 PM	03/30/2018	BAC	16120197
											8.7.24	
Calibration / Check Printout												Lifeloc Technologies

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    ID. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	12/10	34	17AP41035	.100	22		.107			X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														
Cal Standard														
Time: 08:47 AM														
Last Cal														
04/25/2017														
Result														
.107														
Cal Standard														
.100														
Time: 12:11 PM														
Last Check														
03/30/2018														
UOM														
BAC														
Serial No. 16120195														
FC20														
8.7.24														
Calibration / Check Printout														
<b>Lifeloc Technologies</b>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*AL*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/C20      PAS SERIAL NO.: **16640161**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					YES
03/30/2018	12:15	24	17AP41035	.100	21		.103				X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.103	.100	12:16 PM	03/30/2018	BAC	16040161	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 7

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/30/2018	1212 34	17AP41035		.100	22	.105					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:58 AM	04/25/2017	.105	.100	12:13 PM	03/30/2018	BAC	16040140	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

48

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008620

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	0958	34	17AP41035	.100	17							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

03/30/18 10:00  
 NS4 Serial no: 008620  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 01/15/16 11:10 .100  
 Test Result 0001 03/30/18 09:59 .110  
 Push No. 3 to erase,  
 OR Power down to retain











*HS*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 008709

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	0926	34	17AP41035	.100	20					X		M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

03/30/18 09:27 AS4 Serial no: 008709 Version no: 4797  
 No. Date Time ZBAC Calibration 03/17/16 09:40 .100  
 Test Result 0001 03/30/18 09:26 .108  
 Push No. 3 to erase.  
 or Power down to retain



#16

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: 008888

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
03/30/2018	09:18	34	17AP41035	.100	20								M. Tippet	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
Push No. 3 to erase, or power down to retain																
Test Result 0001 03/21/18 02:44 .089 0002 03/21/18 02:46 .096 0003 03/22/18 03:27 .051 0004 03/22/18 03:29 .051 0005 03/22/18 05:09 .047 0006 03/22/18 05:11 .046 0007 03/23/18 01:28 .231 0008 03/23/18 01:32 .227 0009 03/23/18 03:56 .227 0010 03/23/18 03:59 .211 0011 03/27/18 02:13 .113 0012 03/27/18 02:15 .109 0013 03/28/18 02:37 .187 0014 03/28/18 02:39 .164 0015 03/30/18 09:29 .108																
No. Date Time ZBAC Calibration 03/17/16 09:44 .100																
Version not 4797 AS4 Serial no: 008888 03/30/18 09:30																

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#17

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008783

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	0955	34	17AP41035	.100		19		106			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
Push No. 3 to erase, or Power down to retain														
Test Result 0001 03/24/18 00:41 .136 0002 03/24/18 00:43 .144 0003 03/24/18 05:21 .152 0004 03/24/18 05:24 .154 0005 03/29/18 21:35 .120 0006 03/29/18 21:37 .111 0007 03/29/18 23:41 .030 0008 03/29/18 23:43 .029 0009 03/30/18 09:33 .106														
No. Date Time %BAC Calibration 05/14/17 10:07 .100 Version no: 4797 NS4 Serial no: 008783														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#19

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 008810

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	09:58	34	17AP41035		.100	M						X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.															

03/30/18 09:39  
 MS4 Serial no: 008810  
 Version no: 4797  
 No. Date Time ZDAC  
 Calibration 05/21/17 08:20 .100  
 Test Result 0001 03/30/18 09:39 .104  
 Push No. 3 to erase  
 or Power down to retain



#22

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16040162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	12:11	34	17AP41035	.100	22	.098		X		M. Tippett	20137		
Solution prepared by: Cesar A. Mejia													
Date: 1-09-17													
Expiration date: 1-09-19													
Cal Standard					Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
				.100	10:14 AM	08/14/2017	.098	.100	12:12 PM	03/30/2018	BAC	16040162	8.7.24
Calibration / Check Printout													Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16120177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
03/30/2018	12:14	34	17AP41035	.100	22		.103					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.103	.100	12:15 PM	03/30/2018	BAC	16120177	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





#26

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1217	34	17AP41035	.100	21	.105					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.105	.100	12:17 PM	03/30/2018	BAC	16080065	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.











#32

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16880071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	12:19	34	17AD41035	.100		.100					M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 1-09-17  
Expiration date:  
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:41 AM	08/14/2017	.100	.100	12:20 PM	03/30/2018	BAC	16080071	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#34

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 068877

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1010	34	17AP41035	.100	20			X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													
No. Date Time ZBAC Calibration 05/28/17 08:49 .100 Test Result 0001 03/30/18 10:11 .101 Push No. 3 to erase OR Power down to retain													

#36

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Lifeloc/FC20  
PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1	TEST #2 (Optional)	YES	NO				
03/30/2018	12/18	34	17AP41035	.100	23			.097	.100		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia															
Date: 1-09-17															
Expiration date: 1-09-19															
Cal Standard															
Time: 11:02 AM															
Last Cal															
06/07/2017															
Result															
.097															
Cal Standard															
Time: 12:18 PM															
Last Check															
03/30/2018															
UOM															
BAC															
Serial No.															
16040163															
FC20															
8.7.24															
Calibration / Check Printout															
Lifeloc Technologies															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles    AREA NO: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO: 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	12:25	24	17AP41035	.100								M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														
					Cal Standard									
						Time:								
						Last Cal								
						12/19/2017								
					Result									
					Cal Standard									
						Time:								
						Last Check								
						03/30/2018								
					UOM									
						BAC								
						Serial No.								
						16080072								
						FC20								
						8.7.24								
Calibration / Check Printout														
<b>Lifeloc Technologies</b>														

# 38

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

*4/2*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: *018-762*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
03/30/2018	10/8	34	17AP41035	.100	20								M. Tippett	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
Test Result 0001 03/23/18 23:49 .217 0002 03/23/18 23:53 .217 0003 03/24/18 02:07 .153 0004 03/24/18 02:10 .152 0005 03/25/18 02:30 .090 0006 03/25/18 04:09 .064 0007 03/28/18 21:00 .110 0008 03/28/18 21:03 .124 0009 03/30/18 10:19 .102																
No. Date Time ZBAC Calibration 12/28/17 14:01 .100																
NS4 Serial no: 018762 Version no: 4797																
Push No. 3 to erase OR Power down to retain																

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OP1 051

*443*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 ID. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor-IV  
 PAS SERIAL NO.: 018766

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
03/30/2018	102134	17AP41035		.100	20		.102		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
Test Result 0001 03/25/18 04:08 .081 0002 03/25/18 04:12 .066 0003 03/25/18 21:51 .122 0004 03/25/18 21:54 .133 0005 03/30/18 10:21 .102 Calibration 01/22/16 13:16 .100 No. Date Time ZBMC Version no: 4797 NS4 Serial no: 018766 03/30/18 10:22													
Push No. 3 to erase or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 018691

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/30/2018	1022 34	17AP41035		.100	21						M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

03/30/18 10:23  
 NS4 Serial no: 018691  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 05/28/17 08:26 .100  
 Test Result  
 0001 03/30/18 00:50 .147  
 0002 03/30/18 00:52 .142  
 0003 03/30/18 10:22 .101  
 Push No. 3 to erase  
 or Power down to retain

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

*Handwritten signature*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 017990

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1024	34	17AP41035		.100	21		.108		X		M. Tippett	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.															

03/30/18 10:25  
 NS4 Serial no: 017990  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 11/26/03 11:10 .100  
 Test Result 0001 03/30/18 10:24 .108  
 Push No. 3 to erase  
 or Power down to retain







STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#57

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 041495

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1028	34	17AP41035	.100	19		0.102				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
No. Date Time XBAC Calibration 12/28/17 14:18 .100 Test Result 0001 03/30/18 10:28 .102 Push No. 3 to erase, OR Power down to retain														
03/30/18 10:29 NS4 Serial no: 041495 Version no: 4797														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.:

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1725	34	17AP41035	.100	22	.100			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 1-09-17  
 Expiration date:  
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:59 PM	01/29/2018	.100	.100	12:25 PM	03/30/2018	BAC	16080013	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 53

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 ID. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 082654

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
03/30/2018	1030	34	17AP41035	.100	21	.108				X	M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

03/30/18 10:31  
 NS4 Serial no: 082654  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 12/19/16 11:26 .100  
 Test Result 0001 03/30/18 10:31 .108  
 Push No. 3 to erase  
 or power down to retain



*5475 PAS*

AREA NAME: South Los Angeles  
AREA NO: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Aleo-Sensor IV  
PAS SERIAL NO: **008874**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
03/30/2018	1059	34	17AP41035	.100	20						M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time KBAC Calibration 10/24/17 11:04 .100 Test Result 0001 03/24/18 04:58 .000 0002 03/24/18 05:01 .000 0003 03/30/18 10:39 .102 Push No. 3 to erase. OR Power down to retain													
03/30/18 10:41 NS4 Serial) no: 008874 Version no: 4797													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.