

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/01/2018 17:26

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Invalid sample

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 51717	Download date/time: 04/04/2018	11:18	Record ID 9VV2
Record imported (Y/N) N	Source	Exception	

---

Instrument serial No: 68-012340 Accuracy check date/time: 04/04/2018 10:41

First calibration check:  
Second calibration check:

Interferent detected:  
Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom	Pass	Serial No. Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3494
Check sum 11387	Download date/time: 04/04/2018	11:18	Record ID	9VT6			
Record imported (Y/N) N	Source	Diagnostic					

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/04/2018 10:42

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 021

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 47248	Download date/time: 04/04/2018	11:18	Record ID 9VH9
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340 Accuracy check date/time: 04/04/2018 10:44

First calibration check: .140 Interferent detected: No  
Second calibration check: .141 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 33.98  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 021

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 63560	Download date/time: 04/04/2018	11:18	Record ID 9VI2
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/11/2018 10:12

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass	
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass	
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3476	
Check sum	13163	Download date/time:	04/25/2018	10:05		Record ID	9X54	
Record imported (Y/N)				N				Source Diagnostic

Instrument serial No: 68-012340 Accuracy check date/time: 04/11/2018 10:14

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 014

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards		
Prom		Real Time clock		Motor		Printer		
Ram		Auto calibration		Range		Digital voltmeter		
Check sum	98700	Download date/time:	04/25/2018	10:05		Record ID	9WY8	
Record imported (Y/N)				N				Source Calibration

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/11/2018 10:18

First calibration check: .143 Interferent detected: No  
Second calibration check: .141 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 014

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 23317	Download date/time: 04/25/2018	10:05	Record ID 9WZ7
Record imported (Y/N) N	Source Calibration		

---

Instrument serial No: 68-012340 Accuracy check date/time: 04/18/2018 09:13

First calibration check: Interferent detected:  
Second calibration check: Interferent detected:

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator

---

**Diagnostics check**

Eeprom Pass	Serial No. Pass	Temp Pass	Internal standards Pass
Prom Pass	Real Time clock Pass	Motor Pass	Printer Pass
Ram Pass	Auto calibration Pass	Range Pass	Digital voltmeter 3500
Check sum 54981	Download date/time: 04/25/2018	10:05	Record ID 9X69
Record imported (Y/N) N	Source Diagnostic		

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/18/2018 09:16

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 019

Operator

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 49687	Download date/time: 04/25/2018	10:05	Record ID 9X09
Record imported (Y/N) N		Source Calibration	

---

Instrument serial No: 68-012340 Accuracy check date/time: 04/18/2018 09:18

First calibration check: .141 Interferent detected: No  
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 33.97  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 019

Operator G GOSSAGE 7885

---

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 29373	Download date/time: 04/25/2018	10:05	Record ID 9X10
Record imported (Y/N) N		Source Calibration	

---

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/25/2018 09:37

First calibration check:

Interferent detected:

Second calibration check:

Interferent detected:

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy:

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Pass	Serial No.	Pass	Temp	Pass	Internal standards	Pass
Prom	Pass	Real Time clock	Pass	Motor	Pass	Printer	Pass
Ram	Pass	Auto calibration	Pass	Range	Pass	Digital voltmeter	3472
Check sum	22355	Download date/time:	04/25/2018	10:05		Record ID	9X70
Record imported (Y/N) N				Source Diagnostic			

Instrument serial No: 68-012340 Accuracy check date/time: 04/25/2018 09:39

First calibration check: Not performed

Interferent detected: No

Second calibration check: Not performed

Interferent detected: No

Concentration (%W/V)

Solution lot#:

Soln temp C:

Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 022

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom		Serial No.		Temp		Internal standards	
Prom		Real Time clock		Motor		Printer	
Ram		Auto calibration		Range		Digital voltmeter	
Check sum	58310	Download date/time:	04/25/2018	10:05		Record ID	9X25
Record imported (Y/N) N				Source Calibration			

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.

**Long Beach Police Department  
Forensic Science Services Division**

1400 Canal Ave  
Long Beach, Ca 90813  
PH: (562) 570-7680 Fax: (562) 570-7689

**Intoxilyzer 5000EN Periodic Determination of Accuracy / Maintenance Record**

Instrument serial No: 68-012340 Accuracy check date/time: 04/25/2018 09:40

First calibration check: Not performed Interferent detected: No  
Second calibration check: Not performed Interferent detected: No

Concentration (%W/V) Solution lot#: Soln temp C:  
Simulator Ser No.

Number of people tested since last periodic determination of accuracy: 022

Operator \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 22474	Download date/time: 04/25/2018	10:05	Record ID 9X36
Record imported (Y/N) N		Source Calibration	

Instrument serial No: 68-012340 Accuracy check date/time: 04/25/2018 09:42

First calibration check: .138 Interferent detected: No  
Second calibration check: .140 Interferent detected: No

Concentration (%W/V) .133 Solution lot#: IRS042717 Soln temp C: 34.03  
Simulator Ser No. DR3707

Number of people tested since last periodic determination of accuracy: 000

Operator G GOSSAGE 7885 \_\_\_\_\_

**Diagnostics check**

Eeprom	Serial No.	Temp	Internal standards
Prom	Real Time clock	Motor	Printer
Ram	Auto calibration	Range	Digital voltmeter
Check sum 98600	Download date/time: 04/25/2018	10:05	Record ID 9X43
Record imported (Y/N) N		Source Calibration	

This is a true, complete, and accurate report of the periodic determination of accuracy and diagnostic data from the specified Intoxilyzer 5000EN instrument obtained from the CMI COBRA database.