

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#1

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
04/30/2018	0918	34	17AP41035		.100	22							M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UQM	Serial No.	FC20	Calibration / Check Printout
.100	08:16 AM	10/21/2016	.097	.100	09:18 AM	04/30/2018	BAC	16120188	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

2

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0918	24	17AP41035	.100	21								M. Tippert	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:50 AM	01/29/2018	.096	.100	09:20 AM	04/30/2018	BAC	16120194	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#3

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Lifeloc/FC20
PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	09:28	34	17AP41035	.100	21	.097	.100	NO	NO	X	M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 1-09-17
Expiration date:
1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20
.100	10:22 AM	10/21/2016	.097	.100	09:28 AM	04/30/2018	BAC	16120197	8.7.24

Calibration / Check Printout

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 ID. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/F/C20
 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0926	34	17AP41035	.100	21		104					M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
08:47 AM	04/25/2017	.104	.100	09:27 AM	04/30/2018	BAC	16120195	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

HS

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO: 008539

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0958	34		17AP41035	1.00				0.097			X	M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

No. Date Time %BNC
 Calibration 01/15/16 10:40 100
 Test Result 0001 04/30/18 09:59 .097
 Push No. 3 to erase
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#6

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 160 40161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
04/30/2018	0926 34	17AP41035		.100	21		100					X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:56 AM	04/25/2017	.100	.100	09:33 AM	04/30/2018	BAC	16040161	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

166 40140

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifoloc/FC20	PAS SERIAL NO. 166 40140
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	-----------------------------

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0924 34	17AP41035	.100	22		101		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout Lifoloc Technologies
.100	08:58 AM	04/25/2017	.101	.100	09:29 AM	04/30/2018	BAG	16040140	8.7.24	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet path simulator.

#8

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 008620

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/30/2018	1003	34	17AP41035		100								M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Test Result 0001 04/14/18 18:11 .190 0002 04/14/18 18:13 .172 0003 04/30/18 10:03 .102 Calibration 01/15/16 11:10 .100 No. Date Time %BAC Version no: 4797 AS4 Serial no: 008620 04/30/18 10:04															
Push No. 3 to erase. or Power down to retain															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1108.pdf

#9

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082957

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1061	34	17AP41035	.100	21							M. Tippett	20137	
Simulator prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

04/30/18 10:02 NS4 Serial no: 082957
 Version no: 4797
 No. Date Time XBAC
 Calibration 02/17/14 08:52 .100
 Test Result 0001 04/30/18 10:02 .097
 Push No. 3 to erase, or Power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#10

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 077505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	10:34	17AP41035		100	20			097		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19 No. Date Time ZBAC Version not 4297 NS4 Serial) no: 077505 04/30/18 10:05 Calibration 01/18/18 10:56 .100 Test Result 0001 04/30/18 10:04 .097 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 CHP202K_1108.pdf

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 077479

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VAL UE		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1006	34	17AD41035		100								M Tippett	20137	DEVICE NOT IN LOCKER FOR TESTING

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-09) OPI 051

#12

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS WAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	YES	NO				
04/30/2018	1005	34	17AP41035	100	220								M Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Push No. 3 to erase. OR Power down to retain															
04/30/18 10:06 NS4 Serial no: 082653 Version no: 4297 No. Date Time ZBAC Calibration 02/11/16 10:14 .100 Test Result 0001 04/28/18 00:00 .187 0002 04/28/18 00:03 .181 0003 04/28/18 22:35 .236 0004 04/28/18 22:38 .248 0005 04/28/18 22:41 .031 0006 04/30/18 10:06 .101															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 ID. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008 709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	TEST #2 (Optional)			TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
04/30/2018	10:23	34	17AD41035		.100	21			.100				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.																

Test Result
 0001 04/21/18 20:52 .000
 0002 04/22/18 02:49 .147
 0003 04/22/18 02:52 .154
 0004 04/28/18 03:50 .089
 0005 04/29/18 02:11 .067
 0006 04/29/18 02:13 .056
 0007 04/29/18 02:20 .000
 0008 04/29/18 02:59 .154
 0009 04/29/18 03:01 .147
 0010 04/29/18 06:08 .136
 0011 04/29/18 06:11 .135
 0012 04/30/18 10:13 .100

Calibration
 03/17/16 09:40 .100

No. Date Time ZBAC

NS4 Serial no: 008709
 Version no: 4797

Push No. 3 to erase,
 OR Power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#114

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008 714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/30/2018	1614 34		17AP41035	100	21		098			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time ZBAC Calibration 05/14/17 09:58 .100 Test Result 0001 04/30/18 10:17 .098 Version not 4797 NS4 Serial not 008714 04/30/18 10:17													
Push No. 3 to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#16

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008888

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1014 34		17AP41035	.100		2-1		101	X			M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19</p> <p>No. Date Time ZBNC Calibration 03/12/16 09:44 .100 Test Result 04/30/18 10:15 NS4 Serial no: 008888 Version no: 4797</p> <p>PAS RE-TEST RESULTS: 101</p> <p>CALIBRATION REQUIRED: X</p> <p>PAS ACCURACY CHECK RESULTS: TEST #1 (Optional) 101, TEST #2 (Optional) 101</p> <p>PAS TEMP (°C): TEST #1, TEST #2 (Optional)</p> <p>OR Power down to retain. Push No. 3 to erase.</p>														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#117

AREA NAME
 South Los Angeles

AREA NO.
 530

COORDINATOR NAME
 Kirk Lee

I.D. NUMBER
 17046

PAS MAKE/MODEL
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.
 008783

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK/RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS			
04/30/2018	1019 54		17AP41035	.100	19		099					M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
No. Date Time XBAC Calibration 05/14/17 10:07.100 Test Result 0001 04/30/18 10:20.099 Push No. 3 to erase or Power down to retain														
04/30/18 10:21 NS4 Serial no: 006783 Version no: 4797														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1020	34	17AP41035		.100	20		.097		X		M. Tippet	20137		
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.															

04/30/18 18:21
 NS4 Serial no: 082657
 Version no: 4797
 No. Date Time ZBMC
 Calibration 05/21/17 08:18 .100
 Test Result 0001 04/30/18 18:21 .097
 Push No. 3 to erase
 OR Power down to retain

#19

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS WAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008810

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1025	34	17AP41035	.100	20		098		X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

04/30/18 10:26
 NS4 Serial no: 008810
 Version no: 4797
 No. Date Time %BAC
 Calibration 05/21/17 08:28 .100
 Test Result 0001 04/30/18 10:26 .098
 Push No. 3 to erase,
 OR Power down to retain

4-20

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS WAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: *018 684*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1018	34	17AP41035								M Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

04/30/18 10:28
 NS4 Serial no: 018584
 Version no: 4797
 No. Date Time ZBAC
 Calibration 01/12/16 14:36 .093
 Test Result 0001 04/29/18 06:05 .092
 0002 04/29/18 06:08 .089
 0003 04/30/18 10:28 .102
 Push No. 3 to erase
 OR Power down to retain

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

21

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)					
04/30/2018	1030	34	17AP41035	.100						M. Tippet	20137	DEVICE NOT IN WORKING FOR TESTING

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#22

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Lifeloc/FC20	16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0928 34	17AP41035	.100	22		096		X			M. Tippet	20137	

Solution prepared by:

Cesar A. Mejia

Date: 1-09-17

Expiration date:

1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:14 AM	08/14/2017	.096	.100	09:28 AM	04/30/2018	BAC	16040162	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0931 34	17AP41035	.100	22		101				X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	09:50 AM	04/25/2017	.101	.100	09:32 AM	04/30/2018	BAC	16120177	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1050	54	17AP41035	.100	20		TEST #1	TEST #2	NO	NO		M. Tippet	20137	
Simulator prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
04/30/18 09:32 NS4 Serial NO: 008848 Version NO: 4797 No. Date Time %BAC Calibration 05/21/17 09:30 .100 Test Result 0001 04/19/18 11:50 .105 0002 04/30/18 09:31 .100 Push No. 3 to erase. or Power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 077591

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1051	34	17AP41035	.100	20			099				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
<div style="border: 1px solid black; padding: 5px; width: fit-content;"> <p>NS4 Serial no: 077591 Version no: 4797 No. Date Time ZDAC Calibration 05/21/17 08:35 .100 Test Result 0001 04/30/18 10:32 .099 Push No. 3 to erase, OR Power down to retain</p> </div>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#26

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifoloc/FC20
 PAS SERIAL NO.: 16080065

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/30/2018	093634		17AP41035	.100	21		.102				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifoloc Technologies
.100	1:54 PM	12/19/2017	.102	.100	09:37 AM	04/30/2018	BAC	16080065	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 097573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/30/2018	1055 34	17AP41035	.100	21	.094						M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
No. Date Time %BAC Calibration 01/18/18 11:23 .100 Test Result 0001 04/30/18 10:34 .094 Push No. 3 to erase, or power down to retain													
PAS4 Serial no: 077573 Version no: 4797													
04/30/18 10:34													
01/18/18 11:23 .100													
Test Result 0001 04/30/18 10:34 .094													
Push No. 3 to erase, or power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 083051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1041 34		17AP41035	.100	21			094				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19														
No. Date Time XBAC Calibration 12/28/17 13:29 .100 Test Result 0001 04/30/18 10:42 .094 Push No. 3 to erase OR power down to retain														
04/30/18 10:42 N34 Serial no: 083051 Version no: 4797														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 0088947

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

30

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1043	34	17AP41035	100	20		101			X		M. Tippet	20137	

04/30/18 10:47
 NS4 Serial no: 008897
 Version no: 4797
 No. Date Time XBAC
 Calibration 01/18/12 16:13 .100
 Test Result 0001 04/30/18 10:47 .101
 Push No. 2 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS WAKEMODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2					TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1049	34	17A141035			.100	2	1						X	M. Tippett	20137	
<p>Solution prepared by: _____</p> <p>Cesar A. Mejia</p> <p>Date: 1-09-17</p> <p>Expiration date: 1-09-19</p>																	
<p>04/30/18 10:50 NS4 Serial no: 018624 Version no: 4797 No. Date Time KBAC Calibration 02/01/16 11:02 .100 Test Result 0001 04/30/18 10:49 .100 Push No. 3 to erase OR Power down to retain</p>																	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

32

AREA NAME: South Los Angeles | AREA NO: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
04/30/2018	09:39	34	17AP41035	.100	22		.098					X	ML Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	BOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	10:41 AM	08/14/2017	.098	.100	09:40 AM	04/30/2018	BAC	16080071	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#33

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.									
South Los Angeles		530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077498									
DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)	PAS TEST #1 (Optional)	PAS TEST #2 (Optional)	PAS CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
04/30/2018	1047	34	17AP41035	100	20			102			X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															

04/30/18 10:48
NS4 Serial No: B77488
Version no: 4797
No. Date Time XMAC
Calibration 02/11/16 10:52 .100
Test Result 0001 04/30/18 10:47 .102
Push No. 3 to erase
OR Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REPEAT RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	NO	RESULTS					
04/30/2018	1054	34	17AP41035		1.00		20							X	M. Tippett	20137	

04/30/18 10:56
 NS4 Serial no: 008877
 Version no: 4797
 No. Date Time KBAC
 Calibration 05/28/17 08:49 .100
 Test Result 0001 04/30/18 10:55 .094
 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

36

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0932	34	17AP41035	.100	24	.092			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	11:02 AM	06/07/2017	.092	.100	09:33 AM	04/30/2018	BAC	16040163	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME: South Los Angeles | AREA NO: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS WAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO: 018626

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/30/2018	1053	34	17AP41035		.100	2	1	.099			X		M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia</p> <p>Date: 1-09-17</p> <p>Expiration date: 1-09-19</p>															
<p>No. Date Time XMAC Version: 07/15/16 15:21 .100 NS4 Serial) not 018626 Test Result 0001 04/30/18 10:53 .099 Push No. 3 to erase. OR power down to retain</p>															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.099 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

38

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	0944 34		17AP41035	.100	21							M Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	2:05 PM	12/19/2017	.101	.100	09:45 AM	04/30/2018	BAC	16080072	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

39

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alico-Sensor IV	083030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
04/30/2018	1655	34	17AP41035	.100	21									M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																
No. Date Time XMNC Calibration 12/20/17 13:51 .100 Test Result 0001 04/30/18 10:56 .094 Push No. 3 to erase, OR power down to retain																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K_1108.pdf

#410

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV

PAS SERIAL NO: 018692

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
04/30/2018	105834	17AP41035	.100	21		.101		X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

04/30/18 10:59
 #34 Serial no: 018692
 Version no: 4797
 No. Date Time ZBAC
 Calibration 02/01/16 11:30 .100
 Test Result 0001 04/30/18 10:58 .101
 Push No. 3 to erase,
 OR Power down to retain

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#42

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 018762

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	10234		17AP41035	100		21		093				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 1-09-17														
Expiration date: 1-09-19														
No. Date Time ZBNC Version not 4797 NS4 Serial no: 018762 Calibration 12/28/17 14:01 .100 Test Result 0001 04/11/18 00:33 .172 0002 04/11/18 00:36 .179 0003 04/30/18 11:02 .093 Push No. 3 to erase, OR Power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 018 766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	YES	NO				
04/30/2018	1104	34	17AP41035	.100	21								M. Tippett	20137	
Simulator prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Test Result 0001 04/21/18 01:23 .090 0002 04/21/18 01:27 .074 0003 04/21/18 01:30 .072 0004 04/21/18 23:16 .022 0005 04/21/18 23:19 .019 0006 04/22/18 23:15 .136 0007 04/22/18 23:18 .122 0008 04/23/18 02:47 .171 0009 04/23/18 02:50 .154 0010 04/23/18 02:52 .161 0011 04/27/18 20:28 .275 0012 04/27/18 20:32 .215 0013 04/27/18 20:35 .190 0014 04/29/18 04:02 .113 0015 04/29/18 04:05 .127 0016 04/29/18 23:23 .102 0017 04/29/18 23:26 .106 0018 04/30/18 11:05 .094															
Push No. 3 to erase or Power down to retain															
04/30/18 11:05 NS4 Serial no: 018766 Version no: 4797 No. Date Time XBNC Calibration 01/22/16 13:16 .100															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#44

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
ID. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 018691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	YES	NO				
04/30/2018	1106	34	17ADP41035	.100	21								M. Jippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19															
Test Result 0001 04/24/18 00:59 .025 0002 04/24/18 03:31 .173 0003 04/24/18 03:34 .184 0004 04/25/18 02:28 .032 0005 04/25/18 21:04 .092 0006 04/25/18 21:06 .097 0007 04/26/18 01:00 .114 0008 04/26/18 01:09 .129 0009 04/26/18 03:04 .177 0010 04/26/18 03:08 .172 0011 04/30/18 11:00 .094															
Calibration 05/28/17 08:26 .100 No. Date Time XIMC Version: 4797 AS4 Serial no: 018091 04/30/18 11:07															
Push No. 3 to erase OR Power down to retain															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#95

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 017990

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)		YES	NO				
04/30/2018	1108 34		17AP41035	.100	21		101	X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19 No. Date Time ZBAC Calibration 11/26/03 11:10 .100 Test Result 0001 04/30/18 11:09 .101 Push No. 3 to erase OR Power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

446

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO: *16120189*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEMP (°C)	TEST #1	TEST #2 (Optional)	YES				
04/30/2018	<i>1108 34</i>		17AP41035	.100	<i>22</i>		<i>.093</i>			<i>X</i>		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
		09/01/2017	.093	.100	11:08 AM	04/30/2018	BAC	16120189	8.7.24	
	12:17 PM									

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 041486

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
04/30/2018	1113	34	17AP41035	100	21		099		X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

04/30/18 11:17
 NS4 Serial no: 041486
 Version no: 4797
 No. Date Time ZINC
 Calibration 03/08/06 13:56 100
 Test Result 0001 04/30/18 11:16 .099
 Push No. 3 to erase
 or power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#57

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 04/1495

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
04/30/2018	1116	34	17AP41035	.100	20		.095		X			M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19 No. Date Time 2300 Calibration 12/28/17 14:18 .100 Test Result 0001 04/30/18 11:16 .095 Push No. 3 to erase or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#52

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.					TEST #1	TEST #2 (Optional)	YES	NO				
04/30/2018	0945		17AP41035	.100	22								M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 1-09-17
 Expiration date:
 1-09-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:59 PM	01/29/2018	.099	.100	09:45 AM	04/30/2018	BAC	16080013	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS																																																
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES					NO																																															
04/30/2018	118 34	17AP41035	.100	21	.100	X					M. Tippett	20137																																																	
Solution prepared by: Cesar A. Mejia Date: 1-09-17 Expiration date: 1-09-19																																																													
<table border="1"> <tr> <td>No.</td> <td>Date</td> <td>Time</td> <td>ZBAC</td> </tr> <tr> <td>04/30/18</td> <td>11:19</td> <td></td> <td></td> </tr> <tr> <td colspan="4">N54 Serial no: 082654</td> </tr> <tr> <td colspan="4">Version no: 4797</td> </tr> <tr> <td colspan="4">Calibration</td> </tr> <tr> <td>12/19/16</td> <td>11:26</td> <td></td> <td>.100</td> </tr> <tr> <td colspan="4">Test Result</td> </tr> <tr> <td>0001</td> <td>04/21/18</td> <td>01:58</td> <td>.246</td> </tr> <tr> <td>0002</td> <td>04/21/18</td> <td>01:56</td> <td>.234</td> </tr> <tr> <td>0003</td> <td>04/30/18</td> <td>11:19</td> <td>.100</td> </tr> <tr> <td colspan="4">Push No. 3 to erase.</td> </tr> <tr> <td colspan="4">OR Power down to retain</td> </tr> </table>														No.	Date	Time	ZBAC	04/30/18	11:19			N54 Serial no: 082654				Version no: 4797				Calibration				12/19/16	11:26		.100	Test Result				0001	04/21/18	01:58	.246	0002	04/21/18	01:56	.234	0003	04/30/18	11:19	.100	Push No. 3 to erase.				OR Power down to retain			
No.	Date	Time	ZBAC																																																										
04/30/18	11:19																																																												
N54 Serial no: 082654																																																													
Version no: 4797																																																													
Calibration																																																													
12/19/16	11:26		.100																																																										
Test Result																																																													
0001	04/21/18	01:58	.246																																																										
0002	04/21/18	01:56	.234																																																										
0003	04/30/18	11:19	.100																																																										
Push No. 3 to erase.																																																													
OR Power down to retain																																																													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.																																																													

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008874

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

SGT's PAS

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED	PAS RELEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.							
04/30/2018	11:25	34	17AP41035					M. Tippett	20137	
Solution prepared by: Cesar A. Mejia										
Date: 1-09-17										
Expiration date: 1-09-19										
Test Result 0001 04/14/18 00:41 .241 0002 04/16/18 03:41 .108 0003 04/18/18 02:26 .156 0004 04/18/18 02:28 .133 0005 04/18/18 02:28 .133 0006 04/19/18 13:37 .099 0007 04/20/18 04:26 .000 0008 04/20/18 04:28 .000 0009 04/22/18 02:26 .135 0010 04/22/18 02:22 .128 0011 04/23/18 00:34 .167 0012 04/23/18 00:35 .166 0013 04/26/18 04:17 .142 0014 04/26/18 04:20 .104 0015 04/26/18 04:22 .121 0016 04/29/18 04:02 .132 0017 04/29/18 04:04 .100 0018 04/29/18 04:07 .128 0019 04/29/18 21:46 .034 0020 04/29/18 21:49 .033 0021 04/30/18 11:26 .093										
No. Date Time ZINC Calibration 10/24/17 11:04 .100 Version no: 4797 AS4 Serial no: 000074 04/30/18 11:26										
Push No. 3 to erase or Power down to retain										

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.