

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#1

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16120188 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0822 | 39 | 16FP41018 | .100 | 23 | | | | | | Lee | 17046 | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120188 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:20 AM |
| Cal Standard | .100 |
| Result | .104 |
| Last Cal | 10/21/2016 |
| Time: | 08:16 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#2

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16120194 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|---------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------------|----------|----------|---------------------------|
| | | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | | | | |
| 05/18/2018 | 0825 | 34 | 16FP41018 | .100 | 21 | .099 | | | X | | Lee | 17046 | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| | |
|------------------------------|------------|
| Lifeloc Technologies | |
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120194 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:21 AM |
| Cal Standard | .100 |
| Result | .099 |
| Last Cal | 01/29/2018 |
| Time: | 08:50 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of
** If using dry gas for accuracy checks, note the canister seri:

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#3

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16120197 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0827 | 34 | 16FP41018 | .100 | 22 | .099 | | | | X | | Lee | 17046 | |

Solution prepared by:

Cesar A. Mejia

Date: 06-06-16

Expiration date:

06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120197 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:30 AM |
| Cal Standard | .100 |
| Result | .099 |
| Last Cal | 10/21/2016 |
| Time: | 10:22 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the

** If using dry gas for accuracy checks, note the canister serial number.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16120195 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|---------------|-----------------------------|--------------------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | TEMP (#1) | TEMP (#2) | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0829 | 34 | VALUE | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | Lee | 17046 | |
| | | | .100 | 16FP41018 | 22 | | .105 | | | X | | | |

Solution prepared by:

Cesar A. Mejia

Date: 06-06-16

Expiration date:

06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120195 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:28 AM |
| Cal Standard | .100 |
| Result | .105 |
| Last Cal | 04/25/2017 |
| Time: | 08:47 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of t

** If using dry gas for accuracy checks, note the canister seria

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#5

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 008559 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------------------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 0840 | 34 | 16FP41018 | | .100 | 21 | | | | | X | K. Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | Solution prepared by: | | | | | | | | | | |
| | | | | Cesar A. Mejia | | | | | | | | | | |
| | | | | Date: 06-16-16 | | | | | | | | | | |
| | | | | Expiration date: | | | | | | | | | | |
| | | | | 06-06-18 | | | | | | | | | | |

05/18/18 08:44
AS4 Serial not 008559
Version not 4797

No. Date Time ZBAC
Calibration
01/15/16 10:40 .100

Test Result
0001 05/18/18 08:43 .101

Push No. 3 to erase

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#6

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16040161 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|--------------------|-------|---------------|--------------------|-----------------------------|----|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP #1 | TEMP #2 (Optional) | | TEST #1 | TEST #2 (Optional) | YES | NO | YES | NO | | | | |
| 05/18/2018 | 0844 | 34 | 16FP41018 | | .100 | 22 | | 103 | | X | | | Lee | 17046 | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16040161 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:48 AM |
| Cal Standard | .100 |
| Result | .103 |
| Last Cal | 04/25/2017 |
| Time: | 08:56 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed within 150 tests or 10 days.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI051

#7

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16040140 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|-----------------------------------|---------|---------------|-----------------------------|--------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 0846 | 34 | 16FP41018 | .100 | 22 | | | | | | X | Lee | 17046 | |

Solution prepared by:
 Cesar A. Mejia
 Date: 06-06-16
 Expiration date:
 06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16040140 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 08:45 AM |
| Cal Standard | .100 |
| Result | .103 |
| Last Cal | 04/25/2017 |
| Time: | 08:58 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of
 ** If using dry gas for accuracy checks, note the canister seri

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#9

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 082957 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|---|-------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0851 | 34 | 16FP41018 | .100 | 21 | .101 | | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | |

05/18/18 08:52
AS4 Serial no: 082957
Version no: 4797

No. Date Time XBAC
Calibration
02/17/14 08:52 .100

Test Result
0001 05/18/18 08:51 .101

Push No. 3 to erase.

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#10

| AREA NAME | AREA NO. | COORDINATOR NAME | I.D. NUMBER | PAS MAKE/MODEL | PAS SERIAL NO. | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|-------------------|----------|------------------|---|-----------------------------|----------------|-----------------------------------|---------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | | | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| South Los Angeles | 530 | Kirk Lee | 17046 | Intoximeters/Alco-Sensor IV | 077505 | | | | | | | | | | | | |
| 05/18/2018 | 0853 | 34 | 16FP41018 | .100 | 21 | .101 | | | | | | X | | K. Lee | 17046 | | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | | | | |
| | | | <p>Push No. 3 to erase, or power down to retain</p> <hr/> <p>Test Result 0001 05/18/18 08:54 .101</p> <hr/> <p>No. Date Time XBAC Calibration 01/18/18 10:56 .100</p> <hr/> <p>05/18/18 08:54 AS4 Serial no: 077505 Version no: 4797</p> <hr/> <p>or power down to retain</p> | | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#12

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|

082653

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|--|-----------|-------|---------------|--------------------|-----------------------------|----|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP (#1) | TEMP (#2) | | TEST #1 | TEST #2 (Optional) | YES | NO | YES | NO | | | | |
| 05/18/2018 | 0858 | 34 | 16FP41018 | | .100 | 20 | | | | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | | | |
| | | | <p>05/18/18 08:57 A94 Serial no: 082653 Version no: 4797</p> <p>No. Date Time ZBAC Calibration 02/11/16 10:14 .100</p> <p>Test Result 0001 05/18/18 08:57 .106</p> <p>Push No. 3 to erase. or power down to retain</p> | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 008709 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|-----------|---|---------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|---------------------------|
| | | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 0900 | 34 | 16FP41018 | | .100 | 21 | | | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | |
| | | | Push No. 3 to erase, or power down to retain | | | | | | | | | | | |
| | | | Test Result 0001 05/18/18 09:02 .107 | | | | | | | | | | | |
| | | | Calibration 03/17/16 09:40 .100 | | | | | | | | | | | |
| | | | Version no: 4797 A84 Serial no: 008709 05/18/18 09:02 | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#14
008714

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 008714 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|---------------|---------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP (#1) | TEMP (#2) | VALUE | LOT NO. | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0903 | 34 | 21 | .100 | 16FP41018 | 21 | .104 | | X | | K. Lee | 17046 | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

```

05/18/18 09:06
AS4 Serial no: 008714
Version no: 4797

No. Date Time %BAC
Calibration
05/14/17 09:50 .100

Test Result
0001 05/18/18 09:05 .104

Push No. 3 to erase.
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

18

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|-----------|---|---------|-------|--------------------|---------|-----------------------------|-----|----------------------|---|--------------------|----------|----------|----------------------------|--|
| | | | LOT NO. | TEMP #1 | | TEMP #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | | | |
| 05/18/2018 | 0906 | 34 | 16FP41018 | | .100 | 21 | | | | | X | | K. Lee | 17046 | | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | | | | |

05/18/18 09:10
 AS4 Serial no: 082657
 Version no: 4797

No. Date Time ZBAC
 Calibration
 05/21/17 08:18 .100

Test Result
 0001 05/18/18 09:09 .102

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#21

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 008635 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|--|---------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 0914 | 34 | 16FP41018 | .100 | 20 | | 106 | | X | | K. Lee | 17046 | | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | |
| | | | <pre> 05/18/18 09:16 AS4 Serial no: 008635 Version no: 4797 No. Date Time XBAC Calibration 05/21/17 08:23 .100 Test Result 0001 05/11/18 22:19 .112 0002 05/11/18 22:22 .118 0003 05/12/18 04:01 .065 0004 05/12/18 04:04 .061 0005 05/14/18 05:05 .151 0006 05/14/18 05:38 .154 0007 05/14/18 09:15 .106 Push no. 3 to erase. or power down to retain </pre> | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16040162 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | TEMP (#1) | TEMP (#2) | TEST #1 | TEST #2 | TEST #1 | TEST #2 | YES | NO | | | | |
| 05/18/2018 | 0916 | 34 | VALUE | .100 | TEST #1 | 22 | TEST #2 | | TEST #1 | 099 | TEST #2 | Lee | 17046 | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16040162 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 09:12 AM |
| Cal Standard | .100 |
| Result | .099 |
| Last Cal | 08/14/2017 |
| Time: | 10:14 AM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the t
** If using dry gas for accuracy checks, note the canister serial nu

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#24

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 008848 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|---|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0923 | 34 | 16FP41018 | .100 | 20 | | 105 | | X | | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | |

05/18/18 08:26
 AS4 Serial no: 008848
 Version no: 4797

No. Date Time ZBAC
 Calibration
 05/21/17 08:30 .100

Test Result
 0901 05/18/18 08:26 .105

Push no. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#276

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16080065 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|-------|-----------|-----------------------------------|---------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 09:28 | 34 | 16FP41018 | .100 | 21 | .104 | | | X | | Lee | 17046 | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| Lifeloc Technologies | |
|------------------------------|------------|
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080065 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 09:26 AM |
| Cal Standard | .100 |
| Result | .104 |
| Last Cal | 12/19/2017 |
| Time: | 1:54 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the a
** If using dry gas for accuracy checks, note the canister serial nu.....

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#28

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 08305

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|--|-------|-----------|-----------------------------------|-------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 09:32 | 34 | 16FP41018 | .100 | 21 | .098 | | X | | | K. Lee | 17046 | |
| Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | | | |
| <pre> 05/18/18 09:32 PAS Serial no: 083051 Version no: 4797 No. Date Time %BAC Calibration 12/28/17 13:29 .100 Test Result 0981 05/18/18 09:31 .098 Push No. 8 to erase, or power down to retain </pre> | | | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#30
 008897

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|-------|-----------|---|-----------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------------|----------|----------|-------------------------------|
| | | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | | | | |
| 05/18/2018 | 09:34 | 34 | | 16FP41018 | | 20 | .105 | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |
| | | | | | | | | | | | | | |

05/18/18 09:39
 AC4 Serial not 008897
 Version not 4797

No. Date Time XBAC
 Calibration
 01/18/12 16:13 .100

Test Result
 0001 05/18/18 09:39 .105

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#31

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. |
|--------------------------------|-----------------|------------------------------|----------------------|---|----------------|

018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|---|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0939 | 34 | 16FP41018 | .100 | 21 | | | | X | | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | | |

05/18/18 09:42
AS4 Serial no: 018624
Version no: 4797

No. Date Time %BAC
Calibration
02/01/16 11:02 .100

Test Result
001 05/18/18 09:41 .103

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#32

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc/FC20 | PAS SERIAL NO. 16080071 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|---------|-------|---------------|-----------------------------|--------------------|----------------------|--------------------|--------------------|----------|----------|----------------------------|
| | | | TEMP (°C) | LOT NO. | | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | | | | |
| 05/18/2018 | 0941 | 31 | 16FP41018 | | .100 | 22 | | | | | | Lee | 17046 | |

| | | | | | | | | | | | | | |
|------------------------------|--|--|--|--|--|------------|--|--|--|--|--|--|--|
| Lifeloc Technologies | | | | | | | | | | | | | |
| Calibration / Check Printout | | | | | | | | | | | | | |
| FC20 | | | | | | 8.7.24 | | | | | | | |
| Serial No. | | | | | | 16080071 | | | | | | | |
| UOM | | | | | | BAC | | | | | | | |
| Last Check | | | | | | 05/18/2018 | | | | | | | |
| Time: | | | | | | 09:40 AM | | | | | | | |
| Cal Standard | | | | | | .100 | | | | | | | |
| Result | | | | | | .102 | | | | | | | |
| Last Cal | | | | | | 08/14/2017 | | | | | | | |
| Time: | | | | | | 10:41 AM | | | | | | | |
| Cal Standard | | | | | | .100 | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the :
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet gas simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#36

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifeloc FC20 | PAS SERIAL NO. 16040163 |
|--------------------------------|-----------------|------------------------------|----------------------|--------------------------------|----------------------------|

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|---------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP (#1) | LOT NO. | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0949 | 24 | 16PP41018 | .100 | 23 | 1096 | | | | | K. Lee | 17046 | |

| | | | | | | | | | | | | |
|--|------------|--|--|--|--|--|--|--|--|--|--|--|
| <p>Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18</p> | | | | | | | | | | | | |
| <h2>Lifeloc Technologies</h2> <p>Calibration / Check Printout</p> | | | | | | | | | | | | |
| FC20 | 8.7.24 | | | | | | | | | | | |
| Serial No. | 16040163 | | | | | | | | | | | |
| UOM | BAC | | | | | | | | | | | |
| Last Check | 05/18/2018 | | | | | | | | | | | |
| Time: | 09:46 AM | | | | | | | | | | | |
| Cal Standard | .100 | | | | | | | | | | | |
| Result | .096 | | | | | | | | | | | |
| Last Cal | 06/07/2017 | | | | | | | | | | | |
| Time: | 11:02 AM | | | | | | | | | | | |
| Cal Standard | .100 | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the ac
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#37

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---------------------------------|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL ALCOSESION IV | PAS SERIAL NO. 018626 |
|--------------------------------|-----------------|------------------------------|----------------------|---------------------------------|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|---------|---|---------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 0953 34 | 16FP41018 | | .100 | 21 | .102 | | | X | | Lee | 17046 | |
| | | Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18 | | | | | | | | | | | |
| | | <p>Test Result 0091 05/19/18 09:32 .182</p> <p>Calibration 05/19/18 15:21 .180</p> <p>No. Date Time ZBAC 05/18/18 09:53 004 Serial no: 018626 Version no: 4797</p> <p>Each 1.0 to +rase or power down to retain</p> | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16080072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|---------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0955 | 34 | 16FP41018 | | 21 | | | | | X | | Lee | 17046 | |

| | |
|------------------------------|------------|
| Lifeloc Technologies | |
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16080072 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 09:58 AM |
| Cal Standard | .100 |
| Result | .102 |
| Last Cal | 12/19/2017 |
| Time: | 2:05 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the accuracy.
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

39

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 083030 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

| DATE | TIME | TEMPERATURE (°C) | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/COMMENTS |
|------------|------|------------------|-----------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|---------------------------|
| | | TEMP (#1) | TEMP (#2) | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 0957 | 34 | 21 | .100 | 21 | 21 | | | | | | K. Lee | 17046 | |
| | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | |

```

05/18/18 09:59
AS4 Serial no: 083030
Version          no: 4797

No. Date Time ZBAC
Calibration
      17/28/17 13:51 .100

Test Result
9991 17/19/18 09:37 .099

-----
Push No. 8 to erase,
or power down to retain
  
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#42

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 018762 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------------------------------|---|---------------|--------------------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | TEMP (°C) | LOT NO. | VALUE | TEST #1 (Optional) | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 1003 | 34 | 16FP41018 | .100 | 21 | .099 | | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | |

```

05/18/18 10:05
PAS Serial no: 918762
Version: not 4797

Date Time ZBAC
10/20/17 14:01 .100

10/20/17
000: 04/12/16 01:09 .105
000: 04/12/16 01:18 .107
000: 04/12/16 01:24 .088
000: 04/12/16 01:34 .099
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#43

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 018766 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | VALUE | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS | |
|------------|------|-----------|---|---------|-------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|--|
| | | | TEMP (°C) | LOT NO. | | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | | |
| 05/18/2018 | 1005 | 34 | 16FP41018 | | .100 | 21 | | | | | | | K. Lee | 17046 | | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | | | | |

```

05/18/18 10:10
AS4 Serial no: 018766
Version no: 4797

No. Date Time XBAC
Calibration
01/22/16 13:16 .100

Test Result
0001 05/18/18 10:07 .101

Push No. 3 to erase,
or power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#44
018691

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Aico-Sensor IV | PAS SERIAL NO. 018691 |
|--------------------------------|-----------------|------------------------------|----------------------|---|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|---|-------|---------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | LOT NO. | VALUE | | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 1010 | 34 | 16FP41018 | .100 | 21 | .099 | | | X | | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia | | | | | | | | | | |
| | | | Date: 06-16-16 | | | | | | | | | | |
| | | | Expiration date: 06-06-18 | | | | | | | | | | |

05/18/18 10:13
AS4 Serial no: 018691
Version no: 4797

No. Date Time %BAC
Calibration
05/28/17 08:26 .100

Test Result
0001 05/18/18 10:12 .099

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

46

16120189

CHP 202K (Rev. 8-08) OPI 051

| | | | | | |
|--------------------------------|-----------------|------------------------------|----------------------|-------------------------------|----------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | I.D. NUMBER 17046 | PAS MAKE/MODEL Lifloc/FC20 | PAS SERIAL NO. |
|--------------------------------|-----------------|------------------------------|----------------------|-------------------------------|----------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|-----------------------------------|-----------|---------------|--------------------|-----------------------------|--------------------|----------------------|----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | NO | | | | |
| 05/18/2018 | 1015 | 34 | VALUE | 16FP41018 | 23 | .095 | | | | X | Lee | 17046 | | |

Solution prepared by:
Cesar A. Mejia
Date: 06-06-16
Expiration date:
06-06-18

| | |
|------------------------------|------------|
| Lifloc Technologies | |
| Calibration / Check Printout | |
| FC20 | 8.7.24 |
| Serial No. | 16120189 |
| UOM | BAC |
| Last Check | 05/18/2018 |
| Time: | 10:13 AM |
| Cal Standard | .100 |
| Result | .095 |
| Last Cal | 09/01/2017 |
| Time: | 12:17 PM |
| Cal Standard | .100 |

* If the reading is within plus or minus (±) 0.010 percent of the act.....
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#50

| | | | | | |
|--------------------------------|-----------------|------------------------------|---------------------|---|--------------------------|
| AREA NAME South Los Angeles | AREA NO. 530 | COORDINATOR NAME Kirk Lee | ID. NUMBER 17046 | PAS MAKE/MODEL Intoximeters/Alco-Sensor IV | PAS SERIAL NO. 041486 |
|--------------------------------|-----------------|------------------------------|---------------------|---|--------------------------|

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

| DATE | TIME | TEMP (°C) | SIMULATOR SOLUTION (OR DRY GAS)** | | PAS TEMP (°C) | | PAS ACCURACY CHECK RESULTS* | | CALIBRATION REQUIRED | | PAS RETEST RESULTS | OPERATOR | I.D. NO. | REPAIRS/PROBLEMS/ COMMENTS |
|------------|------|-----------|---|---------|---------------|---------|-----------------------------|---------|----------------------|-----|--------------------|----------|----------|----------------------------|
| | | | TEMP | LOT NO. | VALUE | TEST #1 | TEST #2 (Optional) | TEST #1 | TEST #2 (Optional) | YES | | | | |
| 05/18/2018 | 1030 | 34 | 16FP41018 | | .100 | 20 | | | | | X | K. Lee | 17046 | |
| | | | Solution prepared by: Cesar A. Mejia Date: 06-16-16 Expiration date: 06-06-18 | | | | | | | | | | | |
| | | | Push No. 3 to erase, or Power down to retain. | | | | | | | | | | | |
| | | | Test Result 0001 05/18/18 10:38 .102 | | | | | | | | | | | |
| | | | Calibration 03/08/06 13:56 .100 | | | | | | | | | | | |
| | | | No. Date Time %BAC 05/18/18 10:38 AS4 Serial no: 041486 Version no: 4797 | | | | | | | | | | | |
| | | | or Power down to retain | | | | | | | | | | | |

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

