



#2

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Lifeloc/FC20  
PAS SERIAL NO.: 161 20194

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0858	34	16FP41018	100	21	097	100	X			M. Tippett	20137	
Solution prepared by: Cesar A. Mejia													
Date: 06-06-16													
Expiration date: 06-06-18													
		Cal Standard											
				Time: 08:50 AM		Last Cal		01/29/2018		Result			
				Time: 08:59 AM		Last Check		05/29/2018		UOM		BAC	
				Serial No.		16120194		FC20		8.7.24		Calibration / Check Printout	
<b>Lifeloc Technologies</b>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

43

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK/RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEMP (°C)	TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	090734	161	PA1018	100	21	100					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date: 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
100	10:22 AM	10/21/2016	100	100	09:08 AM	05/29/2018	BAC	16120197	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.0990 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#4

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0904	34	16FP41018	.100	21						X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	08:47 AM	04/25/2017	.108	.100	09:05 AM	05/29/2018	BAC	16120195	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.0990 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*AS*

AREA NAME  
South Los Angeles

AREA NO.  
530

COORDINATOR NAME  
Kirk Lee

PAS MAKE/MODEL  
Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
**008559**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	TEST #1				TEST #2	YES				NO						
05/29/2018	1000 34	16FP41018			1.00	21							<b>X</b>			M. Tippet.	20137		
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18																			
<p style="text-align: center;"> <b>MS4 Serial no: 008559</b>                      Version no: 4797                      No. Date Time: KBAC                      Calibration 01/15/16 10:40 .100                      Test Result 0001 05/29/18 10:07 .101                      Push No. 3 to erase,                      or Power down to retain                 </p>																			

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16040161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								YES	NO				
05/29/2018	6:11:34		16FP41018	.100	21		104		X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
	08:56 AM	04/25/2017	.104	.100	09:12 AM	05/29/2018	BAC	16040161	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 7

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 ID. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/29/2018	090734		16FP41018	.100	21			.105					M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date: 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	08:58 AM	04/25/2017	.105	.100	09:07 AM	05/29/2018	BAC	16040140	8.7.24	

Lifeloc Technologies  
 Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 008620

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1011	34	16FP41018	.100	19		.104				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

05/29/18 10:12  
 NS4 Serial no: 008620  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 01/15/16 11:10 .100  
 Test Result  
 0001 05/19/18 18:21 .176  
 0002 05/20/18 17:18 .209  
 0003 05/20/18 17:20 .190  
 0004 05/25/18 16:18 .135  
 0005 05/25/18 16:22 .155  
 0006 05/29/18 10:11 .104  
 Push No. 3 to erase  
 or Power down to retain



#9

AREA NAME  
South Los Angeles

AREA NO. 530  
COORDINATOR NAME Kirk Lee

I.D. NUMBER  
17046

PAS MAKE/MODEL  
Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
082957

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1300	34	16FP41018	.100								M. Tippet	20137	DEVICE NOT IN LOCKER FOR TESTING

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*HP*

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 077505

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1009	34	16FP41018	.100							X	M. Tippett	20137	

05/29/18 10:10  
AS4 Serial no: 077505  
Version no: 4797  
No. Date Time ZBAC  
Calibration 01/18/18 10:56 .100  
Test Result 0001 05/29/18 10:09 .098  
Push No. 3 to erase  
OR Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.099 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions  
Chp202K\_1108.pdf

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	ID. NO.	REPAIRS/PROBLEMS COMMENTS
		TEMP (°C)	LOT NO.	TEST #1 (Optional)	TEST #2	TEST #1 (Optional)	TEST #2	YES	NO	RETEST RESULTS	RETEST RESULTS			
05/29/2018	1610	34	16FP41018									M. Tippett	20137	DEVICE NOT IN LOCKER FOR TESTING

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 082653

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	10/10	34	16FP41018	.100		19		104				M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
Push No. 3 to erase: 0012 05/29/18 10:11 .104 0011 05/27/18 03:49 .189 0010 05/27/18 03:47 .198 0009 05/22/18 01:00 .136 0008 05/27/18 00:57 .122 0007 05/26/18 03:45 .094 0006 05/26/18 03:42 .114 0005 05/25/18 21:41 .000 0004 05/20/18 22:16 .120 0003 05/20/18 22:14 .120 0002 05/20/18 02:03 .118 0001 05/20/18 02:00 .111 Test Result														
No. Date Time %BAC Calibration 02/11/16 10:14 .100 Version no: 4797 NS4 Serial no: 082653 05/29/18 10:11														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#13

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS WAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008709

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1015	34	16FP41018		.100		20					X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18 * If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator. No. Date Time KBAC Calibration 03/17/16 09:40 .100 Test Result 0001 05/29/18 10:16 .104 AS4 Serial no: 008709 Version no: 4797 05/29/18 10:20 Push No. 3 to erase; Or power-down to retain															

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-09) OPI 051

*Handwritten:* #7/4

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: *Handwritten:* 008714

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP. (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.					TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1022	34	16FP41018	.100	20			TEST #1	TEST #2	YES	NO	X	M. Tippet	20137	

05/29/18 10:24  
AS4 Serial no: 008714  
Version no: 4797  
No. Date Time ZMAC  
Calibration 05/14/17 09:58 .100  
Test Result 0001 05/29/18 10:23 .100  
Push No. 3 to erase  
OR power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#17

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO.: 008783

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1024	34	16FP41018	.100	19		N/A				X	M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

05/29/18 10:25  
 MS4 Serial no: 008783  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 05/14/17 10:07.100  
 Test Result 0001 05/29/18 10:24.103  
 Push No. 8 to erase,  
 or Power down to retain



#18

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
ID NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 082657

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1028	34	16FP41018		.100	20		.048				X	M. Tippett	20137	

05/29/18 10:26  
NS4 Serial no: 082657  
Version no: 4797  
No. Date Time ZBAC  
Calibration 05/21/17 08:18 .100  
Test Result  
0001 05/29/18 10:26 .098  
Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions  
Chp202K\_1108.pdf

#19

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008810

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
05/29/2018	103034		16FP41018	.100	19			✓	NO		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18													
No. Date Time %BAC Calibration 05/21/17 08:20 .100 Test Result 0001 05/29/18 10:31 .100 Version not 4797 A94 Serial not 008810 05/29/18 10:31													
Push No. 3 to erase. OF POWER DOWN TO RETAIN													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202K\_1108.pdf

#20

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 018684

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
05/29/2018	1055	34	16FP41018	.100	19		105						M. Tippett	20137	
Simulator prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18															
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.															

05/29/18 10:34  
 AS4 Serial no: 018684  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 01/12/16 14:36 .093  
 Test Result 0001 05/29/18 10:33 .105  
 Push No. 3 to erase  
 or power down to retain

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#21

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: 058635

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.			TEMP #1	TEMP #2		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1234	34	16FP41018	.100									X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia																
Date: 06-06-16																
Expiration date: 06-06-18																

05/29/18 10:35  
 NS4 Serial no: 008635  
 Version no: 4797  
 No. Date Time ZBNC  
 Calibration 05/21/17 08:23 .100  
 Test Result  
 0001 05/19/18 02:11 .183  
 0002 05/19/18 02:17 .126  
 0003 05/19/18 20:08 .208  
 0004 05/19/18 20:11 .227  
 0005 05/26/18 01:58 .202  
 0006 05/26/18 02:02 .205  
 0007 05/26/18 21:09 .158  
 0008 05/26/18 21:01 .151  
 0009 05/21/18 03:07 .105  
 0010 05/21/18 03:40 .111  
 0011 05/22/18 01:02 .218  
 0012 05/22/18 01:05 .212  
 0013 05/22/18 02:01 .212  
 0014 05/22/18 02:04 .207  
 0015 05/22/18 21:48 .111  
 0016 05/22/18 21:51 .120  
 0017 05/18/18 11:34 .103

Flush No. 3 to 0.0350  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#22

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 160 40162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0906	34	16FP41018	.100	21		.099		X		M. Tippet	20137		
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
Cal Standard														
				.100	Time:									
				.100	10:14 AM									
				.099	Last Cal	08/14/2017								
				.100	Result									
				.100	Cal Standard									
				.099	Time:									
				.100	09:06 AM									
				.099	Last Check	05/29/2018								
				.100	UOM	BAC								
				.100	Serial No.	16040162								
				.099	FC20	8.7.24								
				.100	Calibration / Check Printout									Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16120177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0909	34	16FP41018	.100	21						X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	09:50 AM	04/25/2017	.105	.100	09:10 AM	05/29/2018	BAC	16120177	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	YES	NO			
05/29/2018	10:37	34	16FP41018	.100			21				X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18															

No. Date Time K&AC  
 Calibration  
 05/21/17 00:30 .100  
 Test Result  
 0001 05/29/18 09:38 .101  
 Push No. 3 to erase  
 OR POWER DOWN TO RETAIN

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*Handwritten signature*

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 ID NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: 077591

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS/ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
05/29/2018	1040	34	16FP41018	.100	21	.100				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

05/29/18 10:41  
 AS4 Serial no: 077591  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 05/21/17 08:35 .100  
 Test Result 0001 05/29/18 10:40 .100  
 Push No. 3 to erases  
 or power down to retain



#26

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Lifeloc/FC20  
PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0912 34		16FP41018							M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 06-06-16  
Expiration date:  
06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	1:54 PM	12/19/2017	.106	.100	09:12 AM	05/29/2018	BAC	16080065	8.7.24	

Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#27

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: 077573

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1043	34	16FP41018	.100								M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

05/29/18 10:44  
 NS4 Serial no: 077573  
 Version no: 4297  
 No. Date  
 Calibration Time: 2BAC  
 01/18/18 11:23 .100  
 Test Result  
 0001 00/29/18 10:43 .095  
 Push No. 3 to erase  
 or power down to retain

*Handwritten:* #28

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV

PAS SERIAL NO.: 083051

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1044	34	16FP41018	.100								M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

05/29/18 10:45  
 NS4 Serial: not 083051  
 Version: not 4797  
 No. Date Time KBAC  
 Calibration 12/20/17 13:29 .100  
 Test Result 0001 05/29/18 10:45 .094  
 Push No. 3 to erase  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

# 70

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee ID NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: **068897**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1052	34	16FP41018	.100	19						X	M. Tippett	20137	
Simulator Solution (OR DRY GAS)** Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

05/29/18 10:53  
 AS4 Serial no: 008897  
 Version no: 4297  
 No. Date Time XBAC  
 Calibration 01/18/12 16:13 .100  
 Test Result 0001 05/18/18 09:38 .100  
 0002 05/29/18 10:52 .100  
 Test no: 3 to erase  
 or power down to retain

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 018624

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1	TEST #2 (Optional)	YES	NO				
05/29/2018	1732	34	16FP41018	.100	20							X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18															
Push No. 3 to erase OR Power down to retain 05/29/18 12:34 NS4 Serial no: 018624 Version no: 4797 No. Date Time %BAC Calibration 02/01/16 11:02 .100 Test Result 0001 05/29/18 12:33 .100															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles      AREA NO: 530      COORDINATOR NAME: Kirk Lee      ID. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/F/C20      PAS SERIAL NO: 160 80071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0915	34	16FP41018	.100	21		102					M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 06-06-16														
Expiration date: 06-06-18														
Cal Standard														
				.100										
					Time:									
					10:41 AM									
					Last Cal									
					08/14/2017									
					Result			102						
					Cal Standard									
					Time:									
					09:15 AM									
					Last Check									
					05/29/2018									
					UOM									
					BAC									
					Serial No.									
					16080071									
					FC20									
					8.7.24									
					Calibration / Check Printout									
					<b>Lifeloc Technologies</b>									

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#33

AREA NAME: South Los Angeles      AREA NO: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO:

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/29/2018	1230	34	16FP41018	.100		19		104				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18															
Push No. 3 to erase or power down to retain Test Result      0001 05/29/18 12:30 .104 Calibration      02/11/16 10:52 .100 No.      Date      Time      XBAC Version      not 4797 AS4 Serial not 077488 05/29/18 12:31															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME

South Los Angeles

AREA NO. 530  
 COORDINATOR NAME Kirk Lee

I.D. NUMBER 17046

PAS MAKE/MODEL Intoximeters/Alco-Sensor IV

PAS SERIAL NO.

008877

#34

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	123834	16EP41018		100	19	096			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18												
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.												

05/29/18 12:48  
 AS4 Serial no: 008877  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 05/29/17 09:49 .100  
 Test Result  
 0001 05/19/18 05:14 .264  
 0002 05/19/18 15:18 .245  
 0003 05/20/18 12:09 .096  
 Test no. 3 to 0758  
 OR Power down to retain



*436*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: *16040163*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
05/29/2018	0914	34	16FP41018	.100	22								M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	11:02 AM	06/07/2017	.099	.100	09:14 AM	05/29/2018	BAC	16040163	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME  
 South Los Angeles

AREA NO. 530  
 COORDINATOR NAME Kirk Lee

I.D. NUMBER  
 17046

PAS MAKE/MODEL  
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
 018626

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	12:36:34	16FP41018		.100	19							M. Jippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

05/29/18 12:37  
 A64 Serial no: 018626  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 07/15/16 15:21 .100  
 Test Result 0901 9/22/18 12:37 .102  
 Push No. 3 to erase  
 or power down to retain



#39

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 ID NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: 083030

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	12:39	34	16FP41018	.100	19		098		X			M. Tippett	20137	

05/29/18 12:40  
 NS4 Serial no: 083030  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 12/20/17 13:51 .100  
 Test Result  
 0001 05/29/18 12:39 .098  
 Push No. 3 to erase  
 OR POWER DOWN TO RETAIN

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*Handwritten initials*

018692

AREA NAME: South Los Angeles    AREA NO: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV    PAS SERIAL NO: 018692

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.			TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1241	34	16FP41018		.100	19							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18															
Test Result: 85/29/18 12:43 R34 Serial# 018692 Version: 008 4797 No. Date Time ZBAC Calibration: 02/01/16 11:30 .100 Test Result: 8001 24/29/18 12:42 .104 Wash the 5 to 12807 or power down to retain															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#42

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV

PAS SERIAL NO.: 018762

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1246	34	16FP41018	1.00	19		0.95				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 06-06-16														
Expiration date: 06-06-18														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

05/29/18 12:47  
 NS4 Serial no: 018762  
 Version: not 4297  
 No. Date Time XBAC  
 Calibration 12/28/17 14:01 .100

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 018 766

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	1249 34		16FP41018	.100	20				X			M. Tippet	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18</p>														
<p>05/29/18 12:51 MS4 Serial no: 018766 Version no: 4797 No. Date Time XBAC Calibration 01/22/16 13:16 .100 Test Result 0001 05/29/18 12:49 .092 Push No. 3 to erase. OR Power down to retain</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destory Previous Editions

Chp202K\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

*Handwritten signature*

AREA NAME: South Los Angeles      AREA NO: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO: 018691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	12:34	34	16FP41018	.100	20		.095			X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
Push No. 3 to erase, OR POWER DOWN TO RETAIN														
05/29/18 12:54 NS4 Serial no: 018691 Version no: 4797 No. Date Time %BAC Calibration 05/28/17 08:26 .100 Test Result 0001 05/29/18 12:53 .095														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#45

AREA NAME  
 South Los Angeles

AREA NO.  
 530

COORDINATOR NAME  
 Kirk Lee

I.D. NUMBER  
 17046

PAS MAKE/MODEL  
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
 017990

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)		YES	NO				
05/29/2018	1255	34	16FP41018	.100		19				X			
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18													
05/29/18 12:56 NS4 Serial no: 017990 Version no: 4797 No. Date Time %BAC Calibration 11/26/03 11:10 .100 Test Result 0001 05/29/18 12:56 .101 Push No. 3 to erase, or power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

*496*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: *16120189*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
05/29/2018	0919	34	16PP41018	.100		22		.097			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 06-06-16  
 Expiration date:  
 06-06-18

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:17 PM	09/01/2017	.097	.100	09:20 AM	05/29/2018	BAC	16120189	8.7.24		

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor-IV

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.  
 PAS SERIAL NO.: 041495

#57

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	150134		16FP41018	.100	18					X		M. Tippett	20137	
SIMULATOR SOLUTION (OR DRY GAS)** PAS TEMP (°C) Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
PAS MAKE/MODEL: Intoximeters/Alco-Sensor-IV PAS SERIAL NO.: 041495 Version: not 4797 NS4 Serial not 041495 No. Date Time ZBAC Calibration 12/28/17 14:10 .100 Test Result 0001 05/29/18 13:01 .096 Push No. 3 to erase. OR Power down to retain														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO: 16086013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
05/29/2018	0921	34	16FP41018	.100		21		.103			X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														
Cal Standard														
					Time:	Last Cal		Result	Cal Standard					
					Time:	01/29/2018		.103						
					Time:	09:24 AM		.100	Last Check					
					Time:	05/29/2018			UOM					
					Time:	12:59 PM			Serial No.					
					Time:	01/29/2018			FC20					
					Time:				Calibration / Check Printout					

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*SGT's PAS*

AREA NAME: South Los Angeles  
 AREA NO: 530  
 COORDINATOR NAME: Kirk Lee  
 ID NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO: *008874*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2	YES	NO	TEST #1 (Optional)	TEST #2				
05/29/2018	1310	34	16FP41018	.100	19							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 06-06-16 Expiration date: 06-06-18														

Test Result  
 0001 05/20/18 19:59 .053  
 0002 05/21/18 00:54 .145  
 0003 05/21/18 00:58 .148  
 0004 05/21/18 01:59 .029  
 0005 05/21/18 02:01 .025  
 0006 05/21/18 03:16 .121  
 0007 05/21/18 03:18 .113  
 0008 05/21/18 17:41 .000  
 0009 05/25/18 01:33 .208  
 0010 05/25/18 01:35 .204  
 0011 05/25/18 12:30 .170  
 0012 05/25/18 12:57 .254  
 0013 05/25/18 12:59 .212  
 0014 05/25/18 23:55 .000  
 0015 05/27/18 01:09 .167  
 0016 05/27/18 01:12 .140  
 0017 05/28/18 19:52 .179  
 0018 05/28/18 19:54 .179  
 0019 05/29/18 13:11 .097

NS4 Serial no: 008874  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 10/24/17 11:04 .100

Push No. 3 to erase,  
 or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.