





75

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16120197

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1231	34	17GP41049	.100	21		.100				X	M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:22 AM	10/21/2016	.100	.100	12:32 PM	06/07/2018	BAC	16120197	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles    AREA NO.: 530    COORDINATOR NAME: Kirk Lee    ID NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO.: 16120195

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	12:38	34	17GP41049	.100	21	106			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:47 AM	04/25/2017	.106	.100	12:29 PM	06/07/2018	BAC	16120195	8.7.24	Calibration / Check Printout

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*HS*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 208559

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	0939	34	17GP41049	.100	19		1/21				Y		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia												
			Date: 07-24-17												
			Expiration date: 07-24-19												
<p>06/07/18 09:41 MS4 Serial no: 0003559 Version no: 4797</p> <p>No. Date Time ZBAC Calibration 01/15/16 10:40 .100</p> <p>Test Result 0001 06/07/18 09:40 .101</p> <p>Push No. 3 to erase, or power down to retain</p>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OP1 051

#6

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16040161

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	1255	54	17GP41049	.100	21		102					X		20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:56 AM	04/25/2017	.102	.100	12:36 PM	06/07/2018	BAC	16040161	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16040140

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
06/07/2018	1230	34	17GP41049	.100	21		.103				X	M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:58 AM	04/25/2017	.103	.100	12:31 PM	06/07/2018	BAC	16040140	8.7.24	

Calibration / Check Printout  
**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#8

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP. (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.			TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	0944	34	17GP41049	.100	19								M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19															
<div style="border: 1px solid black; padding: 5px; transform: rotate(180deg);"> <p>06/07/18 09:45 NS4 Serial no: 008620 Version no: 4797 No. Date Time %BAC Calibration 01/15/16 11:10 .100 Test Result 0001 06/07/18 09:45 .107 Push No. 3 to erase, or Power down to retain</p> </div>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

*49*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: *082 957*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	YES	NO				
06/07/2018	<i>0941 34</i>		17GP41049	.100	<i>20</i>		<i>.102</i>				<i>X</i>	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
Test Result 0001 05/22/18 02:19 .107 0002 05/22/18 02:22 .104 0003 05/29/18 02:58 .167 0004 05/29/18 02:59 .171 0005 05/30/18 01:06 .026 0006 06/04/18 19:37 .251 0007 06/04/18 19:41 .254 0008 06/05/18 02:43 .082 0009 06/05/18 02:47 .078 0010 06/07/18 09:41 .102 Push No. 3 to erase, or Power down to retain														
No. Date Time %BAC Calibration 02/17/14 08:52 .100 Version no: 4797 NS4 Serial no: 082957 06/07/18 09:42														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#110

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 077505

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/07/2018	0947 34	17GP41049		.100	20	0.101				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													

06/07/18 09:44  
AS4 Serial no: 077505  
Version no: 4797  
No. Date Time %BAC  
Calibration 01/18/18 10:56 .100  
Test Result 0001 06/07/18 09:44 .101  
Push No. 3 to erase,  
or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 077479

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	0946 34		17GP41049	.100								M. Tippett	20137	DEVICES NOT IN WORKING ORDER
														FOR TESTERS

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*4/2*

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: *082653*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/07/2018	0944 34	17GP41049	.100	20	107	Y				M. Tippett	20137		
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Push No. 3 to erase; OR Power down to retain													
Test Result No. Date Time %BAC Calibration 02/11/16 10:14 .100 Version not 4297 NS4 Serial no: 082653 06/07/18 09:49 0001 06/02/18 02:13 .136 0002 06/02/18 02:16 .143 0003 06/02/18 03:17 .056 0004 06/07/18 09:45 .107													

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

#13

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO.: 008709

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
06/07/2018	0955	34	17GP41049	.100	20	0107			NO		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

06/07/18 09:54 NS4 Serial not 008709  
 Version not 4797  
 No. Date Time ZBAC  
 Calibration 03/17/16 09:40 .100  
 Test Result  
 0001 06/03/18 02:45 .218  
 0002 06/03/18 02:47 .240  
 0003 06/03/18 02:50 .251  
 0004 06/07/18 09:53 .107  
 Push No. 3 to erase,  
 or Power down to retain

*#79*

AREA NAME: South Los Angeles | AREA NO: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO: *008714*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	<i>0957</i>	<i>34</i>	17GP41049	.100	<i>20</i>		<i>.105</i>			<i>X</i>		M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
No. Date Time %BAC Calibration 05/14/17 09:50 .100 Test Result 0001 06/07/18 09:57 .105 Push No. 3 to erase, or power down to retain														
PAS RS4 Serial no: 008714 Version no: 4797 06/07/18 09:58 M. Tippet														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.0990 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 008888

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1	TEST #2 (Optional)	YES	NO				
06/07/2018	0952	34	17GP41049	.100	21			108				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19															
Push No. 3 to erase or Power down to retain Test Result 0001 05/20/18 20:52 .035 0002 05/20/18 20:55 .037 0003 05/22/18 01:54 .095 0004 05/22/18 02:00 .096 0005 05/28/18 22:20 .023 0006 05/28/18 22:21 .023 0007 05/28/18 22:50 .040 0008 05/28/18 22:52 .038 0009 05/29/18 01:45 .164 0010 05/29/18 01:47 .154 0011 06/07/18 09:53 .108															
Version: 4797 NS4 Serial no: 008888 No. Date Time %BAC Calibration 03/17/16 09:44 .100															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202k\_1108.pdf

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/07/2018	1000	34	17GP41049	.100	16					X		M. Tippett	20137	
Simulator prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
No. Date Time %BAC Calibration 05/14/17 10:07 .100 Test Result 06/07/18 10:01 A54 Serial no: 008783 Version no: 4797														
Push No. 3 to erase, or Power down to retain														
0001 06/01/18 20:59 .123 0002 06/01/18 21:01 .119 0003 06/02/18 01:15 .050 0004 06/02/18 01:18 .049 0005 06/07/18 10:00 .100														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#118

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 082-657

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	ID NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS				
06/07/2018	1000	74	17GP41049	.100	20								M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												
			Push No. 3 to erase, or Power down to retain Test Result 0001 06/07/18 10:00 .102 Calibration 05/21/17 08:18 .100 No. Date Time ZDAC Version not: 4797 NS4 Serial no: 082657 06/07/18 10:01												

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 008810

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEMP #1 (Optional)	TEMP #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1005	37	117GP41049	19		104		X			ML Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
06/07/18 10:06 NS4 Serial no: 008810 Version no: 4797 No. Date Time %BAC Calibration 05/21/17 08:20 .100 Test Result 0001 06/07/18 10:06 .104 Push No. 3 to erase, or Power down to retain													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canisters serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.







#22

AREA NAME: South Los Angeles    AREA NO: 530    COORDINATOR NAME: Kirk Lee    I.D. NUMBER: 17046    PAS MAKE/MODEL: Lifeloc/FC20    PAS SERIAL NO: 16040162

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	1229	34	17GP41049	100	21	.097					X			

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:14 AM	08/14/2017	.097	.100	12:30 PM	06/07/2018	BAC	16040162	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME: South Los Angeles AREA NO: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO: 16/20177

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (C)	LOT NO	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO	REPAIRS/PROBLEMS/COMMENTS
								TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1233	34	17GP41049	.100	21			.103					M. Tippet	20137	
Solution prepared by: Cesar A. Mejia															
Date: 07-24-17															
Expiration date: 07-24-19															
Cal Standard															
Time: 09:50 AM															
Last Cal															
04/25/2017															
Result															
.103															
Cal Standard															
.100															
Time: 12:34 PM															
Last Check															
06/07/2018															
UOM															
BAC															
Serial No.															
16120177															
FC20															
8.7.24															
Lifeloc Technologies															
Calibration / Check Printout															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





#26

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifloc/FC20      PAS SERIAL NO.: 16080065

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.				TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1235	34	17GP41049	.100	21		.103			X		M. Tippett	20137	

Solution prepared by: Cesar A. Mejia																	
Date: 07-24-17																	
Expiration date: 07-24-19																	
Cal Standard																	
Time:		Last Cal		Result		Cal Standard		Time:		Last Check		UOM		Serial No.		FC20	
1:54 PM		12/19/2017		.103		.100		12:35 PM		06/07/2018		BAC		16080065		8.7.24	
Calibration / Check Printout																	
<b>Lifloc Technologies</b>																	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV      PAS SERIAL NO.: 077573

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		TEMP (°C)	PAS TEMP (°C)		PAS ACCURACY CHECK/RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1219 34	17GP41049		.100	20		.100				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
06/07/18 10:20 NS4 Serial no: 077573 Version no: 4797 No. Date Time KBAC Calibration 01/18/18 11:23 .100 Test Result 0001 06/07/18 10:20 .100 Push No. 3 to erase, or power down to retain														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME  
 South Los Angeles

AREA NO. 530  
 COORDINATOR NAME Kirk Lee

ID. NUMBER  
 17046

PAS MAKE/MODEL  
 Intoximeters/Alco-Sensor IV

PAS SERIAL NO.  
 083051

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.					TEST #1	TEST #2	TEST #1	TEST #2	YES	NO	RETEST RESULTS	RETEST RESULTS				
06/07/2018	1020	34	17GP41049		.100	70					.100					X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19																			
<pre> 06/07/18 10:21 NS4 Serial no: 083051 Version no: 4797 No. Date Time %BAC Calibration 12/28/17 13:29 .100 Test Result 0001 06/07/18 10:21 .100 Push No. 3 to erase OR Power down to retain           </pre>																			

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

430

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. <b>008897</b>
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)*		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/07/2018	1027 34	17GP41049	.100	19		1107			X		M. Tippett	20137	
<p style="text-align: center;">Solution prepared by:            Cesar A. Mejia            Date: 07-24-17            Expiration date:            07-24-19</p>													
<div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p>06/07/18 10:28            NS4 Serial no: 008897            Version no: 4797            No. Date Time XIMC            Calibration 01/18/12 16:13 .100            Test Result 0001 06/07/18 10:27 .107            Push No. 3 to erase            OR Power down to retain</p> </div>													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202k\_1108.pdf

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME **South Los Angeles** AREA NO. **530** COORDINATOR NAME **Kirk Lee** I.D. NUMBER **17046** PAS MAKE/MODEL **Intoximeters/Alco-Sensor IV** PAS SERIAL NO. **018624**

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1051	34	17GP41049	.100	20		105			Y		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
06/07/18 10:32 NS4 Serial no: 018624 Version no: 4797 No. Date Time XMAC Calibration 02/01/16 11:02 .100 Test Result 0001 06/07/18 10:32 .105 Push No. 3 to erase OR Power down to retain														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#32

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16080071

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1237	34	17GP41049	.100	21	101			X	M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:41 AM	08/14/2017	.101	.100	12:38 PM	06/07/2018	BAC	16080071	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#333

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 077488

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK/RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1052	34	17GP41049	.100	19		TEST #1	TEST #2	YES	NO	X	M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia  Date: 07-24-17  Expiration date: 07-24-19</p> <p>06/07/18 10:29  NS4 Serial no: 077488  Version no: 4797  No. Date Time XBAC  Calibration 02/11/16 10:52 .100  Test Result 0001 06/07/18 10:28 .109  Push No. 3 to erase,  or Power down to retain</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
Destroy Previous Editions  
CHP202K\_1108.pdf

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#34

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 008877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	1037	34	17GP41049	.100		19	.100					X	M. Tippett	20137	
Simulator prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19															
Push No. 3 to erase, or power down to retain Test Result 0001 06/07/18 10:37 .100 Calibration 05/28/17 08:49 .100 No. Date Time XBAC Version no: 4797 PAS4 Serial no: 008877 06/07/18 10:40															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.  
 Destroy Previous Editions  
 Chp202K\_1108.pdf



#36

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16040163

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	TEST #1 (Optional)		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO						
06/07/2018	1256	34	17GP41049		.100	22							X	M. Tippett	20137	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

UOM	Serial No.	FC20	Last Check	Time:	Cal Standard	Result	Last Cal	Time:	Cal Standard
BAC	16040163	8.7.24	06/07/2018	12:37 PM	.100	.096	06/07/2017	11:02 AM	.100

Lifeloc Technologies  
Calibration / Check Printout

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



#38

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 ID. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: 16080072

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1245	34	17GP41049	.100	21					X		M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.101	.100	12:45 PM	06/07/2018	BAC	16080072	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





*#42*

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO: *018762*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/07/2018	1202	54	17GP41049		.100								M. Tippett	20137	
Solution prepared by: Cesar A. Mejia															
Date: 07-24-17															
Expiration date: 07-24-19															
<div style="display: flex; justify-content: space-between;"> <div style="width: 45%;"> <p>Test Result</p> <p>0001 06/01/18 23:32 .017</p> <p>0002 06/01/18 23:56 .047</p> <p>0003 06/02/18 02:31 .045</p> <p>0004 06/02/18 02:33 .045</p> <p>0005 06/02/18 03:10 .159</p> <p>0006 06/02/18 03:13 .142</p> <p>0007 06/02/18 03:19 .047</p> <p>0008 06/02/18 04:58 .213</p> <p>0009 06/02/18 05:02 .220</p> <p>0010 06/02/18 20:43 .161</p> <p>0011 06/02/18 20:45 .178</p> <p>0012 06/03/18 04:28 .110</p> <p>0013 06/03/18 04:32 .108</p> <p>0014 06/03/18 20:21 .103</p> <p>0015 06/03/18 20:23 .108</p> <p>0016 06/03/18 20:33 .006</p> <p>0017 06/03/18 22:45 .000</p> <p>0018 06/07/18 12:03 .095</p> </div> <div style="width: 45%; text-align: right;"> <p>12/28/17 14:01 .100</p> <p>Calibration</p> <p>No. Date Time %BAC</p> <p>06/07/18 12:03</p> <p>MS4 Serial no: 018762</p> <p>Version no: 4797</p> </div> </div>															
<p>Push No. 3 to erase</p> <p>OR Power down to retain</p>															

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#43

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: 018766

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/07/2018	1205	74	17GP41049	100	70		096			X		M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

06/07/18 12:06  
 MS4 Serial no: 018766  
 Version no: 4797  
 No. Date Time ZBAC  
 Calibration 01/22/16 13:16 .100  
 Test Result 0001 06/07/18 12:05 .096  
 Push No. 3 to erase  
 or power down to retain





STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

*HS*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: *017990*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	<i>120834</i>		17GP41049	.100	<i>20</i>		<i>.106</i>				<i>X</i>		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19															
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.															

06/07/18 12:09  
 NS4 Serial no: 017990  
 Version no: 4797  
 No. Date Time XBAC  
 Calibration 11/26/03 11:10 .100  
 Test Result 0001 06/07/18 12:09 .106  
 Push No. 3 to erase,  
 or Power down to retain

*496*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Lifeloc/FC20  
 PAS SERIAL NO.: *16120189*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/07/2018	12:41	34	17GP41049	.100	21	.096			X		M. Tippett	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:17 PM	09/01/2017	.096	.100	12:41 PM	06/07/2018	BAC	16120189	8.7.24	Lifeloc Technologies

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



*#51*

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. <i>041495</i>
--------------------------------	-----------------	------------------------------	---------------------	---	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
06/07/2018	1210	34	17GP41049	.100	19						X	ML Tippett	20137	

06/07/18 12:10  
NS4 Serial no: 041495  
Version no: 4297  
No. Date Time ZBMC  
Calibration 12/28/17 14:10 .100  
Test Result 0001 06/07/18 12:10 .101  
Push No. 3 to erase,  
or Power down to Retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles      AREA NO.: 530      COORDINATOR NAME: Kirk Lee      I.D. NUMBER: 17046      PAS MAKE/MODEL: Lifeloc/FC20      PAS SERIAL NO.: 16080013

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS		
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)					YES	NO
06/07/2018	1245	34	17GP41049	100	21	.092					X		M. Tippet	20137	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
100	12:59 PM	01/29/2018	.092	100	12:45 PM	06/07/2018	BAC	16080013	8.7.24	

**Lifeloc Technologies**

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.





*5475 PAS*

AREA NAME: South Los Angeles  
 AREA NO.: 530  
 COORDINATOR NAME: Kirk Lee  
 I.D. NUMBER: 17046  
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
 PAS SERIAL NO.: *008874*

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS	CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)		YES	NO				
06/07/2018	<i>1218</i>	<i>34</i>	17GP41049	.100			<i>100</i>				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Push No. 3 to erase, OR Power down to retain													
Test Result 0001 05/29/18 20:54 .079 0002 05/29/18 20:57 .062 0003 05/29/18 20:59 .057 0004 05/31/18 17:40 .000 0005 06/03/18 02:12 .101 0006 06/03/18 02:15 .094 0007 06/03/18 05:17 .128 0008 06/03/18 05:19 .129 0009 06/04/18 00:55 .054 0010 06/04/18 00:58 .143 0011 06/04/18 01:01 .158 0012 06/07/18 03:48 .000 0013 06/07/18 12:19 .100													
No. Date Time ZBAC Calibration 10/24/17 11:04 .100													
06/07/18 12:19 NS4 Serial no: 008874 Version no: 4797													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.