

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16/20188
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1205	34	17GP41049	22		100			X		M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120188
UOM	BAC
Last Check	06/15/2018
Time:	12:05 PM
Cal Standard	.100
Result	.100
Last Cal	10/21/2016
Time:	08:16 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1205	34	17GP41049	.100	22		.099		X			M. Tippert	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120194
UOM	BAC
Last Check	06/15/2018
Time:	12:05 PM
Cal Standard	.100
Result	.099
Last Cal	01/29/2018
Time:	08:50 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister-serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	12:14	34	17GP41049	21		102			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120197
UOM	BAC
Last Check	06/15/2018
Time:	12:14 PM
Cal Standard	.100
Result	.102
Last Cal	10/21/2016
Time:	10:22 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

Handwritten initials

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	12:11	34	17GP41049	.100	22		1.08		X			M. Tippet	20137	

Solution prepared by:

Cesar A. Mejia

Date: 07-24-17

Expiration date:

07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120195
UOM	BAC
Last Check	06/15/2018
Time:	12:12 PM
Cal Standard	.100
Result	.108
Last Cal	04/25/2017
Time:	08:47 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

HS

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1243B		17GP41049	.100	20					X		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/15/18 12:50
 AS4 Serial no: 008559
 Version not 4797

No. Date Time ZBAC
 Calibration
 01/15/16 10:40 .100

Test Result
 0001 06/15/18 12:49 .102

Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16040161

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	12:19	34	17GP41049	21		.105			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040161
UOM	BAC
Last Check	06/15/2018
Time:	12:20 PM
Cal Standard	.100
Result	.105
Last Cal	04/25/2017
Time:	08:56 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040740
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	12:14	34		17GP41049	.100	22						M. Tippett	20137	

Solution prepared by:

Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040140
UOM	BAC
Last Check	06/15/2018
Time:	12:15 PM
Cal Standard	.100
Result	.107
Last Cal	04/25/2017
Time:	08:58 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#8

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008620

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2	YES				
06/15/2018	1252	34	17GP41049	.100	20	.107			X		M. Tippett	20137	

06/15/18 12:54
 PAS Serial no: 008620
 Version no: 4797

No. Date Time ZBAC
 Calibration
 01/15/16 11:10 .100

Test Result
 0001 06/15/18 12:53 .107

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1249	34	17GP41049	.100	21	102	X				M. Tippett	20137		

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

```

06/15/18 12:50
AS4 Serial no: 082957
Version no: 4797

No. Date Time %BAC
Calibration
    02/17/14 08:52 .100

Test Result
0001 06/14/18 01:35 .184
0002 06/14/18 01:38 .148
0003 06/14/18 01:41 .186
0004 06/15/18 06:15 .000
0005 06/15/18 12:49 .102

Push No. 3 to erase,
or power down to retain
    
```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#10

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077505
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/15/2018	1752	74	17GP41049	.100	21	102				X	M. Tippett	20137	

06/15/18 12:52
AS4 Serial no: 077505
Version no: 4797

No. Date Time ZBAC
Calibration
01/18/18 10:56 .100

Test Result
0001 06/15/18 12:52 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#11

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077479
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/15/2018	12:53		17GP41049	.100							M. Tippet	20137	DEVICE NOT IN LOCKER FOR TESTING

* If the reading is within plus or minus (±) 0.010 percent of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME		AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.							
South Los Angeles		530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	082653							
DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
06/15/2018	7:30	34		17GP41049					X		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date: 07-24-19										

06/15/18 13:04
AS4 Serial no: 082653
Version no: 4797

No. Date Time XBAC
Calibration
02/11/16 10:14 .100

Test Result
0001 06/07/18 22:13 .000
0002 06/08/18 22:14 .000
0003 06/09/18 02:52 .173
0004 06/09/18 02:56 .173
0005 06/15/18 13:04 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				VALUE	TEST #1	TEST #2	YES				
06/15/2018	1308	34	17GP41049	.100						X		M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

06/15/18 13:09
AS4 Serial no: 008709
Version no: 4797

No. Date Time XBAC
Calibration
03/17/16 09:40 .100

Test Result
0001 06/15/18 13:09 .107

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#114

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO.: 008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1321 34	34	17GP41049	.100	21						X	M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19											

06/15/18 13:13
 AS4 Serial no: 008714
 Version no: 4797
 No. Date Time %BAC
 Calibration
 05/14/17 09:50 .100
 Test Result
 0001 06/15/18 13:13 .105
 Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath-simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OP1 051

#1/6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1309	34	17GP41049	.100	Z1					X		M. Tippett	20137	
	1309													
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date:											
			07-24-19											

06/15/18 13:10
AS4 Serial no: 008888
Version no: 4797

No. Date Time %BAC
Calibration
03/17/16 09:44 .100

Test Result
0001 06/12/18 02:11 .017
0002 06/12/18 02:14 .016
0003 06/13/18 04:03 .215
0004 06/13/18 04:06 .209
0005 06/15/18 13:10 .108

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	137	34	17GP41049		.100	19	1/05			X		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/15/18 13:18
AS4 Serial no: 008783
Version no: 4797

No. Date Time %BAC
Calibration
05/14/17 10:07 .100

Test Result
0001 06/09/18 04:10 .128
0002 06/09/18 04:14 .132
0003 06/15/18 13:17 .105

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

South Los Angeles

#18

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alcoo-Sensor IV | PAS SERIAL NO.: 082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1316	34	17GP41049	20		.102			X		M. Tippet	20137	
			Solution prepared by: Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date: 07-24-19										

06/15/18 13:17
AS4 Serial no: 082657
Version no: 4797

No. Date Time %BAC
Calibration
05/21/17 08:18 .100

Test Result
0001 06/15/18 13:17 .102

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	1322 34		17GP41049	.100	20		103		X		M. Tippett	20137	

06/15/18 13:23
AS4 Serial no: 008810
Version no: 4797

No. Date Time ZBAC
Calibration
05/21/17 08:20 .100

Test Result
0001 06/15/18 13:22 .103

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#20

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 018684

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO					
06/15/2018	1324	34		17GP41049	.100	20				X			M. Tippett	20137		
			Solution prepared by:													
			Cesar A. Mejia													
			Date: 07-24-17													
			Expiration date:													
			07-24-19													

06/15/18 13:25
AS4 Serial no: 018684
Version no: 4797

No. Date Time %BAC
Calibration
01/12/16 14:36 .093

Test Result
0001 06/15/18 13:24 .108

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1325 34		17GP41049		20					X		M. Iippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
Push No. 3 to erase, or power down to retain														
Test Result 0001 06/15/18 13:25 .106														
Calibration 05/21/17 08:23 .100														
No. Date Time ZDAC 06/15/18 13:26 AS4 Serial no: 008635 Version no: 4797														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1219	34	17GP41049	.100	21	.100					M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040162
UOM	BAC
Last Check	06/15/2018
Time:	12:14 PM
Cal Standard	.100
Result	.100
Last Cal	08/14/2017
Time:	10:14 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1217	34	17GP41049	22		107					M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120177
UOM	BAC
Last Check	06/15/2018
Time:	12:18 PM
Cal Standard	.100
Result	.107
Last Cal	04/25/2017
Time:	09:50 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#24
#24

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	008848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1328	34	17GP41049	20	20	X					M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 12:29
AS4 Serial no: 008848
Version no: 4797

No. Date Time XBAC
Calibration
05/21/17 08:30 .100

Test Result
0001 06/15/18 12:28 .106

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 160 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1345	34	17GP41049	27	101				X		M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 13:49
AS4 Serial no: 077591
Version no: 4797

No. Date Time %BAC
Calibration
05/21/17 08:35 .100

Test Result
0001 06/08/18 16:10 .018
0002 06/15/18 13:48 .101

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPT 061

#26

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	12:19	34	17GP41049		VALUE	21						M. Tippett	20137	
					.100									

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080065
UOM	BAC
Last Check	06/15/2018
Time:	12:20 PM
Cal Standard	.100
Result	.108
Last Cal	12/19/2017
Time:	1:54 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	077573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	132834		17GP41049	.100	21	.099				X	M. Tippett	20137	

```

06/15/18 13:29
AS4 Serial no: 077573
Version     no: 4797

No.   Date   Time   %BAC
Calibration
      01/18/18 11:23 .100

Test Result
0001 06/15/18 13:28 .099

Push No. 3 to erase,
or Power down to retain

```

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g, with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#28

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 083057
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1329	34	17GP41049	.100	21	1.00	X				M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date: 07-24-19										

06/15/18 13:30
AS4 Serial no: 083051
Version no: 4797

No. Date Time XBAC
Calibration
12/28/17 13:29 .100

Test Result
0001 06/15/18 13:29 .100

Push No. 3 to erase.
or power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#30

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
06/15/2018	1335	34	17GP41049		20		.106				X	M. Tippett	20137	
				Solution prepared by:										
				Cesar A. Mejia										
				Date: 07-24-17										
				Expiration date:										
				07-24-19										

06/15/18 13:37
AS4 Serial no: 008897
Version no: 4797

No. Date Time XBAC
Calibration
01/18/12 16:13 .100

Test Result
0001 06/15/18 13:36 .106

Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	1340	34	17GP41049	.100	Z-1				X		M. Tippett	20137	

06/15/18 13:41
 AS4 Serial no: 018624
 Version no: 4797

 No. Date Time %BAC
 Calibration
 02/01/16 11:02 .100

 Test Result
 0001 06/15/18 13:41 .105

 Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#32
PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	LOT NO.	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	12:22	34	17GP41049		.100	21		6/105			X		M. Tippett	20137	

Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19															
Lifeloc Technologies															
Calibration / Check Printout															
FC20															
8.7.24															
Serial No.															
16080071															
UOM															
BAC															
Last Check															
06/15/2018															
Time:															
12:22 PM															
Cal Standard															
.100															
Result															
.105															
Last Cal															
08/14/2017															
Time:															
10:41 AM															
Cal Standard															
.100															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions
Chp202K_1108.pdf

#34

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008877
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	1:50	34		17GP41049	.100	20				X		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/15/18 13:51
AS4 Serial no: 008877
Version no: 4797

No. Date Time XBRAC
Calibration
05/28/17 08:49 .100

Test Result
0001 06/15/18 13:50 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1221	34	17GP41049	23		.100					M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040163
UOM	BAC
Last Check	06/15/2018
Time:	12:22 PM
Cal Standard	.100
Result	.100
Last Cal	06/07/2017
Time:	11:02 AM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1350 34			17GP41049	20		104		X		M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 13:52
 AS4 Serial no: 018626
 Version no: 4797

No. Date Time XBAC
 Calibration
 07/15/16 15:21 .100

Test Result
 0001 06/15/18 13:51 .104

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 **. If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080072
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1230	34	.100	17GP41049	21							M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080072
UOM	BAC
Last Check	06/15/2018
Time:	12:31 PM
Cal Standard	.100
Result	.106
Last Cal	12/19/2017
Time:	2:05 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/AIco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1353	34	17GP41049	.100	21	.100				X	M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 13:54
 AS4 Serial no: 003030
 Version no: 4797

No. Date Time XBAC
 Calibration
 12/28/17 13:51 .100

Test Result
 0001 06/15/18 13:53 .100

Push No. 3 to erase,
 or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#40

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018692
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
			VALUE		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)						
06/15/2018	1356:34	34	17GP41049		21				X			M. Tippett	20137	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date:											
			07-24-19											

06/15/18 13:57
AS4 Serial no: 018692
Version no: 4797

No. Date Time XBAC
Calibration
02/01/16 11:30 .100

Test Result
0001 06/15/18 13:57 .106

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1358	34	17GP41049	.100	20	.098			X		M. Tippet	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 14:01
 AS4 Serial no: 018762
 Version no: 4797

No. Date Time %BAC
 Calibration
 12/28/17 14:01 .100

Test Result
 0001 06/08/18 01:24 .016
 0002 06/09/18 02:12 .027
 0003 06/15/18 14:01 .098

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#143

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018766
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1402 34		17GP41049	21		.099		X			M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 14:03
 AS4 Serial no: 018766
 Version no: 4797

No. Date Time %BAC
 Calibration
 01/22/16 13:16 .100

Test Result
 0001 06/09/18 02:54 .005
 0002 06/15/18 14:03 .099

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018 691
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/15/2018	1404 34			17GP41049							X	M. Tippett	20137	

06/15/18 14:05
 A64 Serial no: 018691
 Version no: 4797

No. Date Time XBAC
 Calibration
 05/28/17 08:26 .100

Test Result
 0001 06/15/18 14:05 .100

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alcoo-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------	---------------------	--	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1406 34			17GP41049							M. Tippett	20137	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 14:07
 A84 Serial no: 017990
 Version no: 4797

No. Date Time XBAC
 Calibration
 11/26/03 11:10 .100

Test Result
 0001 06/15/18 14:06 .108

Push No. 3 to erase,
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#46

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1227	34	17GP41049	.100	22						X	M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120189
UOM	BAC
Last Check	06/15/2018
Time:	12:27 PM
Cal Standard	.100
Result	.098
Last Cal	09/01/2017
Time:	12:17 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial-number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 041486
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	1408 34			17GP41049	20	105			X		M. Tippett	20137	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19										

06/15/18 14:09
AG4 Serial no: 041486
Version no: 4797

No. Date Time ZBAC
Calibration
03/08/06 13:56 .100

Test Result
0001 06/15/18 14:08 .105

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

AREA NAME		AREA NO.	COORDINATOR NAME		ID. NUMBER	PAS MAKE/MODEL		PAS SERIAL NO.					
South Los Angeles		530	Kirk Lee		17046	Intoximeters/Aleo-Sensor IV		041495					
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/15/2018	140734		17GP41049		28				X		M. Tippett	20137	
		Solution prepared by: Cesar A. Mejia											
		Date: 07-24-17											
		Expiration date: 07-24-19											

06/15/18 14:08
AS4 Serial no: 041495
Version no: 4797

No. Date Time %BAC
Calibration
12/28/17 14:10 .100

Test Result
0001 06/15/18 14:08 .100

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
06/15/2018	1231	34		17GP41049	22				X		M. Tippett	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16080013
UOM	BAC
Last Check	06/15/2018
Time:	12:31 PM
Cal Standard	.100
Result	.105
Last Cal	01/29/2018
Time:	12:59 PM
Cal Standard	.100

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#53

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO.: 082-654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/15/2018	1410	34		17GP41049	.100	21				X		M. Tippett	20137	
Solution prepared by: Cesar A. Mejia														
Date: 07-24-17														
Expiration date: 07-24-19														

06/15/18 14:11
AS4 Serial no: 082654
Version no: 4797

No. Date Time %BAC
Calibration
12/19/16 11:26 .100

Test Result
0001 06/12/18 23:20 .153
0002 06/14/18 01:21 .185
0003 06/14/18 01:25 .186
0004 06/15/18 02:05 .000
0005 06/15/18 02:30 .147
0006 06/15/18 02:33 .151
0007 06/15/18 14:11 .106

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPT 051

SAT'S PAS

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)				
06/15/2018	1420	34		17GP41049							M. Tippett	20137	
										X			
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/15/18 14:21
AS4 Serial no: 008874
Version no: 4797

No. Date Time %BAC
Calibration
10/24/17 11:04 .100

Test Result
0001 06/09/18 23:57 .170
0002 06/10/18 00:00 .166
0003 06/11/18 23:40 .193
0004 06/11/18 23:43 .191
0005 06/15/18 14:20 .099

Push No. 3 to erase,
or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.