

567

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008874
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1302	34		16FP41018	.100	20		099		X			K. Lee	17046	
			Solution prepared by:												
			Cesar A. Mejia												
			Date: 07-24-17												
			Expiration date:												
			07-24-19												

06/25/18 13:05  
 AS4 Serial no: 008874  
 Version no: 4797

No.	Date	Time	%BAC
Calibration			
	10/24/17	11:04	.100
Test Result			
0001	06/16/18	04:59	.129
0002	06/16/18	05:02	.130
0003	06/17/18	00:12	.157
0004	06/17/18	00:15	.153
0005	06/17/18	02:30	.178
0006	06/17/18	02:32	.174
0007	06/18/18	01:12	.109
0008	06/18/18	01:15	.102
0009	06/18/18	04:20	.084
0010	06/18/18	04:23	.064
0011	06/18/18	04:27	.067
0012	06/21/18	03:25	.233
0013	06/21/18	03:27	.233
0014	06/21/18	03:37	.130
0015	06/23/18	01:01	.165
0016	06/23/18	01:03	.163
0017	06/23/18	04:49	.026
0018	06/23/18	04:52	.022
0019	06/23/18	19:51	.255
0020	06/23/18	19:54	.258
0021	06/25/18	13:05	.099

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#1

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120188
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1145	34	16FP41018	24		100			X		Lee	17046	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120188
UOM	BAC
Last Check	06/25/2018
Time:	11:42 AM
Cal Standard	.100
Result	.100
Last Cal	10/21/2016
Time:	08:16 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent  
\*\* If using dry gas for accuracy checks, note the canister st

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI051

#2

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120194
--------------------------------	-----------------	------------------------------	---------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0933	34	16FP41018	.100	22	.101			X		Lee	17046	

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120194
UOM	BAC
Last Check	06/25/2018
Time:	09:33 AM
Cal Standard	.100
Result	.101
Last Cal	01/29/2018
Time:	08:50 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent  
 \*\* If using dry gas for accuracy checks, note the canister s

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#3

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120197
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0945	39	16FP41018	.100	22		.103		X		Lee	17046		

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120197
UOM	BAC
Last Check	06/25/2018
Time:	09:48 AM
Cal Standard	.100
Result	.103
Last Cal	10/21/2016
Time:	10:22 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of  
\*\* If using dry gas for accuracy checks, note the canister ser

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

#4

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120195
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0949 34		16FP41018	.100	22		.107		X			Lee	17046	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120195
UOM	BAC
Last Check	06/25/2018
Time:	09:47 AM
Cal Standard	.100
Result	.107
Last Cal	04/25/2017
Time:	08:47 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of  
\*\* If using dry gas for accuracy checks, note the canister seri

#5  
PAS SERIAL NO. 008559

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008559
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	VALUE	PAS. TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0953	34	16FP41018	.100	20		.102			X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

```

06/25/18 09:58
AS4 Serial no: 008559
Version      no: 4797

No.   Date   Time   ZBAC
Calibration
      01/15/16 10:40 .100

Test Result
0001 06/25/18 09:57 .102

Push No. 3 to erase.
  
```

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#6

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040161
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0956	34	16FP41018	.100	22	.102			X		Lee	17046	

Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
<b>Lifeloc Technologies</b>													
Calibration / Check Printout													
FC20			8.7.24										
Serial No.			16040161										
UOM			BAC										
Last Check			06/25/2018										
Time:			10:00 AM										
Cal Standard			.100										
Result			.102										
Last Cal			04/25/2017										
Time:			08:56 AM										
Cal Standard			.100										

\* If the reading is within plus or minus (±) 0.010 percent c  
\*\* If using dry gas for accuracy checks, note the canister se

#7

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040140
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP ("C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP ("C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	0958	34	16FP41018	22							Lee	17046	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040140
UOM	BAC
Last Check	06/25/2018
Time:	10:08 AM
Cal Standard	.100
Result	.100
Last Cal	04/25/2017
Time:	08:58 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of i  
\*\* If using dry gas for accuracy checks, note the canister seri:



#8

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008620
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1012	34	16FP41018	.100	20	.107			X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date: 07-24-19										

or power down to retain

06/25/18 10:19  
AS4 Serial no: 008620  
Version no: 4797

No. Date Time ZBAC  
Calibration  
01/15/16 11:10 .100

Test Result  
0001 06/25/18 10:18 .107

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082957
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1015	34	16FP41018		.100	21		102		X		K. Lee	17046	
				Solution prepared by:										
				Cesar A. Mejia										
				Date: 07-24-17										
				Expiration date:										
				07-24-19										

06/25/18 10:21  
AS4 Serial no: 082957  
Version no: 4797

No. Date Time %BAC  
Calibration  
02/17/14 08:52 .100

Test Result  
0001 06/15/18 23:19 .017  
0002 06/25/18 10:15 .102

Push No. 3 to erase,

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

# 10

AREA NAME: South Los Angeles  
AREA NO.: 530  
COORDINATOR NAME: Kirk Lee  
I.D. NUMBER: 17046  
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV  
PAS SERIAL NO.: 077505

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1027	34	16FP41018		.100	21				X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

or Power down to retain

06/25/18 10:44  
AS4 Serial no: 077505  
Version no: 4797

No. Date Time %BAC  
Calibration  
01/18/18 10:56 .100

Test Result  
0001 06/25/18 10:27 .100

Push No. 3 to erase.

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
 PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG  
 CHP 202K (Rev. 8-08) OPI051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082653
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

DATE	TIME	TEMP ("C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP ("C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP ("C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
06/25/2018	1046	34	16FP41018		.100	21					X	K. Lee	17046		
				Solution prepared by:											
				Cesar A. Mejia											
				Date: 07-24-17											
				Expiration date:											
				07-24-19											

or power down to retain

06/25/18 10:45  
 AS4 Serial no: 082653  
 Version no: 4797

No. Date Time %BAC  
 Calibration  
 02/11/16 10:14 .100

Test Result

0001 06/17/18 03:52 .056  
 0002 06/23/18 06:32 .060  
 0003 06/23/18 06:35 .000  
 0004 06/23/18 22:57 .110  
 0005 06/23/18 23:02 .112  
 0006 06/23/18 23:07 .000  
 0007 06/25/18 10:44 .106

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008709
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS			
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO		
06/25/2018	1049	34		16FP41018	.100	21		.106				X				17046	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19																	
Push No. 3 to erase, or power down to retain																	
06/25/18 10:51 AS4 Serial no: 008709 Version no: 4797																	
No. Date Time %BAC Calibration 03/17/16 09:40 .100																	
Test Result 0001 06/16/18 02:49 .093 0002 06/16/18 02:51 .091 0003 06/17/18 19:57 .021 0004 06/17/18 20:00 .025 0005 06/18/18 02:11 .164 0006 06/18/18 02:14 .165 0007 06/25/18 10:51 .106																	

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME: South Los Angeles

#14

AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV.	008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1051	34	16FP41018	.100	21					X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

Push No. 3 to erase,  
or power down to retain

Test Result  
0001 06/25/18 10:55 .103

No. Date Time ZBAC  
Calibration  
05/14/17 09:50 .100

06/25/18 10:56  
AS4 Serial no: 008714  
Version no: 4797

or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#16

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/AIco-Sensor IV	PAS SERIAL NO. 008888
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1054	34	16FP41018	.100	21	.106			X		K. Lee	17046	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/25/18 10:51  
 AS4 Serial no: 008888  
 Version no: 4797

No. Date Time %BAC  
 Calibration  
 03/17/16 09:44 .100

Test Result  
 0001 06/19/18 23:07 .243  
 0002 06/19/18 23:10 .247  
 0003 06/20/18 02:15 .224  
 0004 06/20/18 02:17 .213  
 0005 06/21/18 00:24 .202  
 0006 06/21/18 00:27 .207  
 0007 06/25/18 10:50 .106

Push No. 3 to erase,

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#17

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008783
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1056	34	16FP41018		20		103		X		K. Lee	17046	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

or power down to retain

06/25/18 10:59  
AS4 Serial no: 008783  
Version no: 4797

No. Date Time ZBAC  
Calibration  
05/14/17 10:07 .100

Test Result  
0001 06/17/18 22:45 .171  
0002 06/17/18 22:47 .138  
0003 06/17/18 22:49 .171  
0004 06/19/18 22:56 .166  
0005 06/19/18 22:59 .162  
0006 06/22/18 23:32 .000  
0007 06/23/18 01:29 .176  
0008 06/23/18 01:31 .192  
0009 06/24/18 21:39 .256  
0010 06/24/18 21:42 .228  
0011 06/24/18 21:45 .250  
0012 06/25/18 09:51 .000  
0013 06/25/18 10:59 .103

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



### PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#18

AREA NAME		AREA NO.	COORDINATOR NAME		ID. NUMBER	PAS MAKE/MODEL		PAS SERIAL NO.						
South Los Angeles		530	Kirk Lee		17046	Intoximeters/Alco-Sensor IV		082657						
DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
06/25/2018	1058	34	16FP41018	.100	21					X		K. Lee	17046	
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date:											
			07-24-19											
<p>Push No. 3 to erase, or power down to retain.</p> <p>Test Result 0001 06/25/18 10:58 .101</p> <p>Calibration 05/21/17 08:18 .100</p> <p>No. Date Time XBAC 06/25/18 10:58 AS4 Serial no: 082657 Version no: 4797</p>														

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI051

#19

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008810
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1107	34	16FP41018		.100	21					X	K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/25/18 11:11  
 AS4 Serial no: 008810  
 Version no: 4797

No. Date Time XBAC  
 Calibration  
 05/21/17 08:20 .100

Test Result  
 0001 06/25/18 11:10 .101

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/AIco-Sensor IV	PAS SERIAL NO. 17046
--------------------------------	-----------------	------------------------------	----------------------	---	-------------------------

#20  
018684

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1109	34	16FP41018	.100	20					X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/25/18 11:14  
AS4 Serial no: 018684  
Version no: 4797

No. Date Time %BAC  
Calibration  
01/12/16 14:36 .093

Test Result  
0001 06/25/18 11:13 .107

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#21

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008635
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	TEMP (°C)		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1113	34	16FP41018		.100	21					X		K. Lee	17046	
			Solution prepared by:												
			Cesar A. Mejia												
			Date: 07-24-17												
			Expiration date:												
			07-24-19												

06/25/18 11:16  
AS4 Serial no: 008635  
Version no: 4797

No. Date Time %BAC  
Calibration  
05/21/17 08:23 .100

Test Result  
0001 06/25/18 11:15 .104

Push No. 3 to erase,  
or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

# 22

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040162
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1115	34		16FP41018	.100	23		096			X		Lee	17046	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040162
UOM	BAC
Last Check	06/25/2018
Time:	11:12 AM
Cal Standard	.100
Result	.096
Last Cal	08/14/2017
Time:	10:14 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of t  
\*\* If using dry gas for accuracy checks, note the canister serial

#23

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120177
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1119	23	16FP41018	.100	23	.104			X		Lee	17046	

### Lifeloc Technologies

Calibration / Check Printout

FC20	8.7.24
Serial No.	16120177
UOM	BAC
Last Check	06/25/2018
Time:	11:16 AM
Cal Standard	.100
Result	.104
Last Cal	04/25/2017
Time:	09:50 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of tl  
 \*\* If using dry gas for accuracy checks, note the canister seria.

# 24

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/AIco-Sensor IV	PAS SERIAL NO. 008848
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS	
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1					TEST #2 (Optional)
06/25/2018	1121	34	16FP41018		.100		TEST #1 .103					K. Lee	17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19											
			<p style="text-align: center;">06/25/18 10:28 AS4 Serial no: 008848 Version no: 4797</p> <p>No. Date Time %BAC Calibration 05/21/17 08:30 .100</p> <p>Test Result 0001 06/25/18 10:24 .103</p> <p>Push No. 3 to erase, or power down to retain</p>											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

CHP 202K (Rev. 8-08) OPI 051

#25

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077591
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1127	34	16FP41018	.100	20		.102		X		K. Lee	17046	
			Solution prepared by:										
			Cesar A. Mejia										
			Date: 07-24-17										
			Expiration date:										
			07-24-19										

06/25/18 11:28  
AS4 Serial no: 077591  
Version no: 4797

No. Date Time XBAC  
Calibration  
05/21/17 08:35 .100

Test Result  
0001 06/15/18 15:40 .088  
0002 06/15/18 15:43 .085  
0003 06/16/18 06:37 .084  
0004 06/16/18 06:39 .084  
0005 06/16/18 06:40 .083  
0006 06/23/18 15:26 .000  
0007 06/25/18 11:27 .102

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080065
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	TEST #1				
06/25/2018	1129	34	16FP41018		22	103				X	Lee	17046	

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080065
UOM	BAC
Last Check	06/25/2018
Time:	11:26 AM
Cal Standard	.100
Result	.103
Last Cal	12/19/2017
Time:	1:54 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of  
 \*\* If using dry gas for accuracy checks, note the canister seri:

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077573
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1:31	30		16FP41018	.100	22					X	K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/25/18 11:31  
AS4 Serial no: 077573  
Version no: 4797

No. Date Time XBAC  
Calibration  
01/18/18 11:23 .100

Test Result  
0001 06/25/18 11:31 .097

Push No. 3 to erase,

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME	AREA NO.	COORDINATOR NAME	I.D. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/AIco-Sensor IV	#28 083051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1140	34	16FP41018	.100	21		.097		X			K. Lee	17046	

06/25/18 11:39  
 AS4 Serial no: 083051  
 Version no: 4797  
 No. Date Time %BAC  
 Calibration 12/28/17 13:29 .100  
 Test Result 0001 06/25/18 11:39 .097  
 Push No. 3 to erase, or power down to retain  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 008897
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

#30

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1142	34	16FP41018	.100	21		.103		X		K. Lee	17046		
			Solution prepared by:											
			Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date:											
			07-24-19											
			<div style="border: 1px solid black; padding: 5px; margin: 10px 0;">                 No. Date Time ZBAC Calibration                  01/18/12 16:13 .100                   Test Result                  0001 06/25/18 11:46 .103                   Push No. 3 to erase,                  or Power down to retain             </div>											
			06/25/18 11:46 A94 Serial no: 008897 Version no: 4797											

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA  
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
 CHP 202K (Rev. 8-08) OPI 051

#31

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	ID. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018624
--------------------------------	-----------------	------------------------------	---------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1147	34		16FP41018	.100	22	.102		X		K. Lee	17046		

06/25/18 11:53  
 AS4 Serial no: 018624  
 Version no: 4797

No. Date Time XBAC  
 Calibration  
 02/01/16 11:02 .100

Test Result  
 0001 06/25/18 11:50 .102

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#32

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080071
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1151	34	16FP41018	23		.100				X	Lee	17046	

<b>Lifeloc Technologies</b>												
Calibration / Check Printout												
FC20		8.7.24										
Serial No.		16080071										
UOM		BAC										
Last Check		06/25/2018										
Time:		11:49 AM										
Cal Standard		.100										
Result		.100										
Last Cal		08/14/2017										
Time:		10:41 AM										
Cal Standard		.100										

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

\* If the reading is within plus or minus (±) 0.010 percent of t  
\*\* If using dry gas for accuracy checks, note the canister seria

#33

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 077488
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1201	34	16FP41018	.100	22	105			X			K. Lee	17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19											

06/25/18 11:59  
AS4 Serial no: 077488  
Version no: 4797

No. Date Time %BAC  
Calibration  
02/11/16 10:52 .100

Test Result  
0001 06/25/18 11:59 .105

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/AIco-Sensor IV	PAS SERIAL NO. #34
--------------------------------	-----------------	------------------------------	----------------------	---	-----------------------

008877

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1203 34	34	16FP41018		.100	21		099		X			K. Lee	17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												

06/25/18 12:09  
AS4 Serial no: 008877  
Version no: 4797

No. Date Time ZBAC  
Calibration  
05/29/17 08:49 .100

Test Result  
0001 06/25/18 12:09 .099

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.



PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#36

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16040163
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
06/25/2018	1206 34		16FP41018		.100					X	Lee	17046	

Solution prepared by:  
Cesar A. Mejia  
Date: 07-24-17  
Expiration date:  
07-24-19

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16040163
UOM	BAC
Last Check	06/25/2018
Time:	12:01 PM
Cal Standard	.100
Result	.094
Last Cal	06/07/2017
Time:	11:02 AM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of t  
\*\* If using dry gas for accuracy checks, note the canister serial

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#37

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018626
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS			
			TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO							
06/25/2018	120834			16FP41018	.100			21					X		K. Lee	17046		

06/25/18 12:08  
 AS4 Serial no: 018626  
 Version no: 4797

No. Date Time %BAC  
 Calibration  
 07/15/16 15:21 .100

Test Result  
 0001 06/25/18 12:07 .102

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#38

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifoloc/FC20	PAS SERIAL NO. 16080072
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	12:19	34	16FP41018	23	.100	.100	X				Lee	17046	

<b>Lifoloc Technologies</b>												
Calibration / Check Printout												
FC20		8.7.24										
Serial No.		16080072										
UOM		BAC										
Last Check		06/25/2018										
Time:		12:23 PM										
Cal Standard		.100										
Result		.100										
Last Cal		12/19/2017										
Time:		2:05 PM										
Cal Standard		.100										

Solution prepared by:  
 Cesar A. Mejia  
 Date: 07-24-17  
 Expiration date:  
 07-24-19

\* If the reading is within plus or minus (±) 0.010 percent of:  
 \*\* If using dry gas for accuracy checks, note the canister seri:

PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG

#39

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 083030
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	12234			16FP41018	.100	21					X	K. Lee	17046	
				Solution prepared by:										
				Cesar A. Mejia										
				Date: 07-24-17										
				Expiration date:										
				07-24-19										

06/25/18 12:22  
AS4 Serial no: 083030  
Version no: 4797

No. Date Time %BAC  
Calibration  
12/28/17 13:51 .100

Test Result  
0001 06/25/18 12:21 .098

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 17046
--------------------------------	-----------------	------------------------------	----------------------	---	-------------------------

#40

018692

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	12:24	34	16FP41018	.100	22	103				X		17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19										

06/25/18 12:25  
AS4 Serial no: 018692  
Version no: 4797

No. Date Time %BAC  
Calibration  
02/01/16 11:30 .100

Test Result  
0001 06/25/18 12:24 .103

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

South Los Angeles

#42

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 018762
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	122734			16FP41018	.100	22			096		X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												

06/25/18 12:30  
AS4 Serial no: 018762  
Version no: 4797

No. Date Time XBAC  
Calibration  
12/28/17 14:01 .100

Test Result

0001	06/16/18	02:10	.134
0002	06/16/18	02:12	.143
0003	06/17/18	01:29	.028
0004	06/17/18	02:20	.264
0005	06/17/18	02:22	.227
0006	06/17/18	02:24	.261
0007	06/17/18	04:07	.117
0008	06/17/18	04:09	.116
0009	06/22/18	00:08	.198
0010	06/22/18	00:11	.191
0011	06/22/18	03:41	.000
0012	06/23/18	23:04	.000
0013	06/23/18	23:06	.000
0014	06/24/18	02:35	.051
0015	06/24/18	03:33	.112
0016	06/24/18	03:36	.117
0017	06/24/18	03:58	.044
0018	06/24/18	22:56	.042
0019	06/25/18	12:29	.096

Push No. 3 to erase,  
or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. #43 018766
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1224	34	16FP41018	.100	22	.103			X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19										
			06/25/18 12:32 AS4 Serial no: 018766 Version no: 4797										
			No. Date Time %BAC Calibration 01/22/16 13:16 .100										
			Test Result 0001 06/16/18 01:46 .226 0002 06/16/18 01:49 .239 0003 06/16/18 20:11 .084 0004 06/16/18 20:15 .091 0005 06/17/18 01:38 .046 0006 06/17/18 01:42 .054 0007 06/17/18 23:26 .009 0008 06/22/18 03:07 .136 0009 06/22/18 03:11 .156 0010 06/23/18 00:14 .023 0011 06/23/18 00:17 .024 0012 06/23/18 02:42 .007 0013 06/23/18 02:46 .006 0014 06/25/18 12:32 .103										
			Push No. 3 to erase, or Power down to retain										

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

AREA NAME South Los Angeles		AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. #44 018691
--------------------------------	--	-----------------	------------------------------	----------------------	---	---------------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1732 34		16FP41018	.100	TEST #1 22	TEST #2 (Optional) -098	TEST #1 -098	TEST #2 (Optional)			K. Lee	17046	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
06/25/18 12:35 AS4 Serial no: 018691 Version no: 4797  No. Date Time %BAC Calibration 05/28/17 08:26 .100  Test Result 0001 06/25/18 12:34 .098  Push No. 3 to erase.													

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

Chp202K\_1108.pdf



**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPT 051

#45

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 017990
--------------------------------	-----------------	------------------------------	----------------------	---	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS		
			TEMP (°C)	LOT NO.		VALUE	TEST #1	TEST #2 (Optional)	YES					NO	
			TEST #1	TEST #2 (Optional)											
06/25/2018	1230	34	16PP41018		.100							K. Lee	17046		
			Solution prepared by:												
			Cesar A. Mejia												
			Date: 07-24-17												
			Expiration date:												
			07-24-19												

or power down to retain

---

06/25/18 12:36  
 AS4 Serial no: 017990  
 Version no: 4797

No. Date Time %BAC  
 Calibration 11/26/03 11:10 .100

Test Result  
 0001 06/25/18 12:36 .105

Push No. 3 to erase,  
 or power down to retain

\* If the reading is within plus or minus ( $\pm$ ) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
 \*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

#46

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16120189
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
06/25/2018	12:36	34	16FP41018	23			095			X	Lee	17046	

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16120189
UOM	BAC
Last Check	06/25/2018
Time:	12:33 PM
Cal Standard	.100
Result	.095
Last Cal	09/01/2017
Time:	12:17 PM
Cal Standard	.100

TEMP (°C)	LOT NO.	VALUE
		.100
Solution prepared by:		
Cesar A. Mejia		
Date: 07-24-17		
Expiration date:		
07-24-19		

\* If the reading is within plus or minus (±) 0.010 percent c  
\*\* If using dry gas for accuracy checks, note the canister se.....

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

AREA NAME: **South Los Angeles**      AREA NO.: **530**      COORDINATOR NAME: **Kirk Lee**      I.D. NUMBER: **17046**      PAS MAKE/MODEL: **Intoximeters/Alco-Sensor IV**      PAS SERIAL NO.: **17046**

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES					NO
06/25/2018	12:34	36	16FP41018	.100	22	102					X	K. Lee	17046	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														

06/25/18 12:45  
AS4 Serial no: 041486  
Version no: 4797

No.    Date    Time    ZBAC  
Calibration  
03/08/06 13:56 .100

Test Result  
0001 06/25/18 12:45 .102

Push No. 3 to erase,  
or power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**

CHP 202K (Rev. 8-08) OPI 051

#51

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alico-Sensor IV	PAS SERIAL NO. 041495
--------------------------------	-----------------	------------------------------	----------------------	--	--------------------------

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES				
06/25/2018	1250	34	16FP41018	.100	21	.097				X		K. Lee	17046	
			Solution prepared by: Cesar A. Mejia											
			Date: 07-24-17											
			Expiration date: 07-24-19											

06/25/18 12:53  
AS4 Serial no: 041495  
Version no: 4797

No. Date Time %BAC  
Calibration  
12/28/17 14:10 .100

Test Result  
0001 06/25/18 12:52 .097

Push No. 3 to erase,

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

#52

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Lifeloc/FC20	PAS SERIAL NO. 16080013
--------------------------------	-----------------	------------------------------	----------------------	--------------------------------	----------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP ("C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP ("C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
			TEMP	LOT NO.		TEST #1	TEST #2	YES	NO					PAS
06/25/2018	1258	34	16FP41018		23	.100	TEST #1	TEST #2	YES	NO	X	Lee	17046	

Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														

Lifeloc Technologies	
Calibration / Check Printout	
FC20	8.7.24
Serial No.	16080013
UOM	BAC
Last Check	06/25/2018
Time:	12:59 PM
Cal Standard	.100
Result	.099
Last Cal	01/29/2018
Time:	12:59 PM
Cal Standard	.100

\* If the reading is within plus or minus (±) 0.010 percent of  
\*\* If using dry gas for accuracy checks, note the canister seri

STATE OF CALIFORNIA  
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL  
**PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG**  
CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME South Los Angeles    AREA NO. 530    COORDINATOR NAME Kirk Lee    I.D. NUMBER 17046    PAS MAKE/MODEL Intoximeters/Alco-Sensor IV    PAS SERIAL NO. 002654

**PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.**

DATE	TIME	TEMP ("C)	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
06/25/2018	1300	34	16FP41018	.100	22	22	.102	X			K. Lee	17046		

Push No. 3 to erase,  
or power down to retain

No. Date Time %BAC  
Calibration  
12/19/16 11:26 .100

Test Result  
0001 06/20/18 03:28 .044  
0002 06/20/18 03:30 .042  
0003 06/20/18 05:21 .000  
0004 06/21/18 02:46 .000  
0005 06/25/18 13:01 .102

06/25/18 13:01  
AS4 Serial no: 002654  
Version no: 4797

or Power down to retain

\* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).  
\*\* If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.