

#1

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120188

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1008	34	17GP41049	.100	22						X	M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	REPAIRS/PROBLEMS/COMMENTS
.100	08:16 AM	10/21/2016	.100	.100	10:09 AM	07/06/2018	BAC	16120188	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#2

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120194

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	1008	34	17GP41049		100	22	.096	X		M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date: 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:50 AM	01/29/2018	.096	.100	10:08 AM	07/06/2018	BAC	16120194	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

3

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16120197

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS	
							TEST #1 (Optional)	TEST #2 (Optional)						
07/06/2018	1018	34	17GP41049	.100	21		100	101	NO	X	M. Tippet	20137		
Solution prepared by: Cesar A. Mejia														
Date: 07-24-17														
Expiration date: 07-24-19														
Cal Standard					Time:		Last Cal		Result		Cal Standard		Time:	
					10:22 AM		10/21/2016		.101		.100		10:19 AM	
									Last Check		UOM		Serial No.	
									07/06/2018		BAC		16120197	
											FC20		8.7.24	
Calibration / Check Printout														
Lifeloc Technologies														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#4

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/PC20
 PAS SERIAL NO.: 16120195

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	1015	34	17GP41049	.100	22		107		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	08:47 AM	04/25/2017	.107	.100	10:15 AM	07/06/2018	BAC	16120195	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#5

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008559

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)		YES	NO			
07/06/2018	11:21	34	17GP41049	.100	20							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

07/06/18 11:22 AS4 Serial no: 008559
 Version no: 4797
 No. Date Time ZBAC
 Calibration 01/15/16 10:40 .100
 Test Result 0001 07/06/18 11:22 .103
 Push No. 3 to erase,
 OR Power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#6

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: **16040161**

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
				TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	10:26	34	17GP41049	.100	21	.105			X		
Solution prepared by: Cesar A. Mejia											
Date: 07-24-17											
Expiration date: 07-24-19											
			Cal Standard		Time:	Last Cal	04/25/2017	Result	.105	Cal Standard	.100
					Time:	Last Check	07/06/2018			UOM	BAC
					Time:					Serial No.	16040161
					Time:					FC20	8.7.24
										Calibration / Check Printout	
Lifeloc Technologies											

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#7

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16040140

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	10:16	34	17GP41049	.100	22		105			X		M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	08:58 AM	04/25/2017	.105	.100	10:17 AM	07/06/2018	BAC	16040140	8.7.24	

Lifeloc Technologies
 Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#8

AREA NAME
South Los Angeles

AREA NO.
530

COORDINATOR NAME
Kirk Lee

I.D. NUMBER
17046

PAS MAKE/MODEL
Intoximeters/Alco-Sensor-IV

PAS SERIAL NO.
008620

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	112534	17GP41049		.100	20							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
<pre> 07/06/18 11:26 AS4 Serial no: 008620 Version no: 4797 No. Date Time %BAC Calibration 01/15/16 11:10 .100 Test Result 0001 07/06/18 11:26 .108 Push No. 3 to erase, OR Power down to retain </pre>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

9

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. 082 957
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1/21	34	17GP41049	.100	21		103		X			M. Tippert	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
07/06/18 11:22 MS4 Serial no: 082957 Version no: 4797 No. Date Time XBAC Calibration 02/17/14 08:52 .100 Test Result 0001 07/06/18 11:21 .103 Push No. 3 to erase, or power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

Destroy Previous Editions

#110

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 077505

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.				TEMP (°C)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	11:23 34		17GP41049	.100	21		101		X			M. Tippert	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19</p>														
<p>No. Date Time %BAC 07/06/18 11:25 MS4 Serial no: 077505 Version no: 4797 Calibration 01/18/18 10:56 .100 Test Result 0001 07/06/18 11:24 .101 Push No. 3 to erase, OR Power down to retain</p>														

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions
Chp202K_1108.pdf

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#11

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 077479

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
						TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	126	34		17GP41049	.100								M. Tippett	20137	DEVICE NOT IN COLLON FOR TESTING

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#12

AREA NAME South Los Angeles	AREA NO. 530	COORDINATOR NAME Kirk Lee	I.D. NUMBER 17046	PAS MAKE/MODEL Intoximeters/Alco-Sensor IV	PAS SERIAL NO. <i>082653</i>
--------------------------------	-----------------	------------------------------	----------------------	---	---------------------------------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**			PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)				
07/06/2018	<i>1128 34</i>		17GP41049	.100							M. Tippet	20137	<i>Device not in locker from testing</i>

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#13

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 008709

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.				TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	YES	NO				
07/06/2018	1129	34	17GPA1049		.100	25							X	M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19</p> <p>Test Result 0001 07/01/18 01:49 .133 0002 07/01/18 01:52 .160 0003 07/01/18 01:55 .188 0004 07/01/18 02:45 .190 0005 07/01/18 02:47 .177 0006 07/04/18 18:53 .000 0007 07/04/18 18:59 .000 0008 07/06/18 11:30 .106</p> <p>Push No. 3 to erase, or Power down to retain</p> <p>07/06/18 11:31 NS4 Serial no: 006709 Version no: 4797 No. Date Time ZBAC Calibration 03/17/16 09:40 .100</p>																

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#74

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008714

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1/33 34	17GP41049	.100	21							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Push No. 3 to erase: 0003 07/06/18 11:34 .104 0002 07/04/18 19:36 .000 0001 07/04/18 14:14 .000 Test Result Calibration 05/14/17 09:50 .100 No. Date Time ZBMC Version no: 4797 NS4 Serial no: 008714 07/06/18 11:34													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor-IV PAS SERIAL NO.: 008888

#16

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEMP #1	TEMP #2	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1132	34	17GP41049	.100									Mr. Tippett	20137	DEVICE NOT IN LOCKER FOR TESTING	

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g. with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions

#117

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor-IV PAS SERIAL NO.: 008783

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST RESULTS			
07/06/2018	1135	34	17GP41049	.100	20							M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
Test Result: 0001 06/29/18 22:34 .113 0002 06/29/18 22:37 .091 0003 07/01/18 00:26 .065 0004 07/01/18 00:30 .062 0005 07/01/18 01:22 .124 0006 07/01/18 01:26 .104 0007 07/01/18 01:32 .067 0008 07/06/18 11:35 .104														
No. Date Time %BAC Calibration 05/14/17 10:07 .100 Version no: 4797 AS4 Serial no: 008783 07/06/18 11:36														
Push No. 3 to erase, or Power down to retain														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#18

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO.: 082657

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	11:34	34	17GP41049	.100	21		102				X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

07/06/18 11:35
 NS4 Serial no: 082657
 Version no: 4297
 No. Date Time XPRC
 Calibration 05/21/17 08:18 .100
 Test Result 0001 07/06/18 11:34 .102
 Push No. 3 to erase,
 or Power down to Retain

#19

AREA NAME
South Los Angeles

AREA NO.
530

COORDINATOR NAME
Kirk Lee

I.D. NUMBER
17046

PAS MAKE/MODEL
Intoximeters/Alco-Sensor IV

PAS SERIAL NO.
008816

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEMP #1	TEST #1 (Optional)	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1140	34	17GP41049	.100	20			109				M. Tippett	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19</p>														
<p>No. Date Time XPAC Calibration 05/21/17 08:20 .100 Test Result 0001 07/06/18 11:40 .104 Push No. 3 to erase, or Power down to retain</p>														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#20

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 018684

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1142	34	17GP41049	.100	29		106				M. Tippert	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

07/06/18 11:43 NS4 Serial no: 018684
 Version no: 4797
 No. Date Time XBAC
 Calibration 01/12/16 14:36 .093
 Test Result
 0001 06/28/18 06:24 .085
 0002 06/28/18 06:26 .098
 0003 06/28/18 06:27 .095
 0004 07/06/18 11:42 .106
 Push No. 3 to erase.
 or Power down to retain

#21

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008635

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)		LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS		OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP	LOT			TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO	RETEST	RESULTS			
07/06/2018	1143	34		17GPA1049	.100	20				Y				ML Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

07/06/18 11:44
 NS4 Serial no: 008635
 Version: not 4797
 No. Date Time XBAC
 Calibration 05/21/17 08:23 .100
 Test Result 0001 07/06/18 11:43 .106
 Push No. 3 to erase.
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

22

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16040162

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1018	34	17GP41049	.100	22		.098			X		ML Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	10:14 AM	08/14/2017	.098	.100	10:18 AM	07/06/2018	BAC	16040162	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#23

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120177

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	1022	34	17GP41049		100	22	106	X		M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	09:50 AM	04/25/2017	.106	.100	10:22 AM	07/06/2018	BAC	16120177	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#24

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 008848

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1147 34	17GP41049		.100	21						M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Push No. 3 to release or power down to retain													
07/06/18 10:48 NS4 Serial no: 008848 Version no: 4797 No. Date Time ZBAC Calibration 05/21/17 08:30 .100 Test Result 0001 07/06/18 10:48 .105													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202k_1108.pdf

#25

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 077591

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1105	34	17GP41049	.100	29					X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

07/06/18 11:06
 AS4 Serial no: 077591
 Version no: 4797
 No. Date Time %BAC
 Calibration 05/21/17 08:35 .100
 Test Result
 0001 07/03/18 10:26 .068
 0002 07/06/18 11:05 .102
 Push No. 3 to erase,
 or power down to retain

#26

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Lifeloc/FC20
PAS SERIAL NO.: *16080065*

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	<i>162334</i>		17GP41049								20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19												

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Lifeloc Technologies
.100	1:54 PM	12/19/2017	.107	.100	10:24 AM	07/06/2018	BAC	16080065	8.7.24	Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#27

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 077573

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1148	34	17GP41049	.100	22	.099		X			M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

07/06/18 11:49
 MS4 Serial no: 077573
 Version no: 4797
 No. Date Time XBAC
 Calibration 01/18/18 11:23 .100
 Test Result 0001 07/06/18 11:48 .099
 Push/No. 3 to erase
 or power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

28

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 083051

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
					TEST #1	TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	1149	34	17GP41049	.100	21							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

07/06/18 11:50 AS4 Serial no: 083051
 Version no: 4797
 No. Date Time XBAC
 Calibration 12/28/17 13:29 .100
 Test Result 0001 07/06/18 11:50 .099
 Push No. 3 to erase,
 or power down to retain

#30

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV | PAS SERIAL NO.: 002897

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	11:57:34			17GP41049	.100							X	M. Tippett	20137	

07/06/18 11:58
MS4 Serial no: 008897
Version no: 4297
No. Date Time ZBNC
Calibration 01/18/12 16:13 .100
Test Result
0001 07/06/18 11:57 .106
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#31

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 018624

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1159	34	17GP41049	.100	22						X	M. Tippett	20137	

07/06/18 12:00
MS4 Serial no: 018624
Version no: 4797
No. Date Time %BAC
Calibration 02/01/16 11:02 .100
Test Result 0001 07/06/18 11:59 .105
Push No. 3 to erase,
or Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
Destroy Previous Editions
Chp202K_1108.pdf

32

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Lifeloc/FC20
 PAS SERIAL NO.: 16080071

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.		TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	12:26:34		17GP41049	.100	22		.105		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	
.100	10:41 AM	08/14/2017	.105	.100	10:27 AM	07/06/2018	BAC	16080071	8.7.24	

Lifeloc Technologies

Calibration / Check Printout

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#33

AREA NAME	South Los Angeles	AREA NO.	530	COORDINATOR NAME	Kirk Lee	ID. NUMBER	17046	PAS MAKE/MODEL	Intoximeters/Alco-Sensor IV	PAS SERIAL NO.	077488
-----------	-------------------	----------	-----	------------------	----------	------------	-------	----------------	-----------------------------	----------------	--------

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO					
07/06/2018	1156	34	17GP41049		.100	21			108				M. Tippet	20137	
<p>Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19</p>															
<p style="text-align: center;"> No. Date Time ZBAC AS4 Serial no: 077488 Version no: 4797 07/06/18 11:52 Calibration 02/11/16 10:52 .100 Test Result 0001 07/06/18 11:56 .108 Push No. 3 to erase, or Power down to retain </p>															

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#34

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 008877

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1	TEST #2	TEST #1	TEST #2	YES	NO				
07/06/2018	12:34	17GP41049		.100	26									M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19																
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.																

07/06/18 12:07
 AS4 Serial no: 008877
 Version no: 4797
 No. Date Time %BAC
 Calibration 05/28/17 08:49 .100
 Test Result
 0001 07/05/18 06:36 .175
 0002 07/05/18 06:39 .188
 0003 07/06/18 12:06 .101
 Push No. 3 to erase,
 or Power down to retain.

#36

AREA NAME: South Los Angeles | AREA NO.: 530 | COORDINATOR NAME: Kirk Lee | I.D. NUMBER: 17046 | PAS MAKE/MODEL: Lifeloc/FC20 | PAS SERIAL NO.: 16040163

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1 (Optional)	TEST #2 (Optional)					
07/06/2018	1026	34		17GP41049	.100	23		.099		NO	X	M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	11:02 AM	06/07/2017	.099	.100	10:26 AM	07/06/2018	BAC	16040163	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#37

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO.: 018626

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1207 34		17GP41049	.100	21							M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
No. Date Time %BAC Calibration 07/15/16 15:21 .100 Test Result 0001 07/06/18 12:02 .105 Push No. 3 to erase, or Power down to retain														
07/06/18 12:03 AS4 Serial no: 018626 Version no: 4797														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#38

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16086072

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS REJECT RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
		TEMP (°C)	LOT NO.		VALUE	TEST #1 (Optional)	TEST #2 (Optional)	YES				
07/06/2018	1035	34	17GP41049	.100	21	106			X	M. Tippet	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	2:05 PM	12/19/2017	.106	.100	10:36 AM	07/06/2018	BAC	16080072	8.7.24	

Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#39

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV

PAS SERIAL NO.: 083030

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/06/2018	1204	34	17GP41049	.100	21		100				M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

07/06/18 12:05
 AS4 Serial no: 083030
 Version no: 4797
 No. Date Time XMAC
 Calibration 12/28/17 13:51 .100
 Test Result 0001 07/06/18 12:05 .100
 Push No. 3 to erase,
 or power down to retain

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#40

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 018692

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/06/2018	1208	34	17GP41049	.100	2	1	107		NO	X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.													

07/06/18 12:09
 AS4 Serial no: 018692
 Version no: 4797
 No. Date Time XBAC
 Calibration 02/01/16 11:30 .100
 Test Result 0001 07/06/18 12:08 .107
 Push No. 3 to erase,
 or power down to retain

#42

AREA NAME: South Los Angeles
 AREA NO: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO: 018762

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED	PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)					
07/06/2018	1311	34	17GP41049	.100	22		018		X		M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19													
Test Result Calibration 12/28/17 14:01 .100 No. Date Time ZBAC Version no: 4797 AS4 Serial no: 018762 07/06/18 13:12													
Push No. 3 to erase or Power down to retain													

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#43

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 018766

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
									YES	NO				
07/06/2018	1315	34	17GP41049	.100	22						X	M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
Test Result 0001 07/01/18 00:40 .059 0002 07/01/18 00:43 .072 0003 07/06/18 13:14 .098 OR Power-down to retain Push No. 3 to erase.														
No. Date Time XBAC Calibration 01/22/16 13:16 .100 Version no: 4797 AS4 Serial no: 018766 07/06/18 13:14														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

STATE OF CALIFORNIA
DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
CHP 202K (Rev. 8-08) OPI 051

#44

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV PAS SERIAL NO.: 018691

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.			TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1316	34	17GP41049		.100	22							M. Tippett	20137	
Solution prepared by: Cesar A. Mejia															
Date: 07-24-17															
Expiration date: 07-24-19															
<div style="text-align: right;"> 07/06/18 13:17 AS4 Serial no: 018691 Version no: 4797 No. Date Time XBAC Calibration 05/28/17 08:26 .100 Test Result 0001 07/06/18 13:16 .101 Push No. 3 to erase, or power down to retain </div>															

* If the reading is within plus or minus (\pm) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.
 Destroy Previous Editions
 Chp202K_1108.pdf

#45

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 017990

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1317	34	17GP41049	.100	22		109			X		M. Tippet	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.														

07/06/18 13:18
 NS4 Serial no: 017990
 Version no: 4792
 No. Date Time XBAC
 Calibration 11/26/03 11:18 .100
 Test Result 0001 07/06/18 13:17 .109
 Push No. 8 to erase.
 or Power down to retain

46

AREA NAME: South Los Angeles AREA NO.: 530 COORDINATOR NAME: Kirk Lee I.D. NUMBER: 17046 PAS MAKE/MODEL: Lifeloc/FC20 PAS SERIAL NO.: 16120189

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	SIMULATOR SOLUTION (OR DRY GAS)**		PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
		TEMP (°C)	LOT NO.	TEST #1	TEST #2 (Optional)	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	10:32 AM	17.0	17GP41049	22		0.099		X			M. Tippett	20137	

Solution prepared by:
 Cesar A. Mejia
 Date: 07-24-17
 Expiration date:
 07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout
.100	12:17 PM	09/01/2017	.099	.100	10:32 AM	07/06/2018	BAC	16120189	8.7.24	Lifeloc Technologies

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#50

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 041486

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**		VALUE	TEST #1	TEST #2 (Optional)	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
			LOT NO.					TEST #1	TEST #2	TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1319	34	17GP41049		100	22									M. Tippett	20137	
Solution prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19																	
* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110). ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.																	

07/06/18 13:20
 NS4 Serial no: 041486
 Version no: 4797
 No. Date Time %BAC
 Calibration 03/08/06 13:56 .100
 Test Result 0001 07/06/18 13:20 .104
 Push No. 3 to erase.
 or Power down to retain.

#151

AREA NAME: South Los Angeles
AREA NO: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
PAS SERIAL NO: 041495

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RE-TEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/ COMMENTS
							TEST #1	TEST #2	YES	NO				
07/06/2018	1319	34	17GP41049	.100	20	100	TEST #1	TEST #2	YES	NO	X	M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

07/06/18 13:20
NS4 Serial no: 041495
Version no: 4797
No. Date Time ZBAC
Calibration
12/28/17 14:10 .100
Test Result
0001 07/06/18 13:19 .100
Push No. 3 to erase,
OR Power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

#52

AREA NAME: South Los Angeles
AREA NO.: 530
COORDINATOR NAME: Kirk Lee
I.D. NUMBER: 17046
PAS MAKE/MODEL: Lifeloc/FC20
PAS SERIAL NO.: 16080013

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)**	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
								TEST #1	TEST #2 (Optional)	YES	NO				
07/06/2018	1036	34		17GP41049	.100	22		103				X	M. Tippet	20137	

Solution prepared by:
Cesar A. Mejia
Date: 07-24-17
Expiration date:
07-24-19

Cal Standard	Time:	Last Cal	Result	Cal Standard	Time:	Last Check	UOM	Serial No.	FC20	Calibration / Check Printout	Lifeloc Technologies
.100	12:59 PM	01/29/2018	.103	.100	10:37 AM	07/06/2018	BAC	16080013	8.7.24		

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

STATE OF CALIFORNIA
 DEPARTMENT OF CALIFORNIA HIGHWAY PATROL
PRELIMINARY ALCOHOL SCREENING (PAS) DEVICE ACCURACY CHECK/CALIBRATION LOG
 CHP 202K (Rev. 8-08) OPI 051

#53

AREA NAME: South Los Angeles
 AREA NO.: 530
 COORDINATOR NAME: Kirk Lee
 I.D. NUMBER: 17046
 PAS MAKE/MODEL: Intoximeters/Alco-Sensor IV
 PAS SERIAL NO.: 082654

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	LOT NO.	VALUE	TEST #1	TEST #2 (Optional)	PAS ACCURACY CHECK RESULTS*		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	I.D. NO.	REPAIRS/PROBLEMS/COMMENTS
							TEST #1 (Optional)	TEST #2 (Optional)	YES	NO				
07/06/2018	1322	34	17GP41049	.100	22							M. Tippet	20137	
Simulator prepared by: Cesar A. Mejia Date: 07-24-17 Expiration date: 07-24-19														
Push No. 3 to erase. Or power down to retain														
Test Result 0001 06/26/18 00:50 .000 0002 06/26/18 02:08 .138 0003 06/26/18 02:10 .151 0004 06/28/18 00:39 .112 0005 06/28/18 00:42 .115 0006 06/29/18 03:54 .102 0007 06/29/18 03:57 .098 0008 07/06/18 13:23 .105														
No. Date Time XBAC Calibration 12/19/16 11:26 .100														
07/06/18 13:23 NS4 Serial no: 082654 Version no: 4797														

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).
 ** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.

SGTS PAS

AREA NAME	AREA NO.	COORDINATOR NAME	ID. NUMBER	PAS MAKE/MODEL	PAS SERIAL NO.
South Los Angeles	530	Kirk Lee	17046	Intoximeters/Alco-Sensor IV	<i>06888 74</i>

PAS DEVICES SHALL BE CHECKED FOR ACCURACY EVERY 10 DAYS OR EVERY 150 TESTS, WHICHEVER OCCURS FIRST.

DATE	TIME	TEMP (°C)	SIMULATOR SOLUTION (OR DRY GAS)		VALUE	PAS TEMP (°C)		PAS ACCURACY CHECK RESULTS		CALIBRATION REQUIRED		PAS RETEST RESULTS	OPERATOR	ID. NO.	REPAIRS/PROBLEMS/COMMENTS	
			LOT NO.	TEST #1		TEST #2 (Optional)	TEST #1	TEST #2 (Optional)	YES	NO						
07/06/2018	<i>1334 34</i>		17GP41049		.100		<i>20</i>	<i>100</i>				<i>X</i>				
Solution prepared by: Cesar A. Mejia																
Date: 07-24-17																
Expiration date: 07-24-19																

07/06/18 13:35
 NS4 Serial no: 008874
 Version no: 4797
 No. Date Time %BAC
 Calibration 10/24/17 11:04 .100
 Test Result
 0001 06/30/18 01:10 .182
 0002 06/30/18 01:12 .168
 0003 07/02/18 00:19 .168
 0004 07/02/18 00:22 .189
 0005 07/02/18 00:25 .187
 0006 07/02/18 03:12 .127
 0007 07/02/18 03:15 .129
 0008 07/02/18 03:19 .000
 0009 07/04/18 02:12 .068
 0010 07/04/18 02:14 .068
 0011 07/04/18 16:51 .000
 0012 07/04/18 18:52 .000
 0013 07/06/18 13:35 .100
 Push No. 3 to erase
 or power down to retain

* If the reading is within plus or minus (±) 0.010 percent of the actual value of the alcohol solution, no calibration is required (e.g., with a 0.100 percent solution, an acceptable range is 0.090 to 0.110).

** If using dry gas for accuracy checks, note the canister serial number in the lot no. column. If needed, calibration shall be performed with a wet bath simulator.